

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Massachusetts

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2017.

Scope

Mass Echelon III

1000 lb to 0.001 lb

8 oz to 0.03125 oz

50 oz t to 0.005 oz t

12 oz ap to 0.5 gr

Volume Transfer, II

5 gal



2023

A handwritten signature in dark ink, appearing to read "K. Lippa", is written over a horizontal line.

*Katrice Lippa, Chief
NIST Office of Weights and Measures*

Effective Dates: 2023-02-01 to 2024-02-01