NISTIR 7880-11

NIST Micronutrients Measurement Quality Assurance Program Winter 2007 Comparability Studies

Results for Round Robin LXI Fat-Soluble Vitamins and Carotenoids in Human Serum and Round Robin 26 Ascorbic Acid in Human Serum

> David L. Duewer Jeanice B. Thomas



NISTIR 7880-11

NIST Micronutrients Measurement Quality Assurance Program Winter 2007 Comparability Studies

Results for Round Robin LXI Fat-Soluble Vitamins and Carotenoids in Human Serum and Round Robin 26 Ascorbic Acid in Human Serum

> David L. Duewer Jeanice B. Thomas Chemical Sciences Division Materials Measurement Laboratory

> > April, 2013



U.S. Department of Commerce Rebecca Blank, Acting Secretary

National Institute of Standards and Technology Patrick D. Gallagher, Under Secretary of Commerce for Standards and Technology and Director (This page intentionally blank)

Abstract

The National Institute of Standards and Technology coordinates the Micronutrients Measurement Quality Assurance Program (MMQAP) for laboratories that measure fat- and water-soluble vitamins and carotenoids in human serum and plasma. This report describes the design of and results for the Winter 2007 MMQAP measurement comparability improvement studies: 1) Round Robin LXI Fat-Soluble Vitamins and Carotenoids in Human Serum and 2) Round Robin 26 Total Ascorbic Acid in Human Serum. The materials for both studies were shipped to participants in November 2006; participants were requested to provide their measurement results by March 5, 2007.

Keywords

Human Serum Retinol, α-Tocopherol, γ-Tocopherol, Total and *Trans*-β-Carotene Total Ascorbic Acid

Table of Contents

Abstract	iii
Keywords	
Table of Contents	
Introduction	
Round Robin LXI: Fat-Soluble Vitamins and Carotenoids in Human Serum	
Round Robin 26: Vitamin C in Human Serum	2
References	
Appendix A. Shipping Package Inserts for RR61	A1
Appendix B. Final Report for RR61	B1
Appendix C. "All-Lab Report" for RR61	C1
Appendix D. Representative "Individualized Report" for RR61	D1
Appendix E. Shipping Package Inserts for RR26	E1
Appendix F. Final Report for RR26	F1
Appendix G. "All-Lab Report" for RR26	
Appendix H. Representative "Individualized Report" for RR26	

Introduction

Beginning in 1988, the National Institute of Standards and Technology (NIST) has coordinated the Micronutrients Measurement Quality Assurance Program (MMQAP) for laboratories that measure fat- and water-soluble vitamins and carotenoids in human serum and plasma. The MMQAP provides participants with measurement comparability assessment through use of interlaboratory studies, Standard Reference Materials (SRMs) and control materials, and methods development and validation. Serum-based samples with assigned values for the target analytes (retinol, alpha-tocopherol, gamma/beta-tocopherol, *trans*- and total beta-carotene, and total ascorbic acid) and performance-evaluation standards are distributed by NIST to laboratories for analysis.

Participants use the methodology of their choice to determine analyte content in the control and study materials. Participants provide their data to NIST, where it is compiled and evaluated for trueness relative to the NIST value, within-laboratory precision, and concordance within the participant community. NIST provides the participants with a technical summary report concerning their performance for each exercise and suggestions for methods development and refinement. Participants who have concerns regarding their laboratory's performance are encouraged to consult with the MMQAP coordinators.

All MMQAP interlaboratory studies consist of individual units of batch-prepared samples that are distributed to each participant. For historical reasons these studies are referred to as "Round Robins". The MMQAP program and the nature of its studies are described elsewhere. [1,2]

Round Robin LXI: Fat-Soluble Vitamins and Carotenoids in Human Serum

Participants in the MMQAP Fat-Soluble Vitamins and Carotenoids in Human Serum Round Robin LXI comparability study (hereafter referred to as RR61) received one lyophilized and four liquid-frozen human serum test samples for analysis. Unless multiple vials were previously requested, participants received one vial of each serum. These sera were shipped on dry ice to participants in November 2006. The communication materials included in the sample shipment are provided in Appendix A.

Participants are requested to report values for all fat-soluble vitamin-related analytes that are of interest to their organizations. Not all participants report values for the target analytes, and many participants report values for non-target analytes.

The final report delivered to every participant in RR61 consists of three documents:

- A cover letter for the current study, a brief description of the other two documents, and a discussion of our analysis of the overall results that may be of broad interest. This cover letter is reproduced as Appendix B.
- The "All-Lab Report" that lists all of the reported measurement results, a number of consensus statistics for analytes reported by more than one participant, and the mean median and pooled SD from any prior distributions of the serum. This report also provides a numerical "score card" for each participant's measurement comparability for the more commonly reported analytes. This report is reproduced as Appendix C.

• An "Individualized Report" that graphically analyzes each participant's results for all analytes reported by at least five participants. This report also provides a graphical summary of their measurement comparability. The graphical tools used in this report are described in detail elsewhere [3]. An example "Individualized Report" is reproduced as Appendix D.

Round Robin 26: Vitamin C in Human Serum

Participants in the MMQAP Vitamin C in Human Serum Round Robin 26 comparability study (hereafter referred to as RR26) received four frozen serum test samples, one frozen control serum, and a solid ascorbic acid control material for analysis. Unless multiple vials were previously requested, participants received one vial of each material. These sample materials were shipped on dry ice to participants in November 2006. The communication materials included in the sample shipment are provided in Appendix E.

The test and control serum materials were prepared by adding equal volumes of 10 % metaphosphoric acid (MPA) to human serum that had been spiked with ascorbic acid. While these samples contain some dehydroascorbic acid, its content is variable. Therefore, the participants report only total ascorbic acid (TAA, ascorbic acid plus dehydroascorbic acid). Participants are also encouraged to prepare calibration solutions from the supplied solid control to enable calibrating their serum measurements to the same reference standard.

The final report delivered to every participant in RR26 consists of three documents:

- A cover letter for the current study, a brief description of the other two documents, and a discussion of our analysis of overall results that may be of broad interest. This cover letter is reproduced as Appendix F.
- The "All-Lab Report" that summarizes all of the reported measurement results and provides several consensus statistics. This report is reproduced as Appendix G.
- An "Individualized Report" that graphically analyzes each participant's results for TAA, including a graphical summary of their measurement comparability. The graphical tools used in this report are described in detail elsewhere [3]. An example "Individualized Report" is reproduced as Appendix H.

References

- Duewer DL, Brown Thomas J, Kline MC, MacCrehan WA, Schaffer R, Sharpless KE, May WE, Crowell JA. NIST/NCI Micronutrients Measurement Quality Assurance Program: Measurement Repeatabilities and Reproducibilities for Fat-Soluble Vitamin-Related Compounds in Human Sera. Anal Chem 1997;69(7):1406-1413.
- 2 Margolis SA, Duewer DL. Measurement Of Ascorbic Acid in Human Plasma and Serum: Stability, Intralaboratory Repeatability, and Interlaboratory Reproducibility. Clin Chem 1996;42(8):1257-1262.
- 3 Duewer DL, Kline MC, Sharpless KE, Brown Thomas J, Gary KT, Sowell AL. Micronutrients Measurement Quality Assurance Program: Helping Participants Use Interlaboratory Comparison Exercise Results to Improve Their Long-Term Measurement Performance. Anal Chem 1999;71(9):1870-1878.

Appendix A. Shipping Package Inserts for RR61

The following three items were included in each package shipped to an RR61 participant:

- Cover letter
- Datasheet
- Packing List and Shipment Receipt Confirmation Form

The cover letter and datasheet were enclosed in a sealed waterproof bag along with the samples themselves. The packing list was placed at the top of the shipping box, between the cardboard covering and the foam insulation.



UNITED STATES DEPARTMENT OF COMMERCE National Institute of Standards and Technology Geithersburg, Maryland 20899-

November 8, 2006

Dear Colleague:

Enclosed are the samples (Sera 329-333) for the first fat-soluble vitamins and carotenoids in serum round robin study (Round Robin LXI) for the fiscal year (FY) 07 NIST Micronutrients Measurement Quality Assurance Program. You will find one vial of each of four liquid-frozen and one lyophilized serum samples for analysis along with a form for reporting your results. When reporting your results, please submit one value for each analyte for a given serum sample. If a value is obtained below your limit of quantification, please indicate this result on the form by using NQ (*Not Quantified*). Results are due to NIST by **March 5, 2007**. Results received more than two weeks after the due date will not be included in the summary report for this round robin study.

Lyophilized samples should be reconstituted with 1.0 mL of HPLC-grade water or equivalent. We recommend that dissolution be facilitated with 3 to 5 min agitation in an ultrasonic bath or at least 30 min at room temperature with intermittent swirling. (CAUTION: Vigorous shaking will cause foaming and possibly interfere with accurate measurement. The rubber stopper contains phthalate esters that may leach into the sample upon intermittent contact of the liquid sample with the stopper. These esters absorb strongly in the UV region and elute near retinol in most LC systems creating analytical problems.) Pipette a known volume of serum from the vial for analysis. The final volume of the reconstituted sample is greater than 1.0 mL. Water should not be added to the liquid-frozen samples (Sera 329-332).

For consistency, we request that laboratories use the following absorptivities (dL/g•cm): retinol, 1843 at 325 nm (ethanol); retinyl palmitate, 975 at 325 nm (ethanol); α -tocopherol, 75.8 at 292 nm (ethanol); γ -tocopherol, 91.4 at 298 nm (ethanol); α -carotene, 2800 at 444 nm (hexane); β -carotene, 2560 at 450 nm (ethanol), 2592 at 452 nm (hexane); lycopene, 3450 at 472 nm (hexane).

Please mail or fax your results for Round Robin LXI to:

Micronutrients Measurement Quality Assurance Program NIST 100 Bureau Drive Stop 8392 Gaithersburg, MD 20899-8392 Fax: (301) 977-0685

If you have questions or comments regarding this study, please call me at (301) 975-3120; e-mail me at jbthomas@nist.gov; or mailfax queries to the above address.

incerely, DIAD Jeanice Brown Thomas

Research Chemist Analytical Chemistry Division Chemical Science and Technology Laboratory

Enclosures



Date: ____

Round Robin LXI: Human Sera

NIST Micronutrients Measurement Quality Assurance Program

Analyte	329	330	331	332	333	Units*
total retinol						
trans-retinol						
didehydroretinol						
retinyl palmitate						
α -tocopherol						
γ/β-tocopherol						
δ-tocopherol						
total β-carotene						
trans-β-carotene						
total cis-β-carotene						
total α -carotene						
total lycopene						
trans-lycopene						
total β-cryptoxanthin						
total α -cryptoxanthin						
total lutein						
total zeaxanthin						
total lutein&zeaxanthin						
total coenzyme Q10						
ubiquinol (QH ₂)						
ubiquinone (Qox)						
phylloquinone (K1)						
25-hydroxyvitamin D						
Other measurands?		1	1		1	

* we prefer µg/mL

Were the liquid frozen samples #329, 330, 331, and 332 frozen when received? Yes | No

Comments:

Fat-Soluble Vitamins Round Robin LXI

NIST Micronutrients Measurement Quality Assurance Program

Packing List and Shipment Receipt Confirmation Form

This box contains: one vial each of the following five FSV M²QAP sera

Serum	Form	Reconstitute?
#329	Liquid frozen	No
#330	Liquid frozen	No
#331	Liquid frozen	No
#332	Liquid frozen	No
#333	Lyophilized	Yes (1 ml H ₂ O)

Please 1) Open the pack immediately

- 2) Check that it contains all of the above samples
- 3) Check if the vials are intact
- 4) Store the sera at -20 °C or below until analysis
- 5) Complete the following information
- Fax the completed form to us at 301-977-0685 (or email requested information to david.duewer@nist.gov)

1) Date this shipment arrived: _____

2) Are all five sera vials intact? Yes | No If "No", which one(s) were damaged?

- 3) Was there any dry-ice left in cooler? Yes | No
- 4) Did the liquid frozen samples arrive frozen? Yes | No
- 5) At what temperature are you storing the serum samples? _____ °C
- 6) When do you anticipate analyzing these samples?

Your prompt return of this information is appreciated.

The M²QAP Gang

Appendix B. Final Report for RR61

The following three pages are the final report as provided to all participants:

- Cover letter.
- An information sheet that:
 - describes the contents of the "All-Lab" report,
 - describes the content of the "Individualized" report,
 - describes the nature of the test samples and details their previous distributions, if any, and
 - summarizes aspects of the study that we believe may be of interest to the participants.



April 26, 2007

Dear Colleague:

Enclosed is the summary report of the results for round robin LXI (RR61) of the 2007 NIST Micronutrients Measurement Quality Assurance Program (M²QAP) for the fat-soluble vitamins and carotenoids in human serum. Included in this report are: 1) a summary of data and measurement comparability scores for all laboratories, 2) a detailed graphical analysis of your results; and 3) a graphical summary of your measurement comparability.

Data for evaluating laboratory performance in RR61 are provided in text "Score Card" summary, page 7 of the All Lab Report. Laboratory comparability is summarized as follows: results rated 1 to 3 are within 1 to 3 standard deviations of the assigned value, respectively; those rated 4 are >3 standard deviations from the assigned value. Similar information is presented in the graphical "target plot" summary, last page of your Individualized Report.

If you have concerns regarding your laboratory's performance, we suggest that you obtain and analyze a unit of SRM 968c, Fat-Soluble Vitamins, Carotenoids, and Cholesterol in Human Serum. If your measured values do not agree with the certified values, we suggest that you contact us for consultation.

Samples for the second 2007 QA interlaboratory exercise will be shipped starting the week of June 4, 2007. We will send you a reminder via e-mail or fax a week prior to shipment. It is critical that you carefully inspect all samples upon arrival and that you promptly confirm to us that they have arrived. We will replace samples (lost or damaged in shipment or miss-packaged by us) only for participants who report the problem within <u>one calendar week</u> after the package arrives.

We look forward to meeting with you at the Micronutrients Measurement Quality Assurance Workshop that is being held in conjunction with the Experimental Biology meeting on May 2, 2007 at the Convention Center in Washington, DC.

If you have any questions regarding this report, please contact Dave Duewer at david.duewer@nist.gov or me at jbthomas@nist.gov, tel: 301/975-3120, or fax: 301/977-0685.

Sincerely.

Jeanice Brown Thomas Research Chemist Analytical Chemistry Division Chemical Science and Technology Laboratory

Cc: L.C. Sander D.L. Duewer



The NIST M²QAP Round Robin LXI (RR61) report consists of:

Page	"All Lab" Report
1-5	A listing of all results and statistics for all analytes.
6	A legend for the list of results and statistics.
7	The text Comparability Summary ("Score Card") of measurement performance.
Page	"Individualized" Report
	individualized Report
1	Your values, the number of labs reporting values, and our assigned values.

n+1 The graphical Comparability Summary (target plot) of measurement performance.

Serum	Description	Prior Distributions
329	Fresh-frozen, native, single-donor serum prepared in Spring, 2006.	
330	Fresh-frozen, a 15:16 blend of the #329 and #331 sera (<i>i.e.</i> , #330 = $(15 \times #324 + 16 \times #326)/31$, prepared in Spring, 2006.	
331	Fresh-frozen, native, single-donor serum prepared in Spring, 2006.	#326:RR60-9/06
332	Fresh-frozen, native, single-donor, commercially obtained serum prepared in 2002. The same material was used to prepare #333.	#292:RR53-2/03, #301:RR55-3/04, #313:RR57-3/05, #323:RR59-3/06
333	Lyophilized, native, single-donor, commercially obtained serum prepared in 2002. The same material was used to prepare #332.	#290:RR53-2/03, #300:RR55-3/04, #312:RR57-3/05, #322:RR59-3/06

Results

- Sera Stability. There was no significant change in the median level or variability of any measurand in either the fresh-frozen or lyophilized serum of the {332, 333} pair. These materials have been in --80 °C storage for more than four years. The stability of the fresh-frozen material in this and other sera pairs provides strong support for continued use of fresh-frozen rather than lyophilized materials in this program.
- 2) <u>Measurand Additivity.</u> Serum 330 was prepared as a mixture of two native, single-donor sera. Using the correct method of analysis, the median values in Serum 325 are as expected from the Sera 329 and 331 medians and their 15:16 blending ratio. (Previous additivity studies were incorrectly analyzed. The curved lines connecting the three sera in the current graphs are correct for reporting results on a logarithmic y-axis. This topic will be discussed in more detail at the Workshop.)

3) <u>Report of Unusual Performance of High-purity Retinol Standards.</u> Our colleague Fernando Granado-Lorencio, Hospital Universitario Puerta de Hierro (Madrid), reports unusually high chromatographic response per ng injected for recently purchased high-purity retinol standards from Sigma. The new material appears to be of about 20% higher purity than the older material. Similar behavior was noted by other members of the European Standardization Methods Committee during their revision of a candidate reference method for vitamin A in foods.

Appendix C. "All-Lab Report" for RR61

The following eight pages are the "All-Lab Report" as provided to all participants, with two exceptions:

- the participant identifiers (Lab) have been altered.
- the order in which the participant results are listed has been altered.

The data summary in the "All-Lab Report" has been altered to ensure confidentiality of identification codes assigned to laboratories. The only attributed results are those reported by NIST. The NIST results are not used in the assessment of the consensus summary results of the study.

			Total Retinol				tran.	etino				Retiny	Retinyl Palmitate	ate			α-To	Jero		-
Lab	329	330	331	332	333	329	330	331	332	333	329	330	331	332	333	329	330	331	332	333
FSV-BA	0.559	0.433	0.326	0.697	0.709										0.114		15.75			10.16
FSV-BB	0.485	0.388	0.284	0.628	0.617						0.055	0.033	0.011	0.077	0.073	25.17	16.33	9.00	10.42 1	10.07
FSV-BC	0.512	0.396	0.300	0.651	0.619															
FSV-BD	0.536	0.444	0.331	0.696	0.628											24.50				9.80
FSV-BE	0.553	0.390	0.314	0.635	0.607											28.58				11.47
FSV-BF	0.520	0.400	0.290	0.610	0.600											25.00	15.80			10.10
FSV-BG	0.563	0.431	0.316	0.702	0.653						0.106	0.084	0.066	0.172	0.152	26.48	17.00		11.50 1	10.58
FSV-BH	0.422	0.398	0.262	0.584	0.608											21.44	14.39			8.87
FSV-BI	0.472	0.364	0.276	0.577	0.561						0.098	pu	pu	0.110	0.105	23.41	14.99			9.54
FSV-BJ	0.496	0.379	0.280	0.593	0.581							9	2		0.097	27.06	16.60		-	10.07
FSV-BK	0.602	0.427	0.320	0.667	0.637											25.37	14.51			9.28
FSV-BL	0.520	0.370	0.290	0.630	0.600											21.50	14.20	8.20		9.50
FSV-BM	0.519	0.404	0.325	0.634	0.614											26.90	16.30	•	·	10.50
FSV-BN	0.421	0.352	0.251	0.553	0.536						0.050	0.032	0.013	0.100	0.098	22.46	15.93	•	-	10.44
FSV-BO	0.507	0.384	0.315	0.670	0.636												14.10			8.70
FSV-BQ	0.552	0.428	0.352	0.682	0.669												13.10	7.80		9.10
FSV-BR	0.527	0.402	0.305	0.663	0.661											23.06	16.09		`	11.49
FSV-BS	≥0.469	≥ <i>0</i> .366	≥0.274	≥ <i>0</i> .588	≥0.557	0.469	0.366	0.274 (0.588 (0.557						24.82	14.49			9.95
FSV-BT	0.599	0.471	0.342	0.666	0.604											23.26	15.22		`	10.49
FSV-BU	0.329	0 406	0.296	0.565	0 471											23.08				11 69
FSV-RV	0.424	0325	0 262	0.645	0.605											22 95				10 70
		0.000		0.650	0.000							0.050	0100	0 26 0			15.75			
		0.000	0.000		010.0								0.010		002.0	00,00	00.1		`	9.90
F2V-BX	20.57U	≥U.39U	≥U.300	20.047 0 - 00	≥U.033	010.0				0.033						23.09	15.39			10.39
FSV-CC	0.550	0.430	0.320	0.700	0.660	0.540	0.430	0.320 (0.700	0.650						24.10	15.93			9.83
FSV-CE	0.480	0.360	0.300	0.610	0.710											25.00	15.30			5.80
FSV-CF	0.497	0.378	0.282	0.669	0.674											24.20	15.10			11.10
FSV-CG	0.530	0.413	0.322	0.647	0.638											23.43	15.04			9.66
FSV-CI	≥0.462	≥ <i>0</i> .362	≥0.289	≥0.596	≥0.581	0.462	0.362	0.289 (0.596 (0.581	0.045	0.026	0.024	0.092	0.082	23.59	14.61	8.42	9.95	8.98
FSV-CS	0.585	0.437	0.329	0.789	0.650											22.54	15.88		11.40	9.68
FSV-CW	0.513	0.341	0.263	0.611	0.597						0.036 (0.014	0.007	0.091	0.086	25.26	15.70			9.40
FSV-CZ	0.529	0.402	0.304	0.647	0.608											23.80	15.40	8.50 1	10.70 1	10.10
FSV-DD	≥0.600	≥0.340	≥0.380	≥0.730	≥0.680	0.600	0.340	0.380 (0.730 (0.680										
FSV-DF	0.500	0.376	0.303	0.637	0.765															
FSV-DI	0.523	0.418	0.324	0.642	0.624						0.059 (0.033	0.010	0.094	0.091		15.90		-	10.20
FSV-DV	0.568	0.424	0.326	0.696	0.680												13.80	7.60		8.80
FSV-EE	0.454	0.309	0.266	0.550	0.508												13.60	7.10		8.50
z	32	32	32	32	32	2 2									10		33	g		33
Min	0.329	0.309	0.251	0.550	0.471	0.462		0.274 (0.073		13.10			5.80
Median	0.520	0.399	0.304	0.646	0.618	0.510									0.098					9.98
Max	0.602	0.471	0.352	0.789	0.765	0.600	0.430	0.380 (0.730 (0.680					0.230					11.69
SD	0.042	0.035	0.029	0.043	0.038						0.026	0.018	0.010	0.019	0.018	1.56	0.98	0.51	0.62	0.77
C	8	б	10	7	9						45	54	87	19	19	7	9	9	9	80
Npast	0	0	32	33	34	0	0	5	7	7	0	0	ø	12	12	0	0	36	36	37
Medianpast			0.309	0.641	0.605					0.603					0.093				10.39	9.87
SDpast			0.023	0.045	0.044				0.020 (0.020					0.019					0.68
NAV	0.520	0.399	0.304	0.646	0.618	0.489		0.298 (0.622 (0.607	0.057	0.033		0.098	0.098	23.69	15.39		10.50	9.98
NAU	0.042	0.035	0.029	0.051	0.049	0.042		0.029 (0.049	0.026	0.018	0.011	0.026	0.026	1.89	1.19	0.68	0.82	0.78

Page 1/8

All Lab Report

L		Ĕ	pherol		l		0-T	ō-Tocopherol	0			Total	Total β-Carotene	ЭС			trans	trans-β-Carotene	tene	
							330	331	332	333	329	330	331	332	333	329	330		332	333
FSV-BA 1	1.06 2.569 0.90 2.325		3.83 1.	1.94 1	1.87	0.065	0.065	0.079	0.064	0.052	0.580	0.351	0.158 0.130	0.132 0.106	0.127 0.109	0.548	0.330	0.124	0.126	0.121
																				5
					1.95						0.528	0.356	0.169	0.134	0.106					
FSV-BF 0					.77						0.436	0.272	0.121	0.102	0.093					
					.98						0.674	0.410	0.179	0.172	0.177					
	0.67 2.181				1.53						0.580	0.362	0.145	0.119	0.116	0.548	0.344	0.145	0.119	0.116
	.91 2.380			1.95 1	.82						0.544	0.313	0.139	0.111	0.108					
			3.84 1.		.72						0.617	0.359	0.128	0.130	0.126					
FSV-BK																				
					9															
FSV-BN 0	0.90 2.041		3.01	1 707 1	1.48 1.60						0.491	0.358	0.130	0.165	0.109	0.500	0.309	0.150	0.139	0.103
)															
SV-BR																				
	1.24 2.387		3.45 1.	1.80 1	1.84						≥0.554	≥0.345	≥0.167	≥0.134	≥0.129	0.554	0.345	0.167	0.134	0.129
						0.381	0.284	0.282	0.282	0.289	0.520	0.336	0.136	0.135	0.135	0.488	0.317			0.129
FSV-BU 1					.25						0.466	0.278	0.097	0.126	0.110					
					.60						0.584	0.337	0.160	0.121	0.114					
	1.24 3.470				2.39						0.520	0.340	0.160	0.140	0.130					
FSV-BX 1				1.77 1	.72						≥0.527	≥0.307	≥0.139	≥0.119	≥0.110	0.527	0.307	0.139	0.119	0.110
SV-CC																				
FSV-CE											0.780	0.440	0.150	0.220	0.040					
					1.82	0.209	0.260	0.296	0.159	0.121	0.548	0.322	0.143	0.117	0.115	0.507	0.298			0.108
FSV-CI 0					<u>69</u>						20.480	≥0.316	≥0.150	≥0.134 0.136	≥0.139 0.107	0.480	0.316	0.150	0.134	0.139
	1.23 3.0	0.000 4.			20.2						0000	0.310	0.131	0.130	101.0					
			3.80 2.	2.10 2		000.0	0.080	0.120	0.080	0/0.0	150.05 002 0	≥0.284	20.141	≥0.138 0.100	≥0.134 0.134	0.531	0.284	0.141	0.138	0.134
FSV-UZ											0.093	0.390	0.103	0.100	0.171					
FSV-DF																				
	0.90 2.430		3.63 1.	1.84 1	1.78 0	0.033	0.048	0.061	0.042	0.042	0.498	0.315	0.149	0.113	0.111					
FSV-DV FSV-EE																				
z	21	21	21	21	21	9	9	9	9	9	19	19	19	19	19	11	11	11	11	11
			3.01 1.			033	0.038	0.059	0.037	0.032	0.436	0.272	0.097	0.102	0.040	0.453	0.284	0.124	0.102	0.103
Median 0						058	0.073	0.100	0.072	0.061	0.548	0.337	0.145	0.130	0.114	0.507	0.309	0.141	0.126	0.116
	1.25 3.470		5.41 2.		2.39 (0.381	0.284	0.296	0.282	0.289	0.780	0.440	0.179	0.220	0.177	0.554	0.345	0.167	0.139	0.139
SD 0						094	0.121	0.131	0.068	0.047	0.046	0.033	0.019	0.017	0.013	0.034	0.018		0.015	0.015
					12	164	167	131	95	78	8	10	13	13	12	7	9	10	12	13
Npast	0	0		22	22	0	0	5	9	9	0	0	26	26	26	0	0	8	11	11
Medianpast SDpast		ыÖ	3.62 1. 0.24 0.	1.83 1. 0.14 0	1.74 0.11			0.076 0.004	0.063 0.024	0.067 0.023			0.147 0.019	0.120 0.017	0.114 0.016			0.126 0.014	0.116 0.010	0.109 0.009
							0.073	0.100	0.072	0.061	0.548	0.337	0.145	0.130	0.114	0.507	0.309		0.126	0.116
NAU	0.15 0.	0.24 0.3	0.38 0.	0.22 0	0.21	.094	0.121	0.131	0.068	0.047	0.076	0.049	0.023	0.021	0.019	0.053	0.033	0.017	0.015	0.015

Page 2/8

8 0.337 0.339 0 0.274 0.264	0.337 0.274	0.337 0.274 0.250 0.283 0.205	0.337 0.274 0.250 0.283 0.205	0.274	0.337 0.274 0.260 0.283 0.205 0.205	0.337 0.274 0.250 0.283 0.205 0.205 0.205 0.205	0.337 0.274 0.260 0.283 0.283 0.283 0.205 0.283 0.205 0.205 0.205 0.262 0.337 0.032 12 0.262
0.261 0.338	0.261	0.261 0.225 0.159 0.148	0.261 0.225 0.159 0.148	0.261 0.225 0.159 0.148	0.261 0.225 0.159 0.148 0.148	0.261 0.225 0.159 0.148 0.148 0.148 0.223 0.223	0.261 0.225 0.159 0.148 0.148 0.223 0.223 0.223 0.223
0.133	0.133	0.133 0.100 0.085 0.119 0.100	0.133 0.100 0.119 0.100	0.133 0.100 0.085 0.119 0.110	0.133 0.100 0.100 0.100 0.110 0.100	0.133 0.133 0.100 0.100 0.119 0.119 0.110 0.1100 0.100 0.113 0.016	0.133 0.100 0.100 0.119 0.119 0.1100 0.1100 0.1100 0.1133 0.016 0.1133 0.016 0.1133 0.016 0.1133
0.541 0.631 0.566 0.438 0.508 0.508 0.494	0.551 0.631 0.566 0.438 0.508 0.508 0.710 0.710 0.572 0.498 0.508 0.559 0.559	0.551 0.566 0.566 0.438 0.508 0.710 0.710 0.559 0.559 0.559 0.559 0.559 0.559 0.502 0.502 0.502	0.537 0.566 0.566 0.438 0.508 0.508 0.508 0.559 0.559 0.559 0.559 0.502 0.502 0.502 0.537	0.531 0.566 0.438 0.566 0.438 0.508 0.508 0.559 0.559 0.559 0.559 0.502 0.502 0.502 0.502 0.502	0.537 0.531 0.566 0.566 0.566 0.571 0.572 0.559 0.559 0.559 0.559 0.559 0.553 0.502 0.502 0.537 0.537	0.531 0.566 0.566 0.568 0.508 0.5710 0.5710 0.559 0.559 0.559 0.559 0.559 0.559 0.559 0.553 0.553 0.537 0.537 0.537 0.537 0.537 0.537 0.537 0.544	0.531 0.541 0.566 0.566 0.508 0.571 0.572 0.552 0.559 0.552 0.553 0.553 0.537 0.556 0.556 0.556 0.556 0.556 0.556 0.556 0.556 0.556 0.556 0.556 0.5577 0.5577 0.5577 0.55770 0.55770 0.55770000000000
0.421 0.534 0.534 0.534 0.534 0.534 0.429 0.374 0.421 0.421 0.421 0.445 0.445 0.445 0.445	0.421 0.534 0.374 0.374 0.495 0.496 0.496 0.412 0.412 0.410 0.410 0.410	0.421 0.534 0.534 0.374 0.496 0.496 0.412 0.412 0.480 0.412 0.480 0.475 0.480 0.475 0.480	0.421 0.534 0.534 0.499 0.420 0.496 0.412 0.412 0.412 0.412 0.480 0.475 0.480 0.480 0.475 0.433	0.421 0.534 0.534 0.499 0.496 0.496 0.412 0.490 0.412 0.475 0.475 0.475 0.475 0.475 0.475 0.473 0.473 0.433	0.421 0.534 0.499 0.374 0.496 0.496 0.412 0.412 0.412 0.475 0.475 0.475 0.475 0.475 0.433 0.433 0.433	0.421 0.534 0.534 0.499 0.496 0.496 0.412 0.412 0.412 0.412 0.412 0.475 0.475 0.475 0.475 0.477 0.475 0.475 0.433 0.433	0.421 0.534 0.534 0.499 0.496 0.496 0.412 0.412 0.412 0.412 0.412 0.475 0.475 0.475 0.475 0.475 0.433 0.433 0.433 0.433
0.261 0.300 0.280 0.228 0.228 0.269 0.269	0.261 0.300 0.280 0.289 0.269 0.307 0.306 0.307 0.208 0.208	0.261 0.300 0.280 0.228 0.307 0.306 0.307 0.260 0.260 0.260 0.261 0.291 0.201	0.261 0.300 0.280 0.288 0.228 0.228 0.228 0.261 0.261 0.261 0.201 0.201 0.201 0.201 0.278	0.261 0.300 0.280 0.288 0.288 0.228 0.228 0.207 0.201 0.201 0.201 0.201 0.201 0.201 0.278	0.261 0.300 0.280 0.288 0.288 0.228 0.269 0.261 0.278 0.228 0.260 0.278 0.278 0.278 0.278 0.278 0.278 0.278	0.261 0.300 0.280 0.288 0.288 0.288 0.269 0.261 0.288 0.260 0.283 0.269 0.278 0.278 0.278 0.278 0.278 0.278 0.278 0.269 0.278 0.269 0.278 0.278 0.278 0.278 0.278 0.278 0.278 0.278 0.278 0.278 0.269 0.278 0.278 0.269 0.278 0.260 0.269 0.278 0.261 0.278 0.269 0.278 0.2778 0.2780 0.2780 0.2780 0.2780 0.2780 0.2780 0.2780 0.2780 0.2780 0.2780 0.2780 0.2780	0.261 0.300 0.280 0.288 0.288 0.269 0.261 0.261 0.261 0.261 0.263 0.263 0.260 0.260 0.278 0.269 0.278 0.269 0.278 0.269 0.260 0.269 0.278 0.261 0.278 0.261 0.269 0.269 0.269 0.269 0.261 0.278 0.261 0.2780 0.278 0.27800000000000000000000000000000000000
	Л	AI	AI	AI	Al .	AI	A1
	0.014 0.022 0.016 0.012 0.012 0.017 0.017 0.018 0.018 0.018 0.018	0.014 0.016 0.016 0.012 0.0146 0.014 0.013 0.010 0.018 0.018 0.018 0.018 0.018 0.012 0.012	0.014 0.016 0.016 0.012 0.0146 0.014 0.014 0.014 0.012 0.0025 0.00555 0.00555 0.005555 0.005555 0.005555 0.005555 0.005555 0.005	$\begin{array}{c} 0.014\\ 0.016\\ 0.016\\ 0.012\\ 0.012\\ 0.014\\ 0.014\\ 0.012\\ 0.018\\ 0.012\\ 0.012\\ 0.012\\ 0.012\\ 0.012\\ 0.012\\ 0.012\\ 0.012\\ 0.007\\ 0.007\\ 0.0055\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.002\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.007\\ 0.00$	$\begin{array}{c} 0.014 \\ 0.016 \\ 0.016 \\ 0.016 \\ 0.017 \\ 0.017 \\ 0.0146 \\ 0.014 \\ 0.014 \\ 0.012 \\ 0.012 \\ 0.012 \\ 0.012 \\ 0.012 \\ 0.012 \\ 0.012 \\ 0.007 $	0.014 0.016 0.016 0.012 0.012 0.014 0.018 0.018 0.018 0.018 0.018 0.018 0.018 0.018 0.015 0.015 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.005 0.001 0.001 0.012 0.001 0.012 0.001 0.012 0.001 0.012 0.001 0.001 0.002 0.001 0.002 00000000	$\begin{array}{c} 0.014\\ 0.016\\ 0.016\\ 0.012\\ 0.012\\ 0.017\\ 0.017\\ 0.016\\ 0.018\\ 0.016\\ 0.018\\ 0.012\\ 0.016\\ 0.012\\ 0.005\\ 0.005\\ 0.005\\ 0.005\\ 0.005\\ 0.007\\ 0.001\\ 0.007\\ 0.005\\ 0.005\\ 0.007\\ 0.001\\ 0.001\\ 0.005\\ 0.001\\ 0.$
	۸I	004 837 05 837 05 837 05 837 05 837 05 837 05 837 05 837 05 837 05 84 84 84 84 85 85 85 85 85 85 85 85 85 85 85 85 85	004 4 4 222 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5	004 004 005 005 005 005 005 005 005 005	004 004 005 005 005 005 005 005 005 004 004	004 004 005 005 005 005 005 005	004 005 005 005 005 005 005 005
000.0 00000 0000 00000	0.005 0.009 0.005 0.009 0.005 0.003	0.005 0.009 0.005 0.009 0.007 0.007 0.007 0.007	0.005 0.009 0.026 0.005 0.007 0.007 0.007 0.007	0.009 0.009 0.005 0.009 0.007 0.007 0.005 0.007 0.007	0.005 0.009 0.005 0.009 0.007 0.007 0.007 0.007 0.006 0.007 0.005	0.005 0.005 0.005 0.007 0.007 0.007 0.005	0.002 0.009 0.026 0.009 0.005 0.005 0.005 0.007 0.005 0.007 0.006 0.002 0.009 0.002 0.001 37 25 0.007 0.005 0.007 0.005 0.003 0.003
0.040 0.026	0.026 0.026 0.073 0.048 0.073 0.048 0.017 0.017	0.040 0.026 0.073 0.048 0.027 0.017 0.041 0.024	0.040 0.026 0.073 0.048 0.027 0.017 0.041 0.024	0.040 0.026 0.073 0.048 0.027 0.017 0.041 0.024	0.040 0.026 0.073 0.048 0.027 0.017 0.041 0.024 0.026 0.013 0.023 0.021	0.040 0.026 0.073 0.048 0.027 0.017 0.041 0.024 0.026 0.013 0.033 0.021 0.073 0.048 0.008 0.006	0.040 0.026 0.073 0.048 0.027 0.017 0.041 0.024 0.033 0.021 0.033 0.021 0.008 0.006 25 27 0 0 0

Page 3/8

-		Total β-	Total β-Cryptoxanthin	anthin	000	000	Total α-	Total α-Cryptoxanthin	anthin	000		Tot	Total Lutein		000	000	Total	Total Zeaxanthin	hin	000
Lab	329	330	331	332	333	329	330	331	332	333	329	330	331	332	333	329	330	331	332	333
FSV-BA FSV-BB	0.105 0.093	0.121 0.100	0.128 0.104	0.067 0.057	0.064 0.054	0.021 0.021	0.036 0.029	0.045 0.035	0.029 0.026	0.028 0.025	0.161	0.150	0.154	0.082	0.083	0.026	0.056	0.073	0.041	0.045
FSV-BC FSV-BD																				
FSV-BE	0.001	0.005	00100	0.050	0.045															
FSV-BG	0.110	0.120	0.127	0.074	0.070						0.157	0.142	0.128	0.098	0.095	0.028	0.029	0.031	0.030	0.026
FSV-BH	0.119	0.135	0.130	0.065	0.061						0.113	0.109	0.100		0.057	0.023	0.054	0.070	0:030	0.028
FSV-BI	0.100	0.106	0.111	0.057	0.052						0.139	0.132	0.123		0.075	0.026	0.047	0.062	0.031	0.031
	0.105	0.108	0.105	0.053	0.048						0.152	0.149	0.146	0.087	0.083					
FSV-BN	0.079	0.097	0.099	0.051	0.049	0.013	0.026	0.033	0.022	0.021	0.157	0.174	0.167	0.100	0.094	0.032	0.107	0.151	0.041	0.040
	00	211.0	071.0	0000	100.0						0.100	107.0	0.2.0		0	0.000	0.000	10.0	770.0	0000
	0 138	0 130	0 187	0.006	0100															
		201.0		0.000			200	0.00	0.005		0.100	90100	0 1 0 0	0.050	0.055	0,000		010	9000	0.005
	100.0	000.0	0.032	100.0	0.000	020.0	170.0	0.00		0.024	0.120	0.120	0.130		CCD.D	0.010	0.040	0.044	070.0	CZN.0
	0.070	0.074	0.002	100.0	0000															
FSV-BV	0.072	0.074	0.078	0.037	0.035															
FSV-BW																				
FSV-BX	0.074	0.079	0.088	0.048	0.045						0.159	0.154	0.159	0.082	0.077	0.022	0.038	0.052	0.027	0.028
FSV-CC																				
FSV-CE																				
FSV-CF																				
FSV-CG	0.133	0.142	0.150	0.078	0.078															
FSV-CI											0.161	0.190	0.191		0.093	0.030	0.054	0.068	0.037	0.033
FSV-CS	0.082	0.081	0.086	0.054	0.047						0.122	0.115	0.116		0.061	0.026	0.043	0.066	0.042	0.039
FSV-CW	0.097	0.098	0.093	0.055	0.052						0.170	0.151	0.147	0.096	0.092	0.022	0.035	0.041	0.027	0.025
FSV-CZ																				
FSV-DD																				
FSV-DF																				
FSV-DI											0.149	0.166	0.179	060.0	0.085					
FSV-DV																				
	17	17	17	17	17	4	4	4	4	4	13	13	13	13	13	11	11	11	11	1
Min	0.072	0.074	0.062	0.037	0.010	0.013	0.026	0.033	0.022	0.021	0.113	0.109	0.100	0.057	0.055	0.022	0.029	0.031	0.022	0.025
Median	0.097	0.100	0.104	0.057	0.052	0.021	0.028	0.034	0.026	0.025	0.157	0.150	0.147	0.087	0.083	0.026	0.047	0.066	0.030	0.030
Max	0.138	0.142	0.184	0.096	0.078	0.023	0.036	0.045	0.029	0.028	0.199	0.201	0.200	0.105	0.103	0.043	0.107	0.151	0.042	0.045
SD	0.017	0.025	0.026	0.011	0.006	0.002	0.003	0.003	0.002	0.002	0.016	0.025	0.029	0.013	0.013	0.004	0.012	0.018	0.009	0.007
S	18	25	25	19	1	6	1	10	7	8	10	17	20	15	16	16	25	27	29	23
Npast	0	0	23	22	22	0	0	4	5	7	0	0	14	15	15	0	0	11	13	13
Medianpast			0.112	0.055	0.052			0.040	0.027	0.024			0.145	0.080	0.078			0.057	0.032	0.030
SDpast			0.027	0.009	0.008			0.004	0.007	0.007			0.042		0.015			0.018	0.007	0.007
NAV	0.097	0.100	0.104	0.057	0.052	0.021	0.028	0.034	0.026	0.025	0.157	0.150	0.147		0.083	0.026	0.047	0.066	0.030	0.030
NAU	0.022	0.025	0.026	0.014	0.013						0.029	0.027	0.029	0.017	0.016	0.007	0.013	0.018	0.009	0.008

Page 4/8

	0.220	0.252	0.126	333 0.119	329	330	331	332	333	329	330	331	332	333	329	330	331	332 333
0.187	0.206	0.228	0.123	0.128														
0.180 0.167	0.191 0.155	0.195 0.140	0.115 0.112	0.112 0.105														
0.136	0.163	0.170	0.087	0.085														
CO1.7	0.179	C81.0	0.112	0.100	0.781	1.119	1.194	0.804	0.740									
0.189 0.229	0.282 0.256	0.318 0.277	0.141 0.126	0.134 0.133														
0 101	0 2 2 0	0.256	0110	0 1 0 2														
0.164	0.190	0.205	0.102	0.096														
0.152	0.166	0.116	0.101	0.093														
0.201	0.219	0.236	0.127	0.118	0.436	0.595	1.032	0.715	0.709	0.179	0.211	0.814	0.528	0.619	0.257	0.384	0.218 0	0.187 0.090
0.181	0.192	0.212	0.109	0.105														
0.238	0.265	0.293	0.152	0.149	002.0		100	0 EEO	0 2 00									
0.191 0.148	0.158 0.158	0.182 0.182	0.137 0.114	0.120 0.100	0.7.0	0.000		000.0	0.530									
0.192	0.185	0.188	0.124	0.117	0.630	0.730	0.820	0.660	0.500									
					0.680		0.960	0.670	0.690									
											104							
					CI / 0	0.341	080.1	0.740	0.121	0.0.0	0.400	1.20U	0.024	0.7Z1	0.7 80	0.030	n 201.0	U.180 U.ZUI
					0.720	0.911	1.311	0.754	0.726									
17	17	17	17	17	7	7	7	7	7	2	2	2	2	2	2	2	2	
0.136	0.155	0.116	0.087	0.085	0.436	0.595	0.820	0.650	0.500	0.010	0.211	0.814	0.528	0.619	0.257			
0.187	0.192	0.212	0.119	0.112	0.700	0.880	1.090	0.715	0.709	0.095	0.308	1.037	0.676	0.673	0.519			0.187 0.146
0.238	787.0	0.318	ZGL.0	0.149	0.781	911.1	1.311	0.804	0.740	0.179	C04.0	1.26U	0.824	0.121	0.780	0.090	U.Z18 U	187 0.201
0.020 11	0.037 19	0.053 25	0.011 9	0.018 16	0.046	0.101 11	0.112 10	0.064 9	0.064 9									
0	0	22	23	23	0	0	5	5	5	0	0	0	0	0	0	0	0	0
		0.211	0.113	0.108			1.190	0.760	0.699									
		0.056	0.017	0.018			090.0	0.169	0.162									
0.187	0.192	0.212	0.119	0.112	0.700	0 880	1 090	0715	0 700									

All Lab Report

Page 5/8

	Ф.	hylloquir	Phylloquinone (K1) x1000	l) x1000			25-hyc	25-hydroxyvitamin D	min D	
Lab	329	330	331	332	333	329	330	331	332	333
FSV-BA						0.015	0.014	0.008	0.020	0.020
FSV-BB						0.016	0.012	0.010	0.020	0.025
FSV-BC										
FSV-BD										
FSV-BE	1.586	0.995	0.599	0.389	0.357					
FSV-BF										
FSV-BG										
FSV-BH						0.015	0.011	0.007	0.021	0.020
FSV-BI										
FSV-BJ										
ESV-RK										
FSV-BM										
FSV-BN						0.017	0.011	0.008	0.019	0.018
FSV-BO										
FSV-BQ										
FSV-RR										
F5V-B1										
FSV-BU										
FSV-BV										
FSV-BW										
10/10/10/10/10/10/10/10/10/10/10/10/10/1										
122-VS-										
FSV-CE										
FSV-CF										
FSV-CG										
FSV-CI	1.339	0.887	0.530	0.402	0.308					
00-701		0000	0000	10.00	0000					
100-00										
FSV-CZ										
FSV-DD										
FSV-DF										
FSV-DI	0.880	0.700	0.500	0.370	0.330					
ESV-DV										
FSV-FF										
	¢.	¢.	¢.	¢.	۰.	4	4	4	4	4
Min		002.0	0,500	0 2 2 0		0.015	0.01	200.0	0100	a 10 0
INIEGIAN	1.339	0.88/	0.530	0.389	0.330	GL0.0	0.012	0.008	0.020	0.020
Max	1.586	0.995	0.599	0.402	0.357	0.017	0.014	0.010	0.021	0.025
SD						0.001	0.001	0.001	0.001	0.001
S						5	10	7	2	7
Nnast	С	С	С	С	С	C	С	С	ŝ	4
	•	•	•	•	•	•	•	•)	- FC C
Neularipast SDpast										0.004
NAV	1.339	0.887	0.530	0.389	0.330	0.015	0.012	0.008	0.020	0.020
NAU										

Analytes Reported By One Laboratory

Analyte	Code	329	330	331	332	333
Total cis-β-Cryptoxanthin	FSV-BT	0.029	0.032	0.036	0.022	0.022
trans-α-Carotene	FSV-BX	0.285	0.140	0.021	0.081	0.075

.

Term	Legend
N	Number of (non-NIST) quantitative values reported for this analyte
Min	Minimum (non-NIST) quantitative value reported
Median _{part}	Median (non-NIST) quantitative value reported
Max	Maximum (non-NIST) quantitative value reported
SD	Standard deviation for (non-NIST) results: 0.741*(3rd Quartile - 1st Quartile)
CV	Coefficient of Variation for (non-NIST) results: 100*SD/Median
Npast	Mean of N(s) from past RR(s)
Medianpast	Mean of Median(s) from past RR(s)
SD _{past}	Pooled SD from past RR(s)
NAV	NIST Assigned Value
	= (Median _{part} + Mean _{NIST})/2 for analytes reported by NIST analyst(s) = Median _{part} for analytes reported by ≥ 10 labs but not NIST
NAU	NIST Assigned Uncertainty: $(S^2 + S_{btw}^2)^{0.5}$
	S is the maximum of (0.05*NAV, SD, SNIST, eSD) and Sbtw is the standard deviation between Median _{part} and MeanNIST. The expected long-term SD, es is defined in: Duewer, et al. Anal Chem 1997;69(7):1406-1413.
_	Not analyzed
nd	Not detected (i.e., no detectable peak for analyte)
ng	Detected but not quantitatively determined
≤x	Concentration at or below the limit of quantification, x
≥x	Concentration greater than or equal to x
?	Non-quantitative value: extrapolated beyond upper limit of calibration curve
*	Non-quantitative value: heterogeneous serum, damaged sample, malfunctio
italics	Not explictly reported but calculated by NIST from reported values

Comparability Summary

Lab	TR	аT	g/bT	bC	tbC	aC	TLy	TbX	TLu	ΤZ	L&Z	_
FSV-BA	2	1	1	1	1	1	1	1			1	
FSV-BB	1	1	1	1	1	1	1	1	1	2	1	
FSV-BC	1											
FSV-BD	1	1										
FSV-BE	1	3	2	1								
FSV-BF	1	1	1	2		1	1	1			1	
FSV-BG	1	2	2	3		1	1	2	1	2	1	
FSV-BH	2	2	2	1	1	1	1	1	2	1	2	
FSV-BI	2	1	1	1		1	1	1	1	1	1	
FSV-BJ	1	2	1	1		1	1	1	1			
FSV-BK	2	1										
FSV-BL	1	1										
FSV-BM	1	1										
FSV-BN	2	1	2	1	1	1	1	1	1	4	2	
FSV-BO	1	2	1	2	1	1	2	1	2	1	2	
FSV-BQ	2	2										
FSV-BR	1	2										
FSV-BS	2	1	2	1	2	3	1	3			1	
FSV-BT	2	1	2	1	1	1	1	1	2	2	1	
FSV-BU	3	2	2	2		1	2	1			2	
FSV-BV	2	1	1	1		1	1	2			1	
FSV-BW	1	1	4	1		1	2					
FSV-BX	1	1	1	1	1	1		1	1	1	1	
FSV-CC	1	1										
FSV-CE	2	3		4								
FSV-CF	1	2										
FSV-CG	1	1	1	1	1	1	1	2			2	
FSV-CI	1	1	1	1	1	4			2	1	1	
FSV-CS	2	1	2	1		1	1	1	2	1	1	
FSV-CW	2	1	2	1	1	1		1	1	1	1	
FSV-CZ	1	1		3								
FSV-DD	3			-								
FSV-DF	2											
FSV-DI	1	1	1	1		4	1		1			
FSV-DV	2	2										
FSV-EE	2	2										
n	36	33	21	23	11	20	17	17	13	11	17	
				-		-			-			
	TR	аT	g/bT	bC	tbC	aC	TLy	TbX	TLu	ΤZ	L&Z	
% 1	53	64	<u>52</u>	74	91	85	82	76	62	64	71	1
% 2	42	30	43	13	9	0	18	18	38	27	29	
%3	6	6	0	9	0	5	0	6	0	0	0	
% 4	0	0	5	4	Ő	10	0	0	Ő	9	Ő	
, , , ,	-	-	-	•	-	. •	-		-	-	-	J

Label	Definition
Lab	Participant code
TR	Total Retinol
aT	α-Tocopherol
g/bT	γ/β-Tocopherol
bC	Total β-Carotene
tbC	trans-β-Carotene
aC	Total α-Carotene
TLy	Total Lycopene
TbX	Total β-Cryptoxanthin
TLu	Total Lutein
ΤZ	Total Zeaxanthin
L&Z	Total Lutein & Zeaxanthin
n	number of participants providing quantitative data
% 1	Percent of CS = 1 (within 1 SD of medians)
% 2	Percent of CS = 2 (within 2 SD of medians)
% 3	Percent of $CS = 3$ (within 3 SD of medians)

% 4 Percent of CS = 4 (3 or more SD from medians)

"Comparability Score"

The Comparability Score (CS) of summarizes your measurement performance for a given measurand, relative to the consensus medians. CS is the average distance, in standard deviation units, that your measurement performance characteristics are from the consensus performance. CS is calculated when the number of quantitative values you reported for a measurand, N_{you} , is at least two and the measurand has been reported by 10 or more participants.

$$CS = MIN(4, INT(1 + \sqrt{C^2 + AP^2}))$$

$$C = Concordanc e = \sum_{i}^{N_{you}} \frac{You_{i} - Median_{i}}{NAU_{i}} / N_{you}$$

$$AP = Apparent \ Precision = \sqrt{\sum_{i}^{N_{you}} \left(\frac{You_{i} - Median_{i}}{NAU_{i}}\right)^{2} / (N_{you} - 1)}$$

NAU = NIST Assigned Uncertainty, our estimate of the overall measurement standard deviation for each sample. The estimate includes serum heterogeneity, analytical repeatability, and among-participant reproducibility variance components.

For further details, please see: Duewer DL, Kline MC, Sharpless KE, Brown Thomas J, Gary KT. Micronutrients Measurement Quality Assurance Program: Helping participants use interlaboratory comparison exercise results to improve their long-term measurement performance. Anal Chem 1999;71(9):1870-8.

Appendix D. Representative "Individualized Report" for RR61

Each participant in RR61 received an "Individualized Report" reflecting their reported results. Each report included a detailed analysis for analytes that were assayed by at least five participants. The following analytes met this criterion in RR61:

- Total Retinol
- trans-Retinol
- Retinyl Palmitate
- α-Tocopherol
- γ/β -Tocopherol
- δ-Tocopherol
- *trans*-β-Carotene
- Total *cis*-β-Carotene
- Total α-Carotene
- Total Lycopene
- *trans*-Lycopene
- Total β-Cryptoxanthin
- Total Lutein
- Total Zeaxanthin
- Total Lutein & Zeaxanthin
- Coenzyme Q10

The following 14 pages are the "Individualized Report" for the analytes evaluated by participant FSV-BA.

Set 1 of 32

Individualized Round Robin LXI Report: FSV-BA

Summary

	c	32	10	33	21	9	19	1	5	19	17	7	17	4	17	4
Serum 333	NAV	0.618	0.10	9.98	1.780	0.061	0.114	0.116	0.006	0.082	0.496	0.264	0.052	0.025	0.112	0.020
Seru	You	0.709	0.11	10.16	1.867	0.052	0.127	0.121	0.006	0.082	0.486	0.286	0.064	0.028	0.119	0.020
	c	32	10	33	21	9	19	5	9	19	17	~	17	4	17	4
Serum 332				•	1.819											
Seru	You	0.697	0.12	10.59	1.935	0.064	0.132	0.126	0.006	0.084	0.507	0.300	0.067	0.029	0.126	0.020
	c	32	ი	33	21	9	19	7	9	19	17	7	17	4	17	4
Serum 331	NAV	0.304	0.0	8.50	3.710	0.100	0.145	0.141	0.010	0.016	0.583	0.290	0.104	0.034	0.212	0.008
Seru	You				3.829											
	c	32	თ	33	21	9	19	1	2	19	17	2	17	4	17	4
Serum 330	NAV	0.399	0.0	15.39	2.430	0.073	0.337	0.309	0.021	0.151	0.445	0.223	0.100	0.028	0.192	0.012
Seru	You	0.433	0.1	15.75	2.569	0.065	0.351	0.330	0.021	0.154	0.427	0.231	0.121	0.036	0.220	0.014
	c	32	10	33	21	9	19	1	2	19	17	2	17	4	17	4
m 329	NAV	0.520	0.06	23.69	0.902	0.058	0.548	0.507	0.033	0.316	0.269	0.100	0.097	0.021	0.187	0.015
Serum	You	0.559	0.07	24.43	1.061	0.065	0.580	0.548	0.033	0.316	0.262	0.121	0.105	0.021	0.192	0.015
	Analyte	Total Retinol	Retinyl Palmitate	a-Tocopherol	γ/β-Tocopherol	ō-Tocopherol	Total β-Carotene	trans-β-Carotene	Total cis-β-Carotene	Total α-Carotene	Total Lycopene	trans-Lycopene	Total β-Cryptoxanthin	Total α-Cryptoxanthin	Total Lutein&Zeaxanthin	25-hydroxyvitamin D

You : Your reported values for the listed analytes (micrograms/milliliter)


NAV : NIST Assigned Values, here equal to this RR's median

n: Number of non-NIST laboratories reporting quantitative values for this analyte in this serum

Please check our records against your records. Send corrections and/or updates to... Micronutrients Measurement Quality Assurance Program National Institute of Standards and Technology Gaithersburg, MD 20899-8392 USA 100 Bureau Drive Stop 8392

Tel: (301) 975-3935 Fax: (301) 977-0685 Email: david.duewer@nist.gov Page 1 / 14

Individualized Report




 Fresh-frozen: 53:#292, 55:#301, 57:#313, 59:#323
 Native, same single-source as #333

 Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322
 Native, same single-source as #333

#332


#333



#333

Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322

Native, same single-source as #332




Native, single-source 15:16 blend of #329 and #331 Native, single-source Native, same single-source as #333 Native, same single-source as #332

#331

#332

#333




#332

#333


Fresh-frozen: 53:#292, 55:#301, 57:#313, 59:#323

Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322



#333 Lyophilized: 53:#292, 55:#300, 57:#312, 59:#322

Native, same single-source as #332




#332

#333

Fresh-frozen: 53:#292, 55:#301, 57:#313, 59:#323


Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322



#333


Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322

Native, same single-source as #332



Native, single-source Native, same single-source as #333 Native, same single-source as #332

#331




#332


#333

Fresh-frozen: 53:#292, 55:#301, 57:#313, 59:#323

Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322



Comments Native, single-source 15:16 blend of #329 and #331 Native, single-source Native, same single-source as #333 Native, same single-source as #332




15:16 blend of #329 and #331 Native, single-source Native, same single-source as #333 Native, same single-source as #332


#332

#333

Fresh-frozen: 53:#292, 55:#301, 57:#313, 59:#323

Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322






<u>Serum</u>	<u>History</u>	<u>Comments</u>
#329	Fresh frozen: New	Native, single-source
#330	Fresh frozen: New	15:16 blend of #329 and #331
#331	Fresh frozen: 60:#326	Native, single-source
#332	Fresh-frozen: 53:#292, 55:#301, 57:#313, 59:#323	Native, same single-source as #333
#333	Lyophilized: 53:#290, 55:#300, 57:#312, 59:#322	Native, same single-source as #332

Page 14 / 14





Appendix E. Shipping Package Inserts for RR26

The following five items were included in each package shipped to an RR26 participant:

- Cover letter
- Protocol for Preparation and Analysis of the Ascorbic Acid Solid Control Material
- Preparation and Validation of Ascorbic Acid Solid Control Material Datasheet
- Analysis of Control Materials and Test Samples Datasheet
- Packing List and Shipment Receipt Confirmation Form

The cover letter, preparation protocol, and the two datasheets were enclosed in a sealed waterproof bag along with the samples themselves. The packing list was placed at the top of the shipping box, between the cardboard covering and the foam insulation.

November 8, 2006



Dear Colleague:

The samples within this package constitute the first Vitamin C Round Robin (RR26) of the fiscal year (FY) 07 Micronutrients Measurement Quality Assurance Program. RR26 consists of four vials of frozen serum *test* samples (#13, #42, #112, and #113), one vial of ascorbic acid solid control material, and one vial of SRM 970 Ascorbic Acid in Serum-Level I to be used as the serum control material. Please follow the attached protocols when you prepare and analyze these samples. If you cannot prepare the solid control solutions gravimetrically, please prepare equivalent solutions volumetrically and report the exact volumes used. (Routine 0.5 g gravimetric measurements are generally 10-fold more accurate than routine 0.5 mL volumetric measurements.)

Please use SRM 970 to validate the performance of your measurement system <u>before</u> you analyze the *test* samples. The target value and \approx 95% confidence interval for SRM 970-Level I is 8.41 ± 0.61 µmol/L of sample.

Please be aware that sample contact with any oxidant-contaminated surface (vials, glassware, etc.) may degrade your measurement system's performance (SA Margolis and E Park, "Stability of Ascorbic Acid in Solutions in Autosampler Vials", *Clinical Chemistry* 2001, 47(8), 1463-1464). You should suspect such degradation if you observe unusually large variation in replicate analyses.

The report for RR25 was mailed mid-October. If you find your results for RR25 unsatisfactory, we recommend that you obtain **Standard Reference Material** (SRM) 970 Ascorbic Acid in Serum to validate your methodology and value assign in-house control materials. This SRM may be purchased from the Standard Materials Reference Program at NIST (Tel: 301-975-6776, Fax: 301-948-3730, or e-mail: srminfo@nist.gov).

If you have any questions or concerns about the Vitamin C Micronutrients Measurement Quality Assurance Program please contact Jeanice Brown Thomas at tel: 301-975-3120, fax: 301-977-0685, or e-mail: <u>jbthomas@nist.gov</u>.

We ask that you return your results for the RR26 samples by March 5, 2007. We would appreciate receiving your results as soon as they become available. Please use the attached form. Your results will be kept

confidential. Sincerely.

Jeanice Brown Thomas Research Chemist Analytical Chemistry Division Chemical Science and Technology Laboratory

Enclosures: Protocols, Preparation and Analysis of Control Materials and Analysis of Test Samples RR26 Report Form for Ascorbic Acid Solid Control Material Preparation RR26 Report Form for Control Material and Test Sample Analyses



Micronutrient Measurement Quality Assurance Program for Vitamin C

Please Read Through Completely BEFORE Analyzing Samples

Protocol for Preparation and Analysis of the Ascorbic Acid Solid Control Material

The *ascorbic acid solid control material* (in the amber vial) should be prepared and used in the following manner:

- 1) Prepare at least 500 mL of 5% mass fraction metaphosphoric acid (MPA) in distilled water. This solution will be referred to as the "Diluent" below.
- 2) Weigh 0.20 to 0.22 g of the ascorbic acid solid control material to 0.0001 g (if possible), dissolve it in the Diluent in a 100 mL volumetric flask, and dilute with the Diluent to the 100 mL mark. Weigh the amount of Diluent added to 0.1 g. Record the weights. The resulting material will be referred to as the "Stock Solution" below.
- 3) Prepare three dilute solutions of the Stock Solution as follows:

<u>Dilute Solution 1:</u> Weigh 0.500 mL of the Stock Solution to 0.0001 g into a 100 mL volumetric flask; dilute with Diluent to the 100 mL mark. Record the weight.

<u>Dilute Solution 2:</u> Weigh 0.250 mL of the Stock Solution to 0.0001 g into a 100 mL volumetric flask; dilute with Diluent to the 100 mL mark. Record the weight.

<u>Dilute Solution 3:</u> Weigh 0.125 mL of the Stock Solution to 0.0001 g into a 100 mL volumetric flask; dilute with Diluent to the 100 mL mark. Record the weight.

4) Calculate and record the total ascorbic acid concentrations, [TAA], in these Dilute Solutions. If you follow the above gravimetric preparation directions, the [TAA] in µmol/L is calculated:

 $[\mathsf{TAA}]_{\mathsf{DS}} = \frac{(\mathsf{g} \operatorname{Stock} \operatorname{Solution} \operatorname{in} \operatorname{Dilute} \operatorname{Solution}) \cdot (\mathsf{g} \operatorname{AA} \operatorname{in} \operatorname{Stock} \operatorname{Solution}) \cdot (56785 \ \mu \operatorname{mol/g} \cdot \mathsf{L})}{(\mathsf{g} \operatorname{AA} \operatorname{in} \operatorname{Stock} \operatorname{Solution}) + (\mathsf{g} \operatorname{Diluent} \operatorname{in} \operatorname{Stock} \operatorname{Solution})}$

For example, if you prepared the Stock Solution with 0.2000 g of solid ascorbic acid and 103.0 g of Diluent, then 0.5 mL of the Stock Solution should weigh (0.2+103)/200 = 0.52 g and $[TAA]_{DS1} = (0.52 \text{ g})(0.2 \text{ g}) \cdot (56785 \mu \text{mol/g} \cdot \text{L})/(0.2 + 103 \text{ g}) = 57.2 \mu \text{mol/L}$. Likewise, 0.25 mL of the Stock Solution should weigh 0.26 g and $[TAA]_{DS2} = 28.4 \mu \text{mol/L}$ and 0.125 mL should weigh 0.13 g and $[TAA]_{DS3} = 14.2 \mu \text{mol/L}$.

5) Measure the ultraviolet absorbance spectrum of Dilute Solution 1 against the Diluent as the blank using paired 1 cm path length cuvettes. Record the absorbance at 242, 243, 244, and 245 nm. Record the maximum absorbance (A_{max}) within this region. Record the wavelength (λ_{max}) at which this maximum occurs.

The extinction coefficient ($E^{1\%}$) of ascorbic acid at λ_{max} (using a cell with a 1 cm path length) of Dilute Solution #1 can be calculated:

 $E^{1\%}(\frac{dL}{g \cdot cm}) = \frac{(A_{max}) \cdot ((g \text{ AA in Stock Solution}) + (g \text{ Diluent in Stock Solution}))}{(g \text{ Stock Solution in Dilute Solution 1}) \cdot (g \text{ AA in Stock Solution})}$

If your spectrophotometer is properly calibrated, λ_{max} should be between 243 and 244 nm and $E^{1\%}$ should be 550 ± 30 dL/g·cm. If they are not, you should calibrate the wavelength and/or absorbance axes of your spectrophotometer and repeat the measurements.

- 6) Measure and record the concentration of total ascorbic acid in all three dilute solutions and in the 5% MPA Diluent in duplicate using *exactly* the same method that you will use for the serum control materials and test samples, including any enzymatic treatment. We recommend that you analyze these solutions in the following order: Diluent, Dilute Solution 1, Dilute Solution 2, Dilute Solution 3, Dilute Solution 3, Dilute Solution 2, Dilute Solution 1, Diluent.
 - a) Compare the values of the duplicate measurements. *Are you satisfied that your measurement precision is adequate?*
 - b) Compare the measured with the calculated [TAA] values. This is most conveniently done by plotting the measured values on the y-axis of a scatterplot against the calculated values on the x-axis. The line through the four {calculated, measured} data pairs should go through the origin with a slope of 1.0. *Are you satisfied with the agreement between the measured and calculated values?*

Do **<u>not</u>** analyze the serum control materials or test samples until you are satisfied that your system is performing properly!

 Once you have confirmed that your system is properly calibrated, analyze the serum control CS #2 (see protocol below). The target values for this materials is 28.1 ±1.0 µmol/L of sample.

If your measured values are not close to this value, please review your sample preparation procedure and whether you followed *exactly* the same measurement protocol the solutions prepared from the solid control material as you used for these serum controls. If the protocols differ, please repeat from Step 6 using the proper protocol. If the proper protocol was used, your measurement system may not be suitable for MPA-preserved samples. Please contact us: 301-975-3120 or Jeanice.BrownThomas@NIST.gov.

Do <u>not</u> analyze the test samples until you are satisfied that your system is performing properly and is suitable for the analysis of MPA-preserved serum!

Protocol for Analysis of the Serum Control Materials and Test Samples

The *serum control material* and *test samples* are in sealed ampoules. They were prepared by adding equal volumes of 10% MPA to spiked human serum. We have checked the samples for stability and homogeneity. Only the total ascorbic acid is stable. While these samples contain some dehydroascorbic acid, its content is variable. Therefore, only <u>total ascorbic acid</u> should be reported. The *serum control material* and *test samples* should be defrosted by warming at 20 °C for not more than 10 min otherwise some irreversible degradation may occur.

Each *serum test sample* contains between 0.0 and 80.0 μ mol of total ascorbic acid/L of solution. The total ascorbic acid in each ampoule should be measured in duplicate. Please report your results in μ mol/(L of the sample solution) rather than μ mol/(L of serum NIST used to prepare the sample).

Participant #: _____

Date:

Vitamin C Round Robin 26

NIST Micronutrient Measurement Quality Assurance Program

Preparation and Validation of Ascorbic Acid Solid Control Material

STOCK SOLUTION

Mass of ascorbic acid in the Stock Solution	_ g
Mass of 5% MPA Diluent added to the 100 mL volumetric flask	_ g

DILUTE SOLUTION 1

Mass of added stock solution (0.5 mL)	_ g
Mass of 5% MPA Diluent added to the 100 mL volumetric flask	_ g
Absorbance of Dilute Solution 1 at 242 nm	AU
Absorbance of Dilute Solution 1 at 243 nm	AU
Absorbance of Dilute Solution 1 at 244 nm	AU
Absorbance of Dilute Solution 1 at 245 nm	AU
Absorbance of Dilute Solution absorbance maximum	AU
Wavelength of maximum absorbance	_nm
Calculated E ^{1%}	_dL/g·cm
Calculated [TAA] _{DS1}	_μmol/L

DILUTE SOLUTION 2

Mass of added stock solution (0.25 mL)	g
Mass of 5% MPA Diluent added to the 100 mL volumetric flask	g
Calculated [TAA] _{DS2}	_μmol/L

DILUTE SOLUTION 3

Mass of added stock solution (0.125 mL)	g
Mass of 5% MPA Diluent added to the 100 mL volumetric flask	g
Calculated [TAA] _{DS3}	µmol/L

Please return by March 5, 2007

MMQAP
100 Bureau Drive, Stop 8392
Gaithersburg, MD 20899-8392

Fax: 301-977-0685 Email: david.duewer@nist.gov

Date:

Vitamin C Round Robin 26 NIST Micronutrient Measurement Quality Assurance Program

Analysis of Control Materials and Test Samples

Sample	Replicate 1	Replicate 2	Units
Dilute Solution 1			µmol/L of Dilute Solution
Dilute Solution 2			µmol/L of Dilute Solution
Dilute Solution 3			µmol/L of Dilute Solution
5% MPA Diluent			μmol/L of Diluent
SRM 970-Level I			μmol/L of Sample <i>Target:</i> 8.41 ±0.61 μmol/L
Serum Test Sample #13			µmol/L of Sample
Serum Test Sample #42			µmol/L of Sample
Serum Test Sample #112			µmol/L of Sample
Serum Test Sample #113			µmol/L of Sample

Were samples frozen upon receipt? Yes | No

Analysis method: HPLC-EC | HPLC-Fluor DAB | HPLC-OPD | HPLC-UV | AO-OPD | Other If "Other", please describe:

COMMENTS:

Please return by March 5, 2007

Fax: 301-977-0685 Email: david.duewer@nist.gov Vitamin C Round Robin 26 NIST Micronutrients Measurement Quality Assurance Program

Packing List and Shipment Receipt Confirmation Form

This box contains one vial each of the following **six** VitC M^2QAP samples:

Label	Form
VitC #13	Liquid frozen (1:1 serum:10% MPA)
VitC #42	Liquid frozen (1:1 serum:10% MPA)
VitC #112	Liquid frozen (1:1 serum:10% MPA)
VitC #113	Liquid frozen (1:1 serum:10% MPA)
SRM 970 Lv1	Liquid frozen (1:1 serum:10% MPA)
Control	Solid AA

Please 1) Open the pack immediately

- 2) Check that it contains one vial each of the above samples
- 3) Check if the samples arrived frozen
- 4) Store the samples at -20 °C or below until analysis
- 5) Complete the following information
- 6) Fax the completed form to us at 301-977-0685 (or email requested information to david.duewer@nist.gov)

1) Date this shipment arrived: _____

- 2) Are all of the vials intact? Yes | No If "No", which one(s) were damaged?
- 3) Was there any dry-ice left in cooler? Yes | No
- 4) Did the samples arrive frozen? Yes | No
- 5) At what temperature are you storing the samples? _____ °C
- 6) When do you anticipate analyzing these samples? _____

Your prompt return of this information is appreciated.

The M²QAP Gang

Appendix F. Final Report for RR26

The following two pages are the final report as provided to all participants:

- Cover letter.
- An information sheet that:
 - describes the contents of the "All-Lab" report,
 - o describes the content of the "Individualized" report,
 - describes the nature of the test samples and details their previous distributions, if any, and
 - summarizes aspects of the study that we believe may be of interest to the participants.



April 26, 2007

Dear Colleague:

Enclosed is the summary report of the results for Round Robin 26 (RR 26) for the measurement of total ascorbic acid (TAA, ascorbic acid plus dehydroascorbic acid) in human serum. Included in this report are a summary of data for all laboratories and an individualized summary of your laboratory's measurement performance. The robust median is used to estimate the consensus value for all samples, the "median absolute deviation from the median" (MADe) is used to estimate the expected standard deviation, and the coefficient of variation (CV) is defined as 100×MADe/median.

RR 26 consisted of four test samples (13, 42, 112, and 113), one serum control material, and one solid control material for preparation of TAA control solutions. Details regarding the samples can be found in the enclosed report.

If you have concerns regarding your laboratory's performance, we suggest that you obtain and analyze a unit of Standard Reference Material (SRM) 970 Vitamin C in Frozen Human Serum. SRM 970 can be purchased from the NIST SRM Program at phone: 301-975-6776; fax: 301-948-3730. If your measured values do not agree with the certified values, we suggest that you contact us for consultation.

Samples for the second vitamin C round robin (RR 27) of the 2007 M²QAP will be shipped **during the week of June 4, 2007.**

We look forward to meeting with you at the Micronutrients Measurement Quality Assurance Workshop that is being held in conjunction with the Experimental Biology meeting on May 2, 2007 at the Convention Center in Washington, DC.

If you have questions or concerns regarding this report, please contact David Duewer at 301-975-3935; e-mail: david.duewer@nist.gov or me at 301-975-3120; e-mail: jbthomas@nist.gov; or fax: 301-977-0685.

Sincerely, onnel

Jeanide Brown Thomas Research Chemist Analytical Chemistry Division Chemical Science and Technology Laboratory

Enclosures

Cc: L. C. Sander D.L. Duewer



The NIST M²QAP Vitamin C Round Robin 26 (RR26) report consists of

Page		"Individualized" Report							
1		marizes your reported values for the nominal 55 mmol/L solution you prepared from the rbic acid solid control sample, the serum control sample, and the four serum test bles.							
2	Grap	hical summary of your RR26 sample measurements.							
Page		"All Lab" Report							
1	1 A tabulation of results and summary statistics for Total Ascorbic Acid [TAA] in the RR26 samples and control/calibration solutions.								
Serum-ba	ased S	Samples. One serum control and four unknowns were distributed in RR26.							
Contr	ol 1	SRM 970 level 1, ampouled in mid-1998.							
Samp	 le 1 Serum 13, ampouled in late 2001, previously distributed as sample S19:1 (RR19, 203), S21:1 (RR21, Fall 04) and S23:1 (RR23, Fall 05). A "blank" stripped serum. 								
Samp	le 2	Serum 42, ampouled in late 2001, previously distributed as sample S18:2 (RR18, Spring 03), S19:3 (RR19, Fall 03), S21:3 (RR21, Fall 04), S22:3 (RR22, Spring 05) and S24:2 (RR24, Spring 06).							
Samp	le 3	Serum 112, ampouled in 1993, previously distributed as sample 179B in RR4 (1993).							
Samp	le 4	4 Serum 113, ampouled in 1993, previously distributed as sample 180 in RR4 (1993), RR (1994) and RR14 (2001).							

Results.

- All participants who prepared the four 5% MPA control/calibration solutions (the three "Dilute Solutions" and the "Diluent") did so correctly. The criteria used to evaluate this success are: the density of the 5% MPA (≈1.03 g/mL), the observed wavelength maximum of "Dilute Solution 1" (≈244 nm), the observed absorbance at that maximum (≈0.58 OD), the calculated E^{1%} #1"(≈560 dL/g·cm).
- 2) Judging from the calibration parameters calculated for the control/calibration solutions (intercepts close to 0.0, slopes close to 1.0, R² close to 1, and RMS close to 0.0), the measurement systems for all participants are linear and reasonably well calibrated. However, several participants continue to have measurement systems that perform somewhat differently for the control solutions and the test samples.
- 3) Somewhat remarkably, the median and the MADe for sera 112 (S26:3) and 113 (S26:4) are effectively unchanged from the TAA results obtained for these materials in the 1993 RR4 study.

Appendix G. "All-Lab Report" for RR26

The following single page is the "All-Lab Report" as provided to all participants, with two exceptions:

- the participant identifiers (Lab) have been altered.
- the order in which the participant results are listed has been altered.

The data summary in the "All-Lab Report" has been altered to ensure confidentiality of identification codes assigned to laboratories.

Micronutrients Measurement Quality Assurance Program for Total Ascorbic Acid "Round Robin" 26 - March 2007

	JVL	S26:4	21.1	18.9	15.1	21.2	25.1	18.5	22.3	26.4	24.8	22.5	20.9	11	21.5	3.3	15.1	19.9	21.2	23.6	26.4	3.4	16
Calibrated to Gravimetric umol/	tric, µmc	S26:3	79.5	68.4	53.6	69.2	84.0	65.8	71.3	79.6	84.2	73.0	71.5	11	72.7	9.0	53.6	68.8	71.5	79.5	84.2	8.6	12
	Gravimet	S26:2	34.9	32.0	32.1	32.7	37.0	32.0	32.9	34.8	35.0	33.0	30.1	11	33.3	1.9	30.1	32.1	32.9	34.8	37.0	1.4	4
	ated to (S26:1 S	0.0	0.4	1.2	0.3	1.2	0.1	0.0	3.7	0.8	0.4	0.2	11	0.8	1.1	0.0	0.1	0.4	1.0	3.7	0.6	
SS	Calibr	CS#1 S	8.2	8.3	9.2	9.1	10.7	7.9	7.9	11.2	8.2	9.0	7.9	11	8.9	1.1	7.9	8.0	8.3	9.2	11.2	0.6	8
Samples		S26:4 C	21.3	19.4	15.1	21.0	23.8	19.2	22.4	27.8	26.1	23.0	20.2	11	21.8	3.5	15.1	19.8	21.3	23.4	27.8	2.9	14
	٨L	S26:3 S	79.8	71.0	57.2	68.1	82.7	67.0	71.3	83.2	82.0	75.9	68.7	11	73.4	8.2	57.2	68.4	71.3	80.9	83.2	6.8	10
	Measured, µmol/I	S26:2 S;	35.2	33.0	33.7	32.3	35.8	32.9	33.0	36.5	35.7	34.1	29.0	11	33.7	2.1	29.0	32.9	33.7	35.4	36.5	2.1	9
	Measure	S26:1 S2	0.0	0.0	0.0	0.5	0.0	0.6	0.0	4.1	3.5	0.0	0.3	11		1.5				0.5		0.0	
		CS#1 S2	8.5	8.3	8.7	9.1	9.4	8.5	8.0	11.9	10.5	8.9	7.7	11	9.0	1.2	7.7	8.4	8.7	9.3	11.9	0.7	8
		ö	m		m	10	m	6			·	m	C	6	6	4	0	m	m	6		4	-
	y.	E ^{1%}	554.3	560.2	562.3	563.5	571.8	566.6		350.9^{a}	566.6	543.3	542.0		558.9	10.4	542.0	554.3	562.	566.6	571.8	6.4	1.1
Dilute Solution 1	Spectrophotometry	A _{max}	0.5734	0.5630	0.5507	0.5843	0.5910	0.6130		0.369^{a}	0.5948	0.5616	0.5881	6	0.5800	0.0195	0.5507	0.5630	0.5843	0.5910	0.6130	0.0162	2.8
Dil	Spec	λ_{max}	244.	244.	243.	243.	243.6	244.		254^a	244.	243.8	244.	6	243.7	0.4	243.0	243.6	244.0	244.0	244.0	0	0.00
MPA	Density	g/mL	1.033	1.027	1.029	1.028	1.029	1.032	1.031	1.025	1.030	1.029	1.034	11	1.030	0.003	1.025	1.028	1.029	1.031	1.034	0.003	0.26
-	Õ	RMS 0	0.3	0.4	2.2	0.3	1.4	1.0	0.1	0.8	0.1	0.8	0.2	z	Average	SD	Min	%25	Median	%75	Max	eSD	S
	arameters	R ² RI	1.000	000.1	0.995	1.000	0.998	0.999	1.000	0.999	1.000	0.999	000.1		Ave				Me				
	ion Para	Slope F	1.00 1.		-	0.98 1.			•	I.04 0.	0.94 1.		0.96 1.										
S	Calibration Pa	Inter Slo	.27	. 39	.34	.20 (.24	.44	0.10	0.24	2.81 (-0.47	0.13 (
ample			0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	2.8	ې ۲.	.0 0	11	0.2	0.	۲.	0.0	0.	0.0	2.8	0.0	7
ation S		3 MPA										-	0 0	, L		5						0	6
Calibra	d, µmo	Dil:3	7 14.8	9 13.9	0 14.0	3 14.6	3 12.5	16.7		16.1	17.3	15.1	17.0	-	3 15.1	3 1.5	2 12.5	3 14.0	2 14.8	3 16.4	7 17.3	~.	6
Control / Calibration Samples	Measured, µmol/L	Dil:2	\$ 29.7		27.0		27.6			32.4		30.4	30.1	1	29.6	1.6		28.28			32.37	1.8	•
0		Dil:1	58.8	59.4	60.7	57.7	58.1	62.7		62.1	58.9	60.7	59.1	11	59.4	2.1		58.5	59.1	60.7	62.7	1.0	(7)
	mol/L	Dil:3	14.3	14.1	14.2	14.6	15.0	15.0	13.9	15.5	15.3	14.1	17.6	11	14.9	1.0		14.2	14.6	15.2	17.6	0.8	5
	Gravimetric, µmol/I	Dil:2	29.3	28.3	28.2	29.6	29.5	30.9	27.5	29.9	30.1	29.3	31.0	11	29.4	1.1	27.55	28.76	29.51	29.95	31.02	0.8	e
	Gravin	Dil:1	58.7	57.1	55.6	58.9	58.7	61.4	55.2	59.7	59.6	58.7	61.6	11	58.7	2.1	55.2	57.9	58.7	59.7	61.6	1.4	2
ļ		Date	20/03/07	26/11/06	23/01/07	07/03/07	26/01/07	20/02/07	21/02/07	05/02/07	28/02/07	23/03/07	02/03/07	z	Average	SD	Min	%25	Median	%75	Max	MADe	S
		Lab	VC-MA	VC-MB	VC-MC	VC-ME	VC-MG	VC-MH	VC-MI	VC-MJ	VC-MK	VC-MN	VC-MU										

a) 5% Trichloroacetic acid solution

Page 1 / 1

Appendix H. Representative "Individualized Report" for RR26

Each participant in RR26 received an "Individualized Report" reflecting their reported results. The following two pages are the "Individualized Report" for participant "VC-MA".

Vitamin C "Round Robin" 26 Report: Participant VC-MA


			MPA	Dilut	e Solutio	on 1	Co	Control/Calibration Solutions						
			Density	Spect	rophotor	netry	$Y_{meas} = Inter + Slope^* X_{grav}$							
Date	RR	Method	g/mL	λ_{max}	A _{max}	E ^{1%}	Inter	Slope	R^2	SEE				
09/13/04	21	HPLC-EC	1.030	244.0	0.555	562.2	-0.1	0.99	1.000	0.10				
03/08/05	22	HPLC-EC	1.034	243.0	0.559	562.9	0.2	1.06	1.000	0.24				
10/17/05	23	HPLC-EC	1.030	244.0	0.562	567.9	-0.6	1.09	0.998	1.47				
03/09/06	24	HPLC-EC	1.031	244.0	0.568	586.7	0.2	1.13	1.000	0.41				
08/28/06	25	HPLC-EC	1.039	242.0	0.555	557.4	0.8	0.95	0.999	0.92				
03/20/07	26	HPLC-EC	1.033	244.0	0.573	554.3	0.3	1.00	1.000	0.31				
		Mean	1.033	243.5	0.56	565.2				0.57				
		SD	0.003	0.8	0.01	11.5				0.52				
		CV	0.34	0.34	1.3	2.0								

			_	m [AAT]	nmol/Ls	ample						
Date	RR	Sample	Rep ₁	Rep ₂	F_{adj}	Mean	SD_{dup}	N		Mean	SD _{repeat}	SD _{reprod}
09/13/04	21	CS#1	8.1	7.9	1.0	8.0	0.1		7	8.8	0.1	0.6
03/08/05	22	CS#1	8.5	8.7	1.0	8.6	0.1	-				
10/17/05	23	CS#1	9.3	9.5	1.0	9.4	0.1					
10/17/05	23	CS#1	9.3	9.5	1.0	9.4	0.1					
03/09/06	24	CS#1	9.3	9.2	1.0	9.3	0.0					
08/28/06	25	CS#1	8.3	8.6	1.0	8.4	0.2					
03/20/07	26	CS#1	8.6	8.3	1.0	8.5	0.2					
11/13/03	19	S19:1	nd	nd	1.0				2	0.0	0.0	0.0
09/13/04	21	S21:1	nd	nd	1.0							
10/17/05	23	S23:1	0.0	0.0	1.0	0.0	0.0					
03/20/07	26	S26:1	0.0	0.0	1.0	0.0	0.0					
				[
03/20/03	18	S18:2	35.1	36.0	1.0	35.6	0.6		6	35.1	0.3	1.1
11/13/03	19	S19:3	35.9	35.8	1.0	35.9	0.1					
09/13/04	21	S21:3	33.2	32.9	1.0	33.0	0.2					
03/08/05	22	S22:3	35.7	35.6	1.0	35.6	0.1					
03/09/06	24	S24:2	35.8	35.5	1.0	35.6	0.2					
03/20/07	26	S26:2	35.0	35.4	1.0	35.2	0.3					
		0040	0.07	0.0-				-				10.0
09/16/93	04	S04:2	2.27	2.27	28.4	64.4	0.0		2	72.1	0.1	10.8
03/20/07	26	S26:3	79.7	79.8	1.0	79.8	0.1					
00/40/00	0.4	004.0	0.00	0.04	00.4	40.4	0.0		-	40.7	0.4	4.0
09/16/93	04	S04:3	0.66	0.64	28.4	18.4	0.3		4	19.7	0.4	1.3
05/25/95	06	S06:2	0.7	0.7	28.4	19.2	0.2					
09/27/01	14	S14:1	20.4	19.6	1.0	20.0	0.6					
03/20/07	26	S26:4	21.1	21.5	1.0	21.3	0.3					

Please check our records against your records. Send corrections and/or updates to...

Micronutrients Measurement Quality Assurance Program National Institute of Standards and Technology 100 Bureau Drive Stop 8392 Gaithersburg, MD 20899-8392 USA

Fax: (301) 977-0685 Email: david.duewer@nist.gov



Vitamin C "Round Robin" 26 Report: Participant VC-MA

Duewer, Kline, Sharpless, Brown Thomas, Gary, Sowell. Anal Chem 1999;71(9):1870-8.

Sample

<u>Comments</u>

S26:1 VitC #13, a "blank" previously distributed in RRs 16, 19, 21, 22, and 24 S26:2 VitC #42, previously distributed in RRs 18, 19, 21, 22, and 24 S26:3 VitC #112, previous distributed in RR 4 S26:4 VitC #113, previous distributed in RR 4, 6 and 14