

United States Department of Commerce
National Institute of Standards and Technology

~ Certificate of Measurement Traceability ~
for

Tennessee

a participant in the NIST Weights and Measures
Measurement Assurance Program for the Year(s)

2011

The State standards are traceable to NIST and the laboratory has
*conditionally demonstrated** measurement proficiency through
training and interlaboratory comparisons.

***Demonstrated
Areas of Proficiency**

Mass

Echelon III
1000 lb - 0.001 lb
8 oz - 0.03125 oz

Volume

Volume transfer, II
5 gal

*Conditional Status:
Facility has inadequate environ-
mental controls (NIST HB 143
5.3, 6.2). Laboratory lacks ade-
quate administrative proce-
dures. Conditional Recognition
is issued only so the laboratory
can provide services to meet
weights and measures regula-
tions.*



Carol T. Hockert,
for the National Institute of Standards and Technology

Chief, Weights and Measures Division

Title

January 1, 2011 to December 31, 2011

Effective Date

