

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Michigan

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), maintains a documented calibration program, documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, submitted a quality management system, and has demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2017.



2024 to 2025

Scope

Mass Echelon I

20 kg to 1 mg

Mass Echelon II

20 kg to 1 mg

50 lb to 0.001 lb

8 oz to 0.03125 oz

Mass Echelon III

500 kg to 1 mg

1000 lb to 0.001 lb

8 oz to 0.03125 oz

Weight Carts

6000 lb to 3000 lb

Wheel Load Weighers

20 000 lb to 5000 lb

Volume Gravimetric, I

2 L to 100 mL

25 gal to 0.5 pt

Volume Gravimetric, I SVP

30 gal to 5 gal

Volume Transfer, II

2000 gal to 5 gal

Volume Transfer, II LPG

100 gal to 20 gal

*Katrice Lippa, Chief
NIST Office of Weights and Measures*

Effective Dates: 2024-02-01 to 2026-01-31