

Standards Activity for Powered Exoskeletons

Eric Franca, Ph.D.

Lead Reviewer/Biomedical Engineer

CDRH/ODE/DNPMD/NSDP

and

Primary Liaison and Member of JWG36



IEC Joint Working Group 36

- Organizational Structure: TC 62/SC 62D/ JWG 36
 - TC 62 Electrical equipment in medical practice
 - SC 62D Electromedical equipment
 - JWG 36 Medical Robots for Rehabilitation

JWG 36 Scope: To develop IEC 80601-2-78: Medical Electrical Equipment - Part 2-78: Particular requirements for the basic safety and essential performance of medical robots for rehabilitation, compensation or alleviation of disease, injury or disability



Who is involved in JWG 36?

- Chair: Jürgen Stettin (DE) IEC SC 62D
- Secretary: Jeffrey Eggleston (US) IEC SC 62D
- Convenor: Michel Brossoit (CA) CSA
- Project Lead: Burkhard Zimmermann (CH) -Hocoma AG, IISART



Who else is involved in JWG 36?

- Industry representatives
 - International companies with expertise or interest in medical robotics
- Regulatory agency representatives
 - Officially, just the FDA but...
 - Participants from PMDA (Japanese FDA)
 - KFDA (Korean FDA) informed by Korean experts in JWG36
- Standards agency representatives
 - ISO, IEC, AAMI



Development of IEC 80601-78

- Currently a draft standard
- Committee Draft (CD1) should be finalized after next meeting in February
- Commenting Period
 - Should be mid 2017
 - Will last around 2 months
- If comments are extensive, CD2 may be necessary



Why is this Standard Important?

- Global harmonization on a device area with no current standards
- Device-specific safety considerations which may not be covered elsewhere
- Helps build consistency and avoid common design flaws
 - Helps limit avoidable adverse events
- If recognized by FDA, may lead to reduced burdens on Industry and FDA Staff
- Adapt to new technologies in the future with updates



What does IEC 80601-78 cover?

Scope

This International Standard applies to the general requirements for BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ROBOTS that physically interact with a PATIENT to support or perform REHABILITATION, ASSESSMENT, COMPENSATION or ALLEVIATION related to the PATIENT'S MOVEMENT FUNCTIONS following an IMPAIRMENT.

Excluded:

- Robotic (external limb) prosthetics
- Diagnostic imaging equipment
- Robots that don't address impaired body structures or functions



What does IEC 80601-78 include?

NOTE: Content has not been finalized and may change.

- Defines "RACA ROBOT"
 - MEDICAL ROBOT intended to perform REHABILITATION, ASSESSMENT, COMPENSATION or ALLEVIATION comprising an ACTUATED APPLIED PART
- Amends and/or adds RACA ROBOT-specific considerations to IEC 60601-1 clauses.
- May introduce concept of SITUATION AWARENESS



FDA Staff Contacts for JWG 36

Eric Franca, PhD
Biomedical Engineer, Primary Liaison and Member of JWG36
DNPMD/NSDP

Eric.Franca@fda.hhs.gov

Ian Marcus, MS
Biomedical Engineer and Alternate Liaison for JWG36
DNPMD/PMDB

lan.Marcus@fda.hhs.gov



Discussion Questions



- Terminology: What terms are associated with medical exoskeletons, their testing, manufacturing processes, and use?
- Taxonomy: Are there subgroup classification categories that should be considered for a medical exoskeleton standard?
- **Use Cases/Applications**: What are the potential applications for medical exoskeletons and where can these devices be used?

Discussion Questions



Metrics:

- What metrics should be used to evaluate the device performance?
- What measures can be used in standards to assess user safety?

Measurement Tools

- What tools exist/don't exist that are/would be useful to measure the safety and performance of exoskeletons?
- **Test Methods:** What are the key areas of non-clinical testing that can be addressed by standards?
 - What testing methods can be used to characterize device performance? Device safety?
 - What are the scientific and clinical considerations for testing of atypical structures (e.g. soft fabric based exoskeletons)?
 - Should the standard include cybersecurity testing and usability?

Stakeholders

— Who are the stakeholders for these exoskeletons?

