

Detailed Instructions for Installation of metRology for Microsoft Excel

Will Guthrie

NIST Statistical Engineering Division

2013-01-08

Basics


- metRology for Microsoft Excel requires the open-source statistical software R. If you do not already have R on your computer, the installation file will take you to the correct link to install it.
- Alternatively save the file <http://cran.rstudio.com/bin/windows/base/R-2.15.2-win.exe> on your desktop and then double click it to install R.
- You can accept all of the default suggestions when you run the R install script.

Basics Continued

- After installing R, double click on the file metRology-for-Microsoft-Excel-v1.0-setup.exe to configure R and install the user interface.
- You can generally use the defaults throughout this install as well.
- One potential exception is the choice of installation directory metRology for Microsoft Excel. Make sure it is installed in C:\Program Files\metRology for Microsoft Excel

Potential Exception to Use of Defaults During Install

- You are most likely to find the installation script suggesting an alternative installation location if you have a 64-bit machine. If so, just **delete the “(x86)”** in the window at right during the install



Configuring Excel

- When you install metRology for Microsoft Excel, the installation script will most likely open Excel during the install and a menu will appear that asks you to enable content or macros.
- Please click enable to allow installation of the Excel Add-in used to link Excel and R.
- After you enable the content, Excel will close automatically and the installation will finish.

Testing the Software

- After completing the installation first reboot your computer and then you can run a test to make sure everything is working using one of the versions of the file
test-metRology-for-Microsoft-Excel.xls(m).
 - use the [xls](#) file if you have Excel 2003
 - use the [xlsm](#) file if you have Excel 2007 or 2010
- To complete the test open the appropriate test file with Excel, then click the button labeled “Compute Uncertainty”. You should also see a menu item named “metRology” on the menu bar.

Test File – Push Button Circled in Red

test-metRology-for-Microsoft-Excel.xlsm - Microsoft Excel

File Home Insert Page Layout Formulas Data Review View Developer Acrobat metRology

GUM Uncertainty Analysis Consensus Mean Polynomial Regression

GUM Supplement 1 ASTM E691

Uncertainty Analysis Interlaboratory Studies Regression Analysis

E10

	A	B	C	D	E	F	G	H	I	J	K	L	M	N
1		metRology for Microsoft Excel GUM Uncertainty Template for 1 Input												
2														
3			Uncertainty Analysis Input Table											
4		Input Variable Name	x1											
5		Units	mg											
6		Input Value, x_i	1.0000											
7		Standard Uncertainty, u_i	0.0120											
8		Degrees of Freedom, ν_i	60.0											
9		Type of Uncertainty Evaluation	A											
10		Distribution	Normal											
11		Measurement Equation	exp(x1)											
12		← Show/Hide Options												
13														
14			Uncertainty Analysis Results											
15														
16			Output Value	Combined Standard Uncertainty	Effective Degrees of Freedom	Nominal Level of Confidence	Coverage Factor =	Expanded Uncertainty	Approximate Attained Confidence Level					
17			$y=f(x_1, \dots, x_p)$	$u_c(y)$	ν_{eff}	$100(1-\alpha)\%$	k	$U(y)$						
18			x1											
19		Contribution to $[u_c(y)]^2$, %												
20		Sensitivity Coefficient												
21														
22														
23														
24														
25														
26														
27														
28														
29														
30														
31														
32														
33														
34														
35														
36														
37														
38														
39														
40														
41														
42														
43														
44														
45														
46														
47														
48														
49														
50														
51														

← More

← Show/Hide Warnings and Error Messages

GUM Template 1 Input

Ready

93%

Successful Test Results

test-metRology-for-Microsoft-Excel.xlsm - Microsoft Excel

File Home Insert Page Layout Formulas Data Review View Developer Acrobat **metRology**

GUM Uncertainty Analysis Consensus Mean Polynomial Regression
GUM Supplement 1 ASTM E691

Uncertainty Analysis Interlaboratory Studies Regression Analysis

K27

	A	B	C	D	E	F	G	H	I	J	K	L	M	N
1		metRology for Microsoft Excel GUM Uncertainty Template for 1 Input												
2														
3			Uncertainty Analysis Input Table											
4		Input Variable Name	x1											
5		Units	mg											
6		Input Value, x_1	1.0000											
7		Standard Uncertainty, u_1	0.0120											
8		Degrees of Freedom, ν_1	60.0											
9		Type of Uncertainty Evaluation	A											
10		Distribution	Normal											
11		Measurement Equation	exp(x1)											
22		← Show/Hide Options												
23														
24			Uncertainty Analysis Results											
25			Output Value	Combined Standard Uncertainty	Effective Degrees of Freedom	Nominal Level of Confidence	Coverage Factor = Student's t	Expanded Uncertainty	Approximate Attained Confidence Level					
26			$y=f(x_1, \dots, x_p)$	$u_c(y)$	ν_{eff}	$100(1-\alpha)\%$	k	$U(y)$						
27		Compute Uncertainty	2.718	0.0326	60.0	95%	2.00	0.065	93.7%					
28			x1											
29		Contribution to $[u_c(y)]^2$, %	100.0%											
30		Sensitivity Coefficient	2.7183											
31														
32														
33		No errors or warnings.												
34														
35														
48		← More												
49		← Show/Hide Warnings and Error Messages												
50														
51														

GUM Template 1 Input

Ready

93%

More on Test Results and Final Notes

- The numbers in the output area of the test file do not need to match the screenshot exactly
- As long as there are numbers and no apparent errors, all should be well