

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Idaho

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.



Scope

2018 to 2019

Mass Echelon II

30 kg to 1 mg
1000 lb to 1 μ lb
8 oz to 0.03125 oz

Mass Echelon III

30 kg to 1 mg
2500 lb to 0.001 lb
8 oz to 0.03125 oz

Weight Carts

3000 lb to 2000 lb

Volume Gravimetric, I

5 gal

Volume Transfer, II

750 gal to 5 gal
100 gal to 20 gal LPG

Douglas A. Olson, Chief
NIST Office of Weights and Measures

Effective Dates: 2018-01-01 to 2019-12-31