

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Montana

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2017.



2026

Scope

Mass Echelon III

20 kg to 1 mg
2500 lb to 0.001 lb
8 oz to 0.03125 oz

Weight Carts
5000 lb to 2000 lb

Volume Transfer, II

500 gal
100 gal
50 gal
5 gal

Volume Transfer, II LPG

100 gal
25 gal

Katrice Lipa, Chief
NIST Office of Weights and Measures

Effective Dates: 2026-02-01 to 2027-01-31