

# United States Department of Commerce

## National Institute of Standards and Technology

### Certificate of Metrological Traceability For:

# Ohio

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), maintains a documented calibration program, documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, submitted a quality management system, and has demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2017.



2025 to 2026

#### *Scope*

##### **Mass Echelon II**

30 kg to 1 mg  
50 lb to 0.001 lb  
8 oz to 0.001 oz

##### **Mass Echelon III**

500 kg to 1 mg  
5000 lb to 0.001 lb  
8 oz to 0.001 oz  
500 oz t to 1 oz t  
10 dwt to 0.1 dwt

##### **Volume Transfer, II**

1000 gal to 5 gal

##### **Volume Transfer, II LPG**

100 gal

##### **Weight Carts**

6000 lb to 3000 lb

Katrice Lippa, Chief  
NIST Office of Weights and Measures

Effective Dates: 2026-02-01 to 2027-01-30

NOTE: Amended to remove  
Stopwatch measurement scope  
2026-01-30.