

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

South Carolina

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon II

50 kg to 1 mg
2000 lb to 0.001 lb
8 oz to 0.015625 oz

Mass Echelon III

2000 kg to 1 mg
2500 lb to 0.001 lb
8 oz to 0.015625 oz

Weight Carts

6000 lb to 2500 lb
Wheel Load Weighers
40 000 lb to 10 000 lb

Volume Gravimetric, I


5 gal

Volume Transfer, II

20 L to 10 L
1500 gal to 5 gal

Volume Transfer, II LPG

100 gal to 20 gal



2019

A handwritten signature in dark ink, reading "Douglas A. Olson".

Douglas A. Olson, Chief
NIST Office of Weights and Measures

Effective Dates: 2019-02-01 to 2020-02-01