


# National Bureau of Standards Mass Calibration Computer Software

---

R. N. Varner  
R. C. Raybold



---

U.S. DEPARTMENT OF COMMERCE, Philip M. Klutznick, Secretary  
Luther H. Hodges, Jr., Deputy Secretary  
Jordan J. Baruch, Assistant Secretary for Productivity, Technology and Innovation  
NATIONAL BUREAU OF STANDARDS, Ernest Ambler, Director  
Issued July 1980

**National Bureau of Standards Technical Note 1127**

Nat. Bur. Stand. (U.S.), Tech. Note 1127, 164 pages (July 1980)

CODEN: NBTNAE

**U.S. GOVERNMENT PRINTING OFFICE  
WASHINGTON: 1980**

---

For sale by the Superintendent of Documents, U.S. Government Printing Office  
Washington, D.C. Price \$5.00  
(Add 25 percent for other than U.S. mailing)

# CONTENTS

|                                                                                 | <u>Page</u> |
|---------------------------------------------------------------------------------|-------------|
| 1.0 Introduction . . . . .                                                      | 1           |
| 2.0 Calculations . . . . .                                                      | 1           |
| 2.1 Dictionary of Symbols and Corresponding FORTRAN<br>Variable Names . . . . . | 19          |
| 3.0 Information For Users . . . . .                                             | 23          |
| 3.1 Description of Input . . . . .                                              | 23          |
| Figure I - Starting Restraint Identification<br>Number Used at NBS . . . . .    | 29          |
| Figure II - Calibration Design Number Used at NBS .                             | 29          |
| 3.2 Diagnostic and Error Messages . . . . .                                     | 30          |
| 3.3 Description of Data Output Used for Process Control                         | 33          |
| 4.0 Implementation Information . . . . .                                        | 34          |
| 4.1 Description of Main Program and all Subprograms . .                         | 35          |
| 4.1.01 MAIN Program . . . . .                                                   | 35          |
| 4.1.02 BLKDAT Subprogram . . . . .                                              | 35          |
| 4.1.03 READ1 Subprogram . . . . .                                               | 36          |
| 4.1.04 READIT Subprogram . . . . .                                              | 36          |
| 4.1.05 ERROR Subprogram . . . . .                                               | 37          |
| 4.1.06 PRINT1 Subprogram . . . . .                                              | 37          |
| 4.1.07 TEXT1 Subprogram . . . . .                                               | 37          |
| 4.1.08 TEXT2 Subprogram . . . . .                                               | 37          |
| 4.1.09 TEXT3 Subprogram . . . . .                                               | 37          |
| 4.1.10 TEXT4 Subprogram . . . . .                                               | 37          |
| 4.1.11 TEXT5 Subprogram . . . . .                                               | 38          |
| 4.1.12 TEXT6 Subprogram . . . . .                                               | 38          |
| 4.1.13 READ2 Subprogram . . . . .                                               | 38          |
| 4.1.14 SPINV Subprogram . . . . .                                               | 39          |
| 4.1.15 SAVMTX Subprogram . . . . .                                              | 39          |
| 4.1.16 INVCHK Subprogram . . . . .                                              | 39          |
| 4.1.17 PRINT2 Subprogram . . . . .                                              | 39          |
| 4.1.18 CHKLN Subprogram . . . . .                                               | 40          |
| 4.1.19 PGCONT Subprogram . . . . .                                              | 40          |
| 4.1.20 HEADPG Subprogram . . . . .                                              | 40          |
| 4.1.21 FINPRT Subprogram . . . . .                                              | 41          |
| 4.1.22 TEXTS1 Subprogram . . . . .                                              | 41          |

|                                                         | <u>Page</u> |
|---------------------------------------------------------|-------------|
| 4.1.23 TEXTS2 Subprogram . . . . .                      | 41          |
| 4.1.24 DPFD Subprogram . . . . .                        | 41          |
| 4.2 Labeled COMMON . . . . .                            | 42          |
| 4.3 Double Precision Variables . . . . .                | 42          |
| Figure III - Flow Chart of Program . . . . .            | 43          |
| Figure IV - Labeled COMMON Cross Reference Table .      | 44          |
| 4.4 Equivalenced Variables. . . . .                     | 45          |
| 4.5 Parameter Dependent Variables . . . . .             | 45          |
| 4.6 Hardware and System Dependent Variables . . . . .   | 48          |
| 4.7 Required Storage . . . . .                          | 49          |
| Acknowledgments . . . . .                               | 51          |
| References . . . . .                                    | 52          |
| Appendix A.1 -- Sample Input . . . . .                  | 53          |
| Appendix A.2 -- Sample Printed Output . . . . .         | 57          |
| Appendix A.3 -- Sample Process Control Output . . . . . | 94          |
| Appendix B -- Listing of Computer Program . . . . .     | 95          |



National Bureau of Standards Mass Calibration

Computer Software

R. N. Varner

R. C. Raybold

This report describes the FORTRAN computer program used to generate a comprehensive report covering the sequence of operations used to assign mass values to weights submitted to the National Bureau of Standards for calibration. The assignment of these values is accomplished by the method of least squares analysis of the observation of differences between test items and reference items having the same or nearly same density and nominal size. The calculations are defined and the various weighing method options are given. To assist the user, a detailed description of the input data, an input list of error messages, a listing of a sample test case and a listing of the output resulting from the use of the sample test case are given. To assist in the implementation of the computer program, a flow chart, a description of each subprogram, a cross-reference of labeled COMMON, a list of DOUBLE PRECISION variables, a list of EQUIVALENCED variables and other pertinent information is given.

Key words: Calibration report; correction to mass measurements; error checking; FORTRAN program; least squares solution; mass calibration; mass measurements.

## 1.0 Introduction

This report describes the FORTRAN computer program used to generate a comprehensive report covering the sequence of operations used to assign mass values to weights submitted to the National Bureau of Standards for calibration. The assignment of these values is accomplished by the least squares analysis of the observation of differences between test items and reference items having the same or nearly same density and nominal size. The calculations are defined and the various weighing method options are given. To assist the user, a detailed description of the input data, a list of error messages, a listing of a sample test case and a listing of the output resulting from the use of the sample test case are given. To assist in the implementation of the computer software, a flow chart, a description of each subprogram, a cross-reference of labeled COMMON, a list of DOUBLE PRECISION variables, a list of EQUIVALENCED variables and other pertinent information is given.

## 2.0 Calculations

This section describes in detail the calculations used by the computer software to assign mass values to weights which have been submitted for calibration. For a discussion of the philosophy of the procedures used, see references [1,2,6,8]. Section 2.1 lists the definition of the symbols used and their corresponding FORTRAN variable names. The use of  $| |$  in this section denotes the absolute value of the enclosed quantity. The use of  $[ ]$  denotes a matrix operation with the exception of mentioning of a reference. The use of  $'$  denotes the transpose of a matrix. Most, but not all, of the calculations described in this section are performed by the MAIN program.

The average temperature, humidity and pressure are determined from values taken before and after each set of measurements. Because the instruments used to measure these environmental conditions may change with each weighing, the appropriate instrumental correction to the temperature, humidity and pressure are given with each set of weighings as follows:

$$\bar{T} = (T_1 + C_{T1} + T_2 + C_{T2})/2$$

$$\bar{H} = (H_1 + C_{H1} + H_2 + C_{H2})/2$$

$$\bar{P} = (P_1 + C_{P1} + P_2 + C_{P2})/2$$

All input weights (nominals) are converted to grams if they are in pounds.

$$W_j = W_j \times 453.59237 \text{ for } j = 1, \dots, k$$

Accepted mass correction and the volume of the restraint are computed as follows:

If first series

$$C_R = [V'_1 \tau]$$

$$V_R = [V'_1 B]$$

$$\text{where } B_j = \frac{(W_j + .001\tau_j)(1 + \alpha_j \Delta t)}{\rho_j} \text{ for } j = 1, \dots, k$$

and where  $\Delta t = \bar{T} - T_0$ , the difference between observed average temperature and nominal temperature.

If not first series

$$C_R = R_c$$

$$V_R = V_s(1 + S_3 \Delta t)$$

### Calculations for Buoyancy Corrections

The air density using the averaged environmental conditions is computed from the equation given by Bowman and Schoonover [3].

$$\rho_A = \frac{0.464746}{T_K} \left( P - 0.00378029(H)(e_s) \right)$$

where

$\rho_A$  = density of air in mg/cm<sup>3</sup>

$T_K$  = temperature of air in Kelvin

= temperature of air in degrees Celsius + 273.15

H = Relative Humidity in percent  
(i.e. 20% = 20)

P = Barometric Pressure in mm of Mercury

$e_s$  = Saturation Vapor Pressure in mm of Mercury

=  $\text{Exp}(p)/(13.5951 \times 9.80665)$

$p = -4.7406885 \times \ln(T_K)$   
 $-6.8982434 \times 10^3 \times T_K^{-1}$   
 $+0.5938385 \times 10^2$   
 $-0.5797662 \times 10^{-2} \times T_K$   
 $+6.2223854 \times 10^{-6} \times T_K^2$

The above series expansion is based on an equation developed by  
Moxley and Greenspan [9]. Using the above equation three values of air  
density are computed:

$$\rho_a = f(\bar{T}, \bar{P}, \bar{H})$$

$$\rho_{ab} = f(T_1 + C_{T1}, P_1 + C_{P1}, H_1 + C_{H1})$$

$$\rho_{aa} = f(T_2 + C_{T2}, P_2 + C_{P2}, H_2 + C_{H2})$$

The air buoyancy correction is also applied to the sensitivity weight in computing the mass/division factor:

$$S^* = S_w - \rho_a S_v (1 + S_c \Delta t)$$

#### Types of weighings

The following calculations depend upon the type of weighing used. The program has provisions for the following six different types of weighing methods. The values of the input parameters N1 and N2 are designated, in parenthesis, for each method.

1. Single Substitution - one pan (N1 = 2, N2 = 1)
2. Single Substitution - two pan (N1 = 2, N2 = 0)
3. Single Transposition - two pan (N1 = 0, N2 = 0)
4. Double Substitution - one pan (N1 = 3, N2 = 1)
5. Double Substitution - two pan (N1 = 3, N2 = 0)
6. Double Transposition - two pan (N1 = 1, N2 = 0)

The weighing method used depends upon the type of balances and weights available, the requirements of the job at hand and/or the preference of the operator performing the calibrations. In all types of weighing, the observed difference, A-B, where A denotes the unknown weight and B denotes the standard or known weight, and the observed sensitivity, denoted by S, are computed in scale divisions. The drift effect and left-right effect may or may not be calculated depending upon the type of weighing being used. The following descriptions of each weighing method explain the pertinent calculations.

### Single Substitution - one pan balance

Two or three readings are made. The sensitivity weight is not always measured.

$$\begin{aligned}O_1 &= A \\O_2 &= B \\O_3 &= B+S \\(\hat{A}-B) &= O_1 - O_2 \\\hat{S} &= -O_2 + O_3 \quad \text{if } O_3 \neq 0 \\\hat{S} &= 0.0 \quad \text{if } O_3 = 0\end{aligned}$$

There is no drift or left-right effect computed.

### Single Substitution - two pan balance

There are six or nine readings made for each set of A and B weights.

The sensitivity weight is not always measured.

$$\begin{aligned}\left. \begin{aligned}O_1 &= A \\O_2 &= A \\O_3 &= A\end{aligned} \right\} & \begin{aligned} & 3 \text{ readings are taken as the balance} \\ & \text{is approaching a stable condition.}\end{aligned} \\\left. \begin{aligned}O_4 &= B \\O_5 &= B \\O_6 &= B\end{aligned} \right\} & \begin{aligned} & 3 \text{ readings are taken as the balance} \\ & \text{is approaching a stable condition.}\end{aligned} \\\left. \begin{aligned}O_7 &= B+S \\O_8 &= B+S \\O_9 &= B+S\end{aligned} \right\} & \begin{aligned} & 3 \text{ readings are taken as the balance is} \\ & \text{approaching a stable condition. These 3} \\ & \text{readings are sometimes omitted.}\end{aligned}\end{aligned}$$

$$\hat{A}-B = \frac{O_1 + 2O_2 + O_3}{4} - \frac{O_4 + 2O_5 + O_6}{4}$$

$$\hat{S} = \left| \frac{0_7 + 2x0_8 + 0_9}{4} - \frac{0_4 + 2x0_5 + 0_6}{4} \right| \text{ if } 0_7, 0_8 \text{ and } 0_9 \text{ are given}$$

$\hat{S} = 0.0$  if  $0_7, 0_8$  and  $0_9$  are not included. There is no drift of left-right effect computed.

#### Single Transposition - two pan balance

There are six or nine readings made for each set of A and B weights. If a sensitivity reading is not taken zeros must be used as input to the program.

- |                                                                  |                                                                                                                                                                        |
|------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| $\left. \begin{array}{l} 0_1 \\ 0_2 \\ 0_3 \end{array} \right\}$ | <p>With A on one pan and B on the opposite pan,<br/>three readings are taken as the balance<br/>approaches a stable condition.</p>                                     |
| $\left. \begin{array}{l} 0_4 \\ 0_5 \\ 0_6 \end{array} \right\}$ | <p>Weights A and B are interchanged, three readings<br/>are taken as the balance approaches a stable<br/>condition.</p>                                                |
| $\left. \begin{array}{l} 0_7 \\ 0_8 \\ 0_9 \end{array} \right\}$ | <p>The sensitivity weight is added to either pan and<br/>three readings are taken as the balance approaches a<br/>stable condition. These readings may be omitted.</p> |

$$(\hat{A}-\hat{B}) = \frac{1}{2} \left( \frac{0_1 + 2x0_2 + 0_3}{4} + \frac{0_4 + 2x0_5 + 0_6}{4} \right)$$

$$\hat{S} = \left| \frac{0_7 + 2x0_8 + 0_9}{4} - \frac{0_4 + 2x0_5 + 0_6}{4} \right| \text{ if } 0_7, 0_8 \text{ and } 0_9 \text{ are given}$$

$\hat{S} = 0.0$  if  $0_7, 0_8$  and  $0_9$  are not given

$$\hat{LR} = \frac{1}{2} \left( \frac{0_1 + 2x0_2 + 0_3}{4} + \frac{0_4 + 2x0_5 + 0_6}{4} \right)$$

There is no drift factor computed.

#### Double Substitution - one pan balance with drift

Four readings are made for each set of A and B weights where:

$$O_1 = A$$

$$O_2 = B$$

$$O_3 = B+S$$

$$O_4 = A+S$$

The least square solutions for  $(\hat{A}-\hat{B})$ ,  $\hat{S}$ ,  $\hat{\Delta}$  (drift) are:

$$(\hat{A}-\hat{B}) = \frac{O_1 - O_2 - O_3 + O_4}{2}$$

$$\hat{S} = \frac{O_1 - 3O_2 + 3O_3 - O_4}{2}$$

$$\hat{\Delta} = \frac{-O_1 + O_2 - O_3 + O_4}{2}$$

There is no left-right effect computed.

The value  $\hat{\Delta}$  is based on a  $1\Delta$  change in scale readings between each reading.

#### Double Substitution - two pan balance with drift

A total of 12 readings are made for each set of A and B weights where:

|                                                                 |                                                                                           |
|-----------------------------------------------------------------|-------------------------------------------------------------------------------------------|
| $\left. \begin{matrix} O_1 \\ O_2 \\ O_3 \end{matrix} \right\}$ | <p>Three readings are taken of weight A as the balance approaches a stable condition.</p> |
|-----------------------------------------------------------------|-------------------------------------------------------------------------------------------|



$\left. \begin{matrix} 0_4 \\ 0_5 \\ 0_6 \end{matrix} \right\}$  Three readings are taken of weight B as balance approaches a stable condition.

$\left. \begin{matrix} 0_7 \\ 0_8 \\ 0_9 \end{matrix} \right\}$  Three readings are taken of weight B+S as balance approaches a stable condition.

$\left. \begin{matrix} 0_{10} \\ 0_{11} \\ 0_{12} \end{matrix} \right\}$  Three readings are taken of weight A+S as balance approaches a stable condition.

$$(\hat{A}-\hat{B}) = \frac{1}{2} \left( \frac{0_1 + 2x0_2 + 0_3}{4} - \frac{0_4 + 2x0_5 + 0_6}{4} - \frac{0_7 + 2x0_8 + 0_9}{4} + \frac{0_{10} + 2x0_{11} + 0_{12}}{4} \right)$$

$$\hat{S} = \frac{1}{2} \left[ \frac{0_1 + 2x0_2 + 0_3}{4} - \frac{3(0_4 + 2x0_5 + 0_6)}{4} + \frac{3(0_7 + 2x0_8 + 0_9)}{4} - \frac{0_{10} + 2x0_{11} + 0_{12}}{4} \right]$$

$$\hat{\Delta} = \frac{1}{2} \left( -\frac{0_1 + 2x0_2 + 0_3}{4} + \frac{0_4 + 2x0_5 + 0_6}{4} - \frac{0_7 + 2x0_8 + 0_9}{4} + \frac{0_{10} + 2x0_{11} + 0_{12}}{4} \right)$$

The value  $\hat{\Delta}$  is based on a  $1\Delta$  change in scale readings between reading.

# Double Transposition - two pan balance with drift

A total of twelve readings are made for each set of A and B weights  
where:

$\left. \begin{array}{l} 0_1 \\ 0_2 \\ 0_3 \end{array} \right\}$  With A on one pan and B on the opposite pan,  
 three readings are taken as the balance approaches  
 a stable condition.

$\left. \begin{array}{l} 0_4 \\ 0_5 \\ 0_6 \end{array} \right\}$  With A and B interchanged, three readings are  
 taken as the balance approaches a stable  
 condition.

$\left. \begin{array}{l} 0_7 \\ 0_8 \\ 0_9 \end{array} \right\}$  With the sensitivity weight added to either A or B,  
 three readings are taken as the balance approaches  
 a stable condition.

$\left. \begin{array}{l} 0_{10} \\ 0_{11} \\ 0_{12} \end{array} \right\}$  With the sensitivity weight added to the opposite pan  
 from the position of the last 3 readings, 3 readings  
 are taken as the balance approaches a stable condition.

$$(\hat{A}-\hat{B}) = \frac{1}{4} \left( \frac{0_1 + 2 \times 0_2 + 0_3}{4} - \frac{0_4 + 2 \times 0_5 + 0_6}{4} - \frac{0_7 + 2 \times 0_8 + 0_9}{4} + \frac{0_{10} + 2 \times 0_{11} + 0_{12}}{4} \right)$$

$$\hat{S} = \frac{1}{2} \left| \frac{0_1 + 2 \times 0_2 + 0_3}{4} - \frac{3(0_4 + 2 \times 0_5 + 0_6)}{4} + \frac{3(0_7 + 2 \times 0_8 + 0_9)}{4} - \frac{0_{10} + 2 \times 0_{11} + 0_{12}}{4} \right|$$

$$\hat{\Delta} = \frac{1}{2} \left( - \frac{0_1 + 2x0_2 + 0_3}{4} + \frac{0_4 + 2x0_5 + 0_6}{4} - \frac{0_7 + 2x0_8 + 0_9}{4} + \frac{0_{10} + 2x0_{11} + 0_{12}}{4} \right)$$

$$\hat{LR} = \frac{1}{4} \left( \frac{3(0_1 + 2x0_2 + 0_3)}{4} + \frac{0_4 + 2x0_5 + 0_6}{4} + \frac{0_7 + 2x0_8 + 0_9}{4} - \frac{0_{10} + 2x0_{11} + 0_{12}}{4} \right)$$

The value  $\hat{\Delta}$  is based on a 1Δ change in scale readings between each reading.

#### Preparation of data for least squares analysis

The design matrix is used to compute the load  $L_i$  for each observation

$$L_i = \sum_{j=1}^k \frac{|X_{ji}|}{2} w_j \text{ for } i = 1, \dots, n$$

and the maximum load

$$L_{\max} = \max [L]$$

The values  $(\hat{A}-\hat{B})$ ,  $\hat{S}$  and  $\hat{\Delta}$  are converted from scale divisions to units. The following calculations apply to all types of weighings. the average of the observed sensitivity in scale divisions over all of observations which have a constant load is computed.

$$\begin{aligned} \text{a. } \text{SUM} &= \sum \hat{S}_i \text{ for } \hat{S}_i \neq 0 \text{ and until } L_i \text{ changes its value} \\ \hat{S}_i &= S^*/\hat{S}_i \end{aligned}$$

b.  $\bar{D} = \text{SUM}/j$

where  $j$  = number of non zero values in step a.

c.  $S_{\mu} = |S^*/\bar{D}|$

d.  $Y = \frac{(\hat{A}-B) S^*}{\bar{D}}$

e.  $\Lambda_{\mu} = \frac{\hat{\Delta} S^*}{\bar{D}}$

f. Change the sign of  $S_{\mu}$  and  $Y$  if input parameter (N4) indicates a reversed scale.

g. Flag values of

$$\left| \frac{(\hat{A}-B)}{\bar{D}\sqrt{j}} \right| \text{ which are } > .25$$

where  $(\hat{A}-B)$  is the observed deflection and  $\bar{D}\sqrt{j}$  is the sensitivity deflection.

h. Steps a. - g. are repeated for all observations.

Next, the restraint values for the series is computed. If it is the first series the following are computed:

a. Accepted Correction for restraint

$$R_c = [V'_1 \tau]$$

b. Restraint corrected for environmental condition

$$X^* = [V'_1] [\tau - p_a(B)]$$

where  $B$  is defined on page 2.

If not first series, the following are computed:

$$\rho_j = (W_j + .001 R_c) / V_s \text{ for } j=1, \dots, k \text{ and } V_{1j} \neq 0$$

$$\alpha_j = S_3 \text{ for } j=1, \dots, k \text{ and } V_{1j} \neq 0$$

$$X^* = R_c = \rho_a V_s (1 + S_3 \Delta t)$$

In general, let  $E(Y)$  denote the vector of expected values, let  $X$  denote the matrix of coefficients (the design matrix) and let  $\beta$  denote the column vector of parameters to be estimated. Then

$$E(Y) = [X][\beta].$$

Using the matrix of the coefficients from the equations of the expected values, a matrix representing the design and a vector,  $Y$ , for the observed values are set up.

The least squares estimates of the parameters are given by the solution of normal equations.

$$[X'X][\beta] = [X'] [Y]$$

where  $X'X$  is a  $k \times k$  matrix. To bring the equations to full rank the matrix of normal equations is augmented by a restraint vector denoted by  $V_1$ .

Thus:

$$[Z] = \begin{bmatrix} [X'X] & [V_1] & [X' Y] \\ [V_1'] & \lambda & X^* \\ 0 & 0 & -1 \end{bmatrix} \quad \begin{array}{l} \lambda = 1 \text{ if } [V_1] = 0 \\ \lambda = 0 \text{ if } [V_1] \neq 0 \end{array}$$

For more details on the structure of the above matrix see reference [7,10,11]. The inverse of the above matrix  $[Z]$  is calculated. Two are made to determine the success or failure of the inversion.

1. A check is made for singularity.
2. The maximum value of  $[I - ZZ^{-1}]$  should be  $\leq .01 \sigma_w$  where

$I$  is the identity matrix

$\sigma_w$  is the accepted within standard deviation of the process

$Z^{-1}$  is the inverse of  $Z$ .

If either of these tests fail an error message is printed and no further calculations will be made. Control will be sent to the subprogram PRINT2 and no final report will be printed.

If the matrix inversion is successful

$$[Z]^{-1} = \begin{bmatrix} [C] & \vdots & [\hat{\beta}] \\ \dots & \dots & \dots \end{bmatrix} \quad \begin{array}{l} \text{where: } C \text{ is the } k \times k \text{ matrix contain-} \\ \text{ing the elements of the inverse of} \\ \text{matrix of normal equations and } \hat{\beta} \\ \text{contains the parameter estimates} \end{array}$$

The deviations (observed-predicted) are computed. In matrix notation

$$[\delta] = [Y] - [X] [\hat{\beta}]$$

with dimensions  $n \times 1$ ,  $n \times 1$ ,  $n \times k$  and  $k \times 1$ .

The corrections using the estimated parameter values corrected by the air buoyancy correction are computed. An iterative process is used to minimize the error of the estimated values. The calculations are performed using double precision arithmetic.

$$C1_j = \hat{\beta}_j + \rho_a \left( \frac{W_j}{\rho_j} \right) (1 + \alpha_j \Delta t) \quad \text{for } j = 1, \dots, k$$

$$a. \quad C2_j = \hat{\beta}_j + \rho_a \left( \frac{W_j + .001 C1_j}{\rho_j} \right) (1 + \alpha_j \Delta t) \quad \text{for } j = 1, \dots, k$$

b. If all  $|C2_j - C1_j| < .01 \sigma_w$  continue calculations.

If the above condition is not true set

$$C1_j = C2_j \text{ for } j = 1, \dots, k$$

and repeat steps a. and b. above for a maximum of 10 iterations. A

message is printed regarding the number of iterations required.

Determine value of restraint in terms of nominal values.

$$W_R = [V_1' W]$$

Compute volume of unknowns in terms of the estimated correction produced by the above iteration.

$$V_j = \left( \frac{W_j + .001 C2_j}{\rho_j} \right) (1 + \alpha_j \Delta t) \text{ for } j = 1, \dots, k$$

#### Estimates of Uncertainty

Initially, the uncertainty for the starting restraint is entered in two parts: zero for the limit to possible effect of random error, and a value,  $E_s$ , for the uncertainty of the starting standards. All subsequent uncertainty values are based on two components - one due to the uncertainties in starting standards and the other due to the uncertainty from the balance and design used.

The uncertainty in an individual weight consists of the following components:

##### Systematic Error

$$E_j = \left( \frac{W_j}{W_R} \right) E_s \text{ for } j = 1, \dots, k$$

##### Random Error

$$R_j = \sqrt{(3\sigma_w)^2 C_{jj} + \left( \frac{W_j}{W_R} \right)^2 ((3\sigma)^2 + (3\sigma_T)^2)} \text{ for } j = 1, \dots, k$$

where

$\sigma_w$  is the accepted within standard deviation of the

$C_{jj}$  are the diagonal elements of the inverse of the normal equations

$3\sigma$  is random error affecting the restraint\*

$\sigma_T$  is the accepted between variance of the process

The total uncertainty is the sum of the systematic error and the random error.

$$U_j = E_j + R_j \text{ for } j = 1, \dots, k$$

For the outgoing restraint the corresponding values are

Systematic Error

$$E_o = [V_3' E]$$

Random Error

$$R_o = \sqrt{(3\sigma_w)^2 [V_3' C V_3] + \left( \frac{[V_3' W]}{W_R} \right)^2 (3\sigma)^2 + (3\sigma_T)^2}$$

If certain weights are to be used in combination, sets of linear combinations,  $V_5$ , may be specified. If this is the case the following calculations are made:

Correction to nominals in linear combination

$$C_\ell = [V_5' C_2]$$

Systematic error for linear combination

$$E_\ell = [V_5' E]$$

Random error for linear combinations

$$R_\ell = \sqrt{(3\sigma_w)^2 [V_5' C V_5] + \left( \frac{[V_5' W]}{W_R} \right)^2 (3\sigma)^2 + (3\sigma_T)^2}$$

Total uncertainty for linear combination

$$U_\ell = E_\ell + R_\ell$$

Initial restraint, the random error is zero.



Precision control is determined by computing the standard deviation

$$s = \sqrt{\frac{\sum \delta^2}{n-k+1}}$$

and comparing this with the accepted standard deviation,  $\sigma_w$ , by computing the F ratio

$$F = s^2 / \sigma_w^2$$

and comparing it with the critical value for the F distribution. The critical value for the F distribution is given by

$$F_t = \left( 1 - \frac{2}{9(n-k+1)} + 2.32635 \sqrt{\frac{2}{9(n-k+1)}} \right)^3$$

$$\text{for } (n-k+1) \geq 2$$

$$F_t = 6.64 \text{ for } (n-k+1) = 1$$

If  $F > F_t$ , a message is printed, on the last page of printout for a series, stating that the process is not in control.

The observed value for the check standard is given by

$$C_c = [V'_2 C_2]$$

The nominal value of the check standard is given by

$$[V'_2 W]$$

The accepted correction value for the check standard is given by

$$[V'_2 \tau]$$

The difference between the observed and the accepted value is divided by the standard deviation of the check standard to produce the t-ratio

value

$$T = \frac{[V_2' C_2] - [V_2' \tau]}{\sigma_c}$$

where

$$\sigma_c = \sqrt{(\sigma_w)^2 [V_2' C V_2] + \left( \frac{[V_2' W]}{W_R} \right)^2 \sigma^2 + (\sigma_T)^2}$$

This value is compared with  $T_t$  where

$$T_t = \frac{[V_2' E]}{\sigma_c}$$

If  $|T - T_t| \geq 3$  a message is printed on the last page of printout for a series stating that the check standard is not in control.

Compute values for the final summary of the report as requested by the input vector  $V_4$ .

Apparent mass versus brass in milligrams

$$M_a = \left( \frac{(W + .001 C_2) \left(1 - \frac{.0012}{\rho}\right)}{1 - \frac{.0012 (1 + .000054 \times 20)}{8.4}} - W \right) 1000$$

Mass of weight in grams

$$M_T = W + .001 C_2$$

Total uncertainty in grams

$$U_f = .001 U$$

Volume at 20°C

$$V_f = \frac{W + .001 C_2}{\rho}$$

Coefficient of expansion

$$C_f = \alpha$$

Apparent mass versus a standard having a density of 8.0 grams/cm<sup>3</sup> at 20°C

$$CR_f = \left( \frac{(W + .001 C_2) \left( \frac{1 - .0012}{\rho} \right)}{1 - \frac{.0012}{8.0}} - W \right) 1000$$

Set up values to be saved for next series if no errors were made and if another series is requested.

Systematic Error

$$E_s = E_o$$

Random Error

$$3\sigma = R_o$$

$$S_3 = \frac{[V'_3 S_2]}{[V'_3 S_1]}$$

where

$$S_1 = \frac{W + .001 C_2}{\rho}$$

$$S_2 = \alpha \left( \frac{W + .001 C_2}{\rho} \right)$$

Compute restraint for the next series.

$$R_c = [V'_3 C_2]$$

Compute volume of restraint for the next series.

$$V_s = [V'_3] \left( \frac{W_j + .001 C_{2j}}{\rho_j} \right) \text{ for } j = 1, \dots, n$$

## 2.1 Dictionary of Symbols and Corresponding FORTRAN Variable Names

The first column contains the symbol used by this documentation. The second column denotes input (I), output (O) or computed value (C). The third column gives the variable name used by computer program. The fourth column gives a brief description of the parameter. The symbols are listed in the order that they are used in the program.

| <u>Symbol</u> |     | <u>FORTRAN<br/>Name</u> | <u>Description</u>                                                  |
|---------------|-----|-------------------------|---------------------------------------------------------------------|
|               | I/O | B1                      | name or organization                                                |
|               | I/O | B2                      | address of organization                                             |
|               | I/O | B3                      | address of organization                                             |
|               | I/O | B4                      | description of weights being calibrated                             |
|               | I/O | B5                      | serial number of set of weights                                     |
|               | I/O | B6                      | date of report                                                      |
|               | I/O | B7                      | test folder number (used for NBS records of calibrations)           |
| $3\sigma$     | I   | RANERR                  | 3 standard deviation limit for random error affecting the restraint |
| $E_s$         | I   | SYSERR                  | Systematic error in the restraint                                   |
| $T_o$         | I   | TNOM                    | nominal temperature                                                 |
| $T_1$         | I   | T1P                     | temperature reading in Celsius at beginning of measurements         |
| $T_2$         | I   | T2P                     | temperature reading in Celsius at end of measurements               |
| $P_1$         | I   | P1P                     | pressure reading in mm of Mercury at beginning of measurements      |
| $P_2$         | I   | P2P                     | pressure reading in mm of Mercury at end of measurements            |
| $H_1$         | I   | H1P                     | humidity reading in percent at beginning of measurements            |

|                 |     |        |                                                          |
|-----------------|-----|--------|----------------------------------------------------------|
| H <sub>2</sub>  | I   | H2P    | humidity readings in percent at end of measurements      |
| C <sub>P1</sub> | I   | CP1    | correction to "before" pressure reading in mm of Mercury |
| C <sub>P2</sub> | I   | CP2    | correction to "after" pressure readings in mm of Mercury |
| C <sub>T1</sub> | I   | CT1    | correction to "before" temperature reading in Celsius    |
| C <sub>T2</sub> | I   | CT2    | correction to "after" pressure reading in Celsius        |
| C <sub>H1</sub> | I   | CH1    | correction to "before" humidity reading in percent       |
| C <sub>H2</sub> | I   | CH2    | correction to "after" humidity reading in percent        |
| $\bar{T}$       | C   | TBAR   | average corrected temperature in Celsius                 |
| $\bar{P}$       | C   | PBAR   | average corrected pressure in mm of Mercury              |
| $\bar{H}$       | C   | HBAR   | average corrected humidity in percent                    |
| n               | I   | NOBS   | number of observations $\leq 50$                         |
| k               | I   | NUNKN  | number of unknowns $\leq 15$                             |
| $\sigma_w$      | I   | STDEBA | accepted within standard deviation of process in mg.     |
| S <sub>w</sub>  | I   | SWT    | true mass value in mg. of sensitivity weight             |
| S <sub>v</sub>  | I   | VSWT   | volume of sensitivity weight in cm <sup>3</sup> at 20°C  |
| S <sub>c</sub>  | I   | CEXSWT | coefficient of expansion of sensitivity weight           |
| $\sigma_T$      | I   | VARBAL | accepted between standard deviation the process in mg.   |
| W               | I   | ANOM   | nominal value of weight in grams pounds                  |
| $\rho$          | I/C | DENSTY | density of weights in g/cm <sup>3</sup> at               |

|                  |     |        |                                                                                              |
|------------------|-----|--------|----------------------------------------------------------------------------------------------|
| $\alpha$         | I/C | COEFEX | coefficient of expansion of weight                                                           |
| $\tau$           | I   | ACCVAL | accepted correction of weight in mg.                                                         |
| $V_1$            | I   | ARSTIN | vector identifying items in the restraint                                                    |
| $V_2$            | I   | ACKSTD | vector identifying items in the check standard                                               |
| $V_3$            | I   | IRSTOU | vector identifying items to be used as restraint in the next series                          |
| $V_4$            | I   | INPRNT | print vector                                                                                 |
| $X$              | I   | DESMAT | design matrix                                                                                |
| $V_5$            | I   | ALCOM  | linear combination vector                                                                    |
| $O$              | I   | OBSERV | observations                                                                                 |
| $\Delta t$       | C   | TDEL   | difference between observed temperature and nominal temperature                              |
| $C_R$            | C   | CORR   | mass correction for restraint in mg.                                                         |
| $V_R$            | C   | VOLRES | volume of weights in restraint in $\text{cm}^3$                                              |
| $B$              | C   | TEMP   | volume of weights                                                                            |
| $T_K$            | C   | TKEL   | temperature in degrees Kelvin                                                                |
| $\rho_a$         | C   | RHOA   | air density for average environment conditions                                               |
| $\rho_{ab}$      | C   | RHOAB  | air density for "before" environment conditions                                              |
| $\rho_{aa}$      | C   | RHOAA  | air density for "after" environment conditions                                               |
| $L$              | C   | ILOAD  | vector of loads                                                                              |
| $L_{\text{max}}$ | C   | ALOADP | maximum load                                                                                 |
| $S$              | C   | STAR   | the mass of the sensitivity weight with air buoyancy correction applied                      |
| $(A-B)$          | C   | D1     | predicted difference A-B in scale divisions where A is the unknown weight and B the standard |
|                  | C   | DS1    | predicted sensitivity in scale divisions                                                     |

|                |   |        |                                                                               |
|----------------|---|--------|-------------------------------------------------------------------------------|
| $\hat{\Delta}$ | C | DRIFT  | predicted drift in scale divisions                                            |
| $\hat{L}R$     | C | ZERO   | predicted left-right effect in scale division                                 |
| $\overline{D}$ | C | DBAR   | average of observed sensitivities of load of equal size                       |
| $S_{mu}$       | C | SWTPRT | average sensitivity in mass units (mg/division)                               |
| Y              | C | A      | observed (A-B) in mass units where A is the unknown weight and B the standard |
| $\Delta_{mu}$  | C | DRIFT  | drift in mass units                                                           |
| $X^*$          | C | XREST  | accepted mass correction for restraint in mg.                                 |
| $R_c$          | C | TMSUM  | computed restraint correction in mg. for the m+1 series                       |
| Z              | C | Y      | matrix of normal equations                                                    |
| C              | C | Y      | diagonal elements of the inverse of the matrix of normal equations            |
| $\delta$       | C | DELTA  | deviation between observed and predict weight values                          |
| C1             | C | CORRP  | estimated correction to the nominal in grams                                  |
| C2             | C | OBSCOR | observed correction after iteration                                           |
| $\beta$        | C | Y      | estimated values of unknown from $Z^{-1}$                                     |
| $W_R$          | C | WR     | nominal weight of restraint                                                   |
| V              | C | COMVOP | computed volume of unknown using estimated corrections                        |
| E              | C | SERROR | systematic error for each weight                                              |
| R              | C | TRISIC | random error for each weight                                                  |
| U              | C | TOTUN  | total uncertainty of each weight                                              |
| $V_s$          | C | SOLSUM | computed value of volume for m+1 series                                       |
| $C_\ell$       | C | CORR5A | corrections for linear combination                                            |
| $E_\ell$       | C | SER5A  | systematic error for linear combination                                       |

|            |   |        |                                                          |
|------------|---|--------|----------------------------------------------------------|
| $R_L$      | C | SIG35A | random error for linear combination                      |
| $U_L$      | C | UNC5A  | total uncertainty for linear combinations                |
| $E_O$      | C | SERSUM | systematic error for outgoing restraint                  |
| $R_O$      | C | T3SIG  | random error for outgoing restraint                      |
| s          | C | OBSTD  | observed standard deviation of the series                |
| F          | C | FRATIO | f - test ratio                                           |
| $F_t$      | C | PRETST | critical value for the F distribution                    |
| $C_c$      | C | OBCOCK | observed check standard                                  |
| $\sigma_c$ | C | OBSCK  | standard deviation of observed check standard            |
| T          | C | TVAL   | t - test value                                           |
| $M_a$      | C | APPMAS | apparent mass verses brass for final output              |
| $M_T$      | C | TRMASS | mass of weight in grams or pounds for final output       |
| $U_f$      | C | UNCERT | total uncertainty in grams or pounds for final output    |
| $V_f$      | C | VOLPRT | volume at 20°C for final output                          |
| $C_f$      | C | COEPR1 | coefficient of expansion for final output                |
| $CR_f$     | C | CORRB  | apparent mass verses density 8.0 for final output        |
| $S_3$      | C | SUMP   | value used to compute volume of restraint for m+1 series |

### 3.0 Information For Users

This section contains information pertinent to the user of the

#### Description of Input

The input requirements of the computer program are defined in this



section. The data entry column defines each unique data item, not each input record. The data, with a few designated exceptions where a format is given, is recorded in a free field format. The subprogram name and the FORTRAN variable name for each data entry is given along with a brief description of the input parameter. See Appendix A.1 for a listing of sample data.

| <u>DATA ENTRY</u> | <u>FORMAT</u> | <u>SUBPROGRAM</u> | <u>VARIABLE NAMES AND DESCRIPTION</u>                                                                                                                                                                                                                                                                                                                                      |
|-------------------|---------------|-------------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| #1*               | 72A1          | READ1             | B1 - name of company submitting test weight                                                                                                                                                                                                                                                                                                                                |
| #2*               | 72A1          | READ1             | B2 - address (street name and number)                                                                                                                                                                                                                                                                                                                                      |
| #3*               | 72A1          | READ1             | B3 - address (city, state and zip code)                                                                                                                                                                                                                                                                                                                                    |
| #4*               | 72A1          | READ1             | B4 - description of weights to be calibrated                                                                                                                                                                                                                                                                                                                               |
| #5*               | 72A1          | READ1             | B5 - serial number of set of weights                                                                                                                                                                                                                                                                                                                                       |
| #6*               | 72A1          | READ1             | B6 - date of report                                                                                                                                                                                                                                                                                                                                                        |
| #7**              | 72A1          | READ1             | B7 - test folder number (used for NBS records of calibrations)                                                                                                                                                                                                                                                                                                             |
| #8                | Variable      | READ1             | RANERR, SYSERR, TNOM, IBREST<br><br>RANERR - 3 standard deviation limit for random error in the starting restraint<br><br>SYSERR - limit to possible systematic error in the starting restraint<br><br>TNOM - nominal temperature at which apparent mass and volume are reported in degrees Celsius<br><br>***IBREST - starting restraint identification number (2 digits) |

\* Data entries #1-17 are searched for the first non-blank character which time all remaining characters are saved for printing on the document, i.e. leading blanks are eliminated.

\*\* On output, only the first 18 non-blank characters are printed.

\*\*\* See Figure I.

| <u>ENTRY</u> | <u>FORMAT</u> | <u>SUBPROGRAM</u> | <u>VARIABLES NAMES AND DESCRIPTION</u> (continued) |
|--------------|---------------|-------------------|----------------------------------------------------|
|--------------|---------------|-------------------|----------------------------------------------------|

```

#9* Variable READ2 N1, N2, N3, N4 (describe weighing method)

N1=0 single transposition
N1=1 double transposition
N1=2 single transposition
N1=3 double substitution
N2=0 two pan balance
N2=1 one pan balance
N3=0 metric units (grams or milligrams)
N3=1 English units (pounds)
N4=0 regular balance (scale left to right)
N4=1 backwards balance (scale right to left)

#10 Variable READ2 (IDATEi, i=1,2,3), IOP, IBAL, ICKUSD

IDATE1 - month (2 digits)
IDATE2 - day (2 digits)
IDATE3 - year (2 digits)
} date of measurement

IOP - operator number (2 digits)

IBAL - balance number (3 digits)

ICKUSD - check standard identification (3
        digits)

#11 Variable READ2 T1P, T2P, P1P, P2P, H1P, H2P, CP1, CP2, CT1,
CT2, CH1, CH2

T1P, T2P - observed temperature in degrees
           Celsius before and after
           measurements are taken

P1P, P2P - observed pressure in mm of Mercury
           before and after measurements are
           taken

H1P, H2P - observed humidity in percent before
           and after measurements are taken

```

begin reading at this point for each new series.

| <u>DATA</u>  |               |                   |                                                                                                                        |  |
|--------------|---------------|-------------------|------------------------------------------------------------------------------------------------------------------------|--|
| <u>ENTRY</u> | <u>FORMAT</u> | <u>SUBPROGRAM</u> | <u>VARIABLE NAMES AND DESCRIPTION</u> (Continued)                                                                      |  |
|              |               |                   | CP1, CP2 - pressure corrections in mm of Mercury for observed pressure before and after measurements are taken         |  |
|              |               |                   | CT1, CP2 - temperature corrections in degrees Celsius for observed temperature before and after measurements are taken |  |
|              |               |                   | CH1, CH2 - humidity corrections in percent for observed humidity before and after measurements are taken               |  |
| #12          | Variable      | READ2             | NOBS, NUNKN, ICALDS, LINVAR                                                                                            |  |
|              |               |                   | NOBS - number of observations $\leq 50$                                                                                |  |
|              |               |                   | NUNKN - number of unknowns $\leq 15$                                                                                   |  |
|              |               |                   | *ICALDS - calibration design number (3 digits)                                                                         |  |
|              |               |                   | LINVAR - number of linear combinations $\leq 19$                                                                       |  |
| #13          | Variable      | READ2             | STDEBA, SWT, VSWT, CEXSWT, VARBAL                                                                                      |  |
|              |               |                   | STDEBA - accepted within standard deviation of the process                                                             |  |
|              |               |                   | SWT - mass mass value in mg. of the sensitivity weight                                                                 |  |
|              |               |                   | VSWT - volume of sensitivity weight in cm. at 20 degrees Celsius                                                       |  |
|              |               |                   | CEXSWT - coefficient of expansion of sensitivity weight                                                                |  |
|              |               |                   | VARBAL - accepted between standard deviation of the process in mg.                                                     |  |

---

\* See Figure II.

| DATA ENTRY | FORMAT                                                                                  | SUBPROGRAM | VARIABLE NAMES AND DESCRIPTION (Continued)                                                                                                                                                                                                                                                                                                                                                                                                                                                                               |
|------------|-----------------------------------------------------------------------------------------|------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| #14*       | Variable<br>(Exception-<br>the first 15<br>characters<br>are for the<br>identification) | READ2      | <p>AIDCST<sub>j,i</sub>, j=1,...,5, ANOM<sub>i</sub>, DENSTY<sub>i</sub>,<br/>COEFEX<sub>i</sub>, ACCVAL<sub>i</sub>, i=1,...,NUNKN</p> <p>AIDCST<sub>j,i</sub>, j=1,...,5 - weight identification in<br/>positions 1-15 of the input record</p> <p>ANOM<sub>i</sub> - nominal weight in grams or pounds</p> <p>DENSTY<sub>i</sub> - density of weight in g/cm<sup>3</sup> at 20°C</p> <p>COEFEX<sub>i</sub> - coefficient of expansion of weight</p> <p>**ACCVAL<sub>i</sub> - accepted correction to weight in mg.</p> |
| 15         | Variable                                                                                | READ2      | <p>ARSTIN<sub>i</sub>, i=1,...,NUNKN</p> <p>Vector identifying items in the restraint.<br/>Entries may be 0 or 1 only.</p>                                                                                                                                                                                                                                                                                                                                                                                               |
| 16         | Variable                                                                                | READ2      | <p>ACKSTD<sub>i</sub>, i=1,...,NUNKN</p> <p>Vector identifying items in the check,<br/>standard. Entries may be 0, -1, or 1.</p>                                                                                                                                                                                                                                                                                                                                                                                         |
| 17         | Variable                                                                                | READ2      | <p>IRSTOU<sub>i</sub>, i=1,...,NUNKN</p> <p>Vector identifying items in restraint for the<br/>m+1 series. This vector has entries of 0 or 1.<br/>If there is not another series, entries are all<br/>zero.</p>                                                                                                                                                                                                                                                                                                           |
| 18         | Variable                                                                                | READ2      | <p>IPRNT<sub>i</sub>, i=1,...,NUNKN</p> <p>Vector identifying items to be reported in the<br/>summary. A value of 1 means report and 0 means<br/>omit from report.</p>                                                                                                                                                                                                                                                                                                                                                   |
| 19***      | Variable                                                                                | READ2      | <p>****DESMAT<sub>j,i</sub>, j=1,...,NUNKN, i=1,...,NOBS</p> <p>The design matrix consists of entries of 0, 1,<br/>or -1 for the series. Each data item contains<br/>NUNKN values. Repeat NOBS times.</p>                                                                                                                                                                                                                                                                                                                |

Repeat this data entry for each unknown.  
 \*\* This value is always given in mg., even if the nominal is in pounds.  
 \*\*\* Repeat this data entry for each observation.  
 \*\*\*\* The design matrix defines the method which is being used to group and  
 intercompare the unknown weights and the check standard (the known weight).

DATA

ENTRY   FORMAT   SUBPROGRAM   VARIABLE NAMES AND DESCRIPTION   (Continued)

#20   Variable   READ2    $ALCOM_{j,i}$ ,  $j=1,\dots,NUNKN$ ;  $i=1,\dots,LINVAR$ . If  $LINVAR \neq 0$

Enter  $LINVAR$  vectors containing  $NUNKN$  values per vector which describe the required linear combinations. The values in each vector consists 0, 1 or -1.

#21   Variable   READ2    $OBSERV_k$   $k=1,\dots,kk$  where  $kk \leq 600$

Provide readings in scale divisions corresponding to the design and type of weighing as indicated by data entries #9 and #19. The following combinations of  $N1$  and  $N2$  determine the number of entries per record for each observation. If for any reason a reading is not taken, a zero must be used to so indicate.

- a.  $N1=2$  and  $N2=1$  Enter  $0_1, 0_2, 0_3$ ; 2 or 3 values per record
- b.  $N1=2$  and  $N2=0$  Enter  $0_1, 0_2, 0_3, 0_4, 0_5, 0_6, 0_7, 0_8, 0_9$ ; 6 or 9 values per record
- c.  $N1=0$  and  $N2=0$  Enter  $0_1, 0_2, 0_3, 0_4, 0_5, 0_6, 0_7, 0_8, 0_9$ ; 6 or 9 values per record
- d.  $N1=3$  and  $N2=1$  Enter  $0_1, 0_2, 0_3, 0_4$ ; 4 values per record
- e.  $N1=3$  and  $N2=0$  Enter  $0_1, 0_2, 0_3, 0_4, 0_5, 0_6$ ; 6 values per record (2 records  
 $0_7, 0_8, 0_9, 0_{10}, 0_{11}, 0_{12}$
- f.  $N1=1$  and  $N2=0$  Enter same as e. above.

Any other combinations of  $N1$  and  $N2$  assumes f. as defined above.

#22   Variable   READ2   A value (-20000) terminates the reading of observations. It is the responsibility of user to be sure that the number of observations corresponds to the number required by the specified schedule given in data entry #19.

If the vector described in data entry #17 is not equal to zero, continue input of data repeating from data entry #9. If the vector is zero, it terminates the input of data. This flag must appear in position 1 of the input record.

Figure I. Starting Restraint Identification Used at NBS

Given below is an example of frequently used restraint identifications used at NBS. The symbol g denotes grams and kg kilograms.

| <u>Restraint Identification</u> | <u>Weights Used in the Restraint</u>              |
|---------------------------------|---------------------------------------------------|
| 1                               | $N \text{ kg}_1 + N \text{ kg}_2$                 |
| 2                               | $\text{NB}^1 100\text{g} + \text{AA} 100\text{g}$ |
| 4                               | $\text{NB}^1 1\text{g}$                           |

Figure II. Calibration Design Number Used at NBS

Given below is an example of frequently used calibration identifications with the design, the number of weight and the required number of observations for each.

| <u>Design Identification</u> | <u>Design</u> | <u>No. of weights</u> | <u>No. of observations</u> |
|------------------------------|---------------|-----------------------|----------------------------|
| 1                            | 1,1,1         | 3                     | 3                          |
| 16                           | 5,2,2,1,1,1   | 6                     | 8                          |
| 41                           | 1,1,1,1       | 4                     | 6                          |
| 51                           | 1,1,1,1,1     | 5                     | 10                         |
| 53                           | 5,3,2,1,1     | 5                     | 8                          |
| 62                           | 5,3,2,1,1,1   | 6                     | 11                         |

### 3.2 Diagnostic and Error Messages

This section lists all the possible error messages and other informative messages concerning the statistical tests made by the program. The subprogram which contains the message and an explanation of the message is given. The message is given in quotes and an indication (i.e. ---) is given if some computed quantity is also printed.

#### A. "NEG SQRT ARG = ---"

This message is printed by the MAIN program if the value under the radical is negative in the computation of the random error for a weight. If the value under the radical is negative it is assumed to be zero.

B. The following six messages appear in the subprogram ERROR. They are error messages resulting from the execution of subprogram READIT and indicate that an input value is too large or too small for the capacity of the computer being used.

- (1) "\*\*\*\*\* DIAGNOTIC \*\*\*\*\* THE NUMBER OF SIGNIFICANT DIGITS IN A NUMBER HAS REACHED ---. THIS MAY PRODUCE OVERFLOW OR UNDERFLOW"
- (2) "\*\*\*\*\* ERROR \*\*\*\*\* THE NUMBER OF SIGNIFICANT DIGITS IN A NUMBER HAS REACHED ---. THIS WILL PRODUCE OVERFLOW OR UNDERFLOW."
- (3) "\*\*\*\*\* ERROR \*\*\*\*\* NUMBER IS TOO SMALL IN ABSOLUTE VALUE AND WILL PRODUCE UNDERFLOW."
- (4) "\*\*\*\*\* DIAGNOSTIC \*\*\*\*\* NUMBER IS SMALL IN ABSOLUTE VALUE AND PRODUCE UNDERFLOW."
- (5) "\*\*\*\*\* DIAGNOSTIC \*\*\*\*\* NUMBER IS LARGE IN ABSOLUTE VALUE AND PRODUCE OVERFLOW."

(6) "\*\*\*\*\* ERROR \*\*\*\*\* NUMBER IS TOO LARGE IN ABSOLUTE VALUE AND  
WILL PRODUCE OVERFLOW."

C. The following error message which occurs in the subprogram ERROR is printed after each of the six messages stated above. It prints out the data item containing the invalid data value.

"THIS OCCURRED IN CONNECTION WITH READING THE DATA ON THE FOLLOWING  
CARD."

D. The subprogram READIT has an option of specifying an alphanumeric value at the beginning of a data item. If this option does not specify the proper number of characters in the alphanumeric value the following message is printed.

"\*\*\*\*\* ERROR \*\*\*\*\* THE VALUE OF 'KOL' IS --- AND THIS VALUE IS  
INVALID. KOL MUST BE GREATER THAN 0 AND MUST NOT EXCEED 80."

E. If there are problems in the matrix inversion procedure,  
one of the following two messages is printed:

(1) "MATRIX IS SINGULAR"

(2) "ERROR IN INVERSE".

If the first message occurs, it indicates that there is some problem with  
the input data. If the second message is printed it means that the  
condition

$$\max [I-AA^{-1}] \leq .01\sigma_w$$

is not met. In addition to the message; the original matrix, the  
inverted matrix and the  $[I-AA^{-1}]$  matrix is printed. Both of these  
messages are printed by the subprogram PRINT2. After the printing of  
the messages the execution of the program is aborted.



F. "STOPPED AT 10 ITERATIONS"

This message is printed by subprogram PRINT2 and indicates that the iterative process used to compute the observed correction (see page 14 of calculations) was terminated at 10 iterations.

G. "INPUT ERROR IN RESTRAINT. CHECK RESTRAINT VECTOR, NOMINAL VALUE, DENSITY AND COEFFICIENT OF EXPANSION OF RESTRAINT --- MG COMPUTED CORRECTION OF RESTRAINT --- MG"

This diagnostic message is printed by subprogram PRINT2 when the following test fails.

$$|[V_1' \tau] \text{ (accepted restraint)} - [V_1' C_2] \text{ (computed restraint)}| < .1\sigma_w$$

H. One of the following three diagnostic messages concerning the interpretation of the t-test is printed by subprogram PRINT2. See page 17 for calculation of the t-test.

(1) "ABSOLUTE VALUE OF T IS LESS THAN 3. THEREFORE CHECK STANDARD IN CONTROL."

$$|T| < 3.0$$

(2) "ALTHOUGH THE ABSOLUTE VALUE OF T IS GREATER THAN OR EQUAL TO 3, THE T VALUE CORRECTED FOR SYSTEMATIC ERROR IS LESS THAN 3, THEREFORE THE CHECK STANDARD IS IN CONTROL."

$$|T| \geq 3 \text{ AND } (T - T_t) < 3.0$$

(3) "ALTHOUGH THE ABSOLUTE VALUE OF T IS GREATER THAN OR EQUAL TO 3, THE DIFFERENCE IS STILL SIGNIFICANT AFTER ALLOWANCE FOR SYSTEMATIC ERROR, THEREFORE THE CHECK STANDARD IS NOT IN CONTROL."

$$|T| \geq 3 \text{ AND } (T - T_t) \geq 3.0$$

I. One of the following two diagnostic messages concerning the interpretation of the F-test is printed by subprogram PRINT2. The critical value is printed in the space denoted by --. See page 16 for the calculations of the F ratio.

- (1) "F RATIO IS LESS THAN---(CRITICAL VALUE FOR PROBABILITY = .01).  
THEREFORE THE STANDARD DEVIATION IS IN CONTROL."

$$F \leq F_t$$

- (2) "F RATIO IS GREATER THAN---(CRITICAL VALUE FOR PROBABILITY = .01).  
THEREFORE THE STANDARD DEVIATION IS NOT IN CONTROL."

$$F > F_t$$

### 3.3 Description of Data Output Used for Process Control

If neither the t-test or F-test fails, selected values are saved on a unformatted temporary file during the execution of the program. The temporary file is defined by the variable ITMP in the BLKDAT subprogram. See section 4.6. The subprogram FINPRT reads the temporary file, ITMP, and generates a formatted file (IP as defined by subprogram BLKDAT) of the saved parameters.

The table below defines the parameters with their corresponding format. All parameters are contained in an 80 character record. Appendix A.3 lists the process parameters saved from the sample run given in Appendix A.2.

| <u>Variable</u>                                            | <u>Description</u>                                             | <u>Format</u> |
|------------------------------------------------------------|----------------------------------------------------------------|---------------|
| IDATE <sub>1</sub> IDATE <sub>2</sub> , IDATE <sub>3</sub> | date                                                           | 3I2           |
| IBREST                                                     | restraint identification                                       | I2            |
| ICKUSD                                                     | check standard identification                                  | I3            |
| OBCOCK                                                     | observed check standard value                                  | F11.5         |
| IBAL                                                       | balance identification                                         | I3            |
| OBSTD                                                      | observed standard deviation                                    | F9.5          |
| NDGFR                                                      | degrees of freedom                                             | I2            |
| ICALDS                                                     | calibration identification                                     | I3            |
| TBAR                                                       | average corrected temperature                                  | F5.2          |
| DIFT                                                       | difference between "before" and "after"<br>temperature reading | F5.2          |
| PBAR                                                       | average corrected pressure                                     | F6.2          |
| DIFP                                                       | difference between "before" and "after"<br>pressure reading    | F5.2          |
| HBAR                                                       | average corrected humidity                                     | F7.2          |
| DIFH                                                       | difference between "before" and "after"<br>humidity reading    | F4.2          |
| RHOA                                                       | air density as a function of TBAR, PBAR<br>and HBAR            | F6.2          |
| IOP                                                        | operator                                                       | I2            |
|                                                            | denotes standard                                               | I2            |

#### 4.0 Implementation Information

This section describes the information needed for implementing FORTRAN computer software. Information is given concerning the flow function of the MAIN program and all subprograms. In addition, a reference table of labeled COMMON; the use of the DATA, DOUBLE-PRECISION

and EQUIVALENCE statements; the function of switch variables in the various subprogram and other information which may be bothersome on implementation are given. Figure III on page 43 gives a flow chart of the program. Figure IV on page 44 gives a cross reference table of labeled COMMON.

#### 4.1 Description of MAIN program and all Subprograms

The software consists of one main program and 23 subprograms. This section describes briefly the function of each. The subprogram descriptions are listed in the order in which they are called during the execution of the program.

##### 4.1.01 MAIN Program

This program controls the flow of the input, calculations and output. All the calculations described in Section 2.0 are performed in this program with the exception of a few computations which are performed in subprograms PRINT2 and READ2.

##### 4.1.02 BLKDAT Subprogram

This subprogram is a BLOCK DATA subprogram and contains values which may need to be changed to comply with the demands of a specific computer or computer operating system. The DATA statements define the following values:

- a. Machine zero  $1 \times 10^{-8}$  (UNIVAC 1108)
- b. Characters: 0-9, blank, -, ., \*, +, comma, D and E
- c. Input unit number, output unit numbers and number of lines per page
- d. Flag STOP to detect end of data and blank

The number 10 which controls the number of iterations in the MAIN

See page 13.

#### 4.1.03 READ1 Subprogram

This subprogram reads data that is common to all series. Eight data entries are read consisting of administrative data, statistical control parameters, nominal temperature and the starting restraint identification. The first seven data entries consisting of administrative information, company name and address and description of weights being tested, are read with a 72A1 format specification. The information may occur anywhere within the 72 position limit. Leading blanks are eliminated before the information is printed on the report. The eighth data entry gives the values for random error, systematic error, nominal temperature and the starting restraint identification. The values are read by a subprogram, READIT, which permits input in a variable format. Four values are assumed to be given and no check is made for missing values.

#### 4.1.04 READIT Subprogram

This subprogram provides for input in variable format and is used by subprograms READ1 and READ2. The input data is restricted to first 80 positions of the input record. Alphanumeric data may be given in the first n, where n is specified, positions. These characters are saved for output. Numeric values are separated by one or more blanks, a comma, any letter except D or E, or any other permitted character except a plus sign (+), a minus sign (-) or a decimal (.). Numeric values may appear in integer form or floating point form using a decimal point or an exponent in which case the letter D or E must precede the exponent. Values with a D preceding the exponent are accepted only as single precision values, not as double precision values.

#### 4.1.05 ERROR Subprogram

This subprogram is used in conjunction with subprogram READIT described in section 4.1.04. It contains the output statements and corresponding formats for the printing of errors associated with the subprogram READIT's interpretation of meaningless input data.

#### 4.1.06 PRINT1 Subprogram

This subprogram prints the title page and pages 1, 2, 3, 4, 5 and 6 of the report generated for each calibration. This information is pertinent only to the NBS calibration program.

#### 4.1.07 TEXT1 Subprogram

This subprogram contains the output statements and their corresponding formats for the printing of page 1 of the calibration report.

#### 4.1.08 TEXT2 Subprogram

This subprogram contains the output statements and their corresponding format for the printing of page 2 of the calibration report.

#### 4.1.09 TEXT3 Subprogram

This subprogram contains the output statements and their corresponding formats for the printing of page 3 of the calibration report.

#### 4.1.10 TEXT4 Subprogram

This subprogram contains the output statement and their corresponding formats for the printing of page 4 of the calibration report.

#### 4.1.11 TEXT5 Subprogram

This subprogram contains the output statements and their corresponding formats for the printing of page 5 of the calibration report.

#### 4.1.12 TEXT6 Subprogram

This subprogram contains the output statements and their corresponding formats for the printing of page 6 of the calibration report.

#### 4.1.13 READ2 Subprogram

This subprogram uses the subprogram READIT, described in section 4.1.04, to read the following information which is needed for each series.

- a. Description of the weighing method
- b. Administrative data--date, operator, balance identification and check standard identification
- c. Environmental conditions--temperature, pressure, humidity
- d. Number of observations, number of unknowns, design identification and number of linear combinations
- e. Standard deviation of balance, mass of sensitivity weight, volume of sensitivity weight, coefficient of expansion of sensitivity weight, accepted between standard deviation of the process
- f. Information about test item: weight, density, coefficient of expansion and accepted correction
- g. Restraint vector
- h. Check standard vector
- i. Restraint vector for next series
- j. Report vector
- k. Design matrix
- l. Linear combination vector(s)

- m. Observations--Reading of observations is terminated by a -20000 entry.

Information given in pounds is converted to grams. A control parameter designated by the weighing method parameter is defined. The average values for the corrected temperature, pressure and humidity are computed.

#### 4.1.14 SPINV Subprogram

This subprogram inverts the augmented matrix of normal equations. A check is made for singularity. A call is made to subprogram INVCHK for the purpose of checking the success or failure of the inversion.

#### 4.1.15 SAVMTX Subprogram

This subprogram is used by subprogram SPINV described in 4.1.14. The original matrix is saved before the inverse operation of subprogram SPINV begins.

#### 4.1.16 INVCHK Subprogram

This subprogram called by subprogram SPINV makes a check on the success of the matrix inversion. The check  $[I - AA^{-1}] < E$  is made where  $I$  is the identity matrix,  $A$  is the original matrix,  $A^{-1}$  is the inverse of  $A$  and  $E$  is the value  $.01\sigma_w$  where  $\sigma_w$  is the accepted standard deviation of the balance. If the conditions are not met, a flag is defined.

#### 4.1.17 PRINT2 Subprogram

This subprogram makes necessary calculations and contains output statements and their corresponding formats for the printing of the four or five pages of the report associated with each series. The first page of the output contains administrative information: statistical control



values, restraint information, check standard information, test environmental conditions and description of weights being calibrated. The second page of the output contains the design and the observations. The third page of the output contains the computed values for the corrections and their corresponding uncertainties and pertinent information of the restraint for the next series. The fourth or fifth page of the output contains the information concerning the statistical F-test and t-test. If neither the t-test or F-test fails, values are saved for process control. See section 3.3. The fourth page, if linear combinations are requested, contains information related to the requested linear combinations.

#### 4.1.18 CHKLN Subprogram

This subprogram makes a check between the current number of lines on a page of printout and the parameter controlling the number of lines permitted per page. If the maximum is exceeded subprogram PGCONT is called.

#### 4.1.19 PGCONT Subprogram

This subprogram writes the information needed in the case where a continuation page is required due to output page overflow.

#### 4.1.20 HEADPG Subprogram

This subprogram writes the headings on each page. The heading includes the company name and address, a description of the weights tested, the balance, the date, the page number and the series number.

#### 4.1.21 FINPRT Subprogram

This subprogram controls the printing of the four summary pages for each calibration report. If mass was given in English units, the values are converted to grams for the output. The reported corrections are printed in milligrams. The subprogram TEXTS1 and TEXTS2 are called to print the summary text. A subprogram called DPFDD is used to print double precision values of mass and corrections in fixed notations (see Table I and Table II of the output example).

#### 4.1.22 TEXTS1 Subprogram

This subprogram contains the output statements and their corresponding formats for printing the text of the first page of the summary.

#### 4.1.23 TEXTS2 Subprogram

This subprogram contains the output statements and their corresponding formats for printing the text of the second page of the summary.

#### 4.1.24 DPFDD Subprogram

This subprogram converts a double precision value to a string of numeric characters representing its values to be printed as a fixed floating point data type with more than 8 accurate digits, the number permitted on the UNIVAC 1108.

#### 4.2 Labeled COMMON

Figure IV lists all the labeled COMMON areas and the main program and all its subprograms. Check marks indicate which subprograms use which labeled COMMON areas. The numbers in parentheses indicate the amount of storage required.

#### 4.3 Double Precision Variables

The following table defines the subprograms using double precision variables, the variable names and in which labeled COMMON area they appear. N.A. means not applicable.

| <u>Variable</u> | <u>Subprogram</u> | <u>COMMON</u> |
|-----------------|-------------------|---------------|
| VOLP            | MAIN              | N.A.          |
| CORRP           | MAIN              | N.A.          |
| OBSCOR          | MAIN, PRINT2      | COMPUT        |
| TRMASS          | MAIN, FINPRT      | REPRT         |
| APPMAP          | MAIN              | N.A.          |
| APPMAS          | MAIN, FINPRT      | REPRT         |
| CORRBB          | MAIN              | N.A.          |
| CORRB           | MAIN, FINPRT      | REPRT         |
| SUM             | MAIN              | N.A.          |
| SUM1            | MAIN              | N.A.          |
| TEMPAR          | FINPRT            | N.A.          |
| A               | DPFD              | N.A.          |
| X               | DPFD              | N.A.          |

FIGURE III. FLOW CHART OF PROGRAM




FIGURE IV - LABELED COMMON CROSS REFERENCE TABLE

|                  | BLKDAT | CHKLN | DPFD | FINPRT | HEADPG | INVCHK | MAIN | PGCONT | PRINT1 | PRINT2 | READ1 | READ2 | READIT | SAVMTX | SPINV | TEXT1 | TEXT2 | TEXT3 | TEXT4 | TEXT5 | TEXT6 | TEXTS1 | TEXTS2 |
|------------------|--------|-------|------|--------|--------|--------|------|--------|--------|--------|-------|-------|--------|--------|-------|-------|-------|-------|-------|-------|-------|--------|--------|
| CHECK<br>(1544)  |        |       |      |        |        | X      | X    |        |        | X      |       |       |        | X      |       |       |       |       |       |       |       |        |        |
| COMPUT<br>(1057) |        |       |      |        |        |        | X    |        |        | X      |       |       |        |        |       |       |       |       |       |       |       |        |        |
| DPFDVL<br>(22)   | X      |       | X    |        |        |        | X    |        |        | X      |       |       | X      |        |       |       |       |       |       |       |       |        |        |
| INPUT<br>(3531)  |        |       |      | X      | X      |        | X    | X      |        | X      |       | X     |        |        |       |       |       |       |       |       |       |        |        |
| INVCST<br>(1)    | X      |       |      |        |        |        |      |        |        |        |       |       |        |        | X     |       |       |       |       |       |       |        |        |
| ITSTOP<br>(1)    | X      |       |      |        |        |        | X    |        |        |        |       |       |        |        |       |       |       |       |       |       |       |        |        |
| PCHOUT<br>(1)    |        |       |      | X      |        |        |      |        |        | X      |       |       |        |        |       |       |       |       |       |       |       |        |        |
| PRT1<br>(1005)   |        |       |      | X      | X      |        | X    | X      | X      | X      | X     | X     |        |        |       |       |       |       |       |       |       |        |        |
| PRT2<br>(3)      |        | X     |      | X      | X      |        | X    | X      | X      | X      |       |       |        |        |       |       |       |       |       |       |       |        |        |
| PRTL B<br>(17)   |        |       |      |        |        |        |      |        |        | X      |       | X     |        |        |       |       |       |       |       |       |       |        |        |
| RAREA<br>(115)   |        |       |      |        |        |        |      |        |        |        | X     | X     |        |        |       |       |       |       |       |       |       |        |        |
| REPRT<br>(1275)  |        |       |      | X      |        |        | X    |        |        |        |       |       |        |        |       |       |       |       |       |       |       |        |        |
| STOP<br>(5)      | X      |       |      |        |        |        |      |        |        |        | X     |       |        |        |       |       |       |       |       |       |       |        |        |
| UNITIO<br>(5)    | X      | X     |      | X      | X      |        | X    | X      | X      | X      | X     |       | X      |        |       | X     | X     | X     | X     |       |       |        |        |

#### 4.4 Equivalenced Variables

The FORTRAN EQUIVALENCE statement is used only in subprogram READIT.

Following is a table giving the two variables which are equivalenced.

| <u>Variable 1</u> | <u>Variable 2</u> | <u>Constant Value</u> |
|-------------------|-------------------|-----------------------|
| IDIGIT (1)        | KFD (1)           | 0                     |
| IDIGIT (2)        | KFD (2)           | 1                     |
| IDIGIT (3)        | KFD (3)           | 2                     |
| IDIGIT (4)        | KFD (4)           | 3                     |
| IDIGIT (5)        | KFD (5)           | 4                     |
| IDIGIT (6)        | KFD (6)           | 5                     |
| IDIGIT (7)        | KFD (7)           | 6                     |
| IDIGIT (8)        | KFD (8)           | 7                     |
| IDIGIT (9)        | KFD (9)           | 8                     |
| IDIGIT (10)       | KFD (10)          | 9                     |
| IPLUS             | KFD (15)          | +                     |
| IMINUS            | KFD (12)          | -                     |
| ID                | KFD (17)          | D                     |
| IE                | KFD (18)          | E                     |
| IDECML            | KFD (13)          | .                     |
| IBLANK            | KFD (11)          | blank                 |

Parameter Dependent Variables The following table gives a list of

equivalenced variables whose size may be changed due to modification in

number of observations, number of unknowns or number of linear

combinations. The first column gives the variable name and its current

value, where n (the number of observations) = 50, k (the number

of unknowns) = 15, l (the number of linear combinations) = 19, and

m (number of values saved to be printed on summary page) = 50. The second column gives the COMMON area containing the variable. The N.A. entry means that the variable is not in a labeled COMMON area. The third column lists the names of the subprogram(s) containing the variable.

| <u>Variable</u> | <u>Labeled COMMON</u> | <u>Subprograms</u>                          |
|-----------------|-----------------------|---------------------------------------------|
| AIDCST (5,k)    | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| ANOM (k)        | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| DENSTY (k)      | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| COEFEX (k)      | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| ACCVAL (k)      | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| ARSTIN (k)      | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| ACKSTD (k)      | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| IRSTOU (k)      | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| IPRNT (k)       | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| DESMAT (k,n)    | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| OBSERV (12*n)   | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| ALCOM (k,l+1)   | INPUT                 | MAIN, READ2, PRINT2, PGCONT, HEADPG, FINPRT |
| SWTPRT (n)      | COMPUT                | MAIN, PRINT2                                |
| A (n)           | COMPUT                | MAIN, PRINT2                                |
| DELTA (n)       | COMPUT                | MAIN, PRINT2                                |
| OBSCOR (k)      | COMPUT                | MAIN, PRINT2                                |
| COMVOL (k)      | COMPUT                | MAIN, PRINT2                                |
| SERROR (k)      | COMPUT                | MAIN, PRINT2                                |
| TRISIG (k)      | COMPUT                | MAIN, PRINT2                                |
| TOTUN (k)       | COMPUT                | MAIN, PRINT2                                |
| DRIFT (n)       | COMPUT                | MAIN, PRINT2                                |

| <u>Variable</u>        | <u>Labeled COMMON</u> | <u>Subprograms</u>           |
|------------------------|-----------------------|------------------------------|
| ZERO (n)               | COMPUT                | MAIN, PRINT2                 |
| COMVOP (k)             | COMPUT                | MAIN, PRINT2                 |
| CORR5A ( $\ell+1$ )    | COMPUT                | MAIN, PRINT2                 |
| SIG35A ( $\ell+1$ )    | COMPUT                | MAIN, PRINT2                 |
| UNC5A ( $\ell+1$ )     | COMPUT                | MAIN, PRINT2                 |
| IOTSTR (n)             | COMPUT                | MAIN, PRINT2                 |
| SER5A ( $\ell+1$ )     | COMPUT                | MAIN, PRINT2                 |
| DS1 (n)                | COMPUT                | MAIN, PRINT2                 |
| VOLP (k)               | N.A.                  | MAIN                         |
| CORRP (k)              | N.A.                  | MAIN                         |
| TEMP (k)               | N.A.                  | MAIN                         |
| D1 (n)                 | N.A.                  | MAIN                         |
| ILOAD (n)              | N.A.                  | MAIN                         |
| TEMP2 (k)              | N.A.                  | MAIN                         |
| ALOAD (n)              | N.A.                  | MAIN                         |
| AITEM (5,m)            | REPRT                 | MAIN, FINPRT                 |
| APPMAS (m)             | REPRT                 | MAIN, FINPRT                 |
| TRMASS (m)             | REPRT                 | MAIN, FINPRT                 |
| INCRPT (m)             | REPRT                 | MAIN, FINPRT                 |
| VOIPRT (m)             | REPRT                 | MAIN, FINPRT                 |
| COPIB (m)              | REPRT                 | MAIN, FINPRT                 |
| COPIU (m)              | REPRT                 | MAIN, FINPRT                 |
| COPIZ ( $(\ell+2)^2$ ) | CHECK                 | MAIN, SAVMTX, INVCHK, PRINT2 |
| COPIQ ( $(\ell+2)^2$ ) | CHECK                 | MAIN, SAVMTX, INVCHK, PRINT2 |
| COPIK                  | PRTL B                | MAIN, READ2, PRINT2          |



| <u>Variable</u> | <u>Labeled COMMON</u> | <u>Subprograms</u> |
|-----------------|-----------------------|--------------------|
| ITEMP (k)       | N.A.                  | PRINT2             |
| KTEMP (k)       | N.A.                  | PRINT2             |
| JTEMP (k)       | N.A.                  | PRINT2             |
| NNP (m)         | N.A.                  | FINPRT             |
| TEMPAR (m)      | N.A.                  | FINPRT             |
| TRMASX (k)      | N.A.                  | FINPRT             |

In addition to the above dimensioned variables which could be changed, a variable defined as  $NR = n*12$  in subprogram READ2 would have to be changed if the number of observations is increased from the current assigned value of 507.

#### 4.6 Hardware and System Dependent Variables

The following table describes the variables which may present some problem at implementation time. The table lists the variables, the subprograms defining them and the value used for the UNIVAC 1108 at NBS.

| <u>Variable Name</u> | <u>Defining Subprogram</u> | <u>Current Value and Description of Variable</u>                                                        |
|----------------------|----------------------------|---------------------------------------------------------------------------------------------------------|
| ZERMAC               | BLKDAT                     | $1 \times 10^{-8}$ - machine zero                                                                       |
| IR                   | BLKDAT                     | 5 - input unit                                                                                          |
| IW                   | BLKDAT                     | 6 - output (printer) unit                                                                               |
| IP                   | BLKDAT                     | 1 - output (punch) unit                                                                                 |
| IPL                  | BLKDAT                     | 58 - maximum number of lines allowed per printed page                                                   |
| ITMP                 | BLKDAT                     | 7 output (temporary file)                                                                               |
| T                    | READIT                     | This dimensioned variable contains the limits of a real variable beginning at and going up to $10^{38}$ |

| <u>Variable Name</u> | <u>Defining Subprogram</u> | <u>Current Value and Description of Variable</u>                                                                                   |
|----------------------|----------------------------|------------------------------------------------------------------------------------------------------------------------------------|
| IZERO                | READIT                     | 39-number of unique powers of ten represented by the machine range of a real variable given in T above                             |
| MAX                  | READIT                     | 77-number of powers of ten represented by the machine range of a real variable as specified above in T                             |
| KFD                  | BLKDAT                     | Define in hollerith notation the characters 0-9, blank, -, ., *, +, comma, D and E which are used by the subprogram READIT and DPF |
| FS                   | BLKDAT                     | hollerith character S                                                                                                              |
| FT                   | BLKDAT                     | hollerith character T                                                                                                              |
| FO                   | BLKDAT                     | hollerith character O                                                                                                              |
| FP                   | BLKDAT                     | hollerith character P                                                                                                              |
| FB                   | BLKDAT                     | hollerith character Δ (blank)                                                                                                      |

#### 4.7 Required Storage

The following table lists all subprograms and labeled common blocks.

For the subprogram, the number of lines of FORTRAN statements and the number of memory locations for the code and data are given. If the entry is for a labeled COMMON area, a C preceeds the entry and the memory locations are given under the column headed DATA.

FO and FP are used by subprogram READ1 to check for the end of a

| PROGRAM OR<br>COMMON   | LINES OF<br>CODE | MEMORY FOR<br>CODE | MEMORY FOR<br>DATA |
|------------------------|------------------|--------------------|--------------------|
| BLKDAT                 | 41               |                    |                    |
| C CHECK                |                  |                    | 868                |
| CHKLN                  | 11               | 21                 | 5                  |
| C COMPUT               |                  |                    | 559                |
| DPFD                   | 89               | 194                | 40                 |
| C DPFDDL               |                  |                    | 18                 |
| ERROR                  | 56               | 160                | 168                |
| FINPRT                 | 181              | 486                | 363                |
| HEADPG                 | 36               | 41                 | 26                 |
| C INPUT                |                  |                    | 1881               |
| C INVCST               |                  |                    | 1                  |
| INVCHK                 | 69               | 140                | 36                 |
| C ITSTOP               |                  |                    | 1                  |
| MAIN                   | 628              | 1426               | 352                |
| C PCHOUT               |                  |                    | 1                  |
| PGCONT                 | 53               | 68                 | 32                 |
| PRINT1                 | 67               | 192                | 116                |
| PRINT2                 | 670              | 1785               | 1212               |
| C PRT1                 |                  |                    | 517                |
| C PRT2                 |                  |                    | 3                  |
| C PRTL8                |                  |                    | 15                 |
| C RAREA                |                  |                    | 77                 |
| READ1                  | 120              | 170                | 23                 |
| READ2                  | 325              | 516                | 34                 |
| READIT                 | 356              | 586                | 209                |
| C REPRT                |                  |                    | 701                |
| SAVMTX                 | 22               | 22                 | 9                  |
| SPINV                  | 126              | 293                | 43                 |
| C STOP                 |                  |                    | 5                  |
| TEXT1                  | 125              | 64                 | 705                |
| TEXT2                  | 125              | 64                 | 705                |
| TEXT3                  | 125              | 64                 | 705                |
| TEXT4                  | 125              | 64                 | 705                |
| TEXT5                  | 121              | 64                 | 687                |
| TEXT6                  | 40               | 24                 | 170                |
| TEXTS1                 | 124              | 59                 | 705                |
| TEXTS2                 | 30               | 19                 |                    |
| C UNITIO               |                  |                    |                    |
| TOTAL (excluding text) | 2850             | 6100               |                    |
| TOTAL (text)           | 815              | 422                |                    |
| GRAND TOTAL            | 3665             | 6522               |                    |

### Acknowledgements

The authors wish to thank Mr. J. M. Cameron for his helpful suggestions, Mrs. Sue Bussard for her typing skill and patience through many revisions of the document and Laurie Korzendorfer for her help in assembling the appendices. The authors also wish to thank Sally Peavy, Roy Wampler and Clayton Albright for their subprogram contributions.

## REFERENCES

The following references are suggested for detailed description of portions of this report, and for general information concerning the mass measurement process:

1. Pontius, P. E., and Cameron, J. M., "Realistic Uncertainties and the Mass Measurement Process," NBS, (U.S.), Monogr. 103, Aug. 15, 1967.
2. Pontius, P. E., "Measurement Philosophy of the Pilot Program for Mass Calibration," NBS (U.S.), Tech. Note 288, May 6, 1966.
3. Bowman, H. A., and Schoonover, R. M. with Appendix by Mildred Jones, "Procedure for High Precision Density Determinations by Hydrostatic Weighing," J. Res. NBS (U.S.), 71C. Engineering and Instrumentation No. 3, 179-198, July-Aug. 1967.
4. Natrella, M. B., "Experimental Statistics," NBS (U.S.), Handbook 91, Aug. 1, 1963.
5. Ku, H. H., "Precision Measurement and Calibration," Selected NBS papers on Statistical Concepts and Procedures, NBS (U.S.), Spec. Pub. 300, Vol. 1, Feb. 1969.
6. Pontius, P. E., "Mass and Mass Values," NBS (U.S.), Monogr. 133, Jan. 1974.
7. Cameron, J. M., "The Use of the Method of Least Squares in Calibration," NBS (U.S.) NBSIR 74-587, Sept. 1974.
8. Almer, H. W. and Keller, Jerry, "Surveillance Test Procedures, NBS (U.S.) NBSIR 76-999, Feb. 1976.
9. Wexler, A. and Greenspan, L. "Vapor Pressure Equation for Water in the Range 0°C to 100°C," NBS Journal of Research for Physics and Chemistry, Vol. 75A, No. 3, May-June 1971.
10. Cameron, J. M., Croarkin, M. C. and Raybold, R.C., "Designs for Calibration of Standards of Mass," NBS, (U.S.), Tech. Note 952, 1977.
11. Cameron, J. M. and Hailes, G. E., "Designs for the Calibration of Small Groups of Standards in the Presence of Drift," NBS (U.S.) Tech. Note 844, Aug. 1974.
12. Pearson, E. S. and Hartley, H. O., Biometrika Tables for Statisticians, Vol. 1, Cambridge Press, 1956, page 131.

APPENDIX A.1--SAMPLE INPUT

```

X Y Z CORPORATION
SOMEWHERE, U.S.A.
SET OF MASS STANDARDS : 5KG - 100MG
SERIAL NUMBER 12345
MANUFACTURER : TROEMNER, INC.
JUNE 21, 1979
654321
0 .076 20 80
0 0 0 0
5 24 79 84 001 002
21.98 22.22 733.68 734.08 41 41
8 5 53 0
1.15 49.98277 .00301104 .000020 0
      5KG      5000 7.953 .000045
      3KG      3000 7.953 .000045
      2KG      2000 7.9 .000045
      S 1KG-1      1000 8.0064 .000045 11.241
      S 1KG-2      1000 8.0063 .000045 11.825
0 0 0 1 1
0 0 0 1 -1
0 0 0 1 1
1 1 1
1 -1 -1 1 -1
-1 1 1 1 -1
1 -1 -1
-1 1 0 1 1
0 -1 1 1
0 1 -1 0 -1
0 0 1 -1 -1
0 0 0 1 -1
4.7 15.2 5.0 5.4 12.5 5.5 7.3 15.0 7.4
5.8 12.4 6.2 5.6 14.1 5.9 7.8 16.4 8.2
6.4 13.6 6.5 4.4 13.4 4.5 7.2 15.2 7.3
4.8 13.4 5.0 5.4 15.2 5.7 7.3 17.4 7.6
6.0 14.8 6.3 5.3 13.0 5.4 6.7 16.0 7.0
5.2 13.0 5.3 5.4 16.0 5.6 9.2 16.5 9.3
5.9 14.0 6.2 5.9 13.4 6.2 7.8 16.2 8.0
5.9 13.9 6.1 5.0 14.8 5.2 8.6 15.8 8.6
-200000
3 1 0 0
5 23 79 84 003 002
21.91 21.92 736.86 736.76 40 40
6 4 41 0
.028 49.98277 .00301104 .000020 0
      S 1KG-1      1000 8.0064 .000045 11.241
      S 1KG-2      1000 8.0063 .000045 11.825
      1KG      1000 7.953 .000045
      SUM 1KG      1000 7.92641 .000045
1 1
1 1
0 0 0 1

```

```

0 0 1
1 -1
1 0 -1
1 0 0 -1
0 1 -1
0 1 0 -1
0 0 1 -1
16.74 17.34 67.37 66.73
16.85 11.24 61.18 66.77
16.84 13.17 63.11 66.76
17.32 11.13 61.06 67.23
17.22 13.02 62.93 67.16
10.92 12.82 62.86 60.84
-200000
3 1 0 0
5 23 79 84 003 004
21.92 21.96 736.92 736.56 40 40
11 6 62 1
.028 49.98277 .00301104 .000020 0
      500G      500 7.9 .000045
      300G      300 7.953 .000045
      200G      200 7.953 .000045
      100G      100 7.9 .000045
      S      100G      100 7.953 .000045 .9883
SUM 100G      100 7.94234 .000045
1 1 1
0 0 0 0 1
0 0 0 0 0 1
1 1 1 1
1 -1 -1 1 -1
1 -1 -1 0 1 -1
1 -1 -1 -1 0 1
1 -1 -1
1 0 -1 -1 -1 -1
0 1 -1 1 -1 -1
0 1 -1 -1 1 -1
0 1 -1 -1 -1 1
0 0 1 -1 -1
0 0 1 -1 0 -1
0 0 1 0 -1 -1
1 0 0 1
12.78 10.62 60.63 62.86
13.02 12.59 62.54 62.96
14.86 10.75 60.73 64.88
12.86 10.61 60.54 62.82
12.91 13.52 63.54 62.81
10.87 13.44 63.33 60.84
10.94 13.33 63.26 60.82
12.69 11.44 61.38 62.66
11.18 11.71 61.68 61.14
11.18 13.50 63.38 61.11
11.09 13.44 63.40 60.96
-200000
3 1 0 0
5 17 79 84 005 006
21.99 21.96 746.60 746.00 31 31
11 6 62 0
.012 49.98277 .00301104 .000020 0

```

|         |    |         |               |
|---------|----|---------|---------------|
| 50G     | 50 | 7.953   | .000045       |
| 30G     | 30 | 7.953   | .000045       |
| 20G     | 20 | 7.9     | .000045       |
| 10G     | 10 | 7.953   | .000045       |
| S 10G   | 10 | 7.953   | .000045 .0785 |
| SUM 10G | 10 | 7.92641 | .000045       |

```

1 1 1
0 0 0 0 1
0 0 0 0 0 1
1 1 1 1
1 -1 -1 1 -1
1 -1 -1 0 1 -1
1 -1 -1 -1 0 1
1 -1 -1
1 0 -1 -1 -1 -1
0 1 -1 1 -1 -1
0 1 -1 -1 1 -1
0 1 -1 -1 -1 1
0 0 1 -1 -1
0 0 1 -1 0 -1
0 0 1 0 -1 -1
6.18 4.69 54.71 56.20
6.17 4.72 54.71 56.15
6.14 4.76 54.76 56.14
6.19 4.74 54.72 56.18
6.18 4.47 54.45 56.18
4.81 4.44 54.44 54.81
4.79 4.49 54.47 54.75
4.76 4.51 54.49 54.74
4.52 4.55 54.52 54.49
4.48 4.52 54.50 54.47
4.47 4.47 54.46 54.47
-200000
3 1 0 0
5 18 79 84 007 008
21.92 21.88 743.28 742.82 35 35
11 6 62 1
.0017 5.00171 .00185248 .000069 0
5G      5 7.9 .000045
3G      3 7.953 .000045
2G      2 7.953 .000045
1G      1 7.9 .000045
S 1G    1 7.953 .000045 -.0792
SUM 1G  1 16.6 .000020
1 1 1
0 0 0 0 1
0 0 0 0 0 1
1 1 1
1 -1 1 -1
1 -1 0 1 -1
1 -1 -1 0 1
1 -1
1 -1 -1 -1
1 -1 -1
1 1 -1
1 -1 1
-1
0 -1

```



```

0 0 1 0 -1 -1
1 0 0 1
1.084 .926 5.927 6.084
.978 .939 5.931 5.972
.990 1.021 6.021 5.987
1.068 1.013 6.009 6.067
1.068 .875 5.870 6.063
1.062 .862 5.857 6.053
.959 .966 5.957 5.948
.969 .948 5.942 5.960
1.016 .976 5.966 6.006
1.016 .987 5.981 6.006
1.012 .881 5.872 6.004
-200000
3 1 0 0
5 18 79 84 007 008
22.21 22.99 742.52 741.86 36 35
11 6 62 0
.0005 .50156 .0018576 .000069 0
      500MG      .5 16.6 .000020
      300MG      .3 16.6 .000020
      200MG      .2 16.6 .000020
      100MG      .1 16.6 .000020
      S 100MG      .1 16.6 .000020 -.02628
      SUM 100MG      .1 8.17883 .000049
1 1 1
0 0 0 0 1
0
1 1 1 1
1 -1 -1 1 -1
1 -1 -1 0 1 -1
1 -1 -1 -1 0 1
1 -1 -1
1 0 -1 -1 -1 -1
0 1 -1 1 -1 -1
0 1 -1 -1 1 -1
0 1 -1 -1 -1 1
0 0 1 -1 -1
0 0 1 -1 0 -1
0 0 1 0 -1 -1
296.4 224.0 726.2 801.2
285.5 240.3 741.3 788.9
300.5 241.6 744.1 803.7
312.9 253.5 754.8 815.5
312.7 239.8 741.0 815.6
303.8 250.5 751.5 805.7
290.2 264.6 765.4 791.7
303.6 252.3 753.6 806.0
276.7 302.6 804.1 779.3
276.7 315.5 817.1 779.2
276.6 301.0 802.6 778.6
-200000
STOP

```

APPENDIX A.2--SAMPLE PRINTED OUTPUT

U. S. DEPARTMENT OF COMMERCE  
NATIONAL BUREAU OF STANDARDS  
NATIONAL ENGINEERING LABORATORY  
WASHINGTON, D.C. 20234

R E P O R T  
O F  
M A S S     V A L U E S

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
SERIAL NUMBER 12345  
MANUFACTURER : TROEMNER, INC.  
JUNE 21, 1979

TEST NUMBER 654321

FOR THE DIRECTOR,

G. E. MATTINGLY, CHIEF  
FLUID ENGINEERING DIVISION  
CENTER FOR MECHANICAL ENGINEERING  
AND PROCESS TECHNOLOGY  
NATIONAL ENGINEERING LABORATORY

#### INTRODUCTION

THIS DOCUMENT IS A COMPREHENSIVE REPORT COVERING THE SEQUENCE OF OPERATIONS USED TO ASSIGN MASS VALUES TO THE WEIGHTS IDENTIFIED ABOVE. IT INCLUDES A COMPLETE DESCRIPTION OF THE MEASUREMENT METHODS AND PROCEDURES WHICH WERE USED. ALL OF THE DATA, AND THE ANALYSIS OF THIS DATA. THE RESULTS ARE PRESENTED IN SEVERAL FORMATS. ASSIGNED MASS VALUES, DISPLACEMENT VOLUMES, COEFFICIENTS OF EXPANSION, UNCERTAINTIES, TOGETHER WITH THE SUMMED VALUES FOR LINEAR COMBINATIONS OF THE WEIGHTS IN EACH DECADE ARE PRESENTED AT THE END OF THE APPROPRIATE SERIES. THIS INFORMATION SHOULD BE USEFUL TO THOSE WHO MUST ASSIGN MASS VALUES TO OBJECTS OTHER THAN WEIGHTS. FOR CONVENIENCE, THE VALUES AND UNCERTAINTIES, TOGETHER WITH OTHER APPROPRIATE DATA AND COMMENTS ARE ALSO SUMMARIZED IN TABLES I AND II AT THE END OF THE REPORT. CERTAIN INTERMEDIATE PAGES ARE SUMMARIES OF STATISTICAL DATA WHICH RELATE TO THE MASS MEASUREMENT PROCESS USED TO PERFORM THIS WORK. THESE PAGES HAVE BEEN LEFT IN THE REPORT TO RETAIN CONTINUITY. COPIES OF THESE PAGES BECOME PART OF A COLLECTION OF STATISTICAL DATA WHICH REFLECTS THE MEASUREMENT PROCESS PERFORMANCE OVER A PERIOD OF TIME. SUCH A COLLECTION HAS BEEN USED TO ESTABLISH THE CONTROL LIMITS FOR ACCEPTING THE RESULTS OF THIS MEASUREMENT. THESE COLLECTIONS ARE OPEN FOR INSPECTION AT OUR FACILITY.

#### THE MASS MEASUREMENT SYSTEM

THE MASS MEASUREMENT SYSTEM WITHIN THIS COUNTRY CONSISTS OF ALL OF THE MEASUREMENT PROCESSES

WHICH RELY, DIRECTLY OR INDIRECTLY, ON MASS MEASUREMENTS TO ACCOMPLISH A WIDE VARIETY OF ENDEAVORS. IN ORDER FOR THIS SYSTEM TO FUNCTION PROPERLY, EVERYONE WHO MAKES MEASUREMENTS MUST BE ABLE TO VERIFY THAT HIS MEASUREMENT PROCESS PRODUCES CONSISTENT RESULTS WHICH ARE COMPATIBLE WITH HIS PARTICULAR REQUIREMENTS. THE WEIGHTS COVERED BY THIS REPORT, TOGETHER WITH THE ASSIGNED VALUES AND THE APPROPRIATE UNCERTAINTIES FOR THESE VALUES, PROVIDE IN PART A BASIS FOR CONSISTENT MEASUREMENTS WITHIN THIS SYSTEM OF RELATED MEASUREMENT PROCESSES.

APPROPRIATE CHARACTERIZATION OF ANY MEASUREMENT PROCESS IS FUNDAMENTAL TO VERIFYING THAT RESULTS ARE CONSISTENT WITH THE END REQUIREMENT WITH RESPECT TO CORRECTNESS AND ECONOMY OF THE MEASUREMENT EFFORT. WITHOUT THIS INFORMATION, THE BENEFITS OF OWNERSHIP OF THESE WEIGHTS MAY BE COMPLETELY ILLUSORY. THE ASSIGNED UNCERTAINTIES IN THIS REPORT ARE DESCRIPTIVE OF OUR MASS MEASUREMENT PROCESS. EFFECTIVENESS OF THE TRANSFER OF THE UNIT FROM ONE FACILITY TO ANOTHER SHOULD BE VERIFIED BY AN INDEPENDENT TEST. IT IS PRESUMED THAT THESE WEIGHTS WILL BE USED IN A SIMILARLY WELL CHARACTERIZED MEASUREMENT PROCESS SO THAT THE STATISTICAL PARAMETERS OF BOTH PROCESSES CAN BE COMBINED TO PROVIDE A REALISTIC ESTIMATE OF THE UNCERTAINTY OF THE MASS AS ACTUALLY REALIZED IN ANOTHER FACILITY. A COMPREHENSIVE SERIES DIRECTED TOWARD THE EVALUATION OF A PARTICULAR MASS MEASUREMENT PROCESS IS AVAILABLE THROUGH MASS MEASUREMENT ASSURANCE PROGRAMS OF THE NATIONAL BUREAU OF STANDARDS.

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 2

#### WEIGHING DESIGN

ONLY DIFFERENCES IN MASS CAN BE MEASURED. THEREFORE THE MASS VALUES FOR THE 'UNKNOWN' WEIGHTS MUST BE DETERMINED BY COMPARISON WITH OTHER WEIGHTS WHICH HAVE ACCEPTED MASS VALUES. THE 'UNKNOWN' WEIGHTS TOGETHER WITH 'CHECK STANDARDS', ARE GROUPED AND INTERCOMPARED ACCORDING TO THE DESIGN SCHEDULE GIVEN AT THE BEGINNING OF EACH SERIES OF WEIGHINGS. THE FIRST SERIES CONTAINS STANDARDS WHICH PROVIDE THE STARTING VALUES FOR THE SERIES OF WEIGHINGS AND PROVIDE THE TIE POINT FOR CONSISTENCY THROUGHOUT THE MEASUREMENT SYSTEM. THE WEIGHING METHOD USED, I.E., DOUBLE SUBSTITUTION, TRANSPOSITION, ETC., IS INDICATED ALONG WITH THE OBSERVED DATA. IN THE COMPUTATIONS, THE DISPLACEMENT VOLUMES ARE TREATED EXPLICITLY, USING THE DATA LISTED IN THE REPORT. IN ALL CASES, A REDUNDANCY IN THE NUMBER OF MEASUREMENTS PROVIDES A MEANS FOR CHECKING ON THE PRECISION OF THE PROCESS.

WHEN THERE ARE MORE EQUATIONS THAN 'UNKNOWN'S', NOT ALL OBSERVATIONAL EQUATIONS CAN BE SATISFIED EXACTLY AND THE METHOD OF LEAST SQUARES IS USED TO PROVIDE ESTIMATES OF THE 'UNKNOWN' VALUES. THIS METHOD LEADS TO ESTIMATORS WHICH ARE LINEAR FUNCTIONS OF THE DATA AND WHICH HAVE STANDARD DEVIATIONS READILY CALCULATED FROM THE COEFFICIENTS OF THE LINEAR EQUATIONS AND THE STANDARD DEVIATION OF AN INDIVIDUAL MEASUREMENT. THE 'CHECK STANDARD' IS ALSO TREATED AS AN UNKNOWN AND THE DEVIATION OF THE CURRENT RESULT FROM THE ACCEPTED VALUE PROVIDES A MEANS FOR CHECKING THE ADEQUACY OF THE CURRENT DATA. THIS SAME CHECK

STANDARD IS MEASURED WITH EACH TEST OF UNKNOWN'S AND THE COLLECTION OF VALUES OVER TIME IS USED TO EVALUATE THE PERFORMANCE OF THE MEASUREMENT PROCESS.

IN THE CASE OF THE SERIES WHICH INCLUDES THE KNOWN STANDARDS, THE ACCEPTED VALUES OF THESE STANDARDS SERVE AS A RESTRAINT ON THE SOLUTION OF THE EQUATIONS FOR THE VALUES OF ALL OF THE WEIGHTS. THE RESTRAINT FOR THE SOLUTION OF SUBSEQUENT SERIES IS PROVIDED BY THE VALUES ESTABLISHED FOR ONE OR MORE WEIGHTS INCLUDED IN A PREVIOUS SERIES.

ESTIMATED VALUES FOR WEIGHTS WHICH HAVE BEEN GROUPED IN THE SAME SERIES INVOLVE THE SAME OBSERVATIONAL DATA AND ARE, IN ALMOST ALL CASES, CORRELATED. FOR EACH SERIES THERE IS A TABLE OF COMBINATIONS TOGETHER WITH THE APPROPRIATE UNCERTAINTY FOR EACH COMBINATION.

#### PROCESS CONTROL

THE STANDARD DEVIATION, AS COMPUTED FROM THE LEAST SQUARES SOLUTION, PROVIDES A CHECK ON THE SHORT TERM, OR 'WITHIN-RUN' PROCESS PRECISION. AN AVERAGE OF A NUMBER OF THESE STANDARD DEVIATIONS IS TAKEN AS THE ACCEPTED WITHIN-RUN STANDARD DEVIATION OF THE PROCESS AND IS USED AS A REFERENCE VALUE FOR SURVEILLANCE OF THE PROCESS PRECISION. THE VALUES OBTAINED FOR THE 'CHECK STANDARD' PROVIDE, AS TIME GOES ON, A SEQUENCE OF VALUES THAT REALISTICALLY REFLECTS THE VARIATIONS WHICH BESET PRECISE MEASUREMENTS. COLLECTIONS OF VALUES FOR BOTH THE WITHIN-RUN PRECISION AND THE VALUE OBTAINED FOR THE 'CHECK STANDARD' SHOULD

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 3

POSSESS THE PROPERTIES OF RANDOMNESS ASSOCIATED WITH INDEPENDENT MEASUREMENTS FROM A STABLE PROBABILITY DISTRIBUTION. THE REPORTED 'F RATIO' AND 'T VALUE' ARE TESTS OF THE VALUES FROM THE CURRENT RUN FOR CONFORMITY TO THEIR RESPECTIVE DISTRIBUTIONS AND IF SATISFACTORY ARE TAKEN AS EVIDENCE THAT THE PROCESS IS IN CONTROL AND THAT PREDICTIVE STATEMENTS REGARDING UNCERTAINTY ARE VALID.

CONTROL CHARTS ON THE WITHIN-RUN PROCESS PRECISION AND THE VALUES OBTAINED FOR THE CHECK STANDARD ARE KEY ELEMENTS IN MONITORING THE STATE OF CONTROL OF ANY PRECISE MASS MEASUREMENT PROCESS. IN ADDITION TO PROVIDING A BASIS FOR JUDGMENT AS TO THE ADEQUACY OF A GIVEN PROCESS FOR A PARTICULAR REQUIREMENT, THESE DATA PROVIDE A MEANS TO JUDGE THE IMPORTANCE OF LONG TERM, OR 'BETWEEN-RUN' VARIABILITY WHICH CAN BE CHARACTERIZED BY THE STANDARD DEVIATION OF THE VALUES ABOUT THE MEAN. IF THERE IS AN ADDITIONAL COMPONENT OF VARIANCE ENTERING FROM RUN TO RUN, THIS STANDARD DEVIATION WILL BE LARGER THAN CAN BE ACCOUNTED FOR BY THE WITHIN-RUN VARIABILITY. CORRELATION STUDIES, AS WELL AS SUPPLEMENTAL EXPERIMENTS, ARE USED TO DETECT AND REDUCE THE MAGNITUDE OF SIGNIFICANT SYSTEMATIC EFFECTS. APPROPRIATE ACTION, E.G., ADDITIONAL EMPIRICAL CORRECTIONS OR CHANGES IN TECHNIQUE, CAN REDUCE THE EFFECTS FROM KNOWN SOURCES OF SYSTEMATIC VARIABILITY TO A MAGNITUDE WHICH IS NO LONGER IDENTIFIABLE IN THE DATA. IN THE CASES WHERE A SIGNIFICANT LONG TERM, OR BETWEEN-RUN, COMPONENT REMAINS THE UNCERTAINTY HAS BEEN APPROPRIATELY ADJUSTED.

SERIES OF MEASUREMENTS JUDGED AS OUT OF CONTROL RELATIVE TO THE APPROPRIATE PARAMETER ARE CAREFULLY EXAMINED. IF RERUNS WERE NECESSARY IN THE COURSE OF THIS WORK, THE 'OUT OF CONTROL' SERIES, WITH REMARKS AS APPROPRIATE, ARE ATTACHED AT THE END OF THE REPORT FOR YOUR INFORMATION.

#### UNCERTAINTY

IT IS ASSUMED THAT THE PRESENT 'ACCEPTED VALUES' OF TWO NBS STANDARDS AT THE 1 KILOGRAM LEVEL, DESIGNATED N1 AND N2, ARE WITHOUT ERROR. ESTIMATES OF THE UNCERTAINTY OF THE ACCEPTED VALUES OF THE NBS STANDARDS RELATIVE TO THE INTERNATIONAL PROTOTYPE KILOGRAM CAN BE PROVIDED ON REQUEST. HOWEVER, THESE ESTIMATES HAVE NO REAL MEANING IN EITHER NATIONAL OR INTERNATIONAL COMPARISON. THIS IS BECAUSE OF THE LACK OF SUFFICIENT DATA TO PROVIDE A REALISTIC ESTIMATE OF THE UNCERTAINTY IN THE VALUES ASSIGNED TO THE PROTOTYPE KILOGRAMS K20 AND K4, PARTICULARLY IN REGARD TO LONG TERM, OR BETWEEN-RUN VARIABILITY. CHANGES IN THE ACCEPTED VALUES FOR THE NBS STANDARDS AT THE KILOGRAM LEVEL, AS AND WHEN THEY OCCUR, WILL BE REPORTED IN THE SCIENTIFIC PAPERS OF THE BUREAU AND WILL BE GIVEN WIDE DISTRIBUTION. IN CASES WHERE SUCH CHANGES MAY BE OF IMPORTANCE, OR WHERE CONTINUITY IS DESIRED, INSTRUCTIONS WILL BE INCLUDED FOR UP-DATING PREVIOUSLY REPORTED VALUES. WHEN THE VALUES REPORTED ARE BASED ON THE ACCEPTED VALUES OF STANDARDS OTHER THAN STANDARDS N1 AND N2 MENTIONED ABOVE, THE UNCERTAINTY OF THE ACCEPTED VALUE OF THE STANDARD BECOMES A SYSTEMATIC ERROR IN THE ASSIGNMENT OF VALUES TO OTHER STANDARDS AND IS INCLUDED IN THE REPORT.

A BALANCE UNDER STABLE OPERATING CONDITIONS WILL EXHIBIT A CERTAIN CHARACTERISTIC VARIABILITY WHICH CAN BE DESCRIBED BY THE STANDARD DEVIATION FOR SUCH MEASUREMENTS. THE VALUE FOR AN ARTIFICIAL WEIGHT DETERMINED IN REPEATED TESTS WITH THE SAME WEIGHING DESIGN WILL HAVE ITS OWN STANDARD DEVIATION WHICH WILL BE SOME FUNCTION OF THE BALANCE PRECISION AND (POSSIBLY) OF THE BETWEEN-RUN COMPONENT. AS AN OUTER LIMIT OF THE DISTRIBUTION OF RANDOM ERRORS, THREE TIMES THE STANDARD DEVIATION IS USED. SYSTEMATIC ERRORS DUE TO THE PROCEDURES USED OR TO ENVIRONMENTAL EFFECTS ARE LARGELY BALANCED OUT AND CAN USUALLY BE REGARDED AS NEGLIGIBLE. WHEN A NON-NEGLIGIBLE BOUND TO THE POSSIBLE EFFECT FROM KNOWN SOURCES IS AVAILABLE, IT IS CALCULATED AND REPORTED SEPARATELY. E.G., THE UNCERTAINTY OF ACCEPTED VALUE AT OTHER THAN THE 1 KILOGRAM LEVEL. THE DISTRIBUTION IMPLIED BY THE RANDOM ERRORS MAY THUS BE CENTERED SOMEWHERE IN THE RANGE GIVEN BY THE BOUNDS TO THE SYSTEMATIC ERROR. THE TOTAL UNCERTAINTY IS TAKEN AS THE SUM OF THESE TWO COMPONENTS.

THE UNCERTAINTY ASSOCIATED WITH THE ASSIGNED VALUE CAN BE THOUGHT OF AS A BOUND TO THE DEPARTURE OF THE ASSIGNED VALUE FROM A HYPOTHETICAL AVERAGE VALUE THAT WOULD BE OBTAINED IF IT WERE POSSIBLE TO REPEAT THE MEASUREMENT MANY TIMES OVER A WIDE VARIETY OF CONDITIONS. E.G., SUBSTITUTE THE WEIGHT FOR ONE OF THE CHECK STANDARDS. THIS MEANS THAT THE UNCERTAINTY BAND CENTERED ON THE VALUES OBTAINED FROM EACH OF TWO MEASUREMENTS OF THE SAME OBJECT OVER SOME ARBITRARY TIME INTERVAL

SHOULD ALMOST ALWAYS OVERLAP. IN OTHER WORDS, WHILE A SECOND MEASUREMENT WILL PRODUCE A DIFFERENT VALUE, THIS VALUE WILL ONLY RARELY DIFFER FROM THE FIRST VALUE BY MORE THAN THE SUM OF THE TWO UNCERTAINTIES. THE UNCERTAINTY BANDS ARE NOT EXPECTED TO OVERLAP IF SOME EVENT HAS OCCURRED IN THE TIME INTERVAL BETWEEN THE TWO MEASUREMENTS WHICH WILL CHANGE THE MASS OF THE OBJECT, E.G., ABRASIONS, ABUSE, CORROSION, IMPROPER CLEANING AND THE LIKE.

THE UNCERTAINTY IN ASSIGNED VALUE CONTAINED IN THIS REPORT BECOMES A SYSTEMATIC EFFECT FOR THE MEASUREMENT PROCESS IN WHICH THESE WEIGHTS ARE TO BE USED. IN THE ABSENCE OF OTHER SIGNIFICANT SYSTEMATIC EFFECTS IN THE USER'S MEASUREMENT PROCESS (A CONDITION WHICH MUST BE DEMONSTRATED) THE UNCERTAINTY OF THE VALUE ASSIGNED BY THE USER IS AN APPROPRIATE COMBINATION OF THE SYSTEMATIC ERROR IN THE STANDARD AND THE RANDOM COMPONENT ASSOCIATED WITH HIS PROCESS. IF THE MEASUREMENT PROCESSES ARE IN CONTROL AND APPROPRIATE UNCERTAINTIES ARE ASSIGNED, THE VALUES PRODUCED BY DIFFERENT MEASUREMENT FACILITIES WILL HAVE OVERLAPPING UNCERTAINTY BANDS AS DESCRIBED ABOVE. ONE CANNOT DISCUSS DIFFERENCES IN VALUES FOR THE SAME OBJECT OBTAINED BY DIFFERENT FACILITIES WITH ANY DEGREE OF SERIOUSNESS UNLESS EACH VALUE IS ACCOMPANIED BY A REALISTIC UNCERTAINTY STATEMENT.

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 5

#### REFERENCES

THE FOLLOWING REFERENCES ARE SUGGESTED FOR DETAILED DESCRIPTION OF PORTIONS OF THIS REPORT , AND FOR GENERAL INFORMATION CONCERNING THE MASS MEASUREMENT PROCESS:

1. PONTIUS, P. E., AND CAMERON, J. M.  
REALISTIC UNCERTAINTIES AND THE MASS MEASUREMENT PROCESS  
NAT. BUR. STAND. (U.S.), MONOGR. 103  
(AUG. 15, 1967)
2. PONTIUS, P. E.  
MEASUREMENT PHILOSOPHY OF THE PILOT PROGRAM FOR MASS CALIBRATION  
NAT. BUR. STAND. (U.S.) TECH. NOTE 288  
(MAY 6, 1966)
3. BOWMAN, H. A., AND SCHOONOVER, R. M. WITH APPENDIX BY MILORAD JONES  
PROCEDURE FOR HIGH PRECISION DENSITY DETERMINATIONS BY HYDROSTATIC  
WEIGHING  
J. RES. NAT. BUR. STAND. (U.S.) 71C. ENGINEERING AND INSTRUMENTATION  
NO. 3, 179-190 (JULY-AUG. 1967)
4. NATRELLA, M. B.  
EXPERIMENTAL STATISTICS  
NAT. BUR. STAND. (U.S.) HANDBOOK 91  
(AUGUST 1, 1963)
5. KU, H. H.  
PRECISION MEASUREMENT AND CALIBRATION - SELECTED NBS PAPERS ON  
STATISTICAL CONCEPTS AND PROCEDURES  
NAT. BUR. STAND. (U.S.) SPEC. PUBL. 300  
VOL. 1 (FEB. 1969)
6. PONTIUS, P. E.  
MASS AND MASS VALUES  
NAT. BUR. STAND. (U.S.) MONOGR. 133  
(JAN. 1974)
7. CAMERON, J. M., CROARKIN, C. C. AND RAYBOLD, R. C.  
DESIGNS FOR THE CALIBRATION OF STANDARDS OF MASS  
NAT. BUR. STAND. (U.S.) TECH. NOTE 952  
(JUNE 1977)
8. VARNER, R. N., AND RAYBOLD, R. C.  
NATIONAL BUREAU OF STANDARDS MASS CALIBRATION COMPUTER SOFTWARE  
NAT. BUR. STAND. (U.S.) TECH. NOTE  
(IN PROCESS)

X Y Z CORPORATION  
SOMEWHERE, U.S.A.

PAGE 6

SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321  
TO BE PUBLISHED:

9. PONTIUS, P. E.

THE ACCEPTED VALUES AND ASSOCIATED UNCERTAINTY ESTIMATES OF THE NBS  
STANDARDS AT THE 1 KG LEVEL  
NAT. BUR. STAND. (U.S.) TECH. NOTE  
(EXPECTED COMPLETION: 1975)

10. PONTIUS, P. E.

DOCUMENTATION FOR THE MASS MEASUREMENT PROCESS AT NBS  
NAT. BUR. STAND. (U.S.) TECH. NOTE  
(EXPECTED COMPLETION: 1974)



X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 7  
SERIES 1  
5/24/79

BALANCE 1  
OPERATOR 84  
ACCEPTED WITHIN STANDARD DEVIATION OF THE PROCESS 1.15000 MG  
ACCEPTED BETWEEN STANDARD DEVIATION OF THE PROCESS .00000 MG

CALIBRATION DESIGN 53  
RESTRAINT VECTOR 0 0 0 1 1  
MASS CORRECTION OF RESTRAINT 23.06600 MG  
VOLUME OF WEIGHTS BEING USED IN RESTRAINT AT 22.10 C 249.82820 CM3  
SYSTEMATIC ERROR IN THE RESTRAINT .07600 MG  
3 STANDARD DEVIATION LIMIT FOR RANDOM ERROR AFFECTING RESTRAINT .00000 MG

CHECK STANDARD USED 2  
CHECK STANDARD VECTOR 0 0 0 1 -1  
ACCEPTED MASS CORRECTION OF CHECK STANDARD -.58400 MG  
REPORT VECTOR 1 1 1 0 0

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.98   | 22.22   | 22.10   |
| CORRECTED PRESSURE IN MM HG        | 733.680 | 734.080 | 733.880 |
| CORRECTED HUMIDITY IN PERCENT      | 41.00   | 41.00   | 41.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1505  | 1.1501  | 1.1503  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.98   | 22.22   |         |
| OBSERVED PRESSURE IN MM HG         | 733.680 | 734.080 |         |
| OBSERVED HUMIDITY IN PERCENT       | 41.00   | 41.00   |         |

| WEIGHTS BEING TESTED | NOMINAL VALUE G | DENSITY G/CM3 AT 20C | COEFFICIENT OF EXPANSION | ACCEPTED CORRECTION MG |
|----------------------|-----------------|----------------------|--------------------------|------------------------|
| 5KG                  | 5000.0000       | 7.9530               | .000045                  |                        |
| 3KG                  | 3000.0000       | 7.9530               | .000045                  |                        |
| 2KG                  | 2000.0000       | 7.9000               | .000045                  |                        |
| S 1KG-1              | 1000.0000       | 8.0064               | .000045                  | 11.24100               |
| S 1KG-2              | 1000.0000       | 8.0063               | .000045                  | 11.82500               |

X Y Z CORPORATION  
 SOMEWHERE, U.S.A.  
 SET OF MASS STANDARDS : 5KG - 100MG  
 TEST NUMEER 654321

PAGE 8  
 SERIES 1  
 5/24/79

BALANCE 1  
 OPERATOR 84

CALIBRATION DESIGN 53

|   |   | GRAMS |      |      |      |      |
|---|---|-------|------|------|------|------|
|   |   | 5000  | 3000 | 2000 | 1000 | 1000 |
| A | 1 | +     | -    | -    | +    | -    |
| A | 2 | -     | +    | +    | +    | -    |
| A | 3 | +     | -    | -    |      |      |
| A | 4 | -     | +    |      | +    | +    |
| A | 5 |       | -    | +    | +    |      |
| A | 6 |       | +    | -    |      | -    |
| A | 7 |       |      | +    | -    | -    |
| A | 8 |       |      |      | +    | -    |
| R |   |       |      |      | +    | +    |

OBSERVATIONS IN DIVISIONS  
 SINGLE TRANSPCITION TMC PAN BALANCE

|   |   |        |         |        |        |         |        |
|---|---|--------|---------|--------|--------|---------|--------|
| A | 1 | 4.7000 | 15.2000 | 5.0000 | 5.4000 | 12.5000 | 5.5000 |
|   |   | 7.3000 | 15.0000 | 7.4000 |        |         |        |
| A | 2 | 5.8000 | 12.4000 | 6.2000 | 5.6000 | 14.1000 | 5.9000 |
|   |   | 7.8000 | 16.4000 | 8.2000 |        |         |        |
| A | 3 | 6.4000 | 13.6000 | 6.5000 | 4.4000 | 13.4000 | 4.5000 |
|   |   | 7.2000 | 15.2000 | 7.3000 |        |         |        |
| A | 4 | 4.8000 | 13.4000 | 5.0000 | 5.4000 | 15.2000 | 5.7000 |
|   |   | 7.3000 | 17.4000 | 7.6000 |        |         |        |
| A | 5 | 6.0000 | 14.8000 | 6.3000 | 5.3000 | 13.0000 | 5.4000 |
|   |   | 6.7000 | 16.0000 | 7.0000 |        |         |        |
| A | 5 | 5.2000 | 13.0000 | 5.3000 | 5.4000 | 16.0000 | 5.6000 |
|   |   | 9.2000 | 16.5000 | 9.3000 |        |         |        |
| A | 7 | 5.9000 | 14.0000 | 6.2000 | 5.9000 | 13.4000 | 6.2000 |
|   |   | 7.8000 | 16.2000 | 8.0000 |        |         |        |
| A | 8 | 5.9000 | 13.9000 | 6.1000 | 5.0000 | 14.8000 | 5.2000 |
|   |   | 8.6000 | 15.8000 | 8.6000 |        |         |        |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 5  
SERIES 1  
5/24/79

BALANCE 1  
OPERATOR 84

CALIBRATION DESIGN 53

SENSITIVITY WEIGHT  
MASS 49.98277 MG  
VOLUME .00301 CM3 AT 20 C  
COEFFICIENT OF EXPANSION .000020  
S\*=S-PV(S)= 49.97931 MG

|       | A(I)<br>(MG) | DELTA(I)<br>(MG) | AVERAGE<br>SENSITIVITY<br>(MG/DIV) | ZERO(I)<br>(DIV) | OBSERVED<br>SENSITIVITY<br>(MG/DIV) |
|-------|--------------|------------------|------------------------------------|------------------|-------------------------------------|
| A 1   | 11.72699     | .39524           | 22.33712                           | 9.50000          | 22.71787                            |
| A 2   | -8.09721     | 2.22116          | 22.33712                           | 9.56250          | 21.96893                            |
| A 3   | 12.63844     | 1.81339          | 22.97899                           | 9.47500          | 21.73013                            |
| A 4   | -14.07463    | -.01255          | 22.97899                           | 9.76250          | 24.38015                            |
| A 5   | 14.85099     | -2.10971         | 22.84768                           | 9.82500          | 22.21303                            |
| A 6 * | -18.56374    | -2.10973         | 22.84768                           | 9.93750          | 23.51967                            |
| A 7   | 3.22447      | -.01255          | 21.49648                           | 9.87500          | 21.49648                            |
| A 8   | .00000       | -.50669          | 22.21302                           | 9.95000          | 22.21302                            |

\* OBSERVED DEFLECTION IS GREATER THAN OR EQUAL TO ONE FOURTH THE  
SENSITIVITY DEFLECTION

| ITEM<br>(G) | CORRECTION<br>(MG) | VOLUME<br>(AT T)<br>(CM3) | SYSTEMATIC<br>ERROR<br>(MG) | 3 S.D.<br>LIMIT<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|-------------|--------------------|---------------------------|-----------------------------|-------------------------|------------------------------|
| 5000.0000   | 63.07702           | 628.76090                 | .19000                      | 5.45493                 | 5.64493                      |
| 3000.0000   | 24.01883           | 377.25480                 | .11400                      | 3.57109                 | 3.68509                      |
| 2000.0000   | 30.17279           | 253.19230                 | .07600                      | 2.60795                 | 2.68395                      |
| 1000.0000   | 11.78548           | 124.91335                 | .03800                      | .92205                  | .96005                       |
| 1000.0000   | 11.28052           | 124.91485                 | .03800                      | .92205                  | .96005                       |

TEMPERATURE T= 22.10 C

RESTRAINT FOR FOLLOWING SERIES

|                            |   |   |   |           |     |
|----------------------------|---|---|---|-----------|-----|
| RESTRAINT VECTOR           | 0 | 0 | 0 | 1         | 1   |
| MASS CORRECTION            |   |   |   | 23.06600  | MG  |
| VOLUME AT 20 C             |   |   |   | 249.80460 | CM3 |
| SYSTEMATIC ERROR           |   |   |   | .07600    | MG  |
| 3 STANDARD DEVIATION LIMIT |   |   |   | .00000    | MG  |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 10  
SERIES 1  
5/24/79

BALANCE 1  
OPERATOR 84  
MAXIMUM LOAD 6000.0000 G  
STARTING RESTRAINT NUMBER 80

CALIBRATION DESIGN 53

PRECISION CONTROL

OBSERVED STANDARD DEVIATION OF THE PROCESS 2.09386 MG  
ACCEPTED STANDARD DEVIATION OF THE PROCESS 1.15000 MG  
DEGREES OF FREEDOM 4  
F RATIO 3.315

F RATIO IS LESS THAN 3.33 (CRITICAL VALUE FOR PROBABILITY = .01).  
THEREFORE THE STANDARD DEVIATION IS IN CONTROL.

CHECK STANDARD VECTOR 0 0 0 1 -1  
CHECK STANDARD USED 2  
ACCEPTED MASS CORRECTION OF CHECK STANDARD -.58400 MG  
OBSERVED CORRECTION OF CHECK STANDARD .50497 MG  
STANDARD DEVIATION OF THE OBSERVED CORRECTION .61470 MG  
T VALUE 1.77

ABSOLUTE VALUE OF T IS LESS THAN 3.  
THEREFORE CHECK STANDARD IS IN CONTROL.

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.98   | 22.22   | 22.10   |
| CORRECTED PRESSURE IN MM HG        | 733.680 | 734.080 | 733.880 |
| CORRECTED HUMIDITY IN PERCENT      | 41.00   | 41.00   | 41.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1505  | 1.1501  | 1.1503  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.98   | 22.22   |         |
| OBSERVED PRESSURE IN MM HG         | 733.680 | 734.080 |         |
| OBSERVED HUMIDITY IN PERCENT       | 41.00   | 41.00   |         |

X Y Z CORPORATION  
 SOMEWHERE, U.S.A.  
 SET OF MASS STANDARDS : 5KG - 100MG  
 TEST NUMBER 654321

PAGE 11  
 SERIES 2  
 5/23/79

BALANCE 3  
 OPERATOR 84  
 ACCEPTED WITHIN STANDARD DEVIATION OF THE PROCESS .02800 MG  
 ACCEPTED BETWEEN STANDARD DEVIATION OF THE PROCESS .00000 MG

CALIBRATION DESIGN 41  
 RESTRAINT VECTOR 1 1 0 0  
 MASS CORRECTION OF RESTRAINT 23.06600 MG  
 VOLUME OF WEIGHTS BEING USED IN RESTRAINT AT 21.91 C 249.82613 CM3  
 SYSTEMATIC ERROR IN THE RESTRAINT .07600 MG  
 3 STANDARD DEVIATION LIMIT FOR RANDOM ERROR AFFECTING RESTRAINT .00000 MG

CHECK STANDARD USED 2  
 CHECK STANDARD VECTOR 1 -1 0 0  
 ACCEPTED MASS CORRECTION OF CHECK STANDARD -.58400 MG  
 REPORT VECTOR 0 0 1 0

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.91   | 21.92   | 21.91   |
| CORRECTED PRESSURE IN MM HG        | 736.860 | 736.760 | 736.810 |
| CORRECTED HUMIDITY IN PERCENT      | 40.00   | 40.00   | 40.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1559  | 1.1557  | 1.1558  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.91   | 21.92   |         |
| OBSERVED PRESSURE IN MM HG         | 736.860 | 736.760 |         |
| OBSERVED HUMIDITY IN PERCENT       | 40.00   | 40.00   |         |

| WEIGHTS BEING TESTED | NOMINAL VALUE G | DENSITY G/CM3 AT 20C | COEFFICIENT OF EXPANSION | ACCEPTED CORRECTION MG |
|----------------------|-----------------|----------------------|--------------------------|------------------------|
| S 1KG-1              | 1000.0000       | 8.0064               | .000045                  | 11.24100               |
| S 1KG-2              | 1000.0000       | 8.0063               | .000045                  | 11.82500               |
| 1KG                  | 1000.0000       | 7.9530               | .000045                  |                        |
| SUM 1KG              | 1000.0000       | 7.9264               | .000045                  |                        |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 12  
SERIES 2  
5/23/79

BALANCE 3  
OPERATOR 84

CALIBRATION DESIGN 41  
GRAMS  
1000 1000 1000 1000  
A 1 + -  
A 2 + -  
A 3 + -  
A 4 + -  
A 5 + -  
A 6 + -  
R + +

OBSERVATIONS IN DIVISIONS  
DOUBLE SUBSTITUTION ONE PAN BALANCE

|     |         |         |         |         |
|-----|---------|---------|---------|---------|
| A 1 | 16.7400 | 17.2400 | 67.3700 | 66.7300 |
| A 2 | 16.8500 | 11.2400 | 61.1800 | 66.7700 |
| A 3 | 16.8400 | 13.1700 | 63.1100 | 66.7600 |
| A 4 | 17.3200 | 11.1300 | 61.0600 | 67.2300 |
| A 5 | 17.2200 | 13.0200 | 62.9300 | 67.1600 |
| A 6 | 10.9200 | 12.8200 | 62.8600 | 60.8400 |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 13  
SERIES 2  
5/23/79

BALANCE 3  
OPERATOR 84

CALIBRATION DESIGN 41

SENSITIVITY WEIGHT  
MASS 49.98277 MG  
VOLUME .00301 CM3 AT 20 C  
COEFFICIENT OF EXPANSION .00020  
S\*=S-PV(S)= 49.97929 MG

|     | A(I)<br>(MG) | DELTA(I)<br>(MG) | AVERAGE<br>SENSITIVITY<br>(MG/DIV) | DRIFT(I)<br>(MG) | OBSERVED<br>SENSITIVITY<br>(MG/DIV) |
|-----|--------------|------------------|------------------------------------|------------------|-------------------------------------|
| A 1 | -.61998      | -.02625          | .99997                             | -.02000          | .99859                              |
| A 2 | 5.59983      | .00501           | .99997                             | -.01000          | 1.00059                             |
| A 3 | 3.65989      | .02125           | .99997                             | -.01000          | 1.00059                             |
| A 4 | 6.17981      | -.00875          | .99997                             | -.01000          | 1.00079                             |
| A 5 | 4.21487      | -.01750          | .99997                             | .01500           | 1.00169                             |
| A 6 | -1.95994     | -.00375          | .99997                             | -.06000          | .99759                              |

| ITEM<br>(G) | CORRECTION<br>(MG) | VOLUME<br>(AT T)<br>(CM3) | SYSTEMATIC<br>ERROR<br>(MG) | 3 S.D.<br>LIMIT<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|-------------|--------------------|---------------------------|-----------------------------|-------------------------|------------------------------|
| 1000.0000   | 11.23519           | 124.91225                 | .03800                      | .02970                  | .06770                       |
| 1000.0000   | 11.83082           | 124.91388                 | .03800                      | .02970                  | .06770                       |
| 1000.0000   | 6.60911            | 125.75038                 | .03800                      | .05144                  | .08944                       |
| 1000.0000   | 9.05323            | 126.17253                 | .03800                      | .05144                  | .08944                       |

TEMPERATURE T= 21.91 C

RESTRAINT FOR FOLLOWING SERIES  
RESTRAINT VECTOR 0 0 0 1  
MASS CORRECTION 9.05323 MG  
VOLUME AT 20 C 126.16166 CM3  
SYSTEMATIC ERROR .03800 MG  
3 STANDARD DEVIATION LIMIT .05144 MG

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 14  
SERIES 2  
5/23/79

BALANCE 3  
OPERATOR 84  
MAXIMUM LOAD 1000.0000 G  
STARTING RESTRAINT NUMBER 80

CALIBRATION DESIGN 41

PRECISION CONTROL

OBSERVED STANDARD DEVIATION OF THE PROCESS .02282 MG  
ACCEPTED STANDARD DEVIATION OF THE PROCESS .02800 MG  
DEGREES OF FREEDOM 3  
F RATIO .664

F RATIO IS LESS THAN 3.79 (CRITICAL VALUE FOR PROBABILITY = .01).  
THEREFORE THE STANDARD DEVIATION IS IN CONTROL.

CHECK STANDARD VECTOR 1 -1 0 0  
CHECK STANDARD USED 2  
ACCEPTED MASS CORRECTION OF CHECK STANDARD -.58400 MG  
OBSERVED CORRECTION OF CHECK STANDARD -.59562 MG  
STANDARD DEVIATION OF THE OBSERVED CORRECTION .01980 MG  
T VALUE -.59

ABSOLUTE VALUE OF T IS LESS THAN 3.  
THEREFORE CHECK STANDARD IS IN CONTROL.

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.91   | 21.92   | 21.91   |
| CORRECTED PRESSURE IN MM HG        | 736.860 | 736.760 | 736.810 |
| CORRECTED HUMIDITY IN PERCENT      | 40.00   | 40.00   | 40.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1559  | 1.1557  | 1.1558  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.91   | 21.92   |         |
| OBSERVED PRESSURE IN MM HG         | 736.860 | 736.760 |         |
| OBSERVED HUMIDITY IN PERCENT       | 40.00   | 40.00   |         |



X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 15  
SERIES 3  
5/23/79

BALANCE 3  
OPERATOR 84  
ACCEPTED WITHIN STANDARD DEVIATION OF THE PROCESS .02800 MG  
ACCEPTED BETWEEN STANDARD DEVIATION OF THE PROCESS .00000 MG

CALIBRATION DESIGN 62  
RESTRAINT VECTOR 1 1 1 0 0 0  
MASS CORRECTION OF RESTRAINT 9.05323 MG  
VOLUME OF WEIGHTS BEING USED IN RESTRAINT AT 21.94 C 126.17267 CM3  
SYSTEMATIC ERROR IN THE RESTRAINT .03800 MG  
3 STANDARD DEVIATION LIMIT FOR RANDOM ERROR AFFECTING RESTRAINT .05144 MG

CHECK STANDARD USED 4  
CHECK STANDARD VECTOR 0 0 0 0 1 0  
ACCEPTED MASS CORRECTION OF CHECK STANDARD .98830 MG  
REPORT VECTOR 1 1 1 1 0 0

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.92   | 21.96   | 21.94   |
| CORRECTED PRESSURE IN MM HG        | 736.920 | 736.580 | 736.750 |
| CORRECTED HUMIDITY IN PERCENT      | 40.00   | 40.00   | 40.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1560  | 1.1553  | 1.1556  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.92   | 21.96   |         |
| OBSERVED PRESSURE IN MM HG         | 736.920 | 736.580 |         |
| OBSERVED HUMIDITY IN PERCENT       | 40.00   | 40.00   |         |

| WEIGHTS BEING TESTED | NOMINAL VALUE G | DENSITY G/CM3 AT 20C | COEFFICIENT OF EXPANSION | ACCEPTED CORRECTION MG |
|----------------------|-----------------|----------------------|--------------------------|------------------------|
| 500G                 | 500.0000        | 7.9000               | .000045                  |                        |
| 300G                 | 300.0000        | 7.9530               | .000045                  |                        |
| 200G                 | 200.0000        | 7.9530               | .000045                  |                        |
| 100G                 | 100.0000        | 7.9000               | .000045                  |                        |
| S 100G               | 100.0000        | 7.9530               | .000045                  | .58830                 |
| SUM 100G             | 100.0000        | 7.9423               | .000045                  |                        |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 16  
SERIES 3  
5/23/79

BALANCE 3  
OPERATOR 84

CALIBRATION DESIGN 62

|   |    | GRAMS |     |     |     |     |     |
|---|----|-------|-----|-----|-----|-----|-----|
|   |    | 500   | 300 | 200 | 100 | 100 | 100 |
| A | 1  | +     | -   | -   | +   | -   | -   |
| A | 2  | +     | -   | -   | -   | +   | -   |
| A | 3  | +     | -   | -   | -   | -   | +   |
| A | 4  | +     | -   | -   | -   | -   | -   |
| A | 5  | +     | -   | -   | -   | -   | -   |
| A | 6  | -     | +   | -   | +   | -   | -   |
| A | 7  | -     | +   | -   | -   | +   | -   |
| A | 8  | -     | +   | -   | -   | -   | +   |
| A | 9  | -     | -   | +   | -   | -   | -   |
| A | 10 | -     | -   | +   | -   | -   | -   |
| A | 11 | -     | -   | +   | -   | -   | -   |
| R |    | +     | +   | +   |     |     |     |

OBSERVATIONS IN DIVISIONS  
DOUBLE SUBSTITUTION ONE PAN BALANCE

|   |    |         |         |         |         |
|---|----|---------|---------|---------|---------|
| A | 1  | 12.7800 | 10.6200 | 60.6300 | 62.8600 |
| A | 2  | 13.0200 | 12.5900 | 62.5400 | 62.9600 |
| A | 3  | 14.8600 | 10.7500 | 60.7300 | 64.8800 |
| A | 4  | 12.8600 | 10.6100 | 60.5400 | 62.8200 |
| A | 5  | 12.9100 | 13.5800 | 63.5400 | 62.8100 |
| A | 6  | 10.8700 | 13.4400 | 63.3300 | 60.8400 |
| A | 7  | 10.9400 | 13.3300 | 63.2600 | 60.8200 |
| A | 8  | 12.6900 | 11.4400 | 61.3800 | 62.6600 |
| A | 9  | 11.1800 | 11.7100 | 61.6800 | 61.1400 |
| A | 10 | 11.1800 | 13.5000 | 63.3800 | 61.1100 |
| A | 11 | 11.0900 | 13.4400 | 63.4000 | 60.9600 |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 17  
SERIES 3  
5/23/79

BALANCE 3  
OPERATOR 84

CALIBRATION DESIGN 62

SENSITIVITY WEIGHT  
MASS 49.98277 MG  
VOLUME .00301 CM3 AT 20 C  
COEFFICIENT OF EXPANSION .000020  
S\*=S-PV(S)= 49.97929 MG

|      | A(I)<br>(MG) | DELTA(I)<br>(MG) | AVERAGE<br>SENSITIVITY<br>(MG/DIV) | DRIFT(I)<br>(MG) | OBSERVED<br>SENSITIVITY<br>(MG/DIV) |
|------|--------------|------------------|------------------------------------|------------------|-------------------------------------|
| A 1  | 2.19570      | .00092           | 1.00032                            | .03501           | 1.00009                             |
| A 2  | .42514       | -.00556          | 1.00032                            | -.00500          | 1.00049                             |
| A 3  | 4.13132      | -.01335          | 1.00032                            | .02001           | 1.00039                             |
| A 4  | 2.26621      | .00950           | 1.00054                            | .01501           | 1.00129                             |
| A 5  | -.70038      | .00849           | 1.00054                            | -.03002          | .99979                              |
| A 6  | -2.53351     | -.00805          | 1.00139                            | .04006           | 1.00259                             |
| A 7  | -2.41835     | -.01675          | 1.00139                            | -.02503          | 1.00049                             |
| A 8  | 1.26676      | .01632           | 1.00139                            | .01502           | 1.00109                             |
| A 9  | -.53537      | -.03332          | 1.00069                            | -.00500          | 1.00009                             |
| A 10 | -2.29658     | .03149           | 1.00069                            | .02502           | 1.00249                             |
| A 11 | -2.39664     | -.00665          | 1.00069                            | -.04503          | .99949                              |

| ITEM<br>(G) | CORRECTION<br>(MG) | VOLUME<br>(AT T)<br>(CM3) | SYSTEMATIC<br>ERROR<br>(MG) | 3 S.D.<br>LIMIT<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|-------------|--------------------|---------------------------|-----------------------------|-------------------------|------------------------------|
| 500.0000    | 5.89889            | 63.29741                  | .01900                      | .03233                  | .05133                       |
| 300.0000    | 1.75036            | 37.72513                  | .01140                      | .02945                  | .04085                       |
| 200.0000    | 1.40395            | 25.15011                  | .00760                      | .02443                  | .03203                       |
| 100.0000    | 1.01957            | 12.65946                  | .00380                      | .03027                  | .03407                       |
| 100.0000    | .58400             | 12.57509                  | .00380                      | .03027                  | .03407                       |
| 100.0000    | 2.82980            | 12.59220                  | .00380                      | .03027                  | .03407                       |

TEMPERATURE T= 21.94 C

RESTRAINT FOR FOLLOWING SERIES

|                            |   |   |   |          |         |    |
|----------------------------|---|---|---|----------|---------|----|
| RESTRAINT VECTOR           | 0 | 0 | 0 | 0        | 0       | 1  |
| MASS CORRECTION            |   |   |   |          | 2.82980 | MG |
| VOLUME AT 20 C             |   |   |   | 12.59110 | CM3     |    |
| SYSTEMATIC ERROR           |   |   |   | .00380   | MG      |    |
| 3 STANDARD DEVIATION LIMIT |   |   |   | .03027   | MG      |    |

X Y Z CORPORATION  
 SOMEWHERE, U.S.A.  
 SET OF MASS STANDARDS : 5KG - 100MG  
 TEST NUMBER 654321

PAGE 18  
 SERIES 3  
 5/23/79

BALANCE 3  
 OPERATOR 84

CALIBRATION DESIGN 62

| SUM<br>(G) | WEIGHTS USED FOR THE LINEAR COMBINATIONS<br>GRAMS |     |     |     |     |     |
|------------|---------------------------------------------------|-----|-----|-----|-----|-----|
|            | 500                                               | 300 | 200 | 100 | 100 | 100 |
| 600        | +                                                 |     |     | +   |     |     |

VALUES AND UNCERTAINTIES FOR COMBINATIONS OF WEIGHTS  
 UNCERTAINTY IS 3 STANDARD DEVIATION LIMIT PLUS ALLOWANCE FOR  
 (SYSTEMATIC ERROR.)

| SUM<br>(G) | CORR<br>(MG) | SYSTEMATIC<br>(MG) | 3 S.D.<br>ERROR<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|------------|--------------|--------------------|-------------------------|------------------------------|
| 600        | 6.91847      | .02280             | .04750                  | .07030                       |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 19  
SERIES 3  
5/23/79

BALANCE 3  
OPERATOR 84  
MAXIMUM LOAD 600.0000 G  
STARTING RESTRAINT NUMBER 80

CALIBRATION DESIGN 62

PRECISION CONTROL

OBSERVED STANDARD DEVIATION OF THE PROCESS .02284 MG  
ACCEPTED STANDARD DEVIATION OF THE PROCESS .02800 MG  
DEGREES OF FREEDOM 6  
F RATIO .665

F RATIO IS LESS THAN 2.81 (CRITICAL VALUE FOR PROBABILITY = .01).  
THEREFORE THE STANDARD DEVIATION IS IN CONTROL.

CHECK STANDARD VECTOR 0 0 0 0 1 0  
CHECK STANDARD USED 4  
ACCEPTED MASS CORRECTION OF CHECK STANDARD .98830 MG  
OBSERVED CORRECTION OF CHECK STANDARD .98400 MG  
STANDARD DEVIATION OF THE OBSERVED CORRECTION .01009 MG  
T VALUE -.43

ABSOLUTE VALUE OF T IS LESS THAN 3.  
THEREFORE CHECK STANDARD IS IN CONTROL.

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.92   | 21.96   | 21.94   |
| CORRECTED PRESSURE IN MM HG        | 736.920 | 736.580 | 736.750 |
| CORRECTED HUMIDITY IN PERCENT      | 40.00   | 40.00   | 40.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1560  | 1.1553  | 1.1557  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.92   | 21.96   |         |
| OBSERVED PRESSURE IN MM HG         | 736.920 | 736.580 |         |
| OBSERVED HUMIDITY IN PERCENT       | 40.00   | 40.00   |         |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 20  
SERIES 4  
5/17/79

BALANCE 5  
OPERATOR 84  
ACCEPTED WITHIN STANDARD DEVIATION OF THE PROCESS .01200 MG  
ACCEPTED BETWEEN STANDARD DEVIATION OF THE PROCESS .00000 MG

CALIBRATION DESIGN 62  
RESTRAINT VECTOR 1 1 1 0 0 0  
MASS CORRECTION OF RESTRAINT 2.82980 MG  
VOLUME OF WEIGHTS BEING USED IN RESTRAINT AT 21.97 C 12.59222 CM3  
SYSTEMATIC ERROR IN THE RESTRAINT .00380 MG  
3 STANDARD DEVIATION LIMIT FOR RANDOM ERROR AFFECTING RESTRAINT .03027 MG

CHECK STANDARD USED 6  
CHECK STANDARD VECTOR 0 0 0 0 1 0  
ACCEPTED MASS CORRECTION OF CHECK STANDARD .07850 MG  
REPORT VECTOR 1 1 1 1 0 0

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.99   | 21.96   | 21.97   |
| CORRECTED PRESSURE IN MM HG        | 746.600 | 746.000 | 746.300 |
| CORRECTED HUMIDITY IN PERCENT      | 31.00   | 31.00   | 31.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1720  | 1.1712  | 1.1716  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.99   | 21.96   |         |
| OBSERVED PRESSURE IN MM HG         | 746.600 | 746.000 |         |
| OBSERVED HUMIDITY IN PERCENT       | 31.00   | 31.00   |         |

| WEIGHTS BEING TESTED | NOMINAL VALUE G | DENSITY G/CM3 AT 20C | COEFFICIENT OF EXPANSION | ACCEPTED CORRECTION MG |
|----------------------|-----------------|----------------------|--------------------------|------------------------|
| 50G                  | 50.0000         | 7.9530               | .000045                  |                        |
| 30G                  | 30.0000         | 7.9530               | .000045                  |                        |
| 20G                  | 20.0000         | 7.9000               | .000045                  |                        |
| 10G                  | 10.0000         | 7.9530               | .000045                  |                        |
| 5 10G                | 10.0000         | 7.9530               | .000045                  | .07850                 |
| SUM 10G              | 10.0000         | 7.9264               | .000045                  |                        |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMER 654321

PAGE 21  
SERIES 4  
5/17/79

BALANCE 5  
OPERATOR 84

| CALIBRATION DESIGN |   | 62 |    |    |    |    |    |
|--------------------|---|----|----|----|----|----|----|
| GRAMS              |   | 50 | 30 | 20 | 10 | 10 | 10 |
| A 1                | + | -  | -  | -  | +  | -  | -  |
| A 2                | + | -  | -  | -  | -  | +  | -  |
| A 3                | + | -  | -  | -  | -  | -  | +  |
| A 4                | + | -  | -  | -  | -  | -  | -  |
| A 5                | + | -  | -  | -  | -  | -  | -  |
| A 6                |   | +  | -  | +  | -  | -  | -  |
| A 7                |   | +  | -  | -  | +  | -  | -  |
| A 8                |   | +  | -  | -  | -  | -  | +  |
| A 9                |   |    | +  | +  | -  | -  | -  |
| A 10               |   |    | +  | +  | -  | -  | -  |
| A 11               |   |    | +  | +  | -  | -  | -  |
| R                  | + | +  | +  |    |    |    |    |

OBSERVATIONS IN DIVISIONS  
DOUBLE SUBSTITUTION ONE PAN BALANCE

|      |        |        |         |         |
|------|--------|--------|---------|---------|
| A 1  | 6.1800 | 4.6900 | 54.7100 | 56.2000 |
| A 2  | 6.1700 | 4.7200 | 54.7100 | 56.1500 |
| A 3  | 6.1400 | 4.7600 | 54.7600 | 56.1400 |
| A 4  | 6.1900 | 4.7400 | 54.7200 | 56.1800 |
| A 5  | 6.1800 | 4.4700 | 54.4500 | 56.1800 |
| A 6  | 4.8100 | 4.4400 | 54.4400 | 54.8100 |
| A 7  | 4.7900 | 4.4900 | 54.4700 | 54.7500 |
| A 8  | 4.7600 | 4.5100 | 54.4900 | 54.7400 |
| A 9  | 4.5200 | 4.5500 | 54.5200 | 54.4900 |
| A 10 | 4.4800 | 4.5200 | 54.5000 | 54.4700 |
| A 11 | 4.4700 | 4.4700 | 54.4600 | 54.4700 |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 22  
SERIES 4  
5/17/79

BALANCE 5  
OPERATOR 84

CALIBRATION DESIGN 62

SENSITIVITY WEIGHT  
MASS 49.98277 MG  
VOLUME .00301 CM3 AT 20 C  
COEFFICIENT OF EXPANSION .000020  
S=S-PV(S)= 49.97924 MG

|      | A(I)<br>(MG) | DELTA(I)<br>(MG) | AVERAGE<br>SENSITIVITY<br>(MG/DIV) | DRIFT(I)<br>(MG) | OBSERVED<br>SENSITIVITY<br>(MG/DIV) |
|------|--------------|------------------|------------------------------------|------------------|-------------------------------------|
| A 1  | 1.48923      | .00353           | .99948                             | .00000           | .99919                              |
| A 2  | 1.44426      | -.00895          | .99948                             | -.00500          | .99968                              |
| A 3  | 1.37929      | -.00582          | .99948                             | .00000           | .99958                              |
| A 4  | 1.45520      | .01386           | 1.00013                            | .00500           | 1.00008                             |
| A 5  | 1.72023      | -.00262          | 1.00013                            | .01000           | 1.00018                             |
| A 6  | .36992       | .00046           | .99978                             | .00000           | .99958                              |
| A 7  | .28994       | .00920           | .99978                             | -.01000          | .99978                              |
| A 8  | .24995       | -.00704          | .99978                             | .00000           | .99998                              |
| A 9  | -.03000      | .01358           | 1.00005                            | -.00000          | 1.00018                             |
| A 10 | -.03500      | -.00330          | 1.00005                            | .00500           | 1.00008                             |
| A 11 | .00500       | -.00766          | 1.00005                            | .00500           | .99988                              |

| ITEM<br>(G) | CORRECTION<br>(MG) | VOLUME<br>(AT T)<br>(CM3) | SYSTEMATIC<br>ERROR<br>(MG) | 3 S.D.<br>LIMIT<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|-------------|--------------------|---------------------------|-----------------------------|-------------------------|------------------------------|
| 50.0000     | 2.12579            | 6.28776                   | .00190                      | .01731                  | .01921                       |
| 30.0000     | .53569             | 3.77256                   | .00114                      | .01407                  | .01521                       |
| 20.0000     | .16831             | 2.53189                   | .00076                      | .01126                  | .01202                       |
| 10.0000     | .11825             | 1.25751                   | .00038                      | .01314                  | .01352                       |
| 10.0000     | .07388             | 1.25751                   | .00038                      | .01314                  | .01352                       |
| 10.0000     | .06695             | 1.26173                   | .00038                      | .01314                  | .01352                       |

TEMPERATURE T= 21.97 C

CONSTRAINT FOR FOLLOWING SERIES  
CONSTRAINT VECTOR 0 0 0 0 0 1  
CORRECTION .06695 MG  
VOLUME AT 20 C 1.26161 CM3  
SYSTEMATIC ERROR .00038 MG  
STANDARD DEVIATION LIMIT .01314 MG



X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 23  
SERIES 4  
5/17/79

BALANCE 5  
OPERATOR 84  
MAXIMUM LOAD 60.0000 G  
STARTING RESTRAINT NUMBER 80

CALIBRATION DESIGN 62

PRECISION CONTROL

OBSERVED STANDARD DEVIATION OF THE PROCESS .01091 MG  
ACCEPTED STANDARD DEVIATION OF THE PROCESS .01200 MG  
DEGREES OF FREEDOM 6  
F RATIO .826

F RATIO IS LESS THAN 2.81 (CRITICAL VALUE FOR PROBABILITY = .01).  
THEREFORE THE STANDARD DEVIATION IS IN CONTROL.

CHECK STANDARD VECTOR 0 0 0 0 1 0  
CHECK STANDARD USED 6  
ACCEPTED MASS CORRECTION OF CHECK STANDARD .07850 MG  
OBSERVED CORRECTION OF CHECK STANDARD .07388 MG  
STANDARD DEVIATION OF THE OBSERVED CORRECTION .00438 MG  
T VALUE -1.06

ABSOLUTE VALUE OF T IS LESS THAN 3.  
THEREFORE CHECK STANDARD IS IN CONTROL.

| TEST CONDITIONS                    | BEFORE  | AFTER   |
|------------------------------------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.99   | 21.96   |
| CORRECTED PRESSURE IN MM HG        | 746.600 | 746.000 |
| CORRECTED HUMIDITY IN PERCENT      | 31.00   | 31.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1720  | 1.1712  |
| TEMPERATURE CORRECTION             | .000    | .000    |
| PRESSURE CORRECTION                | .000    | .000    |
| HUMIDITY CORRECTION                | .00     | .00     |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.99   | 21.96   |
| OBSERVED PRESSURE IN MM HG         | 746.600 | 746.000 |
| OBSERVED HUMIDITY IN PERCENT       | 31.00   | 31.00   |



X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 24  
SERIES 5  
5/18/79

BALANCE 7  
OPERATOR 84  
ACCEPTED WITHIN STANDARD DEVIATION OF THE PROCESS .00170 MG  
ACCEPTED BETWEEN STANDARD DEVIATION OF THE PROCESS .00000 MG

CALIBRATION DESIGN 62  
RESTRAINT VECTOR 1 1 1 0 0 0  
MASS CORRECTION OF RESTRAINT .06695 MG  
VOLUME OF WEIGHTS BEING USED IN RESTRAINT AT 21.90 C 1.26172 CM3  
SYSTEMATIC ERROR IN THE RESTRAINT .00038 MG  
3 STANDARD DEVIATION LIMIT FOR RANDOM ERROR AFFECTING RESTRAINT .01314 MG

CHECK STANDARD USED 8  
CHECK STANDARD VECTOR 0 0 0 0 1 0  
ACCEPTED MASS CORRECTION OF CHECK STANDARD -.07920 MG  
REPORT VECTOR 1 1 1 1 0 0

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.92   | 21.88   | 21.90   |
| CORRECTED PRESSURE IN MM HG        | 743.280 | 742.820 | 743.050 |
| CORRECTED HUMIDITY IN PERCENT      | 35.00   | 35.00   | 35.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1666  | 1.1660  | 1.1663  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.92   | 21.88   |         |
| OBSERVED PRESSURE IN MM HG         | 743.280 | 742.820 |         |
| OBSERVED HUMIDITY IN PERCENT       | 35.00   | 35.00   |         |

| WEIGHTS BEING TESTED | NOMINAL VALUE G | DENSITY G/CM3 AT 20C | COEFFICIENT OF EXPANSION | ACCEPTED CORRECTION MG |
|----------------------|-----------------|----------------------|--------------------------|------------------------|
| 5G                   | 5.0000          | 7.9000               | .000045                  |                        |
| 3G                   | 3.0000          | 7.9530               | .000045                  |                        |
| 2G                   | 2.0000          | 7.9530               | .000045                  |                        |
| 1G                   | 1.0000          | 7.9000               | .000045                  |                        |
| S 1G                 | 1.0000          | 7.9530               | .000045                  |                        |
| SUM 1G               | 1.0000          | 16.6000              | .000020                  | -.07920                |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 25  
SERIES S  
5/18/79

BALANCE 7  
OPERATOR 84

| CALIBRATION DESIGN |   | 62 |   |   |   |   |   |
|--------------------|---|----|---|---|---|---|---|
| GRAMS              |   | 5  | 3 | 2 | 1 | 1 | 1 |
| A 1                | + | -  | - | - | + | - | - |
| A 2                | + | -  | - | - | - | + | - |
| A 3                | + | -  | - | - | - | - | + |
| A 4                | + | -  | - | - | - | - | - |
| A 5                | + | -  | - | - | - | - | - |
| A 6                |   | +  | - | - | + | - | - |
| A 7                |   | +  | - | - | - | + | - |
| A 8                |   | +  | - | - | - | - | + |
| A 9                |   |    | + | - | - | - | - |
| A 10               |   |    | + | - | - | - | - |
| A 11               |   |    | + | - | - | - | - |
| R                  | + | +  | + |   |   |   |   |

OBSERVATIONS IN DIVISIONS  
DOUBLE SUBSTITUTION ONE PAN BALANCE

|      |        |        |        |        |
|------|--------|--------|--------|--------|
| A 1  | 1.0840 | .9260  | 5.9270 | 6.0840 |
| A 2  | .9780  | .9390  | 5.9310 | 5.9720 |
| A 3  | .9900  | 1.0210 | 6.0210 | 5.9870 |
| A 4  | 1.0680 | 1.0130 | 6.0090 | 6.0670 |
| A 5  | 1.0680 | .8750  | 5.8700 | 6.0630 |
| A 6  | 1.0620 | .8620  | 5.8570 | 6.0530 |
| A 7  | .9590  | .9660  | 5.9570 | 5.9480 |
| A 8  | .9690  | .9480  | 5.9420 | 5.9600 |
| A 9  | 1.0160 | .9760  | 5.9660 | 6.0060 |
| A 10 | 1.0160 | .9870  | 5.9810 | 6.0060 |
| A 11 | 1.0120 | .8810  | 5.8720 | 6.0040 |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 26  
SERIES 5  
5/18/79

BALANCE 7  
OPERATOR 84

CALIBRATION DESIGN 62

SENSITIVITY WEIGHT  
MASS 5.00171 MG  
VOLUME .00185 CM3 AT 20 C  
COEFFICIENT OF EXPANSION .000069  
S\*=S-PV(S)= 4.99955 MG

|      | A(I)<br>(MG) | DELTA(I)<br>(MG) | AVERAGE<br>SENSITIVITY<br>(MG/DIV) | DRIFT(I)<br>(MG) | OBSERVED<br>SENSITIVITY<br>(MG/DIV) |
|------|--------------|------------------|------------------------------------|------------------|-------------------------------------|
| A 1  | .15755       | -.00114          | 1.00031                            | -.00050          | .99961                              |
| A 2  | .04001       | -.00146          | 1.00031                            | .00100           | 1.00171                             |
| A 3  | -.03251      | .00082           | 1.00031                            | -.00150          | .99961                              |
| A 4  | .05655       | .00094           | 1.00096                            | .00150           | 1.00101                             |
| A 5  | .19319       | .00085           | 1.00096                            | .00000           | 1.00091                             |
| A 6  | .19819       | -.00001          | 1.00094                            | -.00200          | 1.00051                             |
| A 7  | -.00801      | -.00004          | 1.00094                            | -.00100          | 1.00151                             |
| A 8  | .01952       | -.00079          | 1.00094                            | -.00150          | 1.00081                             |
| A 9  | .04006       | -.00155          | 1.00148                            | .00000           | 1.00191                             |
| A 10 | .02704       | -.00043          | 1.00148                            | -.00200          | 1.00071                             |
| A 11 | .13169       | .00114           | 1.00148                            | .00050           | 1.00181                             |

| ITEM<br>(G) | CORRECTION<br>(MG) | VOLUME<br>(AT T)<br>(CM3) | SYSTEMATIC<br>ERROR<br>(MG) | 3 S.D.<br>LIMIT<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|-------------|--------------------|---------------------------|-----------------------------|-------------------------|------------------------------|
| 5.0000      | .06375             | .63297                    | .00019                      | .00668                  | .00687                       |
| 3.0000      | .01669             | .37725                    | .00011                      | .00422                  | .00434                       |
| 2.0000      | -.01348            | .25150                    | .00008                      | .00295                  | .00303                       |
| 1.0000      | .02458             | .12660                    | .00004                      | .00224                  | .00228                       |
| 1.0000      | -.07910            | .12574                    | .00004                      | .00224                  | .00228                       |
| 1.0000      | -.14136            | .06023                    | .00004                      | .00224                  | .00228                       |

TEMPERATURE T= 21.90 C

RESTRAINT FOR FOLLOWING SERIES

|                            |   |   |   |   |   |            |
|----------------------------|---|---|---|---|---|------------|
| RESTRAINT VECTOR           | 0 | 0 | 0 | 0 | 0 | 1          |
| MASS CORRECTION            |   |   |   |   |   | -.14136 MG |
| VOLUME AT 20 C             |   |   |   |   |   | .06023 CM3 |
| SYSTEMATIC ERROR           |   |   |   |   |   | .00004 MG  |
| 3 STANDARD DEVIATION LIMIT |   |   |   |   |   | .00224 MG  |

X Y Z CORPORATION  
 SOMEWHERE, U.S.A.  
 SET OF MASS STANDARDS : 5KG - 100MG  
 TEST NUMBER 654321

PAGE 27  
 SERIES 5  
 5/18/79

BALANCE 7  
 OPERATOR 84

CALIBRATION DESIGN 62

| SUM<br>(G) | WEIGHTS USED FOR THE LINEAR COMBINATIONS<br>GRAMS |   |   |   |   |
|------------|---------------------------------------------------|---|---|---|---|
|            | 5                                                 | 3 | 2 | 1 | 1 |
| 6          | +                                                 |   |   | + |   |

VALUES AND UNCERTAINTIES FOR COMBINATIONS OF WEIGHTS  
 (UNCERTAINTY IS 3 STANDARD DEVIATION LIMIT PLUS ALLOWANCE FOR  
 SYSTEMATIC ERROR.)

| SUM<br>(G) | CORR<br>(MG) | SYSTEMATIC<br>(MG) | 3 S.D.<br>ERROR<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|------------|--------------|--------------------|-------------------------|------------------------------|
| 6          | .08873       | .00023             | .00818                  | .00841                       |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 28  
SERIES 5  
5/18/79

BALANCE 7  
OPERATOR 84  
MAXIMUM LCAD 6.0000 G  
STARTING RESTRAINT NUMBER 80

CALIBRATION DESIGN 62

PRECISION CCNTRCL

OBSERVED STANDARD DEVIATION OF THE PROCESS .00131 MG  
ACCEPTED STANDARD DEVIATION OF THE PROCESS .00170 MG  
DEGREES OF FREEDOM 6  
F RATIO .591

F RATIO IS LESS THAN 2.81 (CRITICAL VALUE FOR PROBABILITY = .01).  
THEREFORE THE STANDARD DEVIATION IS IN CONTROL.

CHECK STANDARD VECTOR 0 0 0 0 1 0  
CHECK STANDARD USED 8  
ACCEPTED MASS CORRECTION OF CHECK STANDARD -.07920 MG  
OBSERVED CORRECTION OF CHECK STANDARD -.07910 MG  
STANDARD DEVIATION OF THE OBSERVED CORRECTION .00075 MG  
T VALUE .14

ABSOLUTE VALUE OF T IS LESS THAN 3.  
THEREFORE CHECK STANDARD IS IN CONTROL.

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 21.92   | 21.88   | 21.90   |
| CORRECTED PRESSURE IN MM HG        | 743.280 | 742.820 | 743.050 |
| CORRECTED HUMIDITY IN PERCENT      | 35.00   | 35.00   | 35.00   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1666  | 1.1660  | 1.1663  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 21.92   | 21.88   |         |
| OBSERVED PRESSURE IN MM HG         | 743.280 | 742.820 |         |
| OBSERVED HUMIDITY IN PERCENT       | 35.00   | 35.00   |         |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 29  
SERIES 6  
5/18/79

BALANCE 7  
OPERATOR 84

ACCEPTED WITHIN STANDARD DEVIATION OF THE PROCESS .00050 MG  
ACCEPTED BETWEEN STANDARD DEVIATION OF THE PROCESS .00000 MG

CALIBRATION DESIGN 62

RESTRAINT VECTOR 1 1 1 0 0 0

MASS CORRECTION OF RESTRAINT

VOLUME OF WEIGHTS BEING USED IN RESTRAINT AT 22.60 C

SYSTEMATIC ERROR IN THE RESTRAINT

3 STANDARD DEVIATION LIMIT FOR RANDOM ERROR AFFECTING RESTRAINT

-.14136 MG  
.06024 CM3  
.00004 MG  
.00224 MG

CHECK STANDARD USED 8

CHECK STANDARD VECTOR 0 0 0 0 1 0

ACCEPTED MASS CORRECTION OF CHECK STANDARD -.02628 MG

REPORT VECTOR 1 1 1 1 0 0

TEST CONCTIONS

|                                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 22.21   | 22.99   | 22.60   |
| CORRECTED PRESSURE IN MM HG        | 742.520 | 741.860 | 742.190 |
| CORRECTED HUMIDITY IN PERCENT      | 36.00   | 35.00   | 35.50   |
| COMPUTED AIR DENSITY IN MG/CM3     | 1.1640  | 1.1595  | 1.1620  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 22.21   | 22.99   |         |
| OBSERVED PRESSURE IN MM HG         | 742.520 | 741.860 |         |
| OBSERVED HUMIDITY IN PERCENT       | 36.00   | 35.00   |         |

| WEIGHTS BEING TESTED | NOMINAL VALUE G | DENSITY G/CM3 AT 20C | COEFFICIENT OF EXPANSION | ACCEPTED CORRECTION MG |
|----------------------|-----------------|----------------------|--------------------------|------------------------|
| 500MG                | .5000           | 16.6000              | .000020                  |                        |
| 300MG                | .3000           | 16.6000              | .000020                  |                        |
| 200MG                | .2000           | 16.6000              | .000020                  |                        |
| 100MG                | .1000           | 16.6000              | .000020                  |                        |
| S 100MG              | .1000           | 16.6000              | .000020                  | -.02628                |
| SUM 100MG            | .1000           | 8.1788               | .000049                  |                        |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 30  
SERIES 6  
5/18/79

BALANCE 7  
OPERATOR 84

CALIBRATION DESIGN 62

|   |    | MG  |     |     |     |     |     |
|---|----|-----|-----|-----|-----|-----|-----|
|   |    | 500 | 300 | 200 | 100 | 100 | 100 |
| A | 1  | +   | -   | -   | +   | -   | -   |
| A | 2  | +   | -   | -   | -   | +   | -   |
| A | 3  | +   | -   | -   | -   | -   | +   |
| A | 4  | +   | -   | -   | -   | -   | -   |
| A | 5  | +   | -   | -   | -   | -   | -   |
| A | 6  | -   | +   | -   | +   | -   | -   |
| A | 7  | -   | +   | -   | -   | +   | -   |
| A | 8  | -   | +   | -   | -   | -   | +   |
| A | 9  | -   | -   | +   | -   | -   | -   |
| A | 10 | -   | -   | +   | -   | -   | -   |
| A | 11 | -   | -   | +   | -   | -   | -   |
| R |    | +   | +   | +   |     |     |     |

OBSERVATIONS IN DIVISIONS  
DOUBLE SUBSTITUTION ONE PAN BALANCE

|      |          |          |          |          |
|------|----------|----------|----------|----------|
| A 1  | 296.4000 | 224.0000 | 726.2000 | 801.2000 |
| A 2  | 285.5000 | 240.3000 | 741.3000 | 788.9000 |
| A 3  | 300.5000 | 241.6000 | 744.1000 | 803.7000 |
| A 4  | 312.9000 | 253.5000 | 754.8000 | 815.5000 |
| A 5  | 312.7000 | 239.8000 | 741.0000 | 815.6000 |
| A 6  | 303.8000 | 250.5000 | 751.5000 | 805.7000 |
| A 7  | 290.2000 | 264.6000 | 765.4000 | 791.7000 |
| A 8  | 303.6000 | 252.3000 | 753.6000 | 806.0000 |
| A 9  | 276.7000 | 302.6000 | 804.1000 | 779.3000 |
| A 10 | 276.7000 | 315.5000 | 817.1000 | 779.2000 |
| A 11 | 276.6000 | 301.0000 | 802.6000 | 778.6000 |



X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMER 654321

PAGE 31  
SERIES 6  
5/18/79

BALANCE 7  
OPERATOR 84

CALIBRATION DESIGN 62

SENSITIVITY WEIGHT  
MASS .50156 MG  
VOLUME .00186 CM3 AT 20 C  
COEFFICIENT OF EXPANSION .000069  
S\*=S-PV(S)= .49940 MG

|      | A(I)<br>(MG) | DELTA(I)<br>(MG) | AVERAGE<br>SENSITIVITY<br>(MG/DIV) | DRIFT(I)<br>(MG) | OBSERVED<br>SENSITIVITY<br>(MG/DIV) |
|------|--------------|------------------|------------------------------------|------------------|-------------------------------------|
| A 1  | .07347       | -.00002          | .00100                             | .00130           | .00100                              |
| A 2  | .04626       | -.00026          | .00100                             | .00120           | .00100                              |
| A 3  | .05907       | .00029           | .00100                             | .00035           | .00099                              |
| A 4  | .05992       | .00032           | .00100                             | .00065           | .00100                              |
| A 5  | .07359       | -.00033          | .00100                             | .00085           | .00100                              |
| A 6  | .05362       | .00018           | .00100                             | .00045           | .00100                              |
| A 7  | .02589       | .00024           | .00100                             | .00035           | .00100                              |
| A 8  | .05173       | -.00008          | .00100                             | .00055           | .00100                              |
| A 9  | -.02526      | -.00004          | .00100                             | .00055           | .00100                              |
| A 10 | -.03821      | .00009           | .00100                             | .00045           | .00100                              |
| A 11 | -.02411      | .00029           | .00100                             | .00020           | .00100                              |

| ITEM<br>(G) | CORRECTION<br>(MG) | VOLUME<br>(AT T)<br>(CM3) | SYSTEMATIC<br>ERROR<br>(MG) | 3 S.D.<br>LIMIT<br>(MG) | UNCERTAINTY<br>LIMIT<br>(MG) |
|-------------|--------------------|---------------------------|-----------------------------|-------------------------|------------------------------|
| .5000       | -.04088            | .03012                    | .00002                      | .00117                  | .00119                       |
| .3000       | -.03697            | .01807                    | .00001                      | .00081                  | .00082                       |
| .2000       | -.06351            | .01204                    | .00001                      | .00060                  | .00060                       |
| .1000       | -.01219            | .00602                    | .00000                      | .00058                  | .00058                       |
| .1000       | -.02609            | .00602                    | .00000                      | .00058                  | .00058                       |
| .1000       | -.00580            | .01223                    | .00000                      | .00058                  | .00058                       |

TEMPERATURE T= 22.60 C

X Y Z CCFPCRATION  
 SOMEWHERE, U.S.A.  
 SET OF MASS STANDARDS : 5KG - 100MG  
 TEST NUMBER 654321

PAGE 32  
 SERIES 6  
 5/18/79

BALANCE 7  
 OPERATOR 84  
 MAXIMUM LCAD .6000 G  
 STARTING RESTRAINT NUMBER 80

CALIBRATION DESIGN 62

PRECISION CCNTRCL

OBSERVED STANDARD DEVIATION OF THE PROCESS .00030 MG  
 ACCEPTED STANDARD DEVIATION OF THE PROCESS .00050 MG  
 DEGREES OF FREEDOM 6  
 F RATIO .370

F RATIO IS LESS THAN 2.81 (CRITICAL VALUE FOR PROBABILITY = .01).  
 THEREFORE THE STANDARD DEVIATION IS IN CONTROL.

CHECK STANDARD VECTOR 0 0 0 0 1 0  
 CHECK STANDARD USED 8  
 ACCEPTED MASS CORRECTION OF CHECK STANDARD -.02628 MG  
 OBSERVED CORRECTION OF CHECK STANDARD -.02609 MG  
 STANDARD DEVIATION OF THE OBSERVED CORRECTION .00019 MG  
 T VALUE .97

ABSOLUTE VALUE OF T IS LESS THAN 3.  
 THEREFORE CHECK STANDARD IS IN CONTROL.

| TEST CONDITIONS                    | BEFORE  | AFTER   | AVERAGE |
|------------------------------------|---------|---------|---------|
| CORRECTED TEMPERATURE IN DEGREES C | 22.21   | 22.99   | 22.60   |
| CORRECTED PRESSURE IN MM HG        | 742.520 | 741.860 | 742.190 |
| CORRECTED HUMIDITY IN PERCENT      | 36.00   | 35.00   | 35.50   |
| CORRECTED AIR DENSITY IN MG/CM3    | 1.1640  | 1.1595  | 1.1620  |
| TEMPERATURE CORRECTION             | .000    | .000    |         |
| PRESSURE CORRECTION                | .000    | .000    |         |
| HUMIDITY CORRECTION                | .00     | .00     |         |
| OBSERVED TEMPERATURE IN DEGREES C  | 22.21   | 22.99   |         |
| OBSERVED PRESSURE IN MM HG         | 742.520 | 741.860 |         |
| OBSERVED HUMIDITY IN PERCENT       | 36.00   | 35.00   |         |

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 33  
5/18/79

#### SUMMARY

FOR CONVENIENCE, THE RESULTS OF THIS WORK ARE SUMMARIZED IN TABLES I AND II. THE VALUES ASSIGNED ARE WITH REFERENCE TO THE STANDARDS IDENTIFIED ON THE DATA SHEETS. THE UNCERTAINTY FIGURE IS AN EXPRESSION OF THE OVERALL UNCERTAINTY USING THREE STANDARD DEVIATIONS AS A LIMIT TO THE EFFECT OF RANDOM ERRORS OF THE MEASUREMENT ASSOCIATED WITH THE MEASUREMENT PROCESSES. THE MAGNITUDE OF SYSTEMATIC ERRORS FROM SOURCES OTHER THAN THE USE OF ACCEPTED VALUES FOR CERTAIN STARTING STANDARDS ARE CONSIDERED NEGLIGIBLE. IT SHOULD BE NOTED THAT THE MAGNITUDE OF THE UNCERTAINTY REFLECTS THE PERFORMANCE OF THE MEASUREMENT PROCESS USED TO ESTABLISH THESE VALUES. THE MASS UNIT, AS REALIZABLE IN ANOTHER MEASUREMENT PROCESS, WILL BE UNCERTAIN BY AN AMOUNT WHICH IS A COMBINATION OF THE UNCERTAINTY OF THIS PROCESS AND THE PROCESS IN WHICH THESE STANDARDS ARE USED.

THE ESTIMATED MASS VALUES LISTED IN TABLE I ARE BASED ON AN EXPLICIT TREATMENT OF DISPLACEMENT VOLUMES, E.G., 'TRUE MASS', 'MASS IN VACUO', MASS IN THE NEWTONIAN SENSE. THE DISPLACEMENT VOLUME ASSOCIATED WITH EACH VALUE IS LISTED AS WELL AS THE VOLUMETRIC COEFFICIENT OF EXPANSION. THESE VALUES SHOULD BE USED, TOGETHER WITH APPROPRIATE CORRECTION FOR THE BUOYANT EFFECTS OF THE ENVIRONMENT, TO ESTABLISH CONSISTENT MASS VALUES FOR OBJECTS WHICH DIFFER SIGNIFICANTLY IN DENSITY AND/OR FOR MEASUREMENTS WHICH MUST BE MADE IN DIFFERING ENVIRONMENTS. THE RELATION  $1 \text{ LB AVOG} = .45359237 \text{ KG}$  IS USED AS REQUIRED.

THE ESTIMATED MASS VALUES LISTED IN TABLE II ARE BASED ON AN IMPLICIT TREATMENT OF DISPLACEMENT VOLUMES, E.G., 'APPARENT MASS', 'APPARENT MASS VERSUS BRASS', 'APPARENT MASS VERSUS DENSITY 8.0'. THE VALUES ARE LISTED AS CORRECTIONS TO BE APPLIED TO THE LISTED NOMINAL VALUE (A POSITIVE CORRECTION INDICATES THAT THE MASS IS LARGER THAN THE STATED NOMINAL VALUE BY THE AMOUNT OF THE CORRECTION). THESE VALUES ARE COMPUTED FROM THE VALUES BASED ON AN EXPLICIT TREATMENT OF DISPLACEMENT VOLUMES USING THE FOLLOWING DEFINING RELATIONS AND ARE UNCERTAIN BY THE AMOUNT SHOWN IN TABLE I.

THE ADJUSTMENT OF WEIGHTS TO MINIMIZE THE DEVIATION FROM NOMINAL ON THE BASIS OF 'NORMAL BRASS' (IN ACCORDANCE WITH CCR. A BELOW) IS WIDESPREAD IN THIS COUNTRY AND IN MANY PARTS OF THE WORLD. VALUES STATED ON EITHER BASIS ARE INTERNALLY CONSISTENT AND DEFINITE. THERE IS, HOWEVER, A SYSTEMATIC DIFFERENCE BETWEEN THE VALUES ASSIGNED ON EACH BASIS. THE VALUE ON THE BASIS OF 'DENSITY 8.0' BEING 7 MICROGRAMS/GRAM LARGER THAN THE VALUE ON THE BASIS OF 'NORMAL BRASS'. THIS SYSTEMATIC DIFFERENCE IS CLEARLY DETECTABLE ON MANY DIRECT READING BALANCES.

CORRECTION A - 'APPARENT VERSUS BRASS' OR 'WEIGHT IN AIR' AGAINST BRASS' IS DETERMINED BY A HYPOTHETICAL WEIGHING OF THE OBJECT IN AIR AT 20 CELSIUS IN AIR HAVING A DENSITY OF 1.2 MG/CM<sup>3</sup>, WITH A STANDARD HAVING A DENSITY OF 8.4 G/CM<sup>3</sup> AT 0 CELSIUS WHOSE COEFFICIENT OF VOLUME EXPANSION IS 0.000054 PER CELSIUS, AND WHOSE VALUE IS

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMEER 654321

PAGE 34  
5/18/79

ON ITS TRUE MASS OR WEIGHT IN  
VACUO.

CORRECTION B - 'APPARENT MASS  
VERSUS DENSITY 8.0' IS DETERMINED  
BY A HYPOTHETICAL WEIGHING OF THE

WEIGHT, IN AIR HAVING A DENSITY OF  
1.2 MG/CM<sup>3</sup>, WITH A STANDARD HAVING  
A DENSITY OF 8.0 G/CM<sup>3</sup> AT 20  
CELSIUS, AND WHOSE VALUE IS BASED  
ON ITS TRUE MASS OR WEIGHT IN  
VACUO.

X Y Z CORPORATION  
SOMEWHERE, U.S.A.  
SET OF MASS STANDARDS : 5KG - 100MG  
TEST NUMBER 654321

PAGE 35  
5/18/79

TABLE 1

| ITEM  | MASS<br>(G)   | UNCERTAINTY<br>(G) | VOL AT 20<br>(CM3) | COEF OF EXP |
|-------|---------------|--------------------|--------------------|-------------|
| 5KG   | 5000.06307702 | .00564493          | 628.70150          | .000045     |
| 3KG   | 3000.02401883 | .00368509          | 377.21916          | .000045     |
| 2KG   | 2000.03017279 | .00268395          | 253.16838          | .000045     |
| 1KG   | 1000.00660911 | .00008944          | 125.73955          | .000045     |
| 500G  | 500.00589889  | .00005133          | 63.29189           | .000045     |
| 300G  | 300.00175036  | .00004085          | 37.72183           | .000045     |
| 200G  | 200.00140395  | .00003203          | 25.14792           | .000045     |
| 100G  | 100.00101957  | .00003407          | 12.65836           | .000045     |
| 50G   | 50.00212579   | .00001921          | 6.28720            | .000045     |
| 30G   | 30.00053569   | .00001521          | 3.77223            | .000045     |
| 20G   | 20.00016831   | .00001202          | 2.53167            | .000045     |
| 10G   | 10.00011825   | .00001352          | 1.25740            | .000045     |
| 5G    | 5.00006375    | .00000687          | .63292             | .000045     |
| 3G    | 3.00001669    | .00000434          | .37722             | .000045     |
| 2G    | 1.99998652    | .00000303          | .25148             | .000045     |
| 1G    | 1.00002498    | .00000228          | .12656             | .000045     |
| 500MG | .49995912     | .00000119          | .03012             | .000020     |
| 300MG | .29996303     | .00000082          | .01807             | .000020     |
| 200MG | .19993649     | .00000060          | .01204             | .000020     |
| 100MG | .09998781     | .00000058          | .00602             | .000020     |

X Y Z CORPORATION  
 SOMEWHERE, U.S.A.  
 SET OF MASS STANDARDS : 5KG - 100MG  
 TEST NUMBER 654321

PAGE 36  
 5/18/79

| ITEM  | COR.A (MG) | COR.B (MG) |
|-------|------------|------------|
| 5KG   | 23.69575   | 58.64401   |
| 3KG   | .39018     | 21.35904   |
| 2KG   | 12.39537   | 26.37469   |
| 1KG   | -1.26710   | 5.72251    |
| 500G  | 1.45455    | 4.94937    |
| 300G  | -.61250    | 1.48438    |
| 200G  | -.17129    | 1.22664    |
| 100G  | .13071     | .82967     |
| 50G   | 1.73196    | 2.08145    |
| 30G   | .29940     | .50909     |
| 20G   | -.00946    | .13033     |
| 10G   | .03948     | .10938     |
| 5G    | .01930     | .05425     |
| 3G    | -.00694    | .01403     |
| 2G    | -.02924    | -.01526    |
| 1G    | .01609     | .02308     |
| 500MG | -.00552    | -.00202    |
| 300MG | -.01575    | -.01366    |
| 200MG | -.04937    | -.04797    |
| 100MG | -.00512    | -.00442    |

# APPENDIX A.3--SAMPLE PROCESS CONTROL OUTPUT

SEE SECTION 3.3 FOR A DESCRIPTION OF THIS DATA OUTPUT.

|         |   |         |   |         |   |         |            |                          |             |
|---------|---|---------|---|---------|---|---------|------------|--------------------------|-------------|
| 5247980 | 2 | .50497  | 1 | 2.09386 | 4 | 5322.10 | .24733.88  | .4041.0000               | .01.150384S |
| 5237980 | 2 | -.59562 | 3 | .02282  | 3 | 4121.91 | .01736.81  | -.1040.0000              | .01.155684S |
| 5237980 | 4 | .98400  | 3 | .02284  | 6 | 6221.94 | .04736.75  | -.3440.0000              | .01.155684S |
| 5177980 | 6 | .07388  | 5 | .01091  | 6 | 6221.97 | -.03746.30 | -.6031.0000              | .01.171684S |
| 5187980 | 8 | -.07910 | 7 | .00131  | 6 | 6221.90 | -.04743.05 | -.4635.0000              | .01.166384S |
| 5187980 | 8 | -.02609 | 7 | .00030  | 6 | 6222.60 | .78742.19  | -.6635.5000-1.01.162084S |             |

# APPENDIX B--LISTING OF COMPUTER PROGRAM

```

BLKDAT SUBPROGRAM ---
BLOCK DATA
*****BLD00010
ZERMAL IS MACHINE ZERO THUS THE QUANTITY NEEDS TO BE CHANGED **BLD00030
FOR DIFFERENT COMPLTERS **BLD00040
THIS VALUE IS USED BY SUBPROGRAM SPINV **BLD00050
*****BLD00060
CCMCN /INVCST/ ZERMAL BLD00070
*****BLD00080
KFD(I),I=1,14 CCNTAINS THE DIGITS 0-9 AND CHARACTERS -,* USED **BLD00090
BY SUBPROGRAM DPF0 **BLD00100
WILL NEED TO CHANGE TO COMPLY WITH PROPOSED NEW FORTRAN STANDARD**BLD00110
*****BLD00120
COMMCN /DPFDVL/ KFD(18) BLD00130
*****BLD00140
INPUT-OUTPUT CCTRL PARAMETERS **BLD00150
IR IS THE CARD READER UNIT **BLD00160
IW IS THE LINE PRINTER UNIT **BLD00170
IP IS THE CARD PUNCH UNIT **BLD00180
IPL IS THE NUMBER CF LINES PER PRINTED PAGE **BLD00190
*****BLD00200
COMMON /UNITIO/ IR,IW,IP,IPL,ITMP BLD00210
*****BLD00220
CHARACTERS S T C F TO DETERMINE END OF RUN **BLD00230
*****BLD00240
COMMCN /STOP/ FS,FT,FO,FP,FE BLD00250
*****BLD00260
ITEND IS THE NUMBER OF ITERATIONS ALLOWED IN COMPUTING CBSERVED **BLD00270
CORRECTION TO WEIGHT **BLD00280
IT IS USED IN THE MAIN PROGRAM **BLD00290
*****BLD00300
COMMON /ITSTOP/ ITEND BLD00310
DATA ZERMAL /1.E-8/ BLD00320
DATA KFD(1),KFD(2),KFD(3),KFD(4),KFD(5),KFD(6),KFD(7),KFD(8) /1H0, BLD00330
2 1H1,1H2,1H3,1H4,1H5,1H6,1H7/ BLD00340
DATA KFD(9),KFD(10),KFD(11),KFD(12),KFD(13),KFD(14),KFD(15) /1H8, BLD00350
2 1H9,1H ,1H-,1H.,1H*,1H+/ BLD00360
DATA KFD(16),KFD(17),KFD(18) /1H.,1H0,1H-/ BLD00370
DATA IR,IW,IP,IPL,ITMP /5,6,1,58,7/ BLD00380
DATA FS,FT,FO,FP,FE /1H5,1HT,1H0,1HP,1H / BLD00390
DATA ITEND /10/ BLD00400
END BLD00410

MAIN PROGRAM ---
*****MAN00010
MAIN ROUTINE OF NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **MAN00020
PROGRAM VERSION OF SEPT. 10,1971 **MAN00030
WRITTEN BY ROBERT C. RAYBLD , OFFICE OF MEASUREMENT SERVICES **MAN00040
AND MRS.R.N.VARNER , STATISTICAL ENGINEERING LABORATORY **MAN00050
NATIONAL BUREAU OF STANDARDS , WASHINGTON, D.C. 20234 **MAN00060
MODIFIED BY R. N. VARNER SEPT 1979 **MAN00070
**MAN00080
THE MASS CALIBRATION PROGRAM CONTAINS ONE MAIN PROGRAM **MAN00090
AND 23 SUBPROGRAMS **MAN00100

```



| C** | PROGRAM NAME                                                      | NUMBER OF LINES OF CODE |            |
|-----|-------------------------------------------------------------------|-------------------------|------------|
| C** | MAIN                                                              | 628                     | **MAN00110 |
| C** | BLKDAT                                                            | 41                      | **MAN00120 |
| C** | READ1                                                             | 120                     | **MAN00130 |
| C** | READIT                                                            | 356                     | **MAN00140 |
| C** | ERRCR                                                             | 56                      | **MAN00145 |
| C** | PRINT1                                                            | 67                      | **MAN00150 |
| C** | TEXT1                                                             | 125                     | **MAN00160 |
| C** | TEXT2                                                             | 125                     | **MAN00170 |
| C** | TEXT3                                                             | 125                     | **MAN00180 |
| C** | TEXT4                                                             | 125                     | **MAN00190 |
| C** | TEXT5                                                             | 121                     | **MAN00200 |
| C** | TEXT6                                                             | 40                      | **MAN00210 |
| C** | READ2                                                             | 325                     | **MAN00220 |
| C** | SPINV                                                             | 126                     | **MAN00230 |
| C** | SAVMTX                                                            | 22                      | **MAN00240 |
| C** | INVCHK                                                            | 69                      | **MAN00250 |
| C** | PRINT2                                                            | 670                     | **MAN00260 |
| C** | PGCCNT                                                            | 53                      | **MAN00270 |
| C** | HEADPG                                                            | 36                      | **MAN00280 |
| C** | FINPRT                                                            | 181                     | **MAN00290 |
| C** | TEXTS1                                                            | 124                     | **MAN00300 |
| C** | TEXTS2                                                            | 30                      | **MAN00310 |
| C** | DPFD                                                              | 89                      | **MAN00320 |
| C** | CHKLN                                                             | 11                      | **MAN00330 |
| C** | THIS PROGRAM CONTRCLS THE FLOW OF THE INPUT.                      |                         | **MAN00340 |
| C** | CALCULATIONS AND OUTPUT                                           |                         | **MAN00350 |
| C** | *****                                                             |                         | **MAN00360 |
| C** | DIMENSION FOR COMMON /PRT1/ VARIABLES                             |                         | **MAN00370 |
| C** | *****                                                             |                         | **MAN00380 |
| C** | DIMENSION B1(72),B2(72),B3(72),B4(72),B5(72),B6(72),B7(72),       |                         | **MAN00390 |
| C** | 2 IDATE(3)                                                        |                         | **MAN00400 |
| C** | *****                                                             |                         | **MAN00410 |
| C** | DIMENSION FOR COMMON /INPLT/ VARIABLES                            |                         | **MAN00420 |
| C** | *****                                                             |                         | **MAN00430 |
| C** | DIMENSION AIDCST(5,15),ANCM(15),DENSTY(15),COEFEX(15),ACCV(15),   |                         | **MAN00440 |
| C** | 2 ARSTIN(15),ACKSTD(15),IRSTOU(15),IPRNT(15),DESMAT(15,50),       |                         | **MAN00450 |
| C** | 3 OBSERV(600),ALCOM(15,20)                                        |                         | **MAN00460 |
| C** | *****                                                             |                         | **MAN00470 |
| C** | DIMENSION FOR COMMON /CCMPT/ VARIABLES                            |                         | **MAN00480 |
| C** | *****                                                             |                         | **MAN00490 |
| C** | DIMENSION SWTPRT(50),A(50),DELTA(50),OBSCOR(15),COMVOL(15),       |                         | **MAN00500 |
| C** | 2 SERROR(15),TRISIG(15),TGTUN(15),DRIFT(50),ZERO(50),COMVOP(15),  |                         | **MAN00510 |
| C** | 3 CORR5A(20),SIG35A(20),UNC5A(20),IOTSTR(50),SER5A(20),DS1(50)    |                         | **MAN00520 |
| C** | *****                                                             |                         | **MAN00530 |
| C** | DIMENSION FOR MAIN PROGRAM VARIABLES                              |                         | **MAN00540 |
| C** | *****                                                             |                         | **MAN00550 |
| C** | DIMENSION VOLP(15),CORRP(15),TEMP(15),DI(50),ILOAD(50),TEMP2(15), |                         | **MAN00560 |
| C** | 2 ALCAD(50)                                                       |                         | **MAN00570 |
| C** | *****                                                             |                         | **MAN00580 |
| C** | DIMENSION FOR COMMON /REPT/ VARIABLES                             |                         | **MAN00590 |
| C** | *****                                                             |                         | **MAN00600 |
| C** | DIMENSION AITEM(5,50),APPV(50),TRMASS(50),UNCERT(50),VOLPRT(50)   |                         | **MAN00610 |
| C** | 2 CORRE(50),COEPT(50)                                             |                         | **MAN00620 |
| C** | *****                                                             |                         | **MAN00630 |
| C** | DIMENSION FOR COMMON /CHECK/ VARIABLES                            |                         | **MAN00640 |

```

C** *****MAN00680
    DIMENSION BCHK(289),Y(578)                                MAN00690
C** *****MAN00700
    Labeled COMMON                                           **MAN00710
C** *****MAN00720
    COMMON /UNITIO/ IR,IW,IP,IPL,ITMP                        MAN00730
    COMMON /PRT1/ B1,B2,B3,B4,B5,B6,B7,RANERR,SYSERR,TNCM,L1,L2,L3,L4,MAN00740
    2 L5,L6,IDATE,IBREST                                     MAN00750
    COMMON /PRT2/ IPAGE,NOSER,IPGCT                          MAN00760
    COMMON /INPUT/ TBAR,PBAR,PEAR,STDEBA,SWT,VSWT,CEXSWT,AIDCST,ANOM,MAN00770
    2 DENSITY,COEFEX,ACCVAL,ARSTIN,ACKSTD,DESMAT,CBSERV,VARBAL,ALCOM,T1P,MAN00780
    3,T2F,P1P,P2P,H1P,H2P,CP1,CP2,CT1,CT2,CH1,CH2,OT1P,OT2P,OP1P,OP2P,MAN00790
    4 OH1P,OH2P,IOP,IBAL,NOBS,NUNKN,IRSTGU,IPRNT,ITPOS,ICKUSD,ICALDS,MAN00800
    5 LINVAR,N3,N4                                           MAN00810
    COMMON /COMPUT/ OBSCOR,CORR,VOLRES,RHOA,SWTPRT,A,DELTA,CCNVOL,MAN00820
    2 SEFRR,TRISIG,TCTUN,ACCRR,CORRES,TMSUM,VOLSUM,SERSUM,T3SIG,ALOADP,MAN00830
    3,OBSTD,FRATIO,OBCECK,OBSCK,TVAL,DRIFT,ZERO,V2TAU,STAR,CCMVOP,MAN00840
    4 CORRESA,SIG35A,UNCESA,RHOAA,RHOAB,SER5A,DS1,NDGFR,ISWTCH,IFLAG,MAN00850
    5 IRCUT,IOTSTR,JSTAR                                     MAN00860
    COMMON /REPR/ TRMASS,APPMAS,CORRB,AITEM,UNCERT,VOLPRT,CCEPRT,NPRT,MAN00870
    COMMON /CHECK/ CHCKMA,BCHK,Y                             MAN00880
    COMMON /ITSTOP/ ITEND                                     MAN00890
    COMMON /DPFDVL/ KFD(18)                                  MAN00900
C*****MAN00910
C** TYPE STATEMENTS                                         **MAN00920
C*****MAN00930
    DOUBLE PRECISION VCLP,CORRP,OBSCOR,TRMASS,APPMAP,APPMAS,CORRB,MAN00940
    2 CORRE,SUM,SUM1                                         MAN00950
C*****MAN00960
C** READ IN ADMINISTRATIVE DATA                             **MAN00970
C** DATA COMMON TO ALL SERIES                               **MAN00980
C** INITIALIZE CONTROL VARIALES                               **MAN00990
C*****MAN01000
10  ALL READ1                                              MAN01010
    IXFLAG=0                                                MAN01020
    IXNSIG=0                                                MAN01030
    IPAGE=0                                                  MAN01040
    NPRT=0                                                  MAN01050
    NOSER=0                                                  MAN01060
    ASSIGN 40 TO NCSEF                                       MAN01070
    ASSIGN 410 TO NSEFPP                                     MAN01080
C*****MAN01090
C** PRINT TITLE PAGE FOR COMPLETE RUN                        **MAN01100
C*****MAN01110
    CALL PRINT1                                              MAN01120
C*****MAN01130
C** READ DATA FOR ONE SERIES                                **MAN01140
C*****MAN01150
    20  CALL READ2                                           MAN01160
C*****MAN01170
C** SET UP CHCKMA=STDEEA/100.0 TO CHECK MATRIX INVERSION    **MAN01180
C*****MAN01190
    CHCKMA=STDEBA/100.0                                     MAN01200
C*****MAN01210
C** BEGIN CALCULATIONS FOR ONE SERIES                        **MAN01220
C** CALCULATIONS FOR FIRST PRINTED PAGE OF A SERIES         **MAN01230
C*****MAN01240
    ISTAR=0                                                  MAN01250

```

```

TDELT=TBAR-TNOM
DO 30 I=1,NUNKN
30 TEMP2(I)=1.0+COEFEX(I)*TDELT
GO TO NCSERP, (40,60)
40 ASSIGN 60 TO NCSERP
CORR=0.0
VOLRES=0.0
DO 50 I=1,NUNKN
TEMP(I)=((ANOM(I)+.001*ACCVAL(I))/DENSTY(I))*TEMP2(I)
IF (ARSTIN(I).EQ.0.0) GO TO 50
CORR=CORR+ACCVAL(I)*ARSTIN(I)
VOLRES=VOLRES+TEMP(I)*ARSTIN(I)
50 CCNTINUE
GO TO 70
60 CORR=TMSUM
VOLRES=VOLSUM*(1.0+(SUMP)*TDELT)
C*****
C** CCNVERT TEMPERATURE TO KELVIN
C** COMPUTE VAPOR PRESSURE AND AIR DENSITY
C*****
70 TKEL=TEAR+273.15E0
CCNST=13.5951E0*9.80665E0
E1P=EXP(-4.7406885E0*ALOG(TKEL)-6.8982434E3/TKEL+.5938385E2-0.5797
2662E-2*TKEL+6.2223854E-6*(TKEL**2))
E1=E1P/CCNST
RHOA=(.464746E0*(FEAR-.00378029E0*HEAR*E1))/TKEL
TKEL=T1P+273.15E0
E1P=EXP(-4.7406885E0*ALOG(TKEL)-6.8982434E3/TKEL+.5938385E2-0.5797
2662E-2*TKEL+6.2223854E-6*(TKEL**2))
E1=E1P/CONST
RHOAB=(.464746E0*(P1P-.00378029E0*H1P*E1))/TKEL
TKEL=T2P+273.15E0
E1P=EXP(-4.7406885E0*ALOG(TKEL)-6.8982434E3/TKEL+.5938385E2-0.5797
2662E-2*TKEL+6.2223854E-6*(TKEL**2))
E1=E1P/CONST
RHOAA=(.464746E0*(P2P-.00378029E0*H2P*E1))/TKEL
C*****
C** BEGIN CALCULATIONS FOR SECOND AND THIRD PAGES OF A SERIES
C** THE FOLLOWING OPERATIONS ARE PERFORMED TO CGPE WITH
C** ROUND CFF ERROR
C** COMPUTE LOADS USING DESIGN MATRIX
C*****
DO 130 I=1,NOBS
ILOAD(I)=0
YYYYYY=0.0
DO 80 J=1,NUNKN
YYYYYY=YYYYYY+ABS(CESMAT(J,I))*ANOM(J)/2.0
80 CONTINUE
ZZZZZZ=YYYYYY+.05
IF (ZZZZZ-1.) 100,90,90
C*****
C** LCAD IN GRAMS
C*****
90 ILCAD(I)=INT(YYYYYY+.05)
GO TO 130
C*****
C** LOAD IN MILLIGRAMS OR MILLIFCUNDS
C*****

```

```

100  ZZZZZZ=YYYYYY*1000.+0.05                                MAN01840
    IF (ZZZZZZ-1.) 120,110,110                                MAN01850
110  ILOAD(I)=INT(1000.000*YYYYYY+.05)                        MAN01860
    GO TC 130                                                  MAN01870
C*****MAN01880
C**  LOAD IN MICRO FCUNDS                                     **MAN01890
C*****MAN01900
120  ILOAD(I)=INT(1000000.*YYYYYY+.05)                        MAN01910
130  CONTINUE                                                  MAN01920
    ALOADP=0.0                                                MAN01930
    DO 160 I=1,NOBS                                           MAN01940
        ALOAD(I)=0.0                                          MAN01950
    DO 140 J=1,NUNKN                                           MAN01960
        ALOAD(I)=ALOAD(I)+ABS(DESMAT(J,I))*ANOM(J)/2.0      MAN01970
140  CONTINUE                                                  MAN01980
    IF (ALOADP-ALOAD(I)) 150,160,160                          MAN01990
150  ALOADP=ALOAD(I)                                          MAN02000
160  CONTINUE                                                  MAN02010
C*****MAN02020
C**  COMPUTE S(STAR) SENSITIVITY VALUE (MASS/DIVISION)      **MAN02030
C**  COMPUTE S(*) SENSITIVITY VALUE (MASS/DIVISION)         **MAN02040
C**  COMPUTE A-B , S , DRIFT AND/OR LEFT-RIGHT EFFECT AS REQUIRED BY **MAN02050
C**  WEIGHING METHOD                                          **MAN02060
C*****MAN02070
    STAR=SWT-RHUA*VSWT*(1.0+CEXSWT*(TDELT))
    J=1                                                        MAN02080
    DO 290 I=1,NOBS                                           MAN02090
        GO TO (170,200,230,260,270, 280), ITPOS            MAN02100
C*****MAN02110
C**  SINGLE SUBSTITUTION-SINGLE PAN BALANCE                   **MAN02120
C*****MAN02130
170  D1(I)=CBSEV(J)-CBSEV(J+1)                                MAN02140
    IF (CBSEV(J+2).NE.0.0) GO TO 180                          MAN02150
    DS1(I)=0.0                                                MAN02160
    GO TO 190                                                  MAN02170
180  DS1(I)=CBSEV(J+2)-CBSEV(J+1)                             MAN02180
190  J=J+3                                                    MAN02190
    GO TO 290                                                  MAN02200
C*****MAN02210
C**  SINGLE SUBSTITUTION-TWO-PAN BALANCE                     **MAN02220
C*****MAN02230
200  XP=(CBSEV(J)+2.0*CBSEV(J+1)+CBSEV(J+2))/4.0           MAN02240
    YP=(CBSEV(J+3)+2.0*CBSEV(J+4)+CBSEV(J+5))/4.0           MAN02250
    ZP=(CBSEV(J+6)+2.0*CBSEV(J+7)+CBSEV(J+8))/4.0           MAN02260
    D1(I)=XP-YP                                              MAN02270
    IF (ZP.NE.0.0) GO TO 210                                  MAN02280
    DS1(I)=0.0                                                MAN02290
    GO TO 220                                                  MAN02300
210  DS1(I)=ABS(ZP-YP)                                        MAN02310
220  J=J+9                                                    MAN02320
    GO TC 290                                                  MAN02330
C*****MAN02340
C**  SINGLE TRANSPOSITION-TWO-PAN BALANCE                   **MAN02350
C*****MAN02360
230  XP=(CBSEV(J)+2.0*CBSEV(J+1)+CBSEV(J+2))/4.0           MAN02370
    YP=(CBSEV(J+3)+2.0*CBSEV(J+4)+CBSEV(J+5))/4.0           MAN02380
    ZP=(CBSEV(J+6)+2.0*CBSEV(J+7)+CBSEV(J+8))/4.0           MAN02390
    D1(I)=(XP-YP)/2.0                                         MAN02400

```

```

                ZERC(I)=(XP+YP)/2.0
                IF (ZP.NE.0.0) GO TO 240
                DS1(I)=0.0
                GO TO 250
240    DS1(I)=ABS(ZP-YP)
250    J=J+5
        GO TO 290
C*****
C**  DOUBLE SUBSTITUTION-CNE PAN BALANCE
C*****
260    D1(I)=(CBSERV(J)-CESERV(J+1)-OBSERV(J+2)+OBSERV(J+3))/2.0
        DS1(I)=(CBSERV(J)-3.0*CBSERV(J+1)+3.0*OBSERV(J+2)-OBSERV(J+3))/2.0
        DRIFT(I)=(-OBSERV(J)+OBSERV(J+1)-OBSERV(J+2)+OBSERV(J+3))/2.0
        J=J+4
        GO TO 290
C*****
C**  DOUBLE SUBSTITUTION-TWO PAN BALANCE
C*****
270    XP=(CBSERV(J)+2.0*CBSERV(J+1)+CBSERV(J+2))/4.0
        YP=(CBSERV(J+3)+2.0*CBSERV(J+4)+OBSERV(J+5))/4.0
        ZP=(CBSERV(J+6)+2.0*CBSERV(J+7)+OBSERV(J+8))/4.0
        WP=(CBSERV(J+9)+2.0*CBSERV(J+10)+OBSERV(J+11))/4.0
        D1(I)=(XP-YP-ZP+WP)/2.0
        DS1(I)=ABS((XP-3.*YP+3.*ZP-WP)/2.0)
        DRIFT(I)=(-XP+YP-ZP+WP)/2.0
        J=J+12
        GO TO 290
C*****
C**  DOUBLE TRANSPOSITION-TWO PAN BALANCE
C*****
280    XP=(CBSERV(J)+2.0*CBSERV(J+1)+CBSERV(J+2))/4.0
        YP=(OBSERV(J+3)+2.0*CBSERV(J+4)+OBSERV(J+5))/4.0
        ZP=(CBSERV(J+6)+2.0*CBSERV(J+7)+OBSERV(J+8))/4.0
        WP=(CBSERV(J+9)+2.0*CBSERV(J+10)+OBSERV(J+11))/4.0
        D1(I)=(XP-YP-ZP+WP)/4.0
        DS1(I)=ABS((XP-3.*YP+3.*ZP-WP)/2.0)
        ZERC(I)=(3.0*XP+YP+ZP-WP)/4.0
        DRIFT(I)=(-XP+YP-ZP+WP)/2.0
        J=J+12
290    CCNTINUE
C*****
C**  COMPUTE S(*)/D(S),A(I),DRIFT(I),ZERO(I)
C*****
        KA=1
        SUM=0.0
        J=0
        I=1
        ILCDDP=ILCAD(I)
C*****
C**  CHECK FOR LOAD CHANGES
C*****
300    IF (ILCAD(I).NE.ILCDDP) GO TO 320
        IF (DS1(I).EQ.0.0) GO TO 310
        SUM=SUM+DS1(I)
        DS1(I)=SUM/DS1(I)
        J=J+1
310    I=I+1
        IF (I.GT.NCBS) GO TO 330

```

```

GO TO 300 MAN03000
320 ILODPF=ILOAD(I) MAN03010
330 FLTA=J MAN03020
DEAR=SUM/FLTN MAN03030
FLTNSQ=SQRT(FLTN) MAN03040
DSP=STAR/DEAR MAN03050
KB=I-1 MAN03060
DO 360 K=KA,KB MAN03070
SWTFR1(K)=ABS(DSP) MAN03080
A(K)=D1(K)*DSP MAN03090
DRIFT(K)=DRIFT(K)*DSP MAN03100
IF (N4.EQ.0) GO TO 340 MAN03110
C***** MAN03120
C** CHANGE SIGN FOR REVERSED SCALE **MAN03130
C***** MAN03140
A(K)=-1.*A(K) MAN03150
DRIFT(K)=-1.*DRIFT(K) MAN03160
340 IF (ABS(D1(K)/(DBAR*FLTNSQ)).LE.0.25) GO TO 350 MAN03170
C***** MAN03180
C** IF OBSERVED DEFLECTION IS GREATER THAN OR EQUAL TO **MAN03190
C** .25*SENSITIVITY DEFLECTION A FLAG IS SET UP **MAN03200
C***** MAN03210
ICTSTR(K)=KFD(14) MAN03220
JSTAR=1 MAN03230
GO TO 360 MAN03240
350 ICTSTR(K)=KFD(11) MAN03250
360 CONTINUE MAN03260
IF (I.GT.NCBS) GO TO 370 MAN03270
J=0 MAN03280
SUM=0.0 MAN03290
KA=I MAN03300
GO TO 300 MAN03310
C***** MAN03320
C** COMPUTE DENSITY AND COEFFICIENT OF EXPANSION OF WEIGHTS **MAN03330
C** BEING TESTED **MAN03340
C** COMPUTE XREST (RESTRAINT VALUE) FOR Y(M,N) **MAN03350
C***** MAN03360
370 NXXSIG=0 MAN03370
DO 380 I=1,NUNKN MAN03380
IF (ARSTIN(I).EQ.0.0) GO TO 380 MAN03390
NXXSIG=NXXSIG+1 MAN03400
380 CONTINUE MAN03410
IF (NXXSIG.LT.2) GO TO 400 MAN03420
IF (NXXSIG.GT.1) GO TO 400 MAN03430
DO 390 I=1,NUNKN MAN03440
IF (ARSTIN(I).EQ.0.0) GO TO 390 MAN03450
DENSTY(I)=(ANOM(I)+.001*TMSUM)/VOLSUM MAN03460
COEFEX(I)=SUMP MAN03470
390 CONTINUE MAN03480
400 GO TO NSERPP, (410,430) MAN03490
410 ASSIGN 430 TO NSERPP MAN03500
XREST=0.0 MAN03510
TMSUM=0.0 MAN03520
DO 420 I=1,NUNKN MAN03530
IF (ARSTIN(I).EQ.0.0) GO TO 420 MAN03540
XREST=XREST+(ACCVAL(I)-RHCA*TEMP(I))*ARSTIN(I) MAN03550
TMSUM=TMSUM+ACCVAL(I)*ARSTIN(I) MAN03560
420 CONTINUE MAN03570

```

```

      GO TC 440
430  XREST=TMSUM-RHCA*VOLSUM*(1.0+SUMP*TDELT)
C*****
C**  MULTIPLY TRANSPCSE OF DESIGN MATRIX BY ITSELF AND STORE IN Y
C*****
440  L=0
      DO 470 IA=1,NUNKN
      DO 460 IAA=1,NUNKN
      L=L+1
      SUM=0.0
      DO 450 JA=1,NOBS
      SUM=SUM+DESMAT(IAA,JA)*DESMAT(IA,JA)
450  CONTINUE
      Y(L)=SUM
460  CONTINUE
      L=L+2
      Y(L-1)=ARSTIN(IA)
      Y(L)=0.0
470  CONTINUE
C*****
C**  SET UP REMAINING ELEMENTS OF MATRIX X
C*****
      IRIN=0
      DO 480 I=1,NUNKN
      L=L+1
      Y(L)=ARSTIN(I)
      IF (ARSTIN(I).EQ.0.0) GO TO 480
      IRIN=1.0
480  CONTINUE
      L=L+2
      IF (IRIN.EQ.0) GO TO 490
      Y(L-1)=0.0
      GO TO 500
490  Y(L-1)=1.0
500  Y(L)=0.0
C*****
C**  COMPUTE X'Y
C*****
      DO 520 IAA=1,NUNKN
      L=L+1
      SUM=0.0
      DO 510 JA=1,NOBS
      SUM=SUM+DESMAT(IAA,JA)*A(JA)
510  CCNTINUE
      Y(L)=SUM
520  CONTINUE
      Y(L+1)=XREST
      Y(L+2)=-1.0
C*****
C**  CALL MATRIX INVERSION ROUTINE
C*****
      CALL SPINV (Y,NUNKN+2,NUNKN+2,IFLAG)
C*****
C**  IF MATRIX IS SINGULAR PRINT FIRST TWO PAGES OF SERIES AND
C**  TERMINATE RUN
C**  IFLAG=0  INVERSE SUCCESSFUL
C**  IFLAG=4  MATRIX SINGULAR
C**  IFLAG=-1 I-AA(INV) FAILED

```

```

C*****MAN04160
  IF (IFLAG.EQ.0) GC TC 530                                MAN04170
  CALL PRINT2                                              MAN04180
  STOP                                                    MAN04190
C*****MAN04200
C**  COMPUTE DELTA(I)=A(I)-DESMAT*BETA(I)                **MAN04210
C*****MAN04220
530  L=(NUNKN+2)*(NUNKN+1)+1                                MAN04230
      DO 550 I=1,NOKS                                     MAN04240
      LA=L                                                 MAN04250
      SUM=0.0                                              MAN04260
      DO 540 J=1,NUNKN                                    MAN04270
      SUM=SUM+DESMAT(J,I)*Y(LA)                          MAN04280
      LA=LA+1                                              MAN04290
540  CCNTINUE                                              MAN04300
      DELTA(I)=A(I)-SUM                                    MAN04310
550  CCNTINUE                                              MAN04320
C*****MAN04330
C**  CALCULATE CORRECTION AND VOLUME FOR THIRD OUTPUT PAGE **MAN04340
C**  OF A SERIES                                          **MAN04350
C*****MAN04360
      LA=L                                                 MAN04370
      DO 560 I=1,NUNKN                                    MAN04380
      VOLP(I)=(ANOM(I)/DENSTY(I))*TEMP2(I)               MAN04390
      CORRP(I)=Y(LA)+RHCA*VOLP(I)                        MAN04400
      LA=LA+1                                              MAN04410
560  CONTINUE                                              MAN04420
C*****MAN04430
C**  BEGIN ITERATION AND ITERATE UNTIL DIFFERENCE BETWEEN OLD **MAN04440
C**  AND NEW VALUES ARE LESS THAN .01 STDEBA           **MAN04450
C**  ONLY 10 ITERATIONS ARE ALLOWED                     **MAN04460
C*****MAN04470
      CHCK=.01*STDEEA                                     MAN04480
      ISTCP=0                                              MAN04490
570  ISWTC=0                                              MAN04500
      LA=L                                                 MAN04510
      DO 600 I=1,NUNKN                                    MAN04520
      OBSCOR(I)=Y(LA)+RHCA*((ANOM(I)+.001*CORRP(I))/DENSTY(I))*TEMP2(I) MAN04530
      LA=LA+1                                              MAN04540
      IF (ISWTC.NE.0) GC TO 590                          MAN04550
      IF (SGL(DABS(OBSCOR(I)-CORRP(I)))-CHCK) 590.590.580 MAN04560
580  ISWTC=1                                              MAN04570
590  CORRP(I)=OBSCOR(I)                                  MAN04580
600  CCNTINUE                                              MAN04590
      ISTCP=ISTCP+1                                       MAN04600
      IF (ISTCP.GE.ITEND) GO TC 610                      MAN04610
      IF (ISWTC.NE.0) GC TO 570                          MAN04620
610  WR=0.0                                              MAN04630
      DO 620 I=1,NUNKN                                    MAN04640
      IF (ARSTIN(I).EQ.0.0) GC TC 620                    MAN04650
      WR=WR+ANOM(I)*ARSTIN(I)                            MAN04660
620  CCNTINUE                                              MAN04670
      TRISGR=ANERR**2                                     MAN04680
      TRSGRP=(3.0*STDEBA)**2                             MAN04690
      TRSCVE=(3.0*VAREAL)**2                             MAN04700
      LA=1                                                 MAN04710
      DO 670 I=1,NUNKN                                    MAN04720
      COMVCL(I)=(ANOM(I)+.001*OBSCOR(I))/DENSTY(I)      MAN04730

```



|     |                                                               |            |
|-----|---------------------------------------------------------------|------------|
|     | CCMVOP(I)=COMVOL(I)*TEMP2(I)                                  | MAN04740   |
|     | SERROR(I)=(ANOM(I)/WR)*SYSERR                                 | MAN04750   |
|     | TRTEMP=TRSGRP*Y(LA)+((ANOM(I)/WR)**2)*TRISGR+TRSGVE           | MAN04760   |
|     | IF (TRTEMP) 630,640,640                                       | MAN04770   |
| 630 | WRITE (IW,890) TRTEMP                                         | MAN04780   |
|     | GC TC 660                                                     | MAN04790   |
| 640 | IF (TRTEMP.GE.0.0) GO TO 650                                  | MAN04800   |
|     | TRTEMP=0.0                                                    | MAN04810   |
| 650 | TRISIG(I)=SQRT(TRTEMP)                                        | MAN04820   |
| 660 | TOTUN(I)=SERROR(I)+TRISIG(I)                                  | MAN04830   |
|     | LA=LA+NUNKN+3                                                 | MAN04840   |
| 670 | CCNTINLE                                                      | MAN04850   |
|     | ACORR=TMSUM                                                   | MAN04860   |
|     | CORRES=0.0                                                    | MAN04870   |
|     | TMSUM=0.0                                                     | MAN04880   |
|     | SERSUM=0.0                                                    | MAN04890   |
|     | T3SIG=0.0                                                     | MAN04900   |
|     | DO 690 I=1,NUNKN                                              | MAN04910   |
|     | IF (ARSTIN(I).EQ.0.0) GO TC 680                               | MAN04920   |
|     | CORRES=CORRES+OBSCOR(I)*ARSTIN(I)                             | MAN04930   |
| 680 | IF (IRSTOU(I).EQ.0) GO TC 690                                 | MAN04940   |
|     | C*****MAN04950                                                |            |
|     | C** COMPUTE MASS FOR NEXT RESTRAINT                           | **MAN04960 |
|     | C*****MAN04970                                                |            |
|     | TMSUM=TMSUM+OBSCOR(I)                                         | MAN04980   |
| 690 | CCNTINUE                                                      | MAN04990   |
|     | VOLSUM=0.0                                                    | MAN05000   |
|     | IROUT=0                                                       | MAN05010   |
|     | NXNSIG=0                                                      | MAN05020   |
|     | DO 700 I=1,NUNKN                                              | MAN05030   |
|     | IF (IRSTOU(I).EQ.0) GO TC 700                                 | MAN05040   |
|     | IROUT=1                                                       | MAN05050   |
|     | VOLSUM=VOLSUM+(ANOM(I)+.001*OBSCOR(I))/DENSTY(I)              | MAN05060   |
|     | NXNSIG=NXNSIG+1                                               | MAN05070   |
| 700 | CONTINUE                                                      | MAN05080   |
|     | C*****MAN05090                                                |            |
|     | C** CALCULATIONS FOR THIRD CUTPUT PAGE OF SERIES              | **MAN05100 |
|     | C** COMPUTE SYSTEMATIC AND RANDOM ERROR AND TOTAL UNCERTAINTY | **MAN05110 |
|     | C*****MAN05120                                                |            |
|     | LINVV=LINVAR+1                                                | MAN05130   |
|     | DO 710 K=1,NUNKN                                              | MAN05140   |
| 710 | ALCCM(K,LINVV)=IRSTOU(K)                                      | MAN05150   |
|     | DO 770 L=1,LINVV                                              | MAN05160   |
|     | CORRSA(L)=0.0                                                 | MAN05170   |
|     | SERSA(L)=0.0                                                  | MAN05180   |
|     | DO 720 K=1,NUNKN                                              | MAN05190   |
|     | CORRSA(L)=CORRSA(L)+OBSCOR(K)*ALCOM(K,L)                      | MAN05200   |
|     | SERSA(L)=ABS(SERSA(L)+SEFFOR(K)*ALCOM(K,L))                   | MAN05210   |
| 720 | CCNTINUE                                                      | MAN05220   |
|     | LA=1                                                          | MAN05230   |
|     | DO 740 I=1,NUNKN                                              | MAN05240   |
|     | TEMP(I)=0.0                                                   | MAN05250   |
|     | DO 730 J=1,NUNKN                                              | MAN05260   |
|     | TEMP(I)=TEMP(I)+Y(LA)*ALCCM(J,L)                              | MAN05270   |
| 730 | LA=LA+1                                                       | MAN05280   |
| 740 | LA=LA+2                                                       | MAN05290   |
|     | SUM=0.0                                                       | MAN05300   |
|     | SUM2RP=0.0                                                    | MAN05310   |

```

DO 750 J=1,NUNKN                                MAN05320
SUM2RP=SUM2RP+ANOM(J)*ALCCM(J,L)                 MAN05330
750 SUM=SUM+TEMP(J)*ALCCM(J,L)                   MAN05340
IF (SUM.GE.0.0) GC TC 760                         MAN05350
SUM=0.0                                            MAN05360
760 CONTINUE                                       MAN05370
SIG35A(L)=SQRT(TRSGRP*SUM+((SUM2RP/WR)**2)*TRISGR+TRSQVE) MAN05380
UNCEA(L)=SER5A(L)+SIG35A(L)                     MAN05390
770 CONTINUE                                       MAN05400
SERSUM=SERSA(L INVV)                             MAN05410
T3SIG=SIG35A(L INVV)                             MAN05420
C*****MAN05430
C** BEGIN COMPUTATIONS FOR FOURTH PAGE -- F RATIO AND T-TEST **MAN05440
C*****MAN05450
NDGFR=NCBS-NUNKN+1                                MAN05460
SUM=0.0                                            MAN05470
DO 780 I=1,NOBS                                    MAN05480
SUM=SUM+DELTA(I)**2                                MAN05490
780 CONTINUE                                       MAN05500
OBSTD=SQRT(SUM/FLCAT(NDGFR))                     MAN05510
FRATIO=OBSTD**2/STDEEA**2                        MAN05520
OBCCCK=0.0                                         MAN05530
SUM1=0.0                                           MAN05540
V2TAU=0.0                                          MAN05550
DO 790 I=1,NUNKN                                   MAN05560
IF (ACKSTD(I).EQ.0.0) GC TC 790                   MAN05570
OBCCCK=CBCCCK+CBSCCR(I)*ACKSTD(I)                MAN05580
SUM1=SUM1+ANOM(I)*ACKSTD(I)                       MAN05590
V2TAU=V2TAU+ACCVAl(I)*ACKSTD(I)                  MAN05600
790 CONTINUE                                       MAN05610
LA=1                                               MAN05620
DO 810 J=1,NUNKN                                   MAN05630
TEMP(J)=0.0                                        MAN05640
DO 800 I=1,NUNKN                                   MAN05650
TEMP(J)=TEMP(J)+ACKSTD(I)*Y(LA)                  MAN05660
800 LA=LA+1                                        MAN05670
810 LA=LA+2                                        MAN05680
SUM=0.0                                            MAN05690
DO 820 I=1,NUNKN                                   MAN05700
SUM=SUM+TEMP(I)*ACKSTD(I)                         MAN05710
IF (SUM.GE.0.0) GC TC 830                         MAN05720
SUM=0.0                                            MAN05730
830 CCNTINUE                                       MAN05740
OBSCK=SQRT(STDEBA**2*SUM+(SUM1/WR)**2*(RANERR/3.0)**2+VAREAL**2) MAN05750
IF (OBSCK.NE.0.0) GC TC 840                       MAN05760
TVAL=0.0                                           MAN05770
GO TO 850                                          MAN05780
840 TVAL=(CBCCCK-V2TAU)/OBSCK                     MAN05790
C*****MAN05800
C** SAVE VALUES FOR FINAL RERCRT **MAN05810
C*****MAN05820
950 DO 860 I=1,NUNKN                                MAN05830
IF (IFRNT(I).EQ.0.0) GC TC 860                   MAN05840
NPRT=NPRT+1                                       MAN05850
ITEM(1,NPRT)=AIDCST(1,I)                         MAN05860
ITEM(2,NPRT)=AIDCST(2,I)                         MAN05870
ITEM(3,NPRT)=AIDCST(3,I)                         MAN05880
ITEM(4,NPRT)=AIDCST(4,I)                         MAN05890

```

```

      AITEM(5,NPRT)=AIDCST(5,I)
      APPMAP=((ANOM(I)+.001*CBSCOR(I))*(1.0-.0012/DENSTY(I)))/(1.0-.0012
2*((1.0+.000054*2C.0)/8.4)
      APPMAS(NPRT)=(APPMAP-ANOM(I))*1000.
      TRMASS(NPRT)=ANOM(I)+.001*CBSCCR(I)
      UNCERT(NPRT)=.001*TOTUN(I)
      VOLPRT(NPRT)=CCMVCL(I)
      CORREB=((ANOM(I)+.001*CBSCOR(I))*(1.0-.0012/DENSTY(I)))/(1.0-.0012
2/8.0)
      CORFB(NPRT)=(CORREB-ANOM(I))*1000.
      COEPRT(NPRT)=COEFEX(I)
860  CONTINUE
C*****
C**  SET UP VALUES FOR NEXT SERIES
C**  PRINT FOUR PAGES OF OUTPUT FOR ONE SERIES
C*****
      CALL PRINT2
      SYSERR=SERSUM
      RANERR=T3SIG
      SUM=0.0
      SUM1=0.0
      DO 870 I=1,NUNKN
      IF (IRSTOU(I).EQ.0) GO TO 870
      SUM=SUM+((ANOM(I)+.001*CBSCCR(I))/DENSTY(I)
      SUM1=SUM1+COEFEX(I)*((ANOM(I)+.001*CBSCOR(I))/DENSTY(I))
870  CONTINUE
      IF (SUM.EQ.0.0) GO TO 880
      SUMP=SUM1/SUM
880  CONTINUE
      IF (IRCUT.NE.0) GO TO 20
      IF (NPRT.EQ.0) GO TO 10
      CALL FINPRT
      GO TO 10
C*****
C**  FORMAT STATEMENT
C*****
890  FORMAT (14F,NEG SCFT ARG=E16.7)
      END
--- READ1 SUBPROGRAM ---
      SUBROUTINE READ1
C*****
C**  SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION
C**  PROGRAM VERSION OF SEPT.10,1971      WRITTEN BY R.C.RAYBOLD
C**  AND MRS.R.N.VARNER
C**  MODIFIED BY R. N. VARNER SEPT 1979
C*****
C**  SUBROUTINE TO READ DATA COMMON TO ALL SERIES
C*****
C**  DIMENSION FOR COMMON /PRT1/ VARIABLES
C*****
      DIMENSION B1(72),B2(72),B3(72),B4(72),B5(72),B6(72),B7(72),
2 IDATE(3)
C*****
C**  DIMENSION FOR COMMON /RAREA/ VARIABLES
C*****
      DIMENSION AA(72),AITEM(5)
C*****
C**  LABELED COMMON

```

```

C*****RD100200
  COMMON /PRT1/ B1,B2,B3,B4,B5,B6,B7,RANERR,SYSERR,TNOM,L1,L2,L3,L4,RD100210
  2 L5,L6,IDATE,IBREST RD100220
  COMMON /RAREA/ AA,AAITEM RD100230
  COMMON /UNIT10/ IR,IW,IP,IPL,ITMP RD100240
  COMMON /STCP/ FS,FT,FO,FP,FB RD100250
C*****RD100260
C** READ IN AND SET UP ADMINISTRATIVE DETAILS **RD100270
C** READ 8 CARDS IF ALL 8 ARE NOT USED , ADD BLANK CARDS **RD100280
C** THESE CARDS ARE SEARCHED FOR THE 1ST NON-BLANK CHARACTER AT **RD100290
C** WHICH TIME ALL REMAINING CHARACTERS ARE MOVED **RD100300
C** TO A NEW LOCATION STARTING IN POSITION 1 **RD100310
C** CARD 1 - NAME OF ORGANIZATION **RD100320
C** CARD 2 - ADDRESS STREET NUMBER AND NAME **RD100330
C** CARD 3 - ADDRESS CITY , STATE ZIP CODE **RD100340
C** CARD 4 - TYPE OF SET (EG. 1-KG TO 1-MG) **RD100350
C** CARD 5 - SERIAL NUMBER OF SET OF WEIGHTS BEING TESTED **RD100360
C** CARD 6 - DATE OF THE REPORT **RD100370
C** CARD 7 - TEST NUMBER (MAX OF 18 CHARACTERS) **RD100380
C** EXCEPT FOR THE FIRST PRINTED PAGE ONLY THE **RD100390
C** FIRST 65 CHARACTERS ARE PRINTED ON OUTPUT **RD100400
*****RD100410
DO 10 K=1,72 RD100420
  B1(K)=FB RD100430
  B2(K)=FB RD100440
  B3(K)=FB RD100450
  B4(K)=FB RD100460
  B5(K)=FB RD100470
  B6(K)=FB RD100480
  B7(K)=FB RD100490
10 CONTINUE RD100500
DO 130 I=1,7 RD100510
  READ (IR,140) (AA(L),L=1,72) RD100520
DO 20 J=1,72 RD100530
  IF (AA(J).EQ.FB) GC TO 20 RD100540
  N=72-J RD100550
  JJ=J RD100560
GO TO 30 RD100570
CONTINUE RD100580
GO TO 130 RD100590
DO 120 K=1,N RD100600
*****RD100610
  MOVE NON-BLANK CHARACTERS TO BEGINNING OF FIELD **RD100620
*****RD100630
GO TO (40,50,60,70,80, 90,100), I RD100640
  LI=N RD100650
  B1(K)=AA(JJ) RD100660
  IF (K.NE.4) GO TO 110 RD100670
*****RD100680
TEST TO SEE IF CARD HAS STCP ON IT , IF IT DOES THEN STOP THE **RD100690
PROGRAM OTHERWISE CONTINUE **RD100700
*****RD100710
  IF (B1(1).EQ.FS.AND.B1(2).EQ.FT.AND.B1(3).EQ.FO.AND.B1(4).EQ.FP) SRD100720
  RD100730
  GO TO 110 RD100740
  RD100750
  JJ=AA(JJ) RD100760
  GO TO 110 RD100770

```

```

60      L3=N
        B3(K)=AA(JJ)
        GO TO 110
70      L4=N
        B4(K)=AA(JJ)
        GO TO 110
80      L5=N
        B5(K)=AA(JJ)
        GO TO 110
90      L6=N
        B6(K)=AA(JJ)
        GO TO 110
100     B7(K)=AA(JJ)
110     JJ=JJ+1
120     CONTINUE
130     CONTINUE
*****RD100780
C**      READ RANDOM ERROR LIMIT, SYSTEMATIC ERROR LIMIT AND NOMINAL **RD100790
C**      TEMPERATURE ALL ON ONE CARD FREE FIELD ,FIELDS MUST BE **RD100800
C**      SEPARATED BY A BLANK OR ANY NON-NUMERIC CHARACTER EXCEPT FOR **RD100810
C**      D OR E OR . **RD100820
C** **RD100830
C**      WHERE THE NUMBERS MUST BE IN THE FOLLOWING ORDER **RD100840
C**      RANERR = 3 TIMES RANDOM ERROR IN THE STARTING RESTRAINT IN MG **RD100850
C**      SYSERR = SYSTEMATIC ERROR IN THE STARTING RESTRAINT IN MG **RD100860
C**      TNOM = NOMINAL TEMPERATURE AT WHICH THE APPARENT MASS **RD100870
C**      VALUES ARE REPORTED IN DEGREES C **RD100880
C**      IEREST = STARTING RESTRAINT ID NUMBER **RD100890
C** **RD100900
C**      FOR THE REST OF THE INPLT CARDS FOR EACH SERIES OF WEIGHINGS **RD101000
C**      LOOK AT COMMENTS CARDS IN SUBROUTINE READ2 **RD101010
C*****RD101020
        CALL READIT (AA,1,PAITEM)
        RANERR=AA(1)
        SYSERR=AA(2)
        TNOM=AA(3)
        IBREST=INT(AA(4)+.5)
        RETURN
        RD101030
C*****RD101040
C**      FORMAT STATEMENT **RD101050
C*****RD101060
140     FORMAT (72A1)
        END
        RD101070
--- READIT SUBPRCGRAM ---
        SUPRCUTINE READIT (Z,KCL,A)
C*****RD101080
C**      SUBRCUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **RD101090
C**      PROGRAM VERSION OF SEPT.10,1971 WRITTEN BY R.C.RAYBOLD **RD101100
C**      AND MRS.R.N.VARNER **RD101110
C**      MODIFIED BY R. N. VARNER SEPT 1979 **RD101120
C*****RD101130
C**      ANSI FORTRAN SUBROUTINE TO READ NUMBERS IN ANY FORMAT ANYWHERE **RD101140
C**      ON A CARD (BETWEEN CARD COLUMNS 'KOL' AND 80, INCLUSIVE). **RD101150
C**      IN THIS VERSION OF THE SUBROUTINE, WHEN KOL = 16, CARD COLUMNS **RD101160
C**      1 THROUGH 15 ARE SCANNED TO PICK UP ALPHA-NUMERIC DATA WHICH **RD101170
C**      ARE STORED IN VARIABLE 'A'. **RD101180
C**      WRITTEN BY ROY H. WAMPLER, STATISTICAL ENGINEERING LABORATORY, **RD101190
C**      NATIONAL BUREAU OF STANDARDS, WASHINGTON, D. C. 20234 **RD101200

```

```

*   VERSION OF FEBRUARY 8, 1971
*****
*   NUMBERS TO BE READ BY THIS ROUTINE SHOULD OBEY THE FOLLOWING
*   RULES.
*
*   (1) BETWEEN ANY TWO NUMBERS THERE MUST BE A SEPARATOR. THIS
*   CAN BE ONE OR MORE BLANK SPACES, A COMMA, ANY LETTER EXCEPT D
*   OR E, OR ANY CHARACTER EXCEPT A PLUS SIGN, A MINUS SIGN, OR A
*   DECIMAL.
*
*   (2) NUMBERS CAN APPEAR IN INTEGER FORM. EXAMPLES ARE
*   0 63 -271 +81063 01 2,71,-534,28
*
*   (3) NUMBERS CAN BE WRITTEN WITH A DECIMAL POINT. EXAMPLES ARE
*   0. -1.0 38.1 -63. .00015 +371.286
*
*   (4) NUMBERS CAN BE WRITTEN WITH AN EXPONENT WHICH MUST BE
*   PRECEDED BY A D OR E. (IN THIS ROUTINE D IS CONSIDERED EQUIVA-
*   LENT TO E, AND NUMBERS WITH D ARE NOT INTERPRETED TO BE
*   DOUBLE PRECISION NUMBERS.) EXAMPLES ARE
*   2.1E12 2.1E 12 2.1E+12 2.1E-12
*   -2.1D12 -2.1D 12 -2.1D+12 -2.1D-12
*   0021.E02 .00021E5 2.1E0 2.1E-0
*   21E12 21E+12 21D-12
*   THE LAST THREE EXAMPLES ILLUSTRATE THAT A DECIMAL NEED NOT BE
*   USED IN CONNECTION WITH THE D OR E.
*****
DIMENSION Z(40),A(1),N(60),IDIGIT(10),T(77)
COMMON /DPFDVL/ KFD(18)
COMMON /UNITIO/ IF, Iw, IP, IPL, ITMP
EQUIVALENCE (IDIGIT(1),KFD(1))
EQUIVALENCE (IPLUS,KFD(15))
EQUIVALENCE (IMINUS,KFD(12))
EQUIVALENCE (ID,KFD(17))
EQUIVALENCE (IE,KFD(18))
EQUIVALENCE (IDECML,KFD(13))
EQUIVALENCE (IBLANK,KFD(11))
*****
THE FOLLOWING DIMENSION STATEMENT AND THE THREE DATA STATEMENTS
WHICH FOLLOW THAT ARE MACHINE-DEPENDENT.
T(1) CONTAINS THE MACHINE RANGE OF NUMBERS
IN THIS CASE 1.E-38 TO 1.E38
ZERO = NUMBER OF UNIQUE POWERS OF TEN REPRESENTED BY
THE MACHINE RANGE
MAX = NUMBER OF POWERS OF TEN (NEGATIVE AND POSITIVE)
REPRESENTED BY THE MACHINE RANGE
*****
A T(1),T(2),T(3),T(4),T(5),T(6),T(7),T(8),T(9),T(10),T(11),
(12),T(13),T(14),T(15),T(16),T(17),T(18),T(19),T(20),T(21),T(22)
(23),T(24),T(25),T(26),T(27),T(28),T(29),T(30),T(31),T(32),T(33)
(34),T(35),T(36),T(37),T(38),T(39),T(40),T(41),T(42),T(43),T(44)
(45),T(46),T(47),T(48),T(49),T(50),T(51),T(52),T(53),T(54),T(55)
(56),T(57),T(58),T(59),T(60),T(61),T(62),T(63),T(64),T(65),T(66)
(67),T(68),T(69),T(70),T(71),T(72),T(73),T(74),T(75),T(76),
(77) 1.E-38,1.E-37,1.E-36,1.E-35,1.E-34,1.E-33,1.E-32,1.E-31,
20,1.E-29,1.E-28,1.E-27,1.E-26,1.E-25,1.E-24,1.E-23,1.E-22,
21,1.E-20,1.E-19,1.E-18,1.E-17,1.E-16,1.E-15,1.E-14,1.E-13,
12,1.E-11,1.E-10,1.E-9,1.E-8,1.E-7,1.E-6,1.E-5,1.E-4,1.E-3,

```

```

2 1.E-2,1.E-1,1.,1.E1,1.E2,1.E3,1.E4,1.E5,1.E6,1.E7,1.E8,1.E9,1.E10RDT00730
3,1.E11,1.E12,1.E13,1.E14,1.E15,1.E16,1.E17,1.E18,1.E19,1.E20,1.E21RDT00740
4,1.E22,1.E23,1.E24,1.E25,1.E26,1.E27,1.E28,1.E29,1.E30,1.E31,1.E32RDT00750
5,1.E33,1.E34,1.E35,1.E36,1.E37,1.E38/ RDT00760
DATA IZERO,IMAX /39,77/ RDT00770
C*****RDT00780
C** THE DIMENSIONED VARIABLE T IS USED FOR ENTERING POWERS OF TEN **RDT00790
C** INTO THE PROGRAM. **RDT00800
C** IZERO IS THE SUBSCRIPT OF T SUCH THAT T(IZERO) = 1. (= 1.E0). **RDT00810
C** (ON THE UNIVAC 1108, T(IZERO) = T(39).) **RDT00820
C** IMAX IS THE LARGEST SUBSCRIPT OF T. **RDT00830
C** IN THE PROGRAM IT IS ASSUMED THAT (1 + IMAX)/2 = IZERO. THAT **RDT00840
C** IS, WE ASSUME THAT VALID SINGLE PRECISION NUMBERS RANGE IN **RDT00850
C** ABSOLUTE VALUE FROM 10.** (1 - IZERO) TO 10.** (IZERO - 1), OR ARE **RDT00860
C** EQUAL TO ZERO. **RDT00870
C** NR IS THE COMPUTER'S READING UNIT, AND NW ITS WRITING UNIT. **RDT00880
C*****RDT00890
C** CHARACTERS OF INTEREST ARE IN KFD **RDT00900
C** THEY WILL BE COMPARED WITH N **RDT00910
C*****RDT00920
C*****RDT00930
C** THE CHARACTERS ON ONE CARD ARE READ IN AN A-FORMAT AND STORED **RDT00940
C** IN N. **RDT00950
C*****RDT00960
IF (KOL.GE.1.AND.KCL.LE.80) GO TO 10 RDT00970
CALL ERROR (KOL,A,N,LL,IW,7) RDT00980
RETURN RDT00990
10 IF (KOL.EQ.16) GO TO 20 RDT01000
READ (IR,1010) (N(I),I=1,80) RDT01010
GO TO 30 RDT01020
20 READ (IR,1000) (A(I),I=1,5),(N(I),I=16,80) RDT01030
C*****RDT01040
C** ON SOME COMPUTERS THE FORMAT FOR READING A(I) MAY HAVE TO BE **RDT01050
C** CHANGED. **RDT01060
C*****RDT01070
C** APPROPRIATE VARIABLES ARE INITIALIZED. **RDT01080
C*****RDT01090
30 IDORE=0 RDT01100
IEXP=0 RDT01110
ISIGX=0 RDT01120
K=0 RDT01130
NDE=0 RDT01140
NDEC=0 RDT01150
NODEC=0 RDT01160
NUME=0 RDT01170
NXDIG=0 RDT01180
SIG=0. RDT01190
SIGN=0. RDT01200
DO 40 I=1,40 RDT01210
40 Z(I)=0. RDT01220
C*****RDT01230
C** THE CHARACTERS ON THE CARD ARE EXAMINED. **RDT01240
C** WHEN NUMBERS ARE FOUND THEY ARE STORED IN Z(K). **RDT01250
C*****RDT01260
DO 580 I=KCL,80 RDT01270
C*****RDT01280
C** DETERMINE IF N(I) IS A DIGIT. **RDT01290
C*****RDT01300

```

|        |                                                              |            |
|--------|--------------------------------------------------------------|------------|
|        | IF (N(I).GE.IDIGIT(1).AND.N(I).LE.IDIGIT(10)) GC TO 520      | RDT01310   |
| C***** |                                                              | RDT01320   |
| C**    | N(I) IS NOT A DIGIT.                                         | **RDT01330 |
| C**    | DETERMINE IF N(I) IS A PLUS, MINUS, DECIMAL, D, E, CR BLANK. | **RDT01340 |
| C***** |                                                              | RDT01350   |
|        | IF (NUMB) 140,50,140                                         | RDT01360   |
| 50     | IF (N(I)-IPLUS) 60,130,60                                    | RDT01370   |
| 60     | IF (N(I)-IMINUS) 70,100,70                                   | RDT01380   |
| 70     | IF (N(I)-IDECML) 80,50,80                                    | RDT01390   |
| 80     | NDE=0                                                        | RDT01400   |
|        | SIG=0.                                                       | RDT01410   |
|        | GO TO 780                                                    | RDT01420   |
| 90     | NDE=1                                                        | RDT01430   |
|        | GO TO 780                                                    | RDT01440   |
| 100    | SIG=-1.                                                      | RDT01450   |
| 110    | IF (NDE) 120,780,120                                         | RDT01460   |
| 120    | NDE=0                                                        | RDT01470   |
|        | GO TO 780                                                    | RDT01480   |
| 130    | SIG=1.                                                       | RDT01490   |
|        | GO TO 110                                                    | RDT01500   |
| 140    | IF (IDCRE) 340,150,340                                       | RDT01510   |
| 150    | IF (NDEC) 250,160,250                                        | RDT01520   |
| 160    | IF (N(I)-ID) 180,170,180                                     | RDT01530   |
| 170    | IDCRE=1                                                      | RDT01540   |
|        | NDEC=1                                                       | RDT01550   |
|        | GO TO 780                                                    | RDT01560   |
| 180    | IF (N(I)-IE) 190,170,190                                     | RDT01570   |
| 190    | IF (N(I)-IDECML) 210,200,210                                 | RDT01580   |
| 200    | NDEC=1                                                       | RDT01590   |
|        | GO TO 780                                                    | RDT01600   |
| 210    | IF (N(I)-IPLUS) 230,220,230                                  | RDT01610   |
| 220    | SIG=1.                                                       | RDT01620   |
|        | GO TO 850                                                    | RDT01630   |
| 230    | IF (N(I)-IMINUS) 850,240,850                                 | RDT01640   |
| 240    | SIG=-1.                                                      | RDT01650   |
|        | GO TO 850                                                    | RDT01660   |
| 250    | IF (N(I)-ID) 270,260,270                                     | RDT01670   |
| 260    | IDCRE=1                                                      | RDT01680   |
|        | GO TO 780                                                    | RDT01690   |
| 270    | IF (N(I)-IE) 280,260,280                                     | RDT01700   |
| 280    | IF (N(I)-IDECML) 300,290,300                                 | RDT01710   |
| 290    | NDE=1                                                        | RDT01720   |
|        | GO TO 870                                                    | RDT01730   |
| 300    | IF (N(I)-IPLUS) 320,310,320                                  | RDT01740   |
| 310    | SIG=1.                                                       | RDT01750   |
|        | GO TO 870                                                    | RDT01760   |
| 320    | IF (N(I)-IMINUS) 870,330,870                                 | RDT01770   |
| 330    | SIG=-1.                                                      | RDT01780   |
| 340    | IF (N(I)-IDECML) 370,350,370                                 | RDT01790   |
| 350    | NDE=1                                                        | RDT01800   |
|        | IF (NXDIG) 870,360,870                                       | RDT01810   |
| 360    | IF (NDEC) 870,850,870                                        | RDT01820   |
| 370    | IF (IDCRE=1) 380,440,380                                     | RDT01830   |
| 380    | IF (N(I)-IPLUS) 400,390,400                                  | RDT01840   |
| 390    | SIG=1.                                                       | RDT01850   |
|        | GO TO 420                                                    | RDT01860   |
| 400    | IF (N(I)-IMINUS) 420,410,420                                 | RDT01870   |
| 410    | SIG=-1.                                                      | RDT01880   |



```

420 IF (NXDIG) 870,430,870
430 IF (NDEC) 870,850,870
440 IF (NXDIG) 870,450,870
450 IF (N(I)-IMINUS) 460,510,460
460 IF (N(I)-IPLUS) 470,490,470
470 IF (N(I)-IBLANK) 480,490,480
480 IF (NDEC) 870,850,870
490 ISIGX=1
500 IDCRE=ICORE+1
    GO TO 780
510 ISIGX=-1
    GO TO 500
*****
C** N(I) IS A DIGIT.
*****
520 IF (NUMB) 530,540,530
530 IF (IDCRE) 730,650,730
*****
C** N(I) IS THE FIRST CIGIT OF A NUMBER.
*****
540 IF (SIG) 550,560,550
550 SIGN=SIG
    SIG=0.
    GO TO 570
560 SIGN=1.
570 IF (NDE) 580,590,580
580 NDEC=NDE+1
    NDE=0
590 DO 600 L=1,10
    IF (N(I).NE.IDIGIT(L)) GC TO 600
    IN=L-1
    GO TO 610
600 CCNTINUE
610 ZED=IN
    IF (ZED) 620,630,620
620 LL=1
    GO TO 640
630 LL=0
640 NUME=1
    GO TO 780
*****
C** N(I) IS THE J-TH DIGIT OF A NUMBER WHERE J IS GREATER THAN
C** ONE.
*****
650 IF (NDEC) 660,670,660
660 NDEC=NDEC+1
670 DO 680 L=1,10
    IF (N(I).NE.IDIGIT(L)) GC TO 680
    IN=L-1
    GO TO 690
680 CONTINUE
690 FIN=IN
    ZED=10.*ZED+FIN
    IF (ZED) 700,780,700
700 LL=LL+1
    IF (LL.LT.IZERO) GC TO 780
    IF (LL.EQ.IZERO) GC TO 710
    GO TO 720

```

```

710 CALL ERROR (KOL,A,N,LL,IW,1) RDT02470
    GO TO 780 RDT02480
720 CALL ERROR (KOL,A,N,LL,IW,2) RDT02490
    GO TO 990 RDT02500
*****RDT02510
C** N(I) IS AN EXPONENTIAL DIGIT. **RDT02520
*****RDT02530
730 IF (IDCRE-1) 750,740,750 RDT02540
740 ISIGX=1 RDT02550
750 DO 760 L=1,10 RDT02560
    IF (N(I).NE.IDIGIT(L)) GO TO 760 RDT02570
    IN=L-1 RDT02580
    GO TO 770 RDT02590
760 CONTINUE RDT02600
770 IEXP=10*IEXP+IN RDT02610
    NXDIG=NXDIG+1 RDT02620
    GO TO 780 RDT02630
*****RDT02640
C** DETERMINE IF THE LAST COLUMN OF THE CARD HAS BEEN REACHED. **RDT02650
*****RDT02660
780 IF (I-E0) 980,790,790 RDT02670
*****RDT02680
C** LAST COLUMN HAS BEEN REACHED. **RDT02690
C** END-OF-CARD ROUTINE IS NOW EXECUTED. **RDT02700
*****RDT02710
790 IF (IDCRE) 830,800,830 RDT02720
800 IF (NUMB) 820,810,820 RDT02730
810 SIG=0. RDT02740
    NDE=0 RDT02750
    GO TO 980 RDT02760
820 IF (NDEC) 870,850,870 RDT02770
830 IF (NXDIG) 870,840,870 RDT02780
840 IF (NDEC) 870,850,870 RDT02790
*****RDT02800
C** K-TH NUMBER (WHICH APPEARED IN INTEGER FORM) IS STORED AS **RDT02810
C** Z(K). **RDT02820
*****RDT02830
850 K=K+1 RDT02840
    IF (ZED) 860,960,860 RDT02850
860 Z(K)=SIGN*ZED RDT02860
    GO TO 970 RDT02870
*****RDT02880
C** K-TH NUMBER (WHICH APPEARED IN NON-INTEGER FORM) IS STORED AS **RDT02890
C** Z(K). **RDT02900
*****RDT02910
870 K=K+1 RDT02920
    NDEC=NDEC+NDEEC RDT02930
    IF (ZED) 880,960,880 RDT02940
    KK=LL+ISIGX*IEXP-NDEC+1 RDT02950
    IF (KK.GT.(1-IZERO).AND.KK.LT.IZERO) GO TO 930 RDT02960
    IF (KK.LT.(1-IZERC)) GO TO 890 RDT02970
    IF (KK.EQ.(1-IZERO)) GO TO 900 RDT02980
    IF (KK.EQ.IZERC) GO TO 910 RDT02990
    IF (KK.GT.IZERO) GO TO 920 RDT03000
    CALL ERROR (KOL,A,N,LL,IW,3) RDT03010
    GO TO 930 RDT03020
    CALL ERROR (KOL,A,N,LL,IW,4) RDT03030
    GO TO 930 RDT03040

```

```

910  CALL ERROR (KOL,A,N,LL,IW,5)                                RDT031
      GO TO 930                                                    RDT031
920  CALL ERROR (KOL,A,N,LL,IW,6)                                RDT031
      GO TO 930                                                    RDT031
930  M=KK-LL+IZERO                                                RDT031
      IF (M.LE.0.OR.M.GT.IMAX) GO TO 940                          RDT031
      Z(K)=SIGN*ZED*T(M)                                           RDT031
      GO TO 970                                                    RDT031
940  M=IZERC-LL                                                    RDT031
      IF (M.NE.0) GO TO 950                                        RDT031
C*****RDT031
C**  M IS EQUAL TO ZERC (SPECIAL CASE)                            **RDT031
C*****RDT031
      M=M+1                                                        RDT031
      ZED=ZED*T(M)                                                 RDT031
      M=KK+IZERO-1                                                 RDT031
      Z(K)=SIGN*ZED*T(M)                                           RDT031
      GO TO 970                                                    RDT031
C*****RDT031
C**  M IS NOT EQUAL TO ZERO                                       **RDT031
C*****RDT031
950  ZED=ZED*T(M)                                                  RDT031
      M=KK+IZERO                                                  RDT031
      Z(K)=SIGN*ZED*T(M)                                           RDT031
      GO TO 970                                                    RDT031
960  Z(K)=0.                                                       RDT031
C*****RDT031
C**  APPROPRIATE VARIABLES ARE RE-INITIALIZED.                  **RDT031
C*****RDT031
970  IDORE=0                                                       RDT031
      IEXP=0                                                       RDT031
      ISIGX=0                                                       RDT031
      KK=0                                                         RDT031
      LL=0                                                         RDT031
      M=0                                                         RDT031
      NDEC=0                                                       RDT031
      NODC=0                                                       RDT031
      NUME=0                                                       RDT031
      NXDIG=0                                                       RDT031
      SIGN=0.                                                       RDT031
980  CCNTINUE                                                       RDT031
      RETURN                                                       RDT031
990  KK=0                                                         RDT031
      LL=0                                                         RDT031
      M=0                                                         RDT031
      RETURN                                                       RDT031
C*****RDT031
C**  FORMAT STATEMENTS.                                          **RDT031
C*****RDT031
1000 FORMAT (5A3,65A1)                                           RDT031
1010 FORMAT (80A1)                                               RDT031
      END                                                         RDT031
--- ERROR SUBPROGRAM ---
      SUBROUTINE ERROR (KOL,A,N,LL,IW,KEY)
C*****RDT031
C**  SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION
C**  PROGRAM VERSION OF SEPT.10.1971    WRITTEN BY R.C.RAYEOLD
C**  AND MRS.R.N.VARNER

```

```

C** MODIFIED BY R. N. VARNER SEPT 1979 **ERR00060
C*****ERR00070
C** SUBROUTINE OF THE DIAGNOSTIC OUTPUT FOR READIT SUBROUTINE **ERR00080
C*****ERR00090
    DIMENSION A(1),N(1)
    GO TO (10,20,30,40,50, 60,70), KEY
    WRITE (IW,100) LL
    IF (KOL-16) 90,80,90
    WRITE (IW,110) LL
    IF (KOL-16) 90,80,90
    WRITE (IW,120)
    IF (KOL-16) 90,80,90
    WRITE (IW,130)
    IF (KOL-16) 90,80,90
    WRITE (IW,140)
    IF (KOL-16) 90,80,90
    WRITE (IW,150)
    IF (KOL-16) 90,80,90
    WRITE (IW,160) KOL
    RETURN
    WRITE (IW,180)
    WRITE (IW,170) (A(I),I=1,5),(N(I),I=16,80)
    RETURN
    WRITE (IW,180)
    WRITE (IW,190) (N(I),I=1,80)
    RETURN
C*****ERR00320
C** FORMAT STATEMENTS **ERR00330
C*****ERR00340
100 FORMAT (25H0***** DIAGNOSTIC ***/1X,
2 64H***** THE NUMBER OF SIGNIFICANT DIGITS IN A NUMBER HAS REACHE
3D ,I3,42H. THIS MAY PRODUCE OVERFLOW OR UNDERFLOW.)
110 FORMAT (20H0***** ERROR ***/1X,
2 64H***** THE NUMBER OF SIGNIFICANT DIGITS IN A NUMBER HAS REACHE
3D ,I3,43H. THIS WILL PRODUCE OVERFLOW OR UNDERFLOW.)
120 FORMAT (87H0***** ERROR ***** NUMBER IS TOO SMALL IN ABSOLUTE VERR
2ALUE AND WILL PRODUCE UNDERFLOW.)
130 FORMAT (87H0***** DIAGNOSTIC ***** NUMBER IS SMALL IN ABSOLUTE VERR
2ALUE AND MAY PRODUCE UNDERFLOW.)
140 FORMAT (86H0***** DIAGNOSTIC ***** NUMBER IS LARGE IN ABSOLUTE VERR
2ALUE AND MAY PRODUCE OVERFLOW.)
150 FORMAT (86H0***** ERROR ***** NUMBER IS TOO LARGE IN ABSOLUTE VERR
2ALUE AND WILL PRODUCE OVERFLOW.)
160 FORMAT (44H0***** ERROR ***** THE VALUE OF 'KOL' IS ,I6,
2 27H AND THIS VALUE IS INVALID./1X,
3 50HKOL MUST BE GREATER THAN 0 AND MUST NOT EXCEED 80.)
170 FORMAT (1H ,5A3,65A1)
180 FORMAT (72H THIS OCCURRED IN CONNECTION WITH READING THE DATA ON TERR
2HE FOLLOWING CARD)
190 FORMAT (1H ,80A1)
    END
    PRINT SUBPROGRAM ---
    SUBROUTINE PRINT1
C*****PR100010
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **PR100020
C** PROGRAM VERSION OF SEPT. 10, 1971 WRITTEN BY R.C.RAYECLD **PR100030
C** AND MRS.R.N.VARNER **PR100040
C** MODIFIED BY R. N. VARNER SEPT 1979 **PR100050
C** **PR100060

```

```

C*****PR1000
C** SUBROUTINE TO PRINT TITLE PAGE 1/12/70 **PR1000
C*****PR1000
C** DIMENSION FOR COMMON /PRT1/ VARIABLES **PR1001
C*****PR1001
      DIMENSION B1(72),B2(72),B3(72),B4(72),B5(72),B6(72),B7(72),
      2 IDATE(3)
C*****PR1001
C** LABELED COMMON **PR1001
C*****PR1001
      COMMON /PRT1/ B1,B2,B3,B4,B5,B6,B7,RANERR,SYSEFF,TNOM,L1,L2,L3,L4,
      2 L5,L6,IDATE,IBREST
      COMMON /PRT2/ IPAGE,NOSER,IPGCT
      COMMON /UNITIO/ IR,IW,IP,IPL,ITMP
      WRITE (IW,80)
      WRITE (IW,90) (B1(K),K=1,L1)
      WRITE (IW,90) (B2(K),K=1,L2)
      WRITE (IW,90) (B3(K),K=1,L3)
      WRITE (IW,90) (B4(K),K=1,L4)
      WRITE (IW,90) (B5(K),K=1,L5)
      WRITE (IW,90) (B6(K),K=1,L6)
      WRITE (IW,100) (B7(K),K=1,18)
      DO 70 I=1,6
      IPAGE=IPAGE+1
      WRITE (IW,110) (B1(K),K=1,65),IPAGE
      WRITE (IW,120) (B2(K),K=1,L2)
      WRITE (IW,120) (B3(K),K=1,L3)
      WRITE (IW,130) (B7(K),K=1,16)
      GO TO (10,20,30,40,50, 60), IPAGE
10  CALL TEXT1
      GO TO 70
20  CALL TEXT2
      GO TO 70
30  CALL TEXT3
      GO TO 70
40  CALL TEXT4
      GO TO 70
50  CALL TEXT5
      GO TO 70
60  CALL TEXT6
70  CONTINUE
      RETURN
C*****PR1000
C** FORMAT STATEMENTS **PR1000
C*****PR1000
80  FORMAT (1H1,45X,28FU. S. DEPARTMENT OF COMMERCE/46X,
      2 28FNATIONAL BUREAU OF STANDARDS/45X,
      3 31FNATIONAL ENGINEERING LABORATORY/48X,22FWASHINGTON, D.C. 202
      4//////////59X,11HRE P O R T/55X,3HOF/51X,
      5 22HW A S S V A L U E S/)
90  FORMAT (43X,72A1)
100  FORMAT (//////////26X,11HTEST NUMBER,2X,18A1///60X,
      2 17HFOR THE DIRECTOR,////60X,22HG. E. MATTINGLY, CHIEF/60X,
      3 26HFLUID ENGINEERING DIVISION/60X,
      4 33HCENTER FOR MECHANICAL ENGINEERING/62X,
      5 22HAND PROCESS TECHNOLOGY/60X,
      6 31FNATIONAL ENGINEERING LABORATORY)
110  FORMAT (1H1,65A1,4HPAGE,I3)

```

```

120  FORMAT (1X,72A1)                                PR100650
130  FORMAT (1X,13HTEST NUMBER ,18A1)                PR100660
      END                                              PR100670
--- TEXT1 SUBPROGRAM ---
      SUBROUTINE TEXT1                                TX100010
C*****TX100020
C**  SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **TX100030
C**  PROGRAM VERSION CF SEPT. 10,1971  WRITTEN BY R.C.RAYECLD **TX100040
C**  AND MRS.R.N.VARNER **TX100050
C**  MODIFIED BY R. N. VARNER SEPT 1979 **TX100060
C*****TX100070
      COMMON/UNITIO/IR,IR,IP,IPL,ITMP                TX100080
      WRITE (IW,10)                                    TX100090
      WRITE (IW,20)                                    TX100100
      WRITE (IW,30)                                    TX100110
      WRITE (IW,40)                                    TX100120
      WRITE (IW,50)                                    TX100130
      WRITE (IW,60)                                    TX100140
      WRITE (IW,70)                                    TX100150
      WRITE (IW,80)                                    TX100160
      WRITE (IW,90)                                    TX100170
      WRITE (IW,100)                                   TX100180
      WRITE (IW,110)                                   TX100190
      RETURN                                           TX100200
C *****TX100210
C  FORMAT STATEMENTS                                **TX100220
C *****TX100230
1  FORMAT (/)                                         TX100240
2  FORMAT (1X,36H      INTRODUCTION                 TX100250
236H WHICH RELY, DIRECTLY OR INDIRECT-/             TX100260
31X,36H      ,                                     TX100270
436H LY, ON MASS MEASUREMENTS TO/                   TX100280
51X,36H THIS DOCUMENT IS A COMPREHEN-               TX100290
636H ACCOMPLISH A WIDE VARIETY OF/                  TX100300
71X,36HSIVE REPCRT COVERING THE SEQUENCE            TX100310
836H ENDEAVORS. IN CRDR FOR THIS/                   TX100320
91X,36HCF OPERATIONS USED TO ASSIGN MASS            TX100330
*36H SYSTEM TO FUNCTION PROPERLY.)                 TX100340
      FORMAT (1X,36HVALUES TO THE WEIGHTS IDENTIFIED , TX100350
236H EVERYONE WHO MAKES MEASUREMENTS/               TX100360
31X,36HABOVE. IT INCLUDES A COMPLETE                TX100370
436H MUST BE ABLE TO VERIFY THAT HIS/               TX100380
51X,36HDESCRIPTION OF THE MEASUREMENT              TX100390
636H MEASUREMENT PROCESS PRODUCES/                 TX100400
71X,36HMETHODS AND PROCEDURES WHICH WERE           TX100410
836H CONSISTENT RESULTS WHICH ARE/                  TX100420
91X,36HUSED, ALL OF THE DATA, AND THE              TX100430
*36H COMPATIBLE WITH HIS PARTICULAR)               TX100440
      FORMAT (1X,36HANALYSIS OF THIS DATA. THE ,    TX100450
236H REQUIREMENTS. THE WEIGHTS COVERED/             TX100460
31X,36HRESULTS ARE PRESENTED IN SEVERAL            TX100470
436H BY THIS REPORT, TOGETHER WITH THE/             TX100480
51X,36HFORMATS. ASSIGNED MASS VALUES,             TX100490
636H ASSIGNED VALUES AND THE APPRO-/              TX100500
71X,36HDISPLACEMENT VOLUMES, COEFFICIENTS          TX100510
836H APPROPRIATE UNCERTAINTIES FOR THESE/           TX100520
91X,36HEXPANSION, UNCERTAINTIES, TO-               TX100530
*36H VALUES, PROVIDE IN PART A BASIS)              TX100540

```

|     |                                                   |       |
|-----|---------------------------------------------------|-------|
| 50  | FORMAT (1X,36HGETHER WITH THE SUMMED VALUES FOR . | TX100 |
|     | 236H FOR CONSISTENT MEASUREMENTS WITHIN/          | TX100 |
|     | 31X,36HLINEAR COMBINATIONS OF THE WEIGHTS .       | TX100 |
|     | 436H THIS SYSTEM OF RELATED MEASUREMENT/          | TX100 |
|     | 51X,36HIN EACH DECADE ARE PRESENTED AT .          | TX100 |
|     | 636H PROCESSES. /                                 | TX100 |
|     | 71X,36HTHE END OF THE APPROPRIATE SERIES. .       | TX100 |
|     | 836H /                                            | TX100 |
|     | 91X,36HTHIS INFORMATION SHOULD BE USEFUL .        | TX100 |
|     | *36H APPROPRIATE CHARACTERIZATION)                | TX100 |
| 60  | FORMAT (1X,36HTO THOSE WHO MUST ASSIGN MASS .     | TX100 |
|     | 236H OF ANY MEASUREMENT PROCESS IS/               | TX100 |
|     | 31X,36HVALUES TO OBJECTS OTHER THAN .             | TX100 |
|     | 436H FUNDAMENTAL TO VERIFYING THAT/               | TX100 |
|     | 51X,36HWEIGHTS. FOR CONVENIENCE, THE .            | TX100 |
|     | 636H RESULTS ARE CONSISTENT WITH THE/             | TX100 |
|     | 71X,36HVALUES AND UNCERTAINTIES, TOGETHER .       | TX100 |
|     | 836H END REQUIREMENT WITH RESPECT TO/             | TX100 |
|     | 91X,36HWITH OTHER APPROPRIATE DATA AND .          | TX100 |
|     | *36H CORRECTNESS AND ECONOMY OF THE)              | TX100 |
| 70  | FORMAT (1X,36HCOMMENTS ARE ALSO SUMMARIZED IN .   | TX100 |
|     | 236H MEASUREMENT EFFORT. WITHOUT THIS/            | TX100 |
|     | 31X,36HTABLES I AND II AT THE END OF THE .        | TX100 |
|     | 436H INFORMATION, THE BENEFITS OF/                | TX100 |
|     | 51X,36HREPORT. CERTAIN INTERMEDIATE .             | TX100 |
|     | 636H OWNERSHIP OF THESE WEIGHTS MAY BE/           | TX100 |
|     | 71X,36HPAGES ARE SUMMARIES OF STATISTICAL .       | TX100 |
|     | 836H COMPLETELY ILLUSORY. THE ASSIGNED/           | TX100 |
|     | 91X,36HDATA WHICH RELATE TO THE MASS .            | TX100 |
|     | *36H UNCERTAINTIES IN THIS REPORT ARE)            | TX100 |
| 80  | FORMAT (1X,36HMEASUREMENT PROCESS USED TO .       | TX100 |
|     | 236H DESCRIPTIVE OF OUR MASS MEASURE-/            | TX100 |
|     | 31X,36HPERFORM THIS WORK. THESE PAGES .           | TX100 |
|     | 436H MENT PROCESS. EFFECTIVENESS OF/              | TX100 |
|     | 51X,36HHAVE BEEN LEFT IN THE REPORT TO .          | TX100 |
|     | 636H THE TRANSFER OF THE UNIT FROM ONE/           | TX100 |
|     | 71X,36HRETAIN CONTINUITY. COPIES OF .             | TX100 |
|     | 836H FACILITY TO ANOTHER SHOULD BE/               | TX100 |
|     | 91X,36HTHESE PAGES BECOME PART OF A .             | TX100 |
|     | *36H VERIFIED BY AN INDEPENDENT TEST.)            | TX100 |
| 90  | FORMAT (1X,36HCOLLECTION OF STATISTICAL DATA .    | TX100 |
|     | 236H IT IS PRESUMED THAT THESE WEIGHTS/           | TX100 |
|     | 31X,36HWHICH REFLECTS THE MEASUREMENT .           | TX100 |
|     | 436H WILL BE USED IN A SIMILARLY WELL-/           | TX100 |
|     | 51X,36HPROCESS PERFORMANCE OVER A PERIOD .        | TX100 |
|     | 636H CHARACTERIZED MEASUREMENT PROCESS/           | TX100 |
|     | 71X,36HOF TIME. SUCH A COLLECTION HAS .           | TX100 |
|     | 836H SO THAT THE STATISTICAL PARAMETERS/          | TX100 |
|     | 91X,36HBEEN USED TO ESTABLISH THE CONTROL .       | TX100 |
|     | *36H OF BOTH PROCESSES CAN BE COMBINED)           | TX100 |
| 100 | FORMAT (1X,36HLIMITS FOR ACCEPTING THE RESULTS .  | TX100 |
|     | 236H TO PROVIDE A REALISTIC ESTIMATE OF/          | TX100 |
|     | 31X,36HOF THIS MEASUREMENT. THESE COL- .          | TX100 |
|     | 436H THE UNCERTAINTY OF THE MASS UNIT/            | TX100 |
|     | 51X,36HLECTIONS ARE OPEN FOR INSPECTION .         | TX100 |
|     | 636H AS ACTUALLY REALIZED IN ANOTHER/             | TX100 |
|     | 71X,36HAT OUR FACILITY. .                         | TX100 |
|     | 836H FACILITY. A COMPREHENSIVE SERVICE/           | TX100 |

```

91X,36H
*36H DIRECTED TOWARD THE EVALUATION OF)
110 FORMAT (1X,36H THE MASS MEASUREMENT SYSTEM
236H A PARTICULAR MASS MEASUREMENT/
31X,36H
436H PROCESS IS AVAILABLE THROUGH THE/
51X,36H THE MASS MEASUREMENT SYSTEM
636H MASS MEASUREMENT ASSURANCE PROGRAM/
71X,36H WITHIN THIS COUNTRY CONSISTS OF
836H OF THE NATIONAL BUREAU OF/
91X,36H ALL OF THE MEASUREMENT PROCESSES
*36H STANDARDS.
END
TX101130
TX101140
TX101150
TX101160
TX101170
TX101180
TX101190
TX101200
TX101210
TX101220
TX101230
TX101240
TX101250

--- TEXT2 SUBPROGRAM ---
SUBROUTINE TEXT2
C*****
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **TX200020
C** PROGRAM VERSION OF SEPT. 10, 1971 WRITTEN BY R.C. RAYBOLD **TX200040
C** AND MRS. R.N. VARNER **TX200050
C** MODIFIED BY R. N. VARNER SEPT 1979 **TX200060
C*****
COMMON/UNITIO/IR, IW, IP, IPL, ITMP
WRITE (IW,10)
WRITE (IW,20)
WRITE (IW,30)
WRITE (IW,40)
WRITE (IW,50)
WRITE (IW,60)
WRITE (IW,70)
WRITE (IW,80)
WRITE (IW,90)
WRITE (IW,100)
WRITE (IW,110)
RETURN
TX200010
TX200020
TX200030
TX200040
TX200050
TX200060
TX200070
TX200080
TX200090
TX200100
TX200110
TX200120
TX200130
TX200140
TX200150
TX200160
TX200170
TX200180
TX200190
TX200200
TX200210
C** FORMAT STATEMENTS **TX200220
C*****
10 FORMAT (//)
20 FORMAT (1X,36H WEIGHING DESIGN
236H STANDARD IS MEASURED WITH EACH/
31X,36H
436H TEST OF UNKNOWN AND THE COLLEC-/
51X,36H ONLY DIFFERENCES IN MASS CAN
636H TION OF VALUES OVER TIME IS USED/
71X,36H MEASURED. THEREFORE THE MASS
836H TO EVALUATE THE PERFORMANCE OF THE/
91X,36H VALUES FOR THE 'UNKNOWN' WEIGHTS
*36H MEASUREMENT PROCESS.
FORMAT (1X,36H MUST BE DETERMINED BY COMPARISON
36H /
436H WITH OTHER WEIGHTS WHICH HAVE
51X,36H IN THE CASE OF THE SERIES/
636H ACCEPTED MASS VALUES. THE
71X,36H WHICH INCLUDES THE KNOWN STAND-/
836H 'UNKNOWN' WEIGHTS TOGETHER WITH
91X,36H ARDS, THE ACCEPTED VALUES OF THESE/
*36H CHECK STANDARDS, ARE GROUPED AND
STANDARDS SERVE AS A RESTRAINT ON)
TX200230
TX200240
TX200250
TX200260
TX200270
TX200280
TX200290
TX200300
TX200310
TX200320
TX200330
TX200340
TX200350
TX200360
TX200370
TX200380
TX200390
TX200400
TX200410
TX200420
TX200430
TX200440

```



|    |                                                    |        |
|----|----------------------------------------------------|--------|
| 40 | FORMAT (1X,36HINTERCOMPARED ACCORDING TO THE ,     | TX200, |
|    | 236H THE SOLUTION OF THE EQUATIONS FOR/            | TX200, |
|    | 31X,36HDESIGN SCHEDULE GIVEN AT THE BE- ,          | TX200, |
|    | 436H THE VALUES OF ALL OF THE WEIGHTS./            | TX200, |
|    | 51X,36HGINNING OF EACH SERIES OF WEIGH- ,          | TX200  |
|    | 636H THE RESTRAINT FOR THE SOLUTION OF/            | TX200  |
|    | 71X,36HINGS. THE FIRST SERIES CONTAINS ,           | TX200  |
|    | 836H SUBSEQUENT SERIES IS PROVIDED BY/             | TX200  |
|    | 91X,36HSTANDARDS WHICH PROVIDE THE ,               | TX200  |
|    | *36H THE VALUES ESTABLISHED FOR ONE OR)            | TX200  |
| 50 | FORMAT (1X,36HSTARTING VALUES FOR THE SERIES OF ,  | TX200  |
|    | 236H MORE WEIGHTS INCLUDED IN A/                   | TX200  |
|    | 31X,36HWEIGHINGS AND PROVIDE THE TIE ,             | TX200  |
|    | 436H PREVIOUS SERIES. /                            | TX200  |
|    | 51X,36HPCINT FOR CONSISTENCY THROUGHOUT ,          | TX200  |
|    | 636H /                                             | TX200  |
|    | 71X,36HTHE MEASUREMENT SYSTEM. THE ,               | TX200  |
|    | 836H ESTIMATED VALUES FOR WEIGHTS/                 | TX200  |
|    | 91X,36HWEIGHING METHOD USED, I.E., DOUBLE ,        | TX200  |
|    | *36H WHICH HAVE BEEN GROUPED IN THE)               | TX200  |
| 60 | FORMAT (1X,36HSUBSTITUTION, TRANSPOSITION, ETC., , | TX200  |
|    | 236H SAME SERIES INVOLVE THE SAME/                 | TX200  |
|    | 31X,36HIS INDICATED ALONG WITH THE ,               | TX200  |
|    | 436H OBSERVATIONAL DATA AND ARE, IN/               | TX200  |
|    | 51X,36HCBERVED DATA. IN THE COMPUTA- ,             | TX200  |
|    | 636H ALMOST ALL CASES, CORRELATED, FOR/            | TX200  |
|    | 71X,36HTIONS, THE DISPLACEMENT VOLUMES ,           | TX200  |
|    | 836H EACH SERIES THERE IS A TABLE OF/              | TX200  |
|    | 91X,36HARE TREATED EXPLICITLY, USING THE ,         | TX200  |
|    | *36H COMBINATIONS TOGETHER WITH THE)               | TX200  |
| 70 | FORMAT (1X,36HDATA LISTED IN THE REPORT. IN ALL ,  | TX200  |
|    | 236H APPROPRIATE UNCERTAINTY FOR EACH/             | TX200  |
|    | 31X,36HCASES, A REDUNDANCY IN THE NUMBER ,         | TX200  |
|    | 436H COMBINATION. /                                | TX200  |
|    | 51X,36HOF MEASUREMENTS PROVIDES A MEANS ,          | TX200  |
|    | 636H /                                             | TX200  |
|    | 71X,36HFOR CHECKING ON THE PRECISION OF ,          | TX200  |
|    | 836H PROCESS CONTRL /                              | TX200  |
|    | 91X,36HTHE PROCESS. ,                              | TX200  |
|    | *36H )                                             | TX200  |
| 80 | FORMAT (1X,36H ,                                   | TX200  |
|    | 236H THE STANDARD DEVIATION, AS/                   | TX200  |
|    | 31X,36H WHEN THERE ARE MORE EQUATIONS ,            | TX200  |
|    | 436H COMPUTED FROM THE LEAST SQUARES/              | TX200  |
|    | 51X,36HTHAN 'UNKNOWN', NOT ALL OBSERVA- ,          | TX200  |
|    | 636H SOLUTION, PROVIDES A CHECK ON THE/            | TX200  |
|    | 71X,36HTICNAL EQUATIONS CAN BE SATISFIED ,         | TX200  |
|    | 836H SHORT TERM, OR 'WITHIN-RUN' PRO-/             | TX200  |
|    | 91X,36HEXACTLY AND THE METHOD OF LEAST ,           | TX200  |
|    | *36H CESS PRECISION. AN AVERAGE OF A)              | TX200  |
| 90 | FORMAT (1X,36HSQUARES IS USED TO PROVIDE ,         | TX200  |
|    | 236H NUMBER OF THESE STANDARD DEVIAT-/             | TX200  |
|    | 31X,36HESTIMATES OF THE 'UNKNOWN' VALUES. ,        | TX200  |
|    | 436H TIONS IS TAKEN AS THE ACCEPTED/               | TX200  |
|    | 51X,36HTHIS METHOD LEADS TO ESTIMATORS ,           | TX200  |
|    | 636H WITHIN-RUN STANDARD DEVIATION OF/             | TX200  |
|    | 71X,36HWHICH ARE LINEAR FUNCTIONS OF THE ,         | TX200  |
|    | 836H THE PROCESS AND IS USED AS A/                 | TX200  |

```

91X,36HDATA AND WHICH HAVE STANDARD . TX201030
*36H REFERENCE VALUE FOR SURVEILLANCE) TX201040
100 FORMAT (1X,36HDEVIATIONS READILY CALCULATED FROM . TX201050
236H OF THE PROCESS PRECISION. THE/ TX201060
31X,36HTHE COEFFICIENTS OF THE LINEAR . TX201070
436H VALUES OBTAINED FOR THE 'CHECK/ TX201080
51X,36HFUNCTIONS AND THE STANDARD DEVI- . TX201090
636H STANDARD' PROVIDE, AS TIME GOES/ TX201100
71X,36HTION OF AN INDIVIDUAL MEASUREMENT. . TX201110
836H ON, A SEQUENCE OF VALUES THAT/ TX201120
91X,36HTHE 'CHECK STANDARD' IS ALSO . TX201130
*36H REALISTICALLY REFLECTS THE) TX201140
110 FORMAT (1X,36HTREATED AS AN UNKNOWN AND THE . TX201150
236H VARIATIONS WHICH RESET PRECISE/ TX201160
31X,36HAGREEMENT OF THE CURRENT RESULT . TX201170
436H MEASUREMENTS. COLLECTIONS OF/ TX201180
51X,36HWITH THE ACCEPTED VALUE PROVIDES A . TX201190
636H VALUES FOR BOTH THE WITHIN-RUN/ TX201200
71X,36HTEST OF THE ADEQUACY OF THE CUR- . TX201210
836H PRECISION AND THE VALUE OBTAINED/ TX201220
91X,36HRENT DATA. THIS SAME CHECK . TX201230
*36H FOR THE 'CHECK STANDARD' SHOULD) TX201240
END TX201250

--- TEXT3 SUBPROGRAM ---
SUBROUTINE TEXT3 TX300010
C*****TX300020
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **TX300030
C** PROGRAM VERSION OF SEPT. 10,1971 WRITTEN BY R.C.RAVEOLD **TX300040
C** AND MRS.R.N.VARNER **TX300050
C** MODIFIED BY R. N. VARNER SEPT 1979 **TX300060
C*****TX300070
COMMON/UNITIO/IR, IY, IP, IPL, ITMP
WRITE (IW,10) TX300080
WRITE (IW,20) TX300090
WRITE (IW,30) TX300100
WRITE (IW,40) TX300110
WRITE (IW,50) TX300120
WRITE (IW,60) TX300130
WRITE (IW,70) TX300140
WRITE (IW,80) TX300150
WRITE (IW,90) TX300160
WRITE (IW,100) TX300170
WRITE (IW,110) TX300180
RETURN TX300190
C*****TX300200
C** FORMAT STATEMENTS **TX300210
C*****TX300220
10 FORMAT (/) TX300230
20 FORMAT (1X,36HPOSSESS THE PROPERTIES OF RANDOM- . TX300240
236H SERIES OF MEASUREMENTS JUDGED/ TX300250
31X,36HNESS ASSOCIATED WITH INDEPENDENT . TX300260
436H AS CUT OF CONTROL RELATIVE TO THE/ TX300270
51X,36HMEASUREMENTS FROM A STABLE . TX300280
636H APPROPRIATE PARAMETER ARE CARE-/ TX300290
71X,36HPROBABILITY DISTRIBUTION. THE . TX300300
836H FULLY EXAMINED. IF REGRUS WERE/ TX300310
91X,36HREPORTED 'F RATIO' AND 'T VALUE' . TX300320
*36H NECESSARY IN THE COURSE OF THIS) TX300330

```

30     FORMAT (1X,36HARE TESTS OF THE VALUES FROM THE ,  
       236H WORK, THE 'OUT OF CONTROL' SERIES,/  
       31X,36HCURRENT RUN FOR CONFORMITY TO ,  
       436H WITH REMARKS AS APPROPRIATE, ARE/  
       51X,36HTHEIR RESPECTIVE DISTRIBUTIONS AND ,  
       636H ATTACHED AT THE END OF THE REPORT/  
       71X,36HIF SATISFACTORY ARE TAKEN AS ,  
       836H FOR YOUR INFORMATION. /  
       91X,36HEVIDENCE THAT THE PROCESS IS IN ,  
       \*36H )

40     FORMAT (1X,36HCONTROL AND THAT PREDICTIVE ,  
       236H         UNCERTAINTY /  
       31X,36HSTATEMENTS REGARDING UNCERTAINTY ,  
       436H /  
       51X,36HARE VALID. ,  
       636H IT IS ASSUMED THAT THE PRESENT/  
       71X,36H ,  
       836H 'ACCEPTED VALUES' OF TWO NBS STAN-/  
       91X,36H CONTROL CHARTS ON THE WITHIN- ,  
       \*36H DARDS AT THE 1 KILOGRAM LEVEL.)

50     FORMAT (1X,36HRUN PROCESS PRECISION AND THE ,  
       236H DESIGNATED N1 AND N2, ARE WITHOUT/  
       31X,36HVALUES OBTAINED FOR THE CHECK ,  
       436H ERROR. ESTIMATES OF THE UNCER-/  
       51X,36HSTANDARD ARE KEY ELEMENTS IN ,  
       636H TAINTY OF THE ACCEPTED VALUES OF/  
       71X,36HMONITORING THE STATE OF CONTROL OF ,  
       836H THE NBS STANDARDS RELATIVE TO THE/  
       91X,36HANY PRECISE MASS MEASUREMENT ,  
       \*36H INTERNATIONAL PROTOTYPE KILOGRAM)

60     FORMAT (1X,36HPROCESS. IN ADDITION TO PROVIDING ,  
       236H CAN BE PROVIDED ON REQUEST. /  
       31X,36H A BASIS FOR JUDGMENT AS TO THE ,  
       436H HOWEVER, THESE ESTIMATES HAVE NO/  
       51X,36H ADEQUACY OF A GIVEN PROCESS FOR A ,  
       636H REAL MEANING IN EITHER NATIONAL OR/  
       71X,36H PARTICULAR REQUIREMENT. THESE DATA ,  
       836H INTERNATIONAL COMPARISON. THIS IS/  
       91X,36H PROVIDE A MEANS TO JUDGE THE ,  
       \*36H BECAUSE OF THE LACK OF SUFFICIENT)

70     FORMAT (1X,36HIMPORTANCE OF LONG TERM, OR ,  
       236H DATA TO PROVIDE A REALISTIC/  
       31X,36H 'BETWEEN-RUN' VARIABILITY WHICH ,  
       436H ESTIMATE OF THE UNCERTAINTY IN THE/  
       51X,36H CAN BE CHARACTERIZED BY THE ,  
       636H VALUES ASSIGNED TO THE PROTOTYPE/  
       71X,36H STANDARD DEVIATION OF THE VALUES ,  
       836H KILOGRAMS K20 AND K4, PARTICULARLY/  
       91X,36H ABOUT THE MEAN. IF THERE IS AN ,  
       \*36H IN REGARD TO LONG TERM, OR)

80     FORMAT (1X,36H ADDITIONAL COMPONENT OF VARIANCE ,  
       236H BETWEEN-RUN VARIABILITY. CHANGES/  
       31X,36H ENTERING FROM RUN TO RUN, THIS ,  
       436H IN THE ACCEPTED VALUES FOR THE NBS/  
       51X,36H STANDARD DEVIATION WILL BE LARGER ,  
       636H STANDARDS AT THE KILOGRAM LEVEL,/  
       71X,36H THAN CAN BE ACCOUNTED FOR BY THE ,  
       836H AS AND WHEN THEY OCCUR, WILL BE/

TX300350  
 TX300360  
 TX300370  
 TX300380  
 TX300390  
 TX300400  
 TX300410  
 TX300420  
 TX300430  
 TX300440  
 TX300450  
 TX300460  
 TX300470  
 TX300480  
 TX300490  
 TX300500  
 TX300510  
 TX300520  
 TX300530  
 TX300540  
 TX300550  
 TX300560  
 TX300570  
 TX300580  
 TX300590  
 TX300600  
 TX300610  
 TX300620  
 TX300630  
 TX300640  
 TX300650  
 TX300660  
 TX300670  
 TX300680  
 TX300690  
 TX300700  
 TX300710  
 TX300720  
 TX300730  
 TX300740  
 TX300750  
 TX300760  
 TX300770  
 TX300780  
 TX300790  
 TX300800  
 TX300810  
 TX300820  
 TX300830  
 TX300840  
 TX300850  
 TX300860  
 TX300870  
 TX300880  
 TX300890  
 TX300900  
 TX300910  
 TX300920  
 TX300930  
 TX300940  
 TX300950  
 TX300960  
 TX300970  
 TX300980  
 TX300990  
 TX301000

```

90 91X,36H WITHIN-RUN VARIABILITY. CORRELA- , TX300930
    #36H REPORTED IN THE SCIENTIFIC PAPERS) TX300940
    FORMAT (1X,36HTION STUDIES, AS WELL AS SUPPLE- , TX300950
    236H OF THE BUREAU AND WILL BE GIVEN/ TX300960
    31X,36HMENTAL EXPERIMENTS, ARE USED TO , TX300970
    436H WIDE DISTRIBUTION. IN CASES WHERE/ TX300980
    51X,36HDETECT AND REDUCE THE MAGNITUDE OF , TX300990
    636H SUCH CHANGES MAY BE OF IMPORTANCE./ TX301000
    71X,36HSIGNIFICANT SYSTEMATIC EFFECTS. , TX301010
    836H OR WHERE CONTINUITY IS DESIRED./ TX301020
    91X,36HAPPROPRIATE ACTION. E.G., ADDI- , TX301030
    #36H INSTRUCTIONS WILL BE INCLUDED FOR) TX301040
100 FORMAT (1X,36HTICAL EMPIRICAL CORRECTIONS OR , TX301050
    236H UP-DATING PREVIOUSLY REPORTED/ TX301060
    31X,36HCHANGES IN TECHNIQUE, CAN REDUCE , TX301070
    436H VALUES. WHEN THE VALUES REPORTED/ TX301080
    51X,36HTHE EFFECTS FROM KNOWN SOURCES OF , TX301090
    636H ARE BASED ON THE ACCEPTED VALUES/ TX301100
    71X,36HSYSTEMATIC VARIABILITY TO A , TX301110
    836H OF STANDARDS OTHER THAN STANDARDS/ TX301120
    91X,36HMAGNITUDE WHICH IS NO LONGER , TX301130
    #36H N1 AND N2 MENTIONED ABOVE, THE) TX301140
110 FORMAT (1X,36HIDENTIFIABLE IN THE DATA. IN THE , TX301150
    236H UNCERTAINTY OF THE ACCEPTED VALUE/ TX301160
    31X,36HCASES WHERE A SIGNIFICANT LONG , TX301170
    436H OF THE STANDARD BECOMES A/ TX301180
    51X,36HTERM, OR BETWEEN-RUN, COMPONENT , TX301190
    636H SYSTEMATIC ERROR IN THE ASSIGNMENT/ TX301200
    71X,36HREMAINS THE UNCERTAINTY HAS BEEN , TX301210
    836H OF VALUES TO OTHER STANDARDS AND/ TX301220
    91X,36HAPPROPRIATELY ADJUSTED. , TX301230
    #36H IS INCLUDED IN THE REPORT. ) TX301240
    END TX301250
--- TEXT4 SUBPROGRAM ---
    SUBROUTINE TEXT4 TX400010
C*****TX400020
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **TX400030
C** PROGRAM VERSION OF SEPT. 10, 1971 WRITTEN BY R.C.RAYBOLD **TX400040
C** AND MRS.R.N.VARNER **TX400050
C** MODIFIED BY R. N. VARNER SEPT 1979 **TX400060
C*****TX400070
COMMON/UNITIO/IR,IW,IP,IPL,ITMF TX400080
WRITE (IW,10) TX400090
WRITE (IW,20) TX400100
WRITE (IW,30) TX400110
WRITE (IW,40) TX400120
WRITE (IW,50) TX400130
WRITE (IW,60) TX400140
WRITE (IW,70) TX400150
WRITE (IW,80) TX400160
WRITE (IW,90) TX400170
WRITE (IW,100) TX400180
WRITE (IW,110) TX400190
RETURN TX400200
C*****TX400210
C** FORMAT STATEMENTS **TX400220
C*****TX400230
C** FORMAT (/) TX400240

```

|    |                                                 |          |
|----|-------------------------------------------------|----------|
| 20 | FORMAT (1X,36H A BALANCE UNDER STABLE OPERA-    | TX400250 |
|    | 236H SHOULD ALMOST ALWAYS OVERLAP. IN/          | TX400260 |
|    | 31X,36H TING CONDITIONS WILL EXHIBIT A          | TX400270 |
|    | 436H OTHER WORDS, WHILE A SECOND MEAS-          | TX400280 |
|    | 51X,36H CERTAIN CHARACTERISTIC VARIABILITY      | TX400290 |
|    | 636H MEASUREMENT WILL PRODUCE A DIFFERENT/      | TX400300 |
|    | 71X,36H WHICH CAN BE DESCRIBED BY THE           | TX400310 |
|    | 836H VALUE, THIS VALUE WILL ONLY RARELY/        | TX400320 |
|    | 91X,36H STANDARD DEVIATION FOR SUCH             | TX400330 |
|    | *36H DIFFER FROM THE FIRST VALUE BY)            | TX400340 |
| 30 | FORMAT (1X,36H MEASUREMENTS. THE VALUE FOR A    | TX400350 |
|    | 236H MORE THAN THE SUM OF THE TWO/              | TX400360 |
|    | 31X,36H PARTICULAR WEIGHT DETERMINED IN         | TX400370 |
|    | 436H UNCERTAINTIES. THE UNCERTAINTY/            | TX400380 |
|    | 51X,36H REPEATED TESTS WITH THE SAME            | TX400390 |
|    | 636H BANDS ARE NOT EXPECTED TO OVERLAP/         | TX400400 |
|    | 71X,36H WEIGHING DESIGN WILL HAVE ITS OWN       | TX400410 |
|    | 836H IF SOME EVENT HAS OCCURRED IN THE/         | TX400420 |
|    | 91X,36H STANDARD DEVIATION WHICH WILL BE        | TX400430 |
|    | *36H TIME INTERVAL BETWEEN THE TWO MEAS-)       | TX400440 |
| 40 | FORMAT (1X,36H SOME FUNCTION OF THE BALANCE     | TX400450 |
|    | 236H MEASUREMENTS WHICH WILL CHANGE THE/        | TX400460 |
|    | 31X,36H PRECISION AND (POSSIBLY) OF THE         | TX400470 |
|    | 436H MASS OF THE OBJECT. E.G., AREA-            | TX400480 |
|    | 51X,36H BETWEEN-RUN COMPONENT. AS AN            | TX400490 |
|    | 636H SENS, ABUSE, CORRECTION, IMPROPER/         | TX400500 |
|    | 71X,36H CUTTER LIMIT OF THE DISTRIBUTION OF     | TX400510 |
|    | 836H CLEANING AND THE LIKE.                     | TX400520 |
|    | 91X,36H RANDOM ERRORS, THREE TIMES THE          | TX400530 |
|    | *36H )                                          | TX400540 |
| 50 | FORMAT (1X,36H STANDARD DEVIATION IS USED.      | TX400550 |
|    | 236H THE UNCERTAINTY IN ASSIGNED/               | TX400560 |
|    | 31X,36H SYSTEMATIC ERRORS DUE TO THE            | TX400570 |
|    | 436H VALUE CONTAINED IN THIS REPORT/            | TX400580 |
|    | 51X,36H PROCEDURES USED OR TO ENVIRON-          | TX400590 |
|    | 636H BECOMES A SYSTEMATIC EFFECT FOR/           | TX400600 |
|    | 71X,36H MENTAL EFFECTS ARE LARGELY              | TX400610 |
|    | 836H THE MEASUREMENT PROCESS IN WHICH/          | TX400620 |
|    | 91X,36H BALANCED CUT AND CAN USUALLY BE         | TX400630 |
|    | *36H THESE WEIGHTS ARE TO BE USED. IN)          | TX400640 |
| 60 | FORMAT (1X,36H REGARDED AS NEGLIGIBLE. WHEN A   | TX400650 |
|    | 236H THE ABSENCE OF OTHER SIGNIFICANT/          | TX400660 |
|    | 31X,36H CAN BE NEGLIGIBLE BECAUSE OF THE        | TX400670 |
|    | 436H SYSTEMATIC EFFECTS IN THE USER'S/          | TX400680 |
|    | 51X,36H POSSIBLE EFFECT FROM KNOWN SOURCES      | TX400690 |
|    | 636H MEASUREMENT PROCESS (A CONDITION/          | TX400700 |
|    | 71X,36H IS AVAILABLE, IT IS CALCULATED AND      | TX400710 |
|    | 836H WHICH MUST BE DEMONSTRATED) THE/           | TX400720 |
|    | 91X,36H REPORTED SEPARATELY. E.G., THE          | TX400730 |
|    | *36H UNCERTAINTY OF THE VALUE ASSIGNED)         | TX400740 |
| 70 | FORMAT (1X,36H UNCERTAINTY OF ACCEPTED VALUE AT | TX400750 |
|    | 236H BY THE USER IS AN APPROPRIATE/             | TX400760 |
|    | 31X,36H OTHER THAN THE 1 KILOGRAM LEVEL.        | TX400770 |
|    | 436H COMBINATION OF THE SYSTEMATIC/             | TX400780 |
|    | 51X,36H THE DISTRIBUTION IMPLIED BY THE         | TX400790 |
|    | 636H ERROR IN THE STANDARD AND THE/             | TX400800 |
|    | 71X,36H RANDOM ERRORS MAY THUS BE CENTERED      | TX400810 |
|    | 836H RANDOM COMPONENT ASSOCIATED WITH/          | TX400820 |

|     |                                                                 |            |
|-----|-----------------------------------------------------------------|------------|
|     | 91X,36H SOMEWHERE IN THE RANGE GIVEN BY                         | TX400830   |
|     | *36H HIS PROCESS. IF THE MEASUREMENT                            | TX400840   |
| 80  | FORMAT (1X,36H THE BOUNDS TO THE SYSTEMATIC                     | TX400850   |
|     | 236H PROCESSES ARE IN CONTROL AND/                              | TX400860   |
|     | 31X,36H ERROR. THE TOTAL UNCERTAINTY IS                         | TX400870   |
|     | 436H APPROPRIATE UNCERTAINTIES ARE/                             | TX400880   |
|     | 51X,36H TAKEN AS THE SUM OF THESE TWO                           | TX400890   |
|     | 636H ASSIGNED. THE VALUES PRODUCED BY/                          | TX400900   |
|     | 71X,36H COMPONENTS.                                             | TX400910   |
|     | 836H DIFFERENT MEASUREMENT FACILITIES/                          | TX400920   |
|     | 91X,36H                                                         | TX400930   |
|     | *36H WILL HAVE OVERLAPPING UNCERTAINTY)                         | TX400940   |
| 90  | FORMAT (1X,36H THE UNCERTAINTY ASSOCIATED                       | TX400950   |
|     | 236H BOUNDS AS DESCRIBED ABOVE. ONE/                            | TX400960   |
|     | 31X,36H WITH THE ASSIGNED VALUE CAN BE                          | TX400970   |
|     | 436H CANNOT DISCUSS DIFFERENCES IN/                             | TX400980   |
|     | 51X,36H THOUGHT OF AS A BOUND TO THE                            | TX400990   |
|     | 636H VALUES FOR THE SAME OBJECT/                                | TX401000   |
|     | 71X,36H DEPARTURE OF THE ASSIGNED VALUE                         | TX401010   |
|     | 836H OBTAINED BY DIFFERENT FACILITIES/                          | TX401020   |
|     | 91X,36H FROM A HYPOTHETICAL AVERAGE VALUE                       | TX401030   |
|     | *36H WITH ANY DEGREE OF SERIOUSNESS UN-)                        | TX401040   |
| 100 | FORMAT (1X,36H THAT WOULD BE OBTAINED IF IT WERE                | TX401050   |
|     | 236H LESS EACH VALUE IS ACCOMPANIED BY/                         | TX401060   |
|     | 31X,36H POSSIBLE TO REPEAT THE MEASUREMENT                      | TX401070   |
|     | 436H A REALISTIC UNCERTAINTY STATEMENT./                        | TX401080   |
|     | 51X,36H MANY TIMES OVER A WIDE VARIETY OF                       | TX401090   |
|     | 636H                                                            | TX401100   |
|     | 71X,36H CONDITIONS. E.G., SUBSTITUTE THE                        | TX401110   |
|     | 836H                                                            | TX401120   |
|     | 91X,36H WEIGHT FOR ONE OF THE CHECK                             | TX401130   |
|     | *36H                                                            | TX401140   |
| 110 | FORMAT (1X,36H STANDARDS. THIS MEANS THAT THE                   | TX401150   |
|     | 236H                                                            | TX401160   |
|     | 31X,36H UNCERTAINTY BOUND CENTERED ON THE                       | TX401170   |
|     | 436H                                                            | TX401180   |
|     | 51X,36H VALUES OBTAINED FROM EACH OF TWO                        | TX401190   |
|     | 636H                                                            | TX401200   |
|     | 71X,36H MEASUREMENTS OF THE SAME OBJECT                         | TX401210   |
|     | 836H                                                            | TX401220   |
|     | 91X,36H OVER SOME ARBITRARY TIME INTERVAL                       | TX401230   |
|     | *36H                                                            | TX401240   |
|     | END                                                             | TX401250   |
|     | TEXTS SUBPROGRAM ---                                            |            |
|     | SUBROUTINE TEXTS                                                | TX500010   |
|     | *****                                                           | TX500020   |
|     | SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION | **TX500030 |
|     | PROGRAM VERSION OF SEPT. 10, 1971 WRITTEN BY R.C. RAYECLD       | **TX500040 |
|     | AND MRS. R.N. VARNER                                            | **TX500050 |
|     | MODIFIED BY R. N. VARNER SEPT 1977                              | **TX500060 |
|     | *****                                                           | TX500070   |
|     | COMMON/UNITIO/IR, IR, IP, IPL, ITMP                             | TX500080   |
|     | WRITE (IW,10)                                                   | TX500090   |
|     | WRITE (IW,20)                                                   | TX500100   |
|     | WRITE (IW,30)                                                   | TX500110   |
|     | WRITE (IW,40)                                                   | TX500120   |
|     | WRITE (IW,50)                                                   | TX500130   |
|     | WRITE (IW,60)                                                   | TX500140   |

|                                                   |             |            |
|---------------------------------------------------|-------------|------------|
| WRITE (IW,70)                                     |             | TX500150   |
| WRITE (IW,80)                                     |             | TX500160   |
| WRITE (IW,90)                                     |             | TX500170   |
| WRITE (IW,100)                                    |             | TX500180   |
| WRITE (IW,110)                                    |             | TX500190   |
| RETURN                                            |             | TX500200   |
| C*****                                            |             | TX500210   |
| C** FORMAT STATEMENTS                             |             | **TX500220 |
| C*****                                            |             | TX500230   |
| 10 FORMAT (1X,36H                                 |             | TX500240   |
| 236H                                              | )           | TX500250   |
| 20 FORMAT (1X,36H                                 | REFERENCES, | TX500260   |
| 236H                                              | /           | TX500270   |
| 31X,36H                                           | ,           | TX500280   |
| 436H                                              | /           | TX500290   |
| 51X,36H THE FOLLOWING REFERENCES ARE SUGGEST,     |             | TX500300   |
| 636HED FOR DETAILED DESCRIPTION OF                | /           | TX500310   |
| 71X,36HPORTIONS OF THIS REPORT , AND FOR GE,      |             | TX500320   |
| 836HNERAL INFORMATON CONCERNING THE               | /           | TX500330   |
| 91X,36HMASS MEASUREMENT PROCESS:                  | .           | TX500340   |
| *36H                                              | )           | TX500350   |
| 30 FORMAT (1X,36H                                 |             | TX500360   |
| 236H                                              | /           | TX500370   |
| 31X,36H 1.PCNTIUS, F. E., AND CAMERON, J. M,      |             | TX500380   |
| 436H.                                             | /           | TX500390   |
| 51X,36H REALISTIC UNCERTAINTIES AND THE M,        |             | TX500400   |
| 636HASS MEASUREMENT PROCESS                       | /           | TX500410   |
| 71X,36H NAT. BUR. STAND. (U.S.), MONOGR. ,        |             | TX500420   |
| 836H103                                           | /           | TX500430   |
| 91X,36H (AUG. 15, 1967)                           | .           | TX500440   |
| *36H                                              | )           | TX500450   |
| 40 FORMAT (1X,36H                                 |             | TX500460   |
| 236H                                              | /           | TX500470   |
| 31X,36H 2.PCNTIUS, F. E.                          | ,           | TX500480   |
| 436H                                              | /           | TX500490   |
| 51X,36H MEASUREMENT PHILOSOPHY OF THE PIL,        |             | TX500500   |
| 636HCT PROGRAM FOR MASS CALIBRATION               | /           | TX500510   |
| 71X,36H NAT. BUR. STAND. (U.S.) TECH. NOT,        |             | TX500520   |
| 836HE 288                                         | /           | TX500530   |
| 91X,36H (MAY 6, 1966)                             | .           | TX500540   |
| *36H                                              | )           | TX500550   |
| 50 FORMAT (1X,36H                                 |             | TX500560   |
| 236H                                              | /           | TX500570   |
| 31X,36H 3.BOWMAN, F. A., AND SCHOCNOVER, R.,      |             | TX500580   |
| 436H M. WITH APPENDIX BY MILDRED JONES            | /           | TX500590   |
| 51X,36H PROCEDURE FOR HIGH PRECISION DENS,        |             | TX500600   |
| 636HITY DETERMINATIONS BY HYDROSTATIC             | /           | TX500610   |
| 71X,36H WEIGHING                                  | ,           | TX500620   |
| 836H                                              | /           | TX500630   |
| 91X,36H J. RES. NAT. BUR. STAND. (U.S.) 71C,      |             | TX500640   |
| *36H. ENGINEERING AND INSTRUMENTATION             | )           | TX500650   |
| 60 FORMAT (1X,36H AC. 3, 179-198 (JULY-AUG. 1967) | ,           | TX500660   |
| 236H                                              | /           | TX500670   |
| 31X,36H                                           | .           | TX500680   |
| 436H                                              | /           | TX500690   |
| 51X,36H 4.NATRELLA, M. E.                         | ,           | TX500700   |
| 636H                                              | /           | TX500710   |
| 71X,36H EXPERIMENTAL STATISTICS                   | .           | TX500720   |

|     |                                                                 |   |          |
|-----|-----------------------------------------------------------------|---|----------|
|     | 836H                                                            | / | TX500730 |
|     | 91X,36H NAT. BUR. STAND. (U.S.) HANDBOOK ,                      |   | TX500740 |
|     | *36H91                                                          | ) | TX500750 |
| 70  | FORMAT (1X,36H (AUGUST 1, 1963)                                 |   | TX500760 |
|     | 236H                                                            | / | TX500770 |
|     | 31X,36H                                                         | . | TX500780 |
|     | 436H                                                            | / | TX500790 |
|     | 51X,36H 5.KU, H. H.                                             | . | TX500800 |
|     | 636H                                                            | / | TX500810 |
|     | 71X,36H PRECISION MEASUREMENT AND CALIBRA.                      |   | TX500820 |
|     | 836HTICN - SELECTED NES PAPERS ON                               | / | TX500830 |
|     | 91X,36H STATISTICAL CCNCEPTS AND PROCEDUR.                      |   | TX500840 |
|     | *36HES                                                          | ) | TX500850 |
| 80  | FORMAT (1X,36H NAT. BUR. STAND. (U.S.) SPEC. PUB.               |   | TX500860 |
|     | 236HL. 300                                                      | / | TX500870 |
|     | 31X,36H VOL. 1 (FEB. 1969)                                      | . | TX500880 |
|     | 436H                                                            | / | TX500890 |
|     | 51X,36H                                                         | ) | TX500900 |
| 90  | FORMAT (1X,36H 6.PCNTIUS, P. E.                                 |   | TX500910 |
|     | 236H                                                            | / | TX500920 |
|     | 31X,36H MASS AND MASS VALUES                                    | . | TX500930 |
|     | 436H                                                            | / | TX500940 |
|     | 51X,36H NAT. BUR. STAND. (U.S.) MCNOGR. 1.                      |   | TX500950 |
|     | 636H33                                                          | / | TX500960 |
|     | 71X,36H (JAN. 1974)                                             | . | TX500970 |
|     | 836H                                                            | / | TX500980 |
|     | 91X,36H                                                         | . | TX500990 |
|     | *36H                                                            | ) | TX501000 |
| 100 | FORMAT (1X,36H 7.CAMERON, J. M., CROARKIN, C. C. A.             |   | TX501010 |
|     | 236HND RAYBCLD, R. C.                                           | / | TX501020 |
|     | 31X,36H DESIGNS FOR THE CALIBRATION OF ST,                      |   | TX501030 |
|     | 436HANDARDS OF MASS                                             | / | TX501040 |
|     | 51X,36H NAT. BUR. STAND. (U.S.) TECH. NOT.                      |   | TX501050 |
|     | 536HE 952                                                       | / | TX501060 |
|     | 71X,36H (JUNE 1977)                                             | . | TX501070 |
|     | 836H                                                            | / | TX501080 |
|     | 91X,36H                                                         | . | TX501090 |
|     | *36H                                                            | ) | TX501100 |
| 110 | FORMAT (1X,36H 8.VARNER, R. N., AND RAYBOLD, R. C.,             |   | TX501110 |
|     | 236H                                                            | / | TX501120 |
|     | 31X,36H NATIONAL BUREAU OF STANDARDS MASS,                      |   | TX501130 |
|     | 436H CALIBRATION COMPUTER SOFTWARE                              | / | TX501140 |
|     | 51X,36H NAT. BUR. STAND. (U.S.) TECH. NOT.                      |   | TX501150 |
|     | 636HE                                                           | / | TX501160 |
|     | 71X,36H (IN PROCESS)                                            | . | TX501170 |
|     | 836H                                                            | / | TX501180 |
|     | 91X,36H                                                         | . | TX501190 |
|     | *36H                                                            | ) | TX501200 |
|     | END                                                             |   | TX501210 |
|     | TEXT6 SUBPROGRAM ---                                            |   |          |
|     | SUBROUTINE TEXT6                                                |   | TX600010 |
|     | *****                                                           |   | TX600020 |
|     | SUBROUTINE OF THE NATICNAL BUREAU OF STANDARDS MASS CALIBRATION |   | TX600030 |
|     | PROGRAM VERSION OF SEPT. 10, 1971 WRITTEN BY R.C. RAYEOLD       |   | TX600040 |
|     | AND MRS.R.N. VARNER                                             |   | TX600050 |
|     | MODIFIED BY R. N. VARNER SEPT 1977                              |   | TX600060 |
|     | *****                                                           |   | TX600070 |
|     | COMMON/UNITIO/IR, IW, IP, IPL, ITMP                             |   | TX600080 |



```

WRITE (IW,10)
WRITE (IW,20)
WRITE (IW,30)
RETURN
C*****TX600090
C** FORMAT STATEMENTS**TX600100
C*****TX600110
10 FORMAT (1X,36H TC BE PUBLISHED:TX600120
236H /TX600130
31X,36H /TX600140
436H /TX600150
51X,36H 9.PONTIUS, P. E. /TX600160
636H /TX600170
71X,36H THE ACCEPTED VALUES AND ASSOCIATE, /TX600180
836H UNCERTAINTY ESTIMATES OF THE NBS /TX600190
91X,36H STANDARDS AT THE 1 KG LEVEL ,TX600200
*36H )TX600210
20 FORMAT (1X,36H NAT. BUR. STAND. (U.S.) TECH. NOT, TX600220
236H /TX600230
31X,36H (EXPECTED COMPLETION: 1975) ,TX600240
436H /TX600250
51X,36H /TX600260
636H /TX600270
71X,36H 10.PONTIUS, P. E. /TX600280
836H /TX600290
91X,36H DOCUMENTATION FOR THE MASS MEASUR, TX600300
*36H ELEMENT PROCESS AT NBS )TX600310
30 FORMAT (1X,36H NAT. BUR. STAND. (U.S.) TECH. NOT, TX600320
236H /TX600330
31X,36H (EXPECTED COMPLETION: 1974) ,TX600340
436H )TX600350
END TX600360
--- READ2 SUBPROGRAM ---
SUBROUTINE READ2 RD200010
C*****RD200020
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **RD200030
C** PROGRAM VERSION OF SEPT.10,1971 WRITTEN BY R.C.RAYBOLD **RD200040
C** AND MRS.R.N.VARNER **RD200050
C** MODIFIED BY R. N. VARNER SEPT 1979 **RD200060
C*****RD200070
C** SUBROUTINE TO READ DATA FOR EACH NEW SERIES **RD200080
C*****RD200090
C** DIMENSION FOR COMMON /PRT1/ VARIABLES **RD200100
C*****RD200110
DIMENSION B1(72),B2(72),B3(72),B4(72),B5(72),B6(72),B7(72), RD200120
2 IDATE(3) RD200130
C*****RD200140
C** DIMENSION FOR COMMON /INPL/ VARIABLES **RD200150
C*****RD200160
DIMENSION AIDCST(5,15),ANCM(15),DENSTY(15),COEFEX(15),ACCVAL(15), RD200170
2 ARSTIN(15),ACKSTD(15),IRSTCU(15),IPRNT(15),DESNAT(15,50), RD200180
3 OBSERV(600),ALCOM(15,20) RD200190
C*****RD200200
C** DIMENSION FOR COMMON /RAREA/ VARIABLES **RD200210
C*****RD200220
DIMENSION AA(72),AAITEM(5) RD200230
C*****RD200240
C** DIMENSION FOR COMMON /PRTLE/ VARIABLE **RD200250

```

```

C*****RD200260
  DIMENSION PRTLEX(15)RD200270
C*****RD200280
C**  LABELED COMMONRD200290
C*****RD200300
  COMMON /PRT1/ B1,B2,B3,B4,B5,B6,B7,RANERR,SYSEFF,TNCM,L1,L2,L3,L4,RD200310
  2 L5,L6,IDATE,IERESTRD200320
  COMMON /INPUT/ TEAR,PBAR,FEAR,STDEBA,SWT,VSWT,CEXSWT,AIDCST,ANCM,RD200330
  2 DENSITY,COEFEX,ACCVAL,ARSTIN,ACKSTD,DESMAT,CBSEFF,VARBAL,ALCOM,T1PRD200340
  3,T2P,P1P,P2P,H1P,H2P,CP1,CP2,CT1,CT2,CH1,CH2,OT1P,OT2P,CF1P,OP2P,RD200350
  4 OH1P,CH2P,IOP,IEAL,NCES,NUNKA,IRSTCU,IPRNT,ITPOS,ICKUSD,ICALDS,RD200360
  5 LINVAR,N3,N4RD200370
  COMMON /RARE/ AA,AAITEMRD200380
  COMMON /PRTLB/ PRTLEXRD200390
C*****RD200400
C**  SET UP VARIABLE FOR MAXIMUM NUMBER OF READINGS NR=50*12RD200410
C*****RD200420
  NR=50*12RD200430
C*****RD200440
C**  READ ONE CARD--FREE FIELDRD200450
C**    TYPE OF WEIGHING (N1)RD200460
C**    TYPE OF BALANCE (N2)RD200470
C**    TYPE OF UNITS OF INPUT (N3)RD200480
C**    DIRECTION OF SCALE LEFT TO RIGHT OR RIGHT TO LEFT (N4)RD200490
C**RD200500
C**  WHERERD200510
C**    N1 = 0    SINGLE TRANSFCSTIONRD200520
C**    N1 = 1    DOUBLE TRANSFCSTIONRD200530
C**    N1 = 2    SINGLE SUBSTITUTIONRD200540
C**    N1 = 3    DOUBLE SUBSTITUTIONRD200550
C**RD200560
C**    N2 = 0    TWO PAN BALANCERD200570
C**    N2 = 1    ONE PAN BALANCERD200580
C**RD200590
C**    N3 = 0    METRIC UNITSRD200600
C**    N3 = 1    ENGLISH UNITSRD200610
C**RD200620
C**    N4 = 0    SCALE LEFT TO RIGHT (EG. 0 TO 100 OR -1 TO 1)RD200630
C**    N4 = 1    SCALE RIGHT TO LEFT (EG. 100 TO 0 OR 1 TO -1)RD200640
C*****RD200650
  CALL READIT (AA,1,AAITEM)RD200660
  N1=INT(AA(1)+.1)RD200670
  N2=INT(AA(2)+.1)RD200680
  N3=INT(AA(3)+.1)RD200690
  N4=INT(AA(4)+.1)RD200700
C*****RD200710
  READ DATE OF OBSERVATION,OPERATOR,BALANCE AND CHECK STANDARDRD200720
  USED RD200730
  WHERE  NUMBERS ARE FREE FIELD ON THIS CARD  6 NUMBERSRD200740
  MONTH AND DAY (2 DIGITS) YEAR (4 DIGITS)RD200750
  OPERATOR CODE (INTEGER)RD200760
  BALANCE CODE (INTEGER)RD200770
  CHECK STANDARD CODE (INTEGER)RD200780
C*****RD200790
  CALL READIT (AA,1,AAITEM)RD200800
  IDATE(1)=INT(AA(1)+.1)RD200810
  IDATE(2)=INT(AA(2)+.1)RD200820
  IDATE(3)=INT(AA(3)+.1)RD200830

```

```

IOP=INT(AA(4)+.1)
IBAL=INT(AA(5)+.1)
ICKUSD=INT(AA(6)+.1)
C*****RD200840
C** READ TEMPERATURE,PRESSURE,HUMIDITY AND THEIR CORRECTIONS **RD200850
C** ONE RECCRD FREE FIELD WITH ENTRIES IN THE FOLLOWING ORDER **RD200860
C** 1-TEMPERATURE AT BEGINNING OF THE SERIES IN DEGREES C **RD200870
C** 2-TEMPERATURE AT END OF THE SERIES IN DEGREES C **RD200880
C** 3-PRESSURE AT BEGINNING IN MM HG **RD200890
C** 4-PRESSURE AT END IN MM HG **RD200900
C** 5-HUMIDITY AT BEGINNING IN PERCENT **RD200910
C** 6-HUMIDITY AT END IN PERCENT **RD200920
C** 7-PRESSURE CORRECTION BEGINNING IN MM HG **RD200930
C** 8-PRESSURE CORRECTION END IN MM HG **RD200940
C** 9-TEMPERATURE CORRECTION BEGINNING IN DEGREES C **RD200950
C** 10-TEMPERATURE CORRECTION END IN DEGREES C **RD200960
C** 11-HUMIDITY CORRECTION BEGINNING IN PERCENT **RD200970
C** 12-HUMIDITY CORRECTION END IN PERCENT **RD200980
C*****RD200990
CALL READIT (AA,1,AAITEM)
T1P=AA(1)
T2P=AA(2)
P1P=AA(3)
P2P=AA(4)
H1P=AA(5)
H2P=AA(6)
CP1=AA(7)
CP2=AA(8)
CT1=AA(9)
CT2=AA(10)
CH1=AA(11)
CH2=AA(12)
C*****RD201000
C** SAVE VALUES FOR OUTPUT **RD201010
C** COMPUTE AVERAGE CORRECTED TEMPERATURE, PRESSURE AND HUMIDITY **RD201020
C*****RD201030
OT1F=T1P
OT2F=T2P
OP1P=P1P
OP2P=P2P
OH1P=H1P
OH2P=H2P
T1P=T1P+CT1
T2P=T2P+CT2
P1P=P1P+CP1
P2P=P2P+CP2
H1P=H1P+CH1
H2P=H2P+CH2
TEAR=(T1P+T2P)/2.
PBAR=(P1P+P2P)/2.
HEAR=(H1P+H2P)/2.
C*****RD201040
C** ONE CARD IN THE FOLLOWING ORDER FREE FIELD **RD201050
C** 1. NUMBER OF OBSERVATION ( MAX. = 50 ) **RD201060
C** 2. NUMBER OF UNKNOWNS ( MAX. = 15 ) **RD201070
C** 3. CALIBRATION DESIGN CODE ( INTEGER 0000 TO 9999 ) **RD201080
C** 4. NUMBER OF LINEAR COMBINATION ( MAX. = 20 ) **RD201090
C*****RD201100

```

```

CALL READIT (AA,1,AAITEM) RD201420
NOBS=INT(AA(1)+.1) RD201430
NUNKN=INT(AA(2)+.1) RD201440
ICALDS=INT(AA(3)+.1) RD201450
LINVAR=INT(AA(4)+.1) RD201460
C*****RD201470
C** ONE RECORD-FREE FIELD-WITH ENTRIES IN THE FOLLOWING ORDER **RD201480
C** 1. WITHIN STANDARD DEVIATION OF THE BALANCE IN MG **RD201490
C** 2. MASS VALUE IN MG OF THE SENSITIVITY WEIGHT **RD201500
C** 3. VOLUME OF THE SENSITIVITY WEIGHT IN CM3 **RD201510
C** 4. COEFFICIENT OF EXPANSION OF SENSITIVITY WEIGHT **RD201520
C** 5. BETWEEN STANDARD DEVIATION OF THE BALANCE IN MG **RD201530
C*****RD201540
CALL READIT (AA,1,AAITEM) RD201550
STDEBA=AA(1) RD201560
SWT=AA(2) RD201570
VSWT=AA(3) RD201580
CEXSWT=AA(4) RD201590
VAREAL=AA(5) RD201600
C*****RD201610
C** A GROUP OF N RECORDS (N=NUMBER OF UNKNOWNNS). EACH RECORD HAS **RD201620
C** INFORMATION IN THE FOLLOWING ORDER. ABOUT EACH WEIGHT **RD201630
C** 1. IN COLUMNS 1-15 IDENTIFICATION OF WEIGHTS (ALPHA-NUMERIC) **RD201640
C** **RD201650
C** THE REST OF THE INFORMATION ON EACH RECORD IS FREE FIELD **RD201660
C** **RD201670
C** 2. WEIGHTS NOMINAL VALUE IN GRAMS ( POUNDS IF N3 = 1 ) **RD201680
C** 3. DENSITY OF THE WEIGHT IN CM3 **RD201690
C** 4. COEFFICIENT OF EXPANSION OF THE WEIGHT **RD201700
C** 5. ACCEPTED CORRECTION OF THE WEIGHT IN MG **RD201710
C*****RD201720
DO 20 I=1,NUNKN RD201730
CALL READIT (AA,16,AAITEM) RD201740
AIDCST(1,I)=AAITEM(1) RD201750
AIDCST(2,I)=AAITEM(2) RD201760
AIDCST(3,I)=AAITEM(3) RD201770
AIDCST(4,I)=AAITEM(4) RD201780
AIDCST(5,I)=AAITEM(5) RD201790
ANOM(I)=AA(1) RD201800
IF (N3.EQ.0) GO TO 10 RD201810
PRTLBX(I)=ANOM(I) RD201820
C*****RD201830
C** CONVERT POUNDS TO GRAMS **RD201840
C*****RD201850
ANCM(I)=ANCM(I)*453.59237 RD201860
CONTINUE RD201870
DENSIV(I)=AA(2) RD201880
COEFEX(I)=AA(3) RD201890
ACCVAL(I)=AA(4) RD201900
CONTINUE RD201910
C*****RD201920
C** READ. FREE FIELD. ONE RECORD A VECTOR CONSISTING OF 1'S OR 0'S **RD201930
C** USE A 1 IF WEIGHT IS INCLUDED IN RESTRAINT **RD201940
C** USE A 0 IF WEIGHT IS NOT INCLUDED IN THE RESTRAINT **RD201950
C** FOLLOW ORDER OF INPUT **RD201960
C*****RD201970
CALL READIT (AA,1,AAITEM) RD201980
DO 20 I=1,NUNKN RD201990

```

```

30  ARSTIN(I)=AA(I) RD202
*****RD202
C**  READ, FREE FIELD, ONE RECCRD CONSISTING OF 1'S AND 0'S **RD202
C**  USE 1 IF WEIGHT IS INCLUDED IN CHECK STANDARD **RD202
C**  USE 0 IF WEIGHT IS NOT INCLUDED IN CHECK STANDARD **RD202
C**  FOLLOW CRDR OF INFLT **RD202
*****RD202
CALL READIT (AA,1,AAITEM) RD202
DO 40 I=1,NUNKN RD202
40  ACKSTD(I)=AA(I) RD202
*****RD202
C**  READ, FREE FIELD, ONE RECCRD CONSISTING OF 1'S AND 0'S **RD202
C**  USE 1 IF WEIGHT IS USED IN RESTRAINT FOR NEXT SERIES **RD202
C**  USE 0 IF WEIGHT IS NOT USED IN RESTRAINT FOR NEXT SERIES **RD202
C**  FOLLOW CRDR ON INFLT **RD202
*****RD202
CALL READIT (AA,1,AAITEM) RD202
DO 50 I=1,NUNKN RD202
50 IRSTOU(I)=INT(AA(I)+.1) RD202
*****RD202
C**  READ IN FREE FIELD A CARD WITH A VECTOR OF 1 IF WEIGHT IS ONE **RD202
C**  WHICH IS TO BE PRINTED IN THE REPORT AND 0 IF IT IS NOT, THE **RD202
C**  READ, FREE FIELD, ONE RECCRD CONSISTING OF 1'S AND 0'S **RD202
C**  USE 1 IF WEIGHT IS TO BE PRINTED IN REPORT **RD202
C**  USE 0 IF WEIGHT IS OMITTED IN REPORT **RD202
C**  USE CRDR OF INPUT **RD202
*****RD202
CALL READIT (AA,1,AAITEM) RD202
DO 60 I=1,NUNKN RD202
60  IPRNT(I)=INT(AA(I)+.1) RD202
*****RD202
C**  READ DESIGN MATRIX **RD202
C**  READ IN N (NUMBER OF OBSERVATIONS) RECORDS WITH K (NUMBER **RD202
C**  OF UNKNOWNNS) **RD202
C**  ENTRIES PER RECORD IN THE CRDR OF THE **RD202
C**  IDENTIFICATION CARDS WHERE THE WEIGHING EQUATION ARE SET UP BY **RD202
C**  A SERIES OF 0,-1 AND 1. RD202
*****RD202
DO 80 I=1,NOBS RD202
CALL READIT (AA,1,AAITEM) RD202
DO 70 J=1,NUNKN RD202
DESMAT(J,I)=AA(J) RD202
70  CONTINUE RD202
80  CONTINUE RD202
*****RD202
C**  READ IN LINVAR (NUMBER OF LINEAR COMBINATIONS) RECORDS **RD202
C**  TO COMPUTE VARIANCE OF LINEAR COMBINATIONS OF WEIGHTS **RD202
C**  EACH RECORD CONTAINS K (NUMBER OF UNKNOWNNS) ENTRIES **RD202
C**  ENTRIES OF 0'S, -1'S, AND 1'S ARE USED **RD202
C**  FOLLOW CRDR OF INFLT **RD202
*****RD202
IF (LINVAR.EQ.0) GC TO 110
DO 100 I=1,LINVAR
CALL READIT (AA,1,AAITEM)
DO 90 J=1,NUNKN
ALCCM(J,I)=AA(J)
90  CONTINUE
100 CONTINUE

```

```

C*****RD202580
C** READ OBSERVATIONS FREE FIELD IN THE ORDER THAT THE DESIGN **RD202590
C** MATRIX IS READ **RD202600
C** READ ENTRIES AS DENOTED BY N1 AND N2 BELOW **RD202610
C** **RD202620
C** N1 N2 **RD202630
C** 2 1 X Y Z **RD202640
C** 2 0 X1 X2 X3 Y1 Y2 Y3 Z1 Z2 Z3 **RD202650
C** 0 0 X1 X2 X3 Y1 Y2 Y3 Z1 Z2 Z3 **RD202660
C** 3 1 X Y Z W **RD202670
C** 3 0 2-CARDS X1 X2 X3 Y1 Y2 Y3 **RD202680
C** Z1 Z2 Z3 W1 W2 W3 **RD202690
C** 1 0 2-CARDS X1 X2 X3 Y1 Y2 Y3 **RD202700
C** Z1 Z2 Z3 W1 W2 W3 **RD202710
C** **RD202720
C** **RD202730
C** NOTES **RD202740
C** 1. ANY OTHER COMBINATION OF N1 AND N2 ASSUMES THERE ARE TWO **RD202750
C** RECORDS, SIX ITEMS PER RECORD **RD202760
C** 2. THE MAXIMUM NUMBER OF OBSERVATIONS IS 50 SO YOU MAY HAVE UP TO 600 READINGS GIVEN HERE **RD202770
C** 3. NO MORE OBSERVATIONS ARE STORED WHEN **RD202780
C** -20000 IS ENCOUNTERED **RD202790
C*****RD202800
10 DO 120 K=1,NR RD202810
OBSERV(K)=0.0 RD202820
20 CCNTINUE RD202830
K=1 RD202840
C*****RD202850
C** SET UP FLAGS FOR TYPE OF WEIGHING **RD202860
C*****RD202870
IF (N1.EQ.2.AND.N2.EQ.1) GO TO 130 RD202880
IF (N1.EQ.2.AND.N2.EQ.0) GO TO 140 RD202890
IF (N1.EQ.0.AND.N2.EQ.0) GO TO 150 RD202900
IF (N1.EQ.3.AND.N2.EQ.1) GO TO 160 RD202910
IF (N1.EQ.3.AND.N2.EQ.0) GO TO 170 RD202920
IF (N1.EQ.1.AND.N2.EQ.0) GO TO 180 RD202930
GO TO 180 RD202940
30 IA=3 RD202950
ITPCS=1 RD202960
GO TO 190 RD202970
40 IA=5 RD202980
ITPCS=2 RD202990
GO TO 190 RD203000
50 IA=5 RD203010
ITPCS=3 RD203020
GO TO 190 RD203030
60 IA=4 RD203040
ITPCS=4 RD203050
GO TO 190 RD203060
70 IA=6 RD203070
ITPCS=5 RD203080
GO TO 190 RD203090
80 IA=6 RD203100
ITPCS=6 RD203110
GO TO 190 RD203120
90 *****RD203130
OBSERVATIONS--FREE FIELD INPUT **RD203140
NUMBER -20000 DENOTES LAST CARD FOR A SERIES **RD203150

```

```

C*****RD203
190 CALL READIT (AA,1,AAITEM) RD203
    IF (AA(1).LE.-19999.) GO TO 210 RD203
    DO 200 I=1,IA RD203
    OBSERV(K)=AA(I) RD203
    K=K+1 RD203
200 CONTINUE RD203
    GO TO 190 RD203
210 RETURN RD203
    END RD203
--- SPINV SUBPROGRAM ---
    SUBROUTINE SPINV (A,M,KK,ISIG) INV00
C*****INV00
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **INV00
C** PROGRAM VERSION OF SEPT.10,1971 WRITTEN BY R.C.RAYEOLD **INV00
C** AND MRS.R.N.VARNER **INV00
C** MODIFIED BY R. N. VARNER SEPT 1979 **INV00
C*****INV00
C** MATRIX INVERSION WITH MINIMUM ROUND OFF ERROR ACCUMULATION. **INV00
C** WRITTEN BY MRS.S.T.PEAVY ,STATISTICAL ENGINEERING LABORATORY , **INV00
C** NATIONAL BUREAU OF STANDARDS ,WASHINGTON , D.C. 20234 **INV00
C** A IS THE MATRIX TO BE INVERTED **INV00
C** M IS THE NUMBER OF ROWS IN MATRIX **INV00
C** KK IS THE NUMBER OF COLUMNS IN MATRIX **INV00
C** ISIG=0 INVERSE SUCCESSFUL **INV00
C** ISIG=4 SINGULAR MATRIX **INV00
C** ISIG=-1 CHECK ON I(IDENTITY)-A*A(INVERSE) FAILED **INV00
C*****INV00
    DIMENSION A(1) INV00
    COMMON /INVCST/ ZERMAC INV00
C*****INV00
C** SAVE ORIGINAL MATRIX **INV00
C*****INV00
    CALL SAVMTX (M) INV00
    ISIG=0 INV00
    N=M INV00
    NN=KK INV00
    N2=N+N INV00
    DO 30 J=1,N INV00
    NJCCL=(N+J-1)*NN INV00
    DO 30 I=1,N INV00
    KINJ=NJCCL+I INV00
    IF (I-J) 10,20,10 INV00
10 A(KINJ)=0. INV00
    GO TO 30 INV00
20 A(KINJ)=1. INV00
30 CONTINUE INV00
C*****INV00
C** DETERMINE MAXIMUM ABSOLUTE OF VARIABLE BEING ELIMINATED. **INV00
C*****INV00
    L=0
40 L=L+1
    LCCL=NN*L-NN
    KLL=LCCL+L
    IF (L-N) 50,100,200
C*****
C** FIND THE LARGEST ELEMENT IN THE LTH COLUMN.
C*****

```

```

50      J1=L                                INV00480
        C=AES(A(KLL))                      INV00490
        L1=L+1                            INV00500
        DO 70 I=L1,N                      INV00510
        KIL=LCCL+I                        INV00520
        X=AES(A(KIL))                     INV00530
        IF (C-X) 60,70,70                 INV00540
C***** INV00550
C** RECCRD THE NUMBER CF THE RCW HAVING THE GREATER ELEMENT. ** INV00560
C***** INV00570
60      J1=I                                INV00580
C***** INV00590
C** C BECCMES THE GREATER. ** INV00600
C***** INV00610
        C=X                                INV00620
70      CCNTINUE                            INV00630
C***** INV00640
C** INTERCHANGE ROW J1 WITH RCW L. J1 IS THE ROW WITH THE LARGEST ** INV00650
C** ELEMENT TEST TO SEE IF INTERCHANGING IS NECESSARY. ** INV00660
C***** INV00670
        IF (J1-L) 80,100,80                INV00680
80      DO 90 J=L,N2                      INV00690
        JCCL=NN+J-NN                      INV00700
        KJIJ=JCCL+J1                      INV00710
        HOLD=A(KJIJ)                      INV00720
        KLJ=JCCL+L                        INV00730
        A(KJIJ)=A(KLJ)                    INV00740
        A(KLJ)=HOLD                       INV00750
90      CONTINUE                            INV00760
C***** INV00770
C** IF THE LARGEST ABSOLUTE ELEMENT IN A CCLUMN IS ZERC WE HAVE ** INV00780
C** A SINGULAR MATRIX ** INV00790
C***** INV00800
100     IF (ABS(A(KLL))-ZERMAC) 110,110,120 INV00810
110     ISIG=4                             INV00820
        GO TO 200                          INV00830
C***** INV00840
        ZERC ALL THE ELEMENTS IN THE LTH COLUMN BUT THE PIVOTAL ELEMENT ** INV00850
C***** INV00860
        L1=1                                INV00870
        L2=L-1                              INV00880
        IF (L2) 130,130,150                INV00890
        IF (L-N) 140,170,140               INV00900
        L1=L+1                              INV00910
        L2=N                                INV00920
        DO 160 I=L1,L2                     INV00930
        KIL=LCCL+I                         INV00940
        Z=A(KIL)/A(KLL)                    INV00950
        DO 160 J=L,N2                      INV00960
        JCCL=NN+J-NN                       INV00970
        KJIJ=JCCL+I                       INV00980
        KLJ=JCCL+L                         INV00990
        A(KJIJ)=A(KIJ)+Z*A(KLJ)            INV01000
        IF (L2) 40,40,130                 INV01010
C***** INV01020
        ZER BY DIAGCNAL ELEMENTS. ** INV01030
C***** INV01040
        DO 1=1,N                           INV01050

```



```

      KKK=NN*I-NN+I
      ZZ=A(KKK)
      DO 180 J=1,N2
      KKI=NN*J-NN+I
180   A(KKI)=A(KKI)/ZZ
C*****
C**   RETURN AFTER PUTTING A INVERSE INTO B
C*****
      DO 190 J=1,N
      JCCL=NN*J-NN
      NJCCL=NN*N+JCCL
      DO 190 I=1,N
      KIJ=JCCL+I
      KINJ=NJCCL+I
190   A(KIJ)=A(KINJ)
C*****
C**   CHECK SUCCESS OR FAILURE OF MATRIX INVERSION
C*****
      CALL INVCHK (A,M,KK,ISIG)
200   RETURN
      END
--- SAVMTX SUBPROGRAM ---
      SUBROUTINE SAVMTX (M)
C*****
C**   SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION
C**   PROGRAM VERSION OF SEPT.10,1971      WRITTEN BY R.C.RAYECLD
C**   AND MRS.R.N.VARNER
C**   MODIFIED BY R. N. VARNER SEPT 1979
C*****
C**   SUBROUTINE TO SAVE ORIGINAL MATRIX BEFORE INVERSION
C**   MATRIX IS IN Y AND IS STORED IN B
C*****
C**   DIMENSION FOR COMMON /CHECK/ VARIABLES
C*****
      DIMENSION B(289),Y(578)
C*****
C**   LABELED COMMON
C*****
      COMMON /CHECK/ CHKMA,B,Y
      MA=N*M
      DO 10 I=1,MA
10    B(I)=Y(I)
      RETURN
      END
--- INVCHK SUBPROGRAM ---
      SUBROUTINE INVCHK (A,M,KK,ISIG)
C*****
C**   SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION
C**   PROGRAM VERSION OF SEPT.10,1971      WRITTEN BY R.C.RAYECLD
C**   AND MRS.R.N.VARNER
C**   MODIFIED BY R. N. VARNER SEPT 1979
C*****
C**   SUBROUTINE TO CHECK I-AA(INV)
C**   A IS INVERTED MATRIX
C**   B IS ORIGINAL MATRIX
C**   M IS NUMBER OF ROWS IN MATRIX
C**   KK IS NUMBER OF COLUMNS IN MATRIX
C**   ISIG=-1 IF I-A*A(INV) FAILS

```

```

C***** IVC00140
      DIMENSION A(1) IVC00150
C***** IVC00160
C** DIMENSION FOR CCMCN /CHECK/ ** IVC00170
C***** IVC00180
      DIMENSION B(289),Y(578) IVC00190
C***** IVC00200
C** LABELED COMMON ** IVC00210
C** ** IVC00220
      COMMON /CHECK/ CHCKMA,E,Y IVC00230
      N=M IVC00240
      NN=KK IVC00250
C***** IVC00260
C** RESTORE ORIGINAL MATRIX ** IVC00270
C***** IVC00280
      NP=N*N IVC00290
      DO 10 J=1, NP IVC00300
      NP2=NP+J IVC00310
10 A(NP2)=E(J) IVC00320
C***** IVC00330
C** B=I-AA(INV) ** IVC00340
C***** IVC00350
      IBP=1 IVC00360
      K=1 IVC00370
      DO 60 J=1, NN IVC00380
      IAP=N*N+1 IVC00390
      DO 50 KA=1, N IVC00400
      IA=IAP IVC00410
      IB=IBP IVC00420
      B(K)=0.0 IVC00430
      DO 20 I=1, NN IVC00440
      B(K)=B(K)+A(IA)*A(IB) IVC00450
      IB=IB+1 IVC00460
      IA=IA+N IVC00470
      IF (J.EQ.KA) GO TO 30 IVC00480
      B(K)=-B(K) IVC00490
      GO TO 40 IVC00500
      B(K)=1.0-B(K) IVC00510
      IAP=IAP+1 IVC00520
      K=K+1 IVC00530
      IBP=IBP+N IVC00540
C***** IVC00550
      PICK UP THE LARGEST ABSCLTE VALUE FROM I-AA(INV) ** IVC00560
C***** IVC00570
      BIG=AES(D(1)) IVC00580
      J=N+N IVC00590
      DO 70 I=2, J IVC00600
      BIG=AMAX1(BIG, AES(E(I))) IVC00610
C***** IVC00620
      IF CONDITION IS NOT MET SET SIGNAL (SIG)=-1 ** IVC00630
      CHCKMA=.01*S.D. OF EALANCE(INPUT VALUE) ** IVC00640
C***** IVC00650
      IF (EIG.LE.CHCKMA) RETURN IVC00660
      SIG=-1 IVC00670
      RETURN IVC00680
      END IVC00690
      SUBPROGRAM ---
      SUBROUTINE PRINT2

```

PR200010

```

*****PR2000
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **PR2000
C** PROGRAM VERSION OF SEPT. 10, 1971 WRITTEN BY R.C. RAYECLD **PR2000
C** AND MRS. R. N. VARNER **PR2000
C** MODIFIED BY R. N. VARNER SEPT 1979 **PR2000
*****PR2000
C** SUBROUTINE TO PRINT PAGES FOR CNE SERIES **PR2000
*****PR2000
C** DIMENSION FOR COMMON /PRT1/ VARIABLES **PR2000
*****PR2000
C** DIMENSION E1(72), E2(72), B3(72), B4(72), B5(72), B6(72), B7(72), PR2001
2 IDATE(3) PR2001
*****PR2000
C** DIMENSION FOR COMMON /INPUT/ VARIABLES **PR2000
*****PR2000
C** DIMENSION AIDCST(5), ANCM(15), DENSTY(15), COEFEX(15), ACCVAL(15), PR2001
2 ARSTIN(15), ACKSTD(15),IRSTCU(15), IPRNT(15), DESMAT(15,50), PR200
3 CBSEFV(600), ALCOM(15,20) PR200
*****PR2000
C** DIMENSION FOR COMMON /COMPT/ VARIABLES **PR2000
*****PR2000
C** DIMENSION SWTPRT(50), A(50), DELTA(50), OBSCOR(15), COMVOL(15), PR200
2 SERROR(15), TRISIG(15), TCTUN(15), DRIFT(50), ZERO(50), COMVGF(15), PR200
3 CORRESA(20), SIG35A(20), UNCSA(20), IOTSTR(50), SER5A(20), DS1(50) PR200
*****PR2000
C** DIMENSION FOR SUBROUTINE PRINT2 **PR2000
*****PR2000
C** DIMENSION ITEMP(15), KTEMP(15), JTEMP(20) PR2000
*****PR2000
C** DIMENSION FOR COMMON /CHECK/ VARIABLES **PR2000
*****PR2000
C** DIMENSION B(289), Y(578) PR2000
*****PR2000
C** DIMENSION FOR COMMON /PRTLB/ VARIABLES **PR2000
*****PR2000
C** DIMENSION PRTLEX(15) PR2000
*****PR2000
C** LAELED COMMON **PR2000
*****PR2000
COMMON /PRT1/ E1, E2, B3, B4, B5, B6, B7, RANERR, SYSERR, TNOM, L1, L2, L3, L4, PR200
2 L5, L6, IDATE, IEREST PR200
COMMON /PRT2/ IPAGE, NOSER, IPGCT PR200
COMMON /INPUT/ TEAR, PBAR, FBAR, STDEB, SWT, VSWT, CEXSWT, AIDCST, ANCM, PR200
2 DENSTY, COEFEX, ACCVAL, ARSTIN, ACKSTD, DESMAT, CBSEFV, VARBAL, ALCOM, T1PR200
3, T2F, P1P, P2P, H1P, H2P, CP1, CP2, CT1, CT2, CF1, CF2, OT1P, OT2P, CF1P, DP2P, PR200
4 OH1P, OH2P, IOP, IEAL, NOES, NUNKN, IRSTCU, IPRNT, ITPOS, ICKUSD, ICALDS, PR200
5 LINVAR, N3, N4 PR200
COMMON /COMPT/ CESSCR, CORR, VOLRES, RHOA, SWTPRT, A, DELTA, COMVOL, PR200
2 SERROR, TRISIG, TCTUN, ACCRF, CRRRES, TMSUM, VGLSUM, SERSUM, T3SIG, ALLOADPR200
3, OBSTD, FRATIO, CBCCCK, OBSCK, TVAL, DRIFT, ZERO, V2TAU, STAR, COMVGF, PR200
4 CORRESA, SIG35A, UNCSA, RHOAA, RHOAB, SER5A, DS1, NDGFR, ISWTCH, IFLAG, PR200
5 IRCUT, IOTSTR, JSTAR PR200
COMMON /PRTLB/ PRTLBX PR200
COMMON /CHECK/ CHCKMA, B, Y PR200
COMMON /UNITID/ IR, IW, IP, IPL, ITEMP PR200
COMMON /DPFDVL/ KFD(18) PR200
COMMON /PCHOUT/ NTCF PR200

```

```

C*****PR200600
C** TYPE STATEMENT **PR200610
C*****PR200620
DOUBLE PRECISION CESCOR PR200630
C*****PR200640
NOSER=NCSE+1 PR200650
ITST=1 PR200660
IF (NOSER.NE.1) GC TO 10 PR200670
NTOP=0 PR200680
REWIND ITMP PR200690
C*****PR200700
C** WRITE FIRST PAGE OF SERIES **PR200710
C*****PR200720
10 CALL HEADPG PR200730
CALL CHKLN (2) PR200740
WRITE (IW,1140) STDEBA,VAREAL PR200750
CALL CHKLN (3) PR200760
WRITE (IW,1150) ICALDS PR200770
DO 15 I=1,NUNKN PR200780
15 ITEM(I)=ARSTIN(I) PR200785
CALL CHKLN (1) PR200790
WRITE (IW,1160) (ITEM(I),I=1,NUNKN) PR200800
CALL CHKLN (6) PR200810
WRITE (IW,1170) CCRR,TEAR,VOLRES,SYSERR,RANERR PR200820
DO 20 I=1,NUNKN PR200830
20 ITEM(I)=ACKSTD(I) PR200840
CALL CHKLN (1) PR200850
WRITE (IW,1190) ICKUSD PR200860
CALL CHKLN (1) PR200870
WRITE (IW,1180) (ITEM(I),I=1,NUNKN) PR200880
CALL CHKLN (1) PR200890
WRITE (IW,1200) V2TAU PR200900
DO 30 I=1,NUNKN PR200910
30 ITEM(I)=IPRNT(I) PR200920
CALL CHKLN (1) PR200930
WRITE (IW,1040) (ITEM(I),I=1,NUNKN) PR200940
CALL CHKLN (15) PR200950
WRITE (IW,1210) T1P,T2P,TEAR,P1P,P2P,PBAR,H1P,H2P,HEAR,RT CAB,RHOAA PR200960
2,RHCA,CT1,CT2,CP1,CP2,CH1,CH2,CT1P,OT2P,OP1P,OP2P,OH1P,OH2P PR200970
CALL CHKLN (3) PR200980
IF (N3.EQ.0) GO TO 70 PR200990
C*****PR201000
C** OUTPUT IN ENGLISH (POUNDS) **PR201010
C*****PR201020
WRITE (IW,1220) PR201030
DO 60 I=1,NUNKN PR201040
CALL CHKLN (1) PR201050
IF (ACCVAL(I)) 40,50,40 PR201060
WRITE (IW,1230) (AIDCST(IU,I),IU=1,5),PRTLEX(I),DENSTY(I),COEFEX(I) PR201070
2) ACCVAL(I) PR201080
GO TO 60 PR201090
WRITE (IW,1230) (AIDCST(IU,I),IU=1,5),PRTLEX(I),DENSTY(I),COEFEX(I) PR201100
2) PR201110
CONTINUE PR201120
GO TO 110 PR201130
C*****PR201140
C** OUTPUT IN METRIC (GRAMS) **PR201150
C*****PR201160

```

```

70  WRITE (IW,1250)
    DO 100 I=1,NUNKN
    CALL CHKLN (1)
    IF (ACCVAl(I)) 80,50,80
80  WRITE (IW,1240) (AIDCST(IU,I),IU=1,5),ANOM(I),DENSTY(I),CCFEX(I),
    2ACCVAl(I)
    GO TO 100
90  WRITE (IW,1240) (AIDCST(IU,I),IU=1,5),ANOM(I),DENSTY(I),CCFEX(I)
100 CCNTINUE
*****
C**  WRITE SECCND PAGE CF SERIES
*****
110  CALL HEADPG
    CALL CHKLN (3)
    WRITE (IW,1150) ICALDS
    CALL CHKLN (2)
    IF (N3.EQ.0) GO TO 210
*****
C**  OUTPUT IN ENGLISH UNITS
C**  JFLAG=1 PCUNDS
C**  JFLAG=2 MILLI-PCUNDS
C**  JFLAG=3 MICRO-PCUNDS
*****
    JFLAG=1
    IF (PRTL BX(1).LT.0.001) GO TO 120
    IF (PRTL BX(1).LT.1.0) JFLAG=2
    GO TO 130
120  JFLAG=3
130  DO 170 I=1,NUNKN
    GO TO (140,150,160), JFLAG
140  KTEMP(I)=INT(PRTL BX(I)+.5)
    GO TO 170
150  KTEMP(I)=INT(1000.*PRTL BX(I)+.5)
    GO TO 170
160  KTEMP(I)=INT(100000.*PRTL BX(I)+.5)
170  CONTINUE
    GO TO (180,190,200), JFLAG
180  WRITE (IW,1050) (KTEMP(I),I=1,NUNKN)
    GO TO 270
190  WRITE (IW,1060) (KTEMP(I),I=1,NUNKN)
    GO TO 270
200  WRITE (IW,1070) (KTEMP(I),I=1,NUNKN)
    GO TO 270
*****
C**  OUTPUT IN METRIC
C**  JFLAG=1 GRAMS
C**  JFLAG=2 MILLIGRAMS
*****
210  JFLAG=1
    IF (ANGM(1).LT.1.0) JFLAG=2
    DO 240 I=1,NUNKN
    GO TO (220,230), JFLAG
220  KTEMP(I)=INT(ANGM(I)+.5)
    GO TO 240
230  KTEMP(I)=INT(ANGM(I)*1000.+5)
240  CCNTINUE
    GO TO (250,260), JFLAG
250  WRITE (IW,1260) (KTEMP(I),I=1,NUNKN)

```

```

GO TO 270
260 WRITE (IW,1270) (ITEMP(I),I=1,NUNKN)
C*****
C** SET UP AND PRINT DESIGN MATRIX
C*****
270 DO 310 I=1,NOBS
      DO 300 J=1,NUNKN
        IF (DESMAT(J,I).EQ.0.0) GC TO 280
        IF (DESMAT(J,I).EQ.1.0) GC TO 290
        ITEMP(J)=KFD(12)
        GO TO 300
280 ITEMP(J)=KFD(11)
      GO TO 300
290 ITEMP(J)=KFD(15)
300 CCNTINUE
      CALL CHKLN (1)
      WRITE (IW,1280) I,(ITEMP(J),J=1,NUNKN)
310 CONTINUE
      DO 340 J=1,NUNKN
        IF (ARSTIN(J).EQ.0.0) GO TO 320
        IF (ARSTIN(J).EQ.1.0) GO TO 330
        ITEMP(J)=KFD(12)
        GO TO 340
320 ITEMP(J)=KFD(11)
      GO TO 340
330 ITEMP(J)=KFD(15)
340 CONTINUE
      CALL CHKLN (1)
      WRITE (IW,1290) (ITEMP(J),J=1,NUNKN)
C*****
C** WRITE OBSERVATIONS AS READ
C*****
      CALL CHKLN (3)
      WRITE (IW,1300)
      J=1
      CALL CHKLN (2)
      GO TO (350,390,400,450,470, 480), ITPOS
350 WRITE (IW,1310)
      DO 380 I=1,NOBS
        CALL CHKLN (1)
        IF (DS1(I).EQ.0.0) GO TO 360
        WRITE (IW,1370) I,CBSERV(J),OBSERV(J+1),OBSERV(J+2)
        GC TO 370
360 WRITE (IW,1370) I,CBSERV(J),OBSERV(J+1)
370 J=J+3
380 CONTINUE
      GO TO 510
C*****
      WRITE (IW,1320)
      GC TO 410
400 WRITE (IW,1330)
      DO 440 I=1,NOBS
        CALL CHKLN (2)
        IF (DS1(I).EQ.0.0) GC TO 420
        WRITE (IW,1380) I,CBSERV(J),OBSERV(J+1),OBSERV(J+2),OBSERV(J+3),OBSERV(J+4),
        CBSERV(J+5),CBSERV(J+6),OBSERV(J+7),CESERV(J+8)
        GC TO 430
410 WRITE (IW,1380) I,CBSERV(J),OBSERV(J+1),OBSERV(J+2),OBSERV(J+3),OBSERV(J+4),
        CBSERV(J+5),CBSERV(J+6),OBSERV(J+7),CESERV(J+8)

```

```

2SERV(J+4),CBSERV(J+5)
430 J=J+9
440 CCNTINUE
GO TO 510
C*****
450 WRITE (IW,1340)
DO 460 I=1,NOBS
CALL C+KLN (1)
WRITE (IW,1390) I,CBSERV(J),OBSERV(J+1),OBSERV(J+2),OBSERV(J+3)
J=J+4
460 CCNTINUE
GO TO 510
C*****
470 WRITE (IW,1350)
GO TO 490
480 WRITE (IW,1360)
490 JB=12
DO 500 I=1,NOBS
CALL C+KLN (2)
WRITE (IW,1400) I,(OBSERV(JA),JA=J,JB)
JB=JB+12
J=J+12
500 CCNTINUE
510 IF (IFLAG) 530,540,520
C*****
C** RETURN IF MATRIX IS SINGULAR
C** OR INVERSE IS IN ERROR
C*****
520 CALL C+KLN (3)
WRITE (IW,1410)
RETURN
530 CALL C+KLN (3)
WRITE (IW,1420)
M=NUNKN+2
MA=M*M+1
MB=MA+M*M-1
ABC=FLCAT(MB-MA+1)/5.0+3.0
IABC=(MB-MA+1)/5+3
IF (ABC.GT.FLOAT(IABC)) IABC=IABC+1
CALL C+KLN (IABC)
WRITE (IW,1430) (Y(I),I=MA,MB)
MA=1
MB=M*M
ABC=FLCAT(MB-MA+1)/5.0+3.0
IABC=(MB-MA+1)/5+3
IF (ABC.GT.FLOAT(IABC)) IABC=IABC+1
CALL C+KLN (IABC)
WRITE (IW,1440) (Y(I),I=MA,ME)
MB=M*M
ABC=FLCAT(MB)/5.0+3.0
IABC=MB/5+3
IF (ABC.GT.FLOAT(IABC)) IABC=IABC+1
CALL C+KLN (IABC)
WRITE (IW,1450) (E(I),I=1,MB)
RETURN
C*****
C** WRITE THIRD PAGE CF SERIES
C*****

```

```

540  CALL HEADPG                                PR202910
      CALL CHKLN (3)                            PR202920
      WRITE (IW,1150) ICALDS                    PR202930
      CALL CHKLN (5)                            PR202940
      WRITE (IW,1460) SWT,VSWT,CEXSWT,STAR      PR202950
      CALL CHKLN (5)                            PR202960
      GO TO (550,550,580,610,610, 630), ITPQS  PR202970
550  WRITE (IW,1470)                            PR202980
      DO 570 I=1,NOBS                           PR202990
      CALL CHKLN (1)                            PR203000
      IF (DS1(I).EQ.0.0) GO TC 560               PR203010
      WRITE (IW,1540) I,IOTSTR(I),A(I),DELTA(I),SWTPRT(I),DS1(I) PR203020
      GO TC 570                                  PR203030
560  WRITE (IW,1540) I,ICTSTR(I),A(I),DELTA(I),SWTPRT(I) PR203040
570  CCNTINUE                                   PR203050
      GO TC 650                                  PR203060
C*****PR203070
580  WRITE (IW,1480)                            PR203080
      CALL CHKLN (1)                            PR203090
      DO 600 I=1,NOBS                           PR203100
      IF (DS1(I).EQ.0.0) GO TO 590               PR203110
      WRITE (IW,1540) I,ICTSTR(I),A(I),DELTA(I),SWTPRT(I),ZERC(I),DS1(I) PR203120
      GO TC 600                                  PR203130
590  WRITE (IW,1540) I,IOTSTR(I),A(I),DELTA(I),SWTPRT(I),ZERC(I) PR203140
600  CCNTINUE                                   PR203150
      GO TO 650                                  PR203160
C*****PR203170
610  WRITE (IW,1490)                            PR203180
      DO 620 I=1,NOBS                           PR203190
      CALL CHKLN (1)                            PR203200
      WRITE (IW,1540) I,ICTSTR(I),A(I),DELTA(I),SWTPRT(I),DRIFT(I),DS1(I) PR203210
      2)                                         PR203220
620  CONTINUE                                   PR203230
      GO TC 650                                  PR203240
C*****PR203250
630  WRITE (IW,1500)                            PR203260
      DO 640 I=1,NOBS                           PR203270
      CALL CHKLN (1)                            PR203280
      WRITE (IW,1540) I,IOTSTR(I),A(I),DELTA(I),SWTPRT(I),DRIFT(I),ZERO(PR203290
      2I),DS1(I)                               PR203300
640  CCNTINUE                                   PR203310
650  IF (JSTAR.EQ.0) GO TC 660                  PR203320
      CALL CHKLN (3)                            PR203330
      WRITE (IW,1510)                            PR203340
      CALL CHKLN (6)                            PR203350
      IF (N3.EQ.0) GO TC 680                    PR203360
C*****PR203370
C** ENGLISH UNITS **PR203380
C*****PR203390
      WRITE (IW,1530)                            PR203400
      DO 670 I=1,NUNKN                          PR203410
      CALL CHKLN (1)                            PR203420
      FCBSCR=CBSCOR(I)                          PR203430
      WRITE (IW,1560) PRTEX(I),FCBSCR,COMVOP(I),SERROR(I),TRISIG(I),TOTPR203440
      2UN(I)                                     PR203450
670  CONTINUE                                   PR203460
      GO TC 700                                  PR203470
C*****PR203480

```



```

C** METRIC UNITS **PR203490
C*****PR203500
680 WRITE (IW,1520) PR203510
      DQ 690 I=1,NUNKN PR203520
      CALL CHKLN (1) PR203530
      FCBSCF=QBSCOR(I) PR203540
      WRITE (IW,1550) ANCM(I),FCBSCF,COMVGP(I),SEFFOR(I),TRISIG(I),TOTUNPR203550
      2(I) PR203560
690 CONTINUE PR203570
700 IF (ISWICH.EQ.0) GO TO 710 PR203580
      CALL CHKLN (3) PR203590
      WRITE (IW,1570) PR203600
710 CALL CHKLN (3) PR203610
      WRITE (IW,1580) TEAR PR203620
      XXXX=0.1*STDEBA PR203630
      XXXX=ABS(ACORR-CORRES) PR203640
      IF (XXXX.LE.XXXX) GO TO 720 PR203650
      CALL CHKLN (7) PR203660
      WRITE (IW,1590) ACCRR,CORRES PR203670
720 IF (IRCUT.EQ.0.0) GO TO 730 PR203680
      CALL CHKLN (6) PR203690
      WRITE (IW,1600) (IRSTOU(I),I=1,NUNKN) PR203700
      WRITE (IW,1610) TMSUM,VCLSUM,SERSUM,TJSIG PR203710
C*****PR203720
C** WRITE LAST PAGE OF SERIES IF LINVAR NOT EQUAL TO ZERO **PR203730
C*****PR203740
730 IF (LINVAR.EQ.0) GO TO 930 PR203750
      CALL HEADPG PR203760
      CALL CHKLN (3) PR203770
      WRITE (IW,1150) ICALDS PR203780
      CALL CHKLN (3) PR203790
      WRITE (IW,1620) PR203800
      CALL CHKLN (1) PR203810
      IF (N3.EQ.0) GO TO 770 PR203820
C*****PR203830
C** ENGLISH UNITS **PR203840
C*****PR203850
      GO TO (740,750,760), JFLAG PR203860
740 WRITE (IW,1080) PR203870
      GO TO 800 PR203880
750 WRITE (IW,1090) PR203890
      GO TO 800 PR203900
760 WRITE (IW,1100) PR203910
      GO TO 800 PR203920
C*****PR203930
C** METRIC UNITS **PR203940
C*****PR203950
770 GO TO (780,790), JFLAG PR203960
780 WRITE (IW,1630) PR203970
      GO TO 800 PR203980
790 WRITE (IW,1640) PR203990
800 CALL CHKLN (1) PR204000
      WRITE (IW,1650) (KTEMP(I),I=1,NUNKN) PR204010
      DO 840 I=1,LINVAR PR204020
      JTEMP(I)=0 PR204030
      DO 830 J=1,NUNKN PR204040
      JTEMP(I)=JTEMP(I)+KTEMP(J)*INT(ALCOM(J,I)) PR204050
      IF (ALCOM(J,I).EQ.0.0) GO TO 810 PR204060

```

|     |                                                                |            |
|-----|----------------------------------------------------------------|------------|
|     | IF (ALCCM(J,I).EQ.1.0) GC TC 820                               | PR204070   |
|     | ITEMP(J)=KFD(12)                                               | PR204080   |
|     | GO TO 830                                                      | PR204090   |
| 810 | ITEMP(J)=KFD(11)                                               | PR204100   |
|     | GC TC 830                                                      | PR204110   |
| 820 | ITEMP(J)=KFD(15)                                               | PR204120   |
| 830 | CONTINUE                                                       | PR204130   |
|     | CALL CHKLN (1)                                                 | PR204140   |
|     | WRITE (IW,1660) JTEMP(I),(ITEMP(K),K=1,NUNKN)                  | PR204150   |
| 840 | CONTINUE                                                       | PR204160   |
|     | CALL CHKLN (8)                                                 | PR204170   |
|     | WRITE (IW,1670)                                                | PR204180   |
|     | WRITE (IW,1680)                                                | PR204190   |
|     | CALL CHKLN (2)                                                 | PR204200   |
|     | IF (N3.EQ.0) GO TO 880                                         | PR204210   |
|     | GO TO (850,860,870), JFLAG                                     | PR204220   |
| 850 | WRITE (IW,1110)                                                | PR204230   |
|     | GO TO 910                                                      | PR204240   |
| 860 | WRITE (IW,1120)                                                | PR204250   |
|     | GO TO 910                                                      | PR204260   |
| 870 | WRITE (IW,1130)                                                | PR204270   |
|     | GO TO 910                                                      | PR204280   |
| 880 | GO TO (890,900), JFLAG                                         | PR204290   |
| 890 | WRITE (IW,1690)                                                | PR204300   |
|     | GO TO 910                                                      | PR204310   |
| 900 | WRITE (IW,1700)                                                | PR204320   |
| 910 | DO 520 I=1,LINVAR                                              | PR204330   |
|     | CALL CHKLN (1)                                                 | PR204340   |
|     | WRITE (IW,1710) JTEMP(I),CORR5A(I),SER5A(I),SIG35A(I),UNCEA(I) | PR204350   |
| 920 | CONTINUE                                                       | PR204360   |
|     | *****PR204370                                                  |            |
|     | C** WRITE LAST PAGE OF SERIES                                  | **PR204380 |
|     | *****PR204390                                                  |            |
| 930 | CALL HEADPG                                                    | PR204400   |
|     | CALL CHKLN (2)                                                 | PR204410   |
|     | WRITE (IW,1720) ALCADP,IBREST                                  | PR204420   |
|     | CALL CHKLN (3)                                                 | PR204430   |
|     | WRITE (IW,1150) ICALDS                                         | PR204440   |
|     | CALL CHKLN (9)                                                 | PR204450   |
|     | WRITE (IW,1730) OBSTD,STDEEA,NDGFR,FRATIO                      | PR204460   |
|     | *****PR204470                                                  |            |
|     | C** COMPUTE F - VALUE                                          | **PR204480 |
|     | *****PR204490                                                  |            |
|     | PRETST=6.64                                                    | PR204500   |
|     | CALL CHKLN (4)                                                 | PR204510   |
|     | IF (NDGFR.EQ.1) GC TO 940                                      | PR204520   |
|     | ANDGFR=NDGFR                                                   | PR204530   |
| 940 | PRETST=(1.0-2.0/(9.*ANDGFR)+2.32635*SQRT(2.0/(9.0*ANDGFR)))*3  | PR204540   |
|     | IF (FRATIO.GT.PRETST) GO TO 950                                | PR204550   |
|     | WRITE (IW,1740) PRETST                                         | PR204560   |
|     | GO TO 960                                                      | PR204570   |
| 950 | WRITE (IW,1750)                                                | PR204580   |
|     | WRITE (IW,1760) PRETST                                         | PR204590   |
|     | WRITE (IW,1750)                                                | PR204600   |
| 960 | ITST=0                                                         | PR204610   |
| 970 | DO 570 I=1,NUNKN                                               | PR204620   |
|     | ITEMP(I)=ACKSTD(I)                                             | PR204630   |
|     | CALL CHKLN (3)                                                 | PR204640   |

```

WRITE (IW,1810)
CALL CHKLN (1)
WRITE (IW,1180) (ITEMP(I),I=1,NUNKN)
CALL CHKLN (1)
WRITE (IW,1190) ICKUSD
CALL CHKLN (1)
WRITE (IW,1200) V2TAU
CALL CHKLN (6)
WRITE (IW,1770) OBCOCK,OBESCK,TVAL
CALL CHKLN (1)
IF (OBESCK.NE.0.0) GO TO 980
WRITE (IW,1820)
980 IF (ABS(TVAL).GE.3.) GO TO 990
CALL CHKLN (3)
WRITE (IW,1780)
GO TO 1030
990 WRITE (IW,1750)
ITST=0
YXYX=0.0
DO 1000 I=1,NUNKN
1000 YXYX=YXYX+(ACKSTD(I)*SERRCR(I))
XYYX=ABS(TVAL)-(YXYX/OBESCK)
CALL CHKLN (3)
IF (XYYX.GE.3.) GO TO 1010
WRITE (IW,1790)
GO TO 1020
1010 WRITE (IW,1800)
ITST=0
1020 CONTINUE
CALL CHKLN (16)
WRITE (IW,1750)
1030 WRITE (IW,1210) T1P,T2P,TEAR,P1P,P2P,PBAR,H1P,H2P,HEAR,RHCA8,RHOAA,
2,RHCA,CT1,CT2,CP1,CP2,CH1,CH2,CT1P,CT2P,OP1P,OP2P,OH1P,CH2P
DIFT=T2P-T1P
DIFP=P2P-P1P
DIFH=H2P-H1P
C*****PR205010
C** PUNCH CNTRCL DATA RECORDS **PR205020
C*****PR205030
NTOP=NTOP+ITST
IF (ITST.EQ.0) RETURN
WRITE (ITMP) (IDATE(K),K=1,3),IBREST,ICKUSC,OBCOCK,IBAL,OBSTD,NDGFP
2R,ICALDS,TBAR,DIFT,PBAR,DIFP,HEAR,DIFH,RHOA,IOP
NTCF=NTOP+1
RETURN
C*****PR205100
C** FORMAT STATEMENTS **PR205110
C*****PR205120
1040 FORMAT (1X,13HREPCRT VECTCR,1X,15I5)
1050 FORMAT (6X,3H LB/6X,15I6)
1060 FORMAT (6X,9H MILLI-LB/6X,15I6)
1070 FORMAT (6X,9H MICRC-LB/6X,15I6)
1080 FORMAT (2X,4H(LB),6X,2H(LB))
1090 FORMAT (1X,8HMILLI-LB,6X,8HMILLI-LB)
1100 FORMAT (1X,8HMICRC-LB,6X,8HMICRO-LB)
1110 FORMAT (2X,4H(LB),6X,4H(MG),7X,4H(MG),10X,4H(MG),10X,4H(MG)/)
1120 FORMAT (1X,10H(MILLI-LB),3X,4H(MG),7X,4H(MG),10X,4H(MG),10X,4H(MG)
2/)

```

```

1130 FORMAT (1X,10H(MICRO-LE),3X,4H(MG),7X,4H(MG),10X,4H(MG),10X,4H(MG) PR205230
2/) PR205240
1140 FORMAT (50H ACCEPTED WITHIN STANDARD DEVIATION OF THE PROCESS,2X, PR205250
2 F13.5,3H MG/51H ACCEPTED BETWEEN STANDARD DEVIATION OF THE PROCESS PR205260
3S,1X,F13.5,3H MG) PR205270
1150 FORMAT (//20H CALIBRATION DESIGN ,14) PR205280
1160 FORMAT (17H RESTRAINT VECTOR,1X,15I5) PR205290
1170 FORMAT (31H MASS CORRECTION OF RESTRAINT ,31X,F13.5,3H MG/ PR205300
2 46H VOLUME OF WEIGHTS BEING USED IN RESTRAINT AT F6.2,3H C ,7X, PR205310
3 F13.5,4H CM3/36H SYSTEMATIC ERROR IN THE RESTRAINT ,26X,F13.5, PR205320
4 3H MG/65H 3 STANDARD DEVIATION LIMIT FOR RANDOM ERROR AFFECTING RPR205330
5 RESTRAINT ,F10.5,3F MG//) PR205340
1180 FORMAT (22H CHECK STANDARD VECTOR,1X,15I5) PR205350
1190 FORMAT (22H CHECK STANDARD USED ,14) PR205360
1200 FORMAT (44H ACCEPTED MASS CORRECTION OF CHECK STANDARD ,F13.5, PR205370
2 3H MG) PR205380
1210 FORMAT (//16H TEST CONDITIONS,29X,6HBEFORE,7X,5HAFTER,7X,7HAVERAGE PR205390
2/36H CORRECTED TEMPERATURE IN DEGREES C ,3F13.2/ PR205400
3 28H CORRECTED PRESSURE IN MM HG,9X,3F13.3/ PR205410
4 30H CORRECTED HUMIDITY IN PERCENT,6X,3F13.2/ PR205420
5 31H COMPUTED AIR DENSITY IN MG/CM3,7X,3F13.4/ PR205430
6 23H TEMPERATURE CORRECTION,14X,2F13.3/20H PRESSURE CORRECTION,17X PR205440
7,2F13.3/20H HUMIDITY CORRECTION,16X,2F13.2/ PR205450
8 34H OBSERVED TEMPERATURE IN DEGREES C,2X,2F13.2/ PR205460
9 28H OBSERVED PRESSURE IN MM HG ,9X,2F13.3/ PR205470
* 30H OBSERVED HUMIDITY IN PERCENT ,6X,2F13.2//) PR205480
1220 FORMAT (4X,13HWEIGHTS BEING,5X,7HNMOMINAL,6X,7HDENSITY,5X, PR205490
2 11H-COEFFICIENT,3X,8HACCEPTED/8X,6HTESTED,7X,8HVALUE LB,3X, PR205500
3 12H-G/CM3 AT 20C,2X,12HCF EXPANSION,2X,13HCCORRECTION MG/) PR205510
1230 FORMAT (1X,5A3,1X,F12.7,3X,F7.4,6X,F7.6,3X,F13.5) PR205520
1240 FORMAT (1X,5A3,1X,F12.4,3X,F7.4,6X,F7.6,3X,F13.5) PR205530
1250 FORMAT (4X,13HWEIGHTS BEING,5X,7HNMOMINAL,6X,7HDENSITY,5X, PR205540
2 11H-COEFFICIENT,3X,8HACCEPTED/8X,6HTESTED,7X,8HVALUE G ,3X, PR205550
3 12H-G/CM3 AT 20C,2X,12HOF EXPANSION,2X,13HCCORRECTION MG/) PR205560
1260 FORMAT (6X,6H GRAMS/6X,15I6) PR205570
1270 FORMAT (6X,3H MG/6X,15I6) PR205580
1280 FORMAT (3H A ,12,1X,15(5X,A1)) PR205590
1290 FORMAT (3H R 3X,15(5X,A1)) PR205600
1300 FORMAT (//26H OBSERVATIONS IN DIVISIONS) PR205610
1310 FORMAT (39H SINGLE SUBSTITUTION SINGLE PAN BALANCE/) PR205620
1320 FORMAT (36H SINGLE SUBSTITUTION TWO PAN BALANCE/) PR205630
1330 FORMAT (37H SINGLE TRANSPOSITION TWO PAN BALANCE/) PR205640
1340 FORMAT (36H DOUBLE SUBSTITUTION ONE PAN BALANCE/) PR205650
1350 FORMAT (36H DOUBLE SUBSTITUTION TWO PAN BALANCE/) PR205660
1360 FORMAT (37H DOUBLE TRANSPOSITION TWO PAN BALANCE/) PR205670
1370 FORMAT (3H A ,12,1X,3F11.4) PR205680
1380 FORMAT (3H A ,12,1X,6F11.4/6X,3F11.4) PR205690
1390 FORMAT (3H A ,12,1X,4F11.4) PR205700
1400 FORMAT (3H A ,12,1X,6F11.4/6X,6F11.4) PR205710
1410 FORMAT (//19H MATRIX IS SINGULAR) PR205720
1420 FORMAT (//17H1ERROR IN INVERSE) PR205730
1430 FORMAT (//4H A =/(5E16.8)) PR205740
1440 FORMAT (//12H A(INVERSE)=/(5E16.8)) PR205750
1450 FORMAT (//15H I-AA(INVERSE)=/(5E16.8)) PR205760
1460 FORMAT (//19H SENSITIVITY WEIGHT/5H MASS,2X,F13.5,3H MG/7H VOLUME, PR205770
2X,F13.5,12H CM3 AT 20 C/25H COEFFICIENT OF EXPANSION,2X,F8.6/5X, PR205780
1470 S=S-PV(S)=,1X,F13.5,3H MG) PR205790
1480 FORMAT (/36X,7HAVERAGE,6X,8HOBERVED/13X,4H-A(I),6X,8HDELTA(I),3X, PR205800

```

```

2 11HSENSITIVITY,2X,11HSENSITIVITY/13X,4H(MG),8X,4H(MG),6X, PR2058
3 8H(MG/DIV),6X,8H(MG/DIV//) PR2058
1480 FORMAT (/35X,7HAVERAGE,16X,8HCBSERVED/13X,4HA(I),6X,8HDELTA(I),3X, PR2058
2 11HSENSITIVITY,4X,7HZERC(I),2X,11HSENSITIVITY/13X,4H(MG),9X, PR2058
3 4H(MG),5X,8H(MG/DIV),6X,5H(DIV),6X,8H(MG/DIV//) PR2058
1490 FORMAT (/37X,7HAVERAGE,16X,8HCBSERVED/13X,4HA(I),5X,8HDELTA(I),5X, PR2058
2 11HSENSITIVITY,2X,8HDRIIFT(I),2X,11HSENSITIVITY/13X,4H(MG),7X, PR2058
3 4H(MG),8X,8H(MG/DIV),6X,4H(MG),6X,8H(MG/DIV//) PR2058
1500 FORMAT (/37X,7HAVERAGE,27X,8HCBSERVED/13X,4HA(I),6X,8HDELTA(I),4X, PR2058
2 11HSENSITIVITY,2X,8HDRIIFT(I),3X,7HZERO(I),3X,11HSENSITIVITY/13X, PR2058
3 4H(MG),8X,4H(MG),7X,8H(MG/DIV),5X,4H(MG),6X,5H(DIV),6X,8H(MG/DIV) PR2058
4/) PR2058
1510 FORMAT (/6X,64H* CBSERVED DEFLECTION IS GREATER THAN OR EQUAL TO QPR2058
2NE FOURTH THE/8X,22HSENSITIVITY DEFLECTION) PR2058
1520 FORMAT (/31X,6HVCLUME,5X,10HSYSTEMATIC,3X,6H3 S.O.,3X, PR2058
2 11HUNCERTAINTY/5X,4HITEM,8X,10HCORRECTION,4X,6H(AT T),8X,5HERRORPR2058
3,6X,5HLIMIT,6X,5HLIMIT/6X,3H(G),12X,4H(MG),7X,5H(CM3),9X,4H(MG),7XPR2058
4,4H(MG),7X,4H(MG//) PR2058
1530 FORMAT (/31X,6HVCLUME,5X,10HSYSTEMATIC,3X,6H3 S.O.,3X, PR2058
2 11HUNCERTAINTY/5X,4HITEM,8X,10HCORRECTION,4X,6H(AT T),8X,5HERROR,PR2058
3 6X,5HLIMIT,6X,5HLIMIT/5X,4H(LB),12X,4H(MG),7X,5H(CM3),9X,4H(MG), PR2058
4 7X,4H(MG),7X,4H(MG//) PR2058
1540 FORMAT (3H A ,I2,1X,A1,6F12.5) PR2058
1550 FORMAT (1X,F12.4,2F13.5,3F11.5) PR2058
1560 FORMAT (1X,F12.7,2F13.5,3F11.5) PR2058
1570 FORMAT (/25H STOPPED AT 10 ITERATIONS/) PR2058
1580 FORMAT (/15H TEMPERATURE T=,F6.2,3H C/) PR2058
1590 FORMAT (1X,72H*****PR2058
2*****/1X, PR2058
3 72HINPUT ERROR IN RESTRAINT. CHECK RESTRAINT VECTOR, NEMINAL VALUPR2058
4E, DENSITY/1X, PR2058
5 72HANC COEFFICIENT OF EXPANSION IN THIS AND PREVIOUS SERIES. PR2058
6 /2X,30H INPUT CCRRECTION OF RESTRAINT,4X,F11.4,3H MG/2X,PR2058
7 33H CCMPUTED CORRECTION OF RESTRAINT,1X,F11.4,3H MG/1X, PR2058
8 72H*****PR2058
9*****/) PR2058
1600 FORMAT (31H RESTRAINT FOR FCLLCWING SERIES/17H RESTRAINT VECTOR,1XPR2058
2,15I5) PR2058
1610 FORMAT (16H MASS CCRRECTION,17X,F13.5,3H MG/15H VOLUME AT 20 C,18XPR2058
2,F13.5,4H CM3/17H SYSTEMATIC ERROR,16X,F13.5,3H MG/ PR2058
3 27H 3 STANDARD DEVIATION LIMIT,6X,F13.5,3H MG) PR2058
1620 FORMAT (/3X,3HSUM,6X,40HWEIGHTS USED FOR THE LINEAR COMEINATIONS)PR2058
1630 FORMAT (3X,3H(G),6X,5HGRAMS) PR2058
1640 FORMAT (2X,4H(MG),6X,2HMG) PR2058
1650 FORMAT (6X,15I5) PR2058
1660 FORMAT (1X,15,15(4X,A1)) PR2058
1670 FORMAT (/10X, PR2058
2 52HVALUES AND UNCERTAINTIES FOR COMBINATIONS OF WEIGHTS/1X, PR2058
3 61H(UNCERTAINTY IS 3 STANDARD DEVIATION LIMIT PLUS ALLCWANCE FORPR2058
4 19H SYSTEMATIC ERROR.)) PR2058
1680 FORMAT (/39X,6H3 S.O.,5X,11HUNCERTAINTY/3X,3HSUM,8X,4HCORR,4X, PR2058
2 10HSYSTEMATIC,7X,5HEPRCOR,8X,5HLIMIT) PR2058
1690 FORMAT (3X,3H(G),8X,4H(MG),7X,4H(MG),10X,4H(MG),10X,4H(MG//) PR2058
1700 FORMAT (2X,4H(MG),8X,4H(MG),7X,4H(MG),10X,4H(MG),10X,4H(MG//) PR2058
1710 FORMAT (1X,15,4F13.5) PR2058
1720 FORMAT (13H MAXIMUM LOAD,F15.4,2H G/1X, PR2058
2 25HSTARTING RESTRAINT NUMBER,2X,12) PR2058
1730 FORMAT (/18H PRECISION CONTROL//) PR2058

```

```

2 43H OBSERVED STANDARD DEVIATION OF THE PROCESS,F12.5,4H MG/ PR206390
3 43H ACCEPTED STANDARD DEVIATION OF THE PROCESS,F12.5,4H MG/ PR206400
4 19H DEGREES OF FREEDOM,15/8H F RATIO,F12.3) PR206410
1740 FORMAT (/21H F RATIO IS LESS THAN,F6.2, PR206420
2 40H (CRITICAL VALUE FOR PROBABILITY = .01)./ PR206430
3 48H THEREFORE THE STANDARD DEVIATION IS IN CONTROL./) PR206440
1750 FORMAT (80H ***** PR206450
2***** PR206460
1760 FORMAT (24H F RATIO IS GREATER THAN,F6.2, PR206470
2 40H (CRITICAL VALUE FOR PROBABILITY = .01)./ PR206480
3 52H THEREFORE THE STANDARD DEVIATION IS NOT IN CONTROL.) PR206490
1770 FORMAT (38H OBSERVED CORRECTION OF CHECK STANDARD,F19.5,3H MG/ PR206500
2 46H STANDARD DEVIATION OF THE OBSERVED CORRECTION,F11.5,3H MG/ PR206510
3 8H T VALUE,F8.2///) PR206520
1780 FORMAT (36H ABSOLUTE VALUE OF T IS LESS THAN 3./ PR206530
2 40H THEREFORE CHECK STANDARD IS IN CONTROL./) PR206540
1790 FORMAT (1X,72HALTHOUGH THE ABSOLUTE VALUE OF T IS GREATER THAN OR PR206550
2EQUAL TO 3 , /1X, PR206560
3 72HTHE T VALUE CORRECTED FOR SYSTEMATIC ERROR IS LESS THAN 3, PR206570
4 /1X, PR206580
5 72HTHEREFORE THE CHECK STANDARD IS IN CONTROL . PR206590
6 ) PR206600
1800 FORMAT (1X,72HALTHOUGH THE ABSOLUTE VALUE OF T IS GREATER THAN OR PR206610
2EQUAL TO 3 , /1X, PR206620
3 72HTHE DIFFERENCE IS STILL SIGNIFICANT AFTER ALLOWANCE FOR SYSTEMPR206630
4ATIC /1X, PR206640
5 72HERROR , THEREFORE THE CHECK STANDARD IS NOT IN CONTROL . PR206650
6 ) PR206660
1810 FORMAT (//) PR206670
1820 FORMAT (52H T VALUE EQUALS ZERO--CHECK STANDARD VECTOR IS ZERO.) PR206680
END PR206690
--- PGCONT SUBPROGRAM ---
SUBROUTINE PGCONT PG000090
C*****PG000020
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **PG000030
C** PROGRAM VERSION OF SEPT. 10,1971 WRITTEN BY R.C.RAYECLD **PG000040
C** AND MRS.R.N.VARNER **PG000050
C** MODIFIED BY R. N. VARNER SEPT 1979 **PG000060
C*****PG000070
C** SUBROUTINE TO WRITE CONTINUATION PAGE **PG000080
C*****PG000090
C** DIMENSION FOR COMMON /PRT1/ VARIABLES **PG000100
C*****PG000110
DIMENSION B1(72),B2(72),B3(72),B4(72),B5(72),B6(72),B7(72), PG000120
2 IDATE(3) PG000130
C*****PG000140
C** DIMENSION FOR COMMON /INPUT/ VARIABLES **PG000150
C*****PG000160
DIMENSION AIDCST(5,15),ANCM(15),DENSTY(15),COEFEX(15),ACCVAL(15), PG000170
2 ARSTIN(15),ACKSTD(15),IRSTCU(15),IPRNT(15),DESMAT(15,50), PG000180
3 OBSERV(600),ALCOM(15,20) PG000190
C*****PG000200
C** LABELED COMMON **PG000210
C*****PG000220
COMMON /PRT1/ B1,B2,B3,B4,B5,B6,B7,RANERR,SYSERR,TNOM,L1,L2,L3,L4, PG000230
2 L5,L6,IDATE,IBREST PG000240
COMMON /PRT2/ IPAGE,NOSER,IPGCT PG000250
COMMON /INPUT/ TEAR,PBAR,TEAR,STDEBA,SWT,VSWT,CEXSWT,AIDCST,ANCM, PG000260

```

```

2 DENSTY,COEFEX,ACCVAL,ARSTIN,ACKSTD,DESMAT,CBSEVR,VARBAL,ALCOM,T1PPGC00270
3,T2F,P1P,P2P,H1P,H2P,CP1,CP2,CT1,CT2,CH1,CH2,OT1P,OT2P,OP1P,OP2P,PGC00280
4 OH1P,OH2P,IOP,IBAL,NOBS,NUNKN,IRSTOU,IPRNT,ITPOS,ICKUSD,ICALDS,PGC00290
5 LINVAR,N3,N4PGC00300
COMMON /UNITIO/ IR,IW,IP,IPL,ITMPPGC00310
WRITE (IW,30) IPAGEPGC00320
*****PGC00330
C** NOSER IS SET EQUAL TO 200 IN FINPRTPGC00340
*****PGC00350
IF (NOSER.LT.200) GO TO 10PGC00360
IPAGE=IPAGE+1PGC00370
WRITE (IW,50) (B1(K),K=1,65),IPAGE,(B2(K),K=1,65),(IDATE(K),K=1,3)PGC00380
2,(B3(K),K=1,65)PGC00390
GO TO 20PGC00400
10 IPAGE=IPAGE+1PGC00410
WRITE (IW,40) (B1(K),K=1,65),IPAGE,(B2(K),K=1,65),NCSER,(B3(K),K=1,3)PGC00420
2,65),(IDATE(K),K=1,3)PGC00430
20 IPGCT=6PGC00440
RETURNPGC00450
*****PGC00460
C** FORMAT STATEMENTSPGC00470
*****PGC00480
30 FORMAT (1H1,30X,19+CONTINUED FROM PAGE13)PGC00490
40 FORMAT (1H,65A1,4HPAGE,I3/1X,65A1,6HSERIES,I2/1X,65A1,I2,1H/,I2,PGC00500
2 1H/,I2//)PGC00510
50 FORMAT (1X,65A1,4HPAGE,I3/1X,65A1,I2,1H/,I2,1H/,I2/1X,65A1//)PGC00520
ENDPGC00530
--- HEADPG SUBPROGRAM ---
SUBROUTINE HEADPGHDP00010
*****HDP00020
C** SUBROUTINE TO PRINT PAGE HEADINGS**HDP00030
C** ADDED BY R. N. VARNER SEPT 1975**HDP00040
*****HDP00050
C** DIMENSION FOR COMMON /PRT1/ VARIABLESP**HDP00060
*****HDP00070
DIMENSION B1(72),B2(72),B3(72),B4(72),B5(72),B6(72),B7(72),HDP00080
2 IDATE(3)HDP00090
*****HDP00100
C** DIMENSION FOR COMMON /INPLT/ VARIABLESP**HDP00110
*****HDP00120
DIMENSION AIDCST(5,15),ANCM(15),DENSTY(15),COEFEX(15),ACCVAL(15),HDP00130
2 ARSTIN(15),ACKSTD(15),IRSTOU(15),IPRNT(15),DESMAT(15,50),HDP00140
3 CBSEVR(600),ALCOM(15,20)HDP00150
*****HDP00160
C** LABELED COMMON**HDP00170
*****HDP00180
COMMON /PRT1/ B1,B2,B3,B4,B5,B6,B7,RANERR,SYSERR,TNOM,L1,L2,L3,L4,HDP00190
2 L5,L6,IDATE,IERESTHDP00200
COMMON /PRT2/ IPAGE,NOSER,IPGCTHDP00210
COMMON /INPUT/ TBAR,PBAR,HEAR,STDEBA,SWT,VSWT,CEXSWT,AIDCST,ANCM,HDP00220
2 DENSTY,COEFEX,ACCVAL,ARSTIN,ACKSTD,DESMAT,CBSEVR,VARBAL,ALCOM,T1PHDP00230
3,T2F,P1P,P2P,H1P,H2P,CP1,CP2,CT1,CT2,CH1,CH2,OT1P,OT2P,CP1P,OP2P,HDP00240
4 OH1P,OH2P,IOP,IBAL,NOBS,NUNKN,IRSTOU,IPRNT,ITPOS,ICKUSD,ICALDS,HDP00250
5 LINVAR,N3,N4HDP00260
COMMON /UNITIO/ IR,IW,IP,IPL,ITMPHDP00270
IPAGE=IPAGE+1HDP00280
WRITE (IW,10) (B1(K),K=1,65),IPAGE,(B2(K),K=1,65),NCSER,(B3(K),K=1,3)HDP00290
2,65),(IDATE(K),K=1,3),(B7(K),K=1,18),IBAL,ICP

```

```

      IPGCT=8                                FDP00310
      RETURN                                FDP00320
10    FORMAT (1H1,65A1,4HPAGE,13/1X,65A1,6HSERIES,12/1X,65A1,12,1H/12,  FDP00330
      2 1H/12/1X,11HTEST NUMBER,2X,18A1///8H BALANCE,1X,13/9H OPERATOR,1XHDP00340
      3,12)                                FDP00350
      END                                  FDP00360
--- FINPR SUBPROGRAM ---
      SUBROUTINE FINPRT                                FPR00010
C*****FPR00020
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **FPR00030
C** PROGRAM VERSION OF SEPT. 10,1971 WRITTEN BY R.C.RAYEOLC **FPR00040
C** AND MRS.R.N.VARNER **FPR00050
C** MODIFIED BY R. N. VARNER SEPT 1979 **FPR00060
C*****FPR00070
C** SUBROUTINE TO PRINT REPORT **FPR00080
C*****FPR00090
C** DIMENSION FOR COMMON /REPT/ VARIABLES **FPR00100
C*****FPR00110
      DIMENSION AITEM(5,50),APPMAS(50),TRMASS(50),UNCERT(50),VCLPRT(50),FPR00120
      2 COEFRT(50),CORRB(50) FPR00130
C*****FPR00140
C** DIMENSION FOR SUBROUTINE FINPRT **FPR00150
C*****FPR00160
      DIMENSION TRMASX(15),AAAMAS(13),BBBMAS(13),NNP(50),TEMPAR(50) FPR00170
      DIMENSION IAP(9),TAP(9) FPR00180
C*****FPR00190
C** DIMENSION FOR COMMON /PRT1/ VARIABLES **FPR00200
C*****FPR00210
      DIMENSION B1(72),B2(72),B3(72),B4(72),B5(72),B6(72),B7(72), FPR00220
      2 IDATE(3) FPR00230
C*****FPR00240
C** DIMENSION FOR COMMON /INPUT/ VARIABLES **FPR00250
C*****FPR00260
      DIMENSION AIDCST(5,15),ANCM(15),DENSTY(15),COEFEX(15),ACCVAL(15), FPR00270
      2 ARSTIN(15),ACKSTD(15),IRSTCU(15),IPRNT(15),DESMAT(15,50), FPR00280
      3 CBSEFV(600),ALCOM(15,20) FPR00290
C*****FPR00300
C** LABELED COMMON **FPR00310
C*****FPR00320
      COMMON /PRT1/ B1,B2,B3,B4,B5,B6,B7,RANERR,SYSERR,TNCM,L1,L2,L3,L4,FPR00330
      2 L5,L6,IDATE,IBREST FPR00340
      COMMON /PRT2/ IPAGE,NOSER,IPGCT FPR00350
      COMMON /REPT/ TRMASS,APPMAS,CORRB,AITEM,UNCERT,VCLPRT,COEFRT,NPRTFPR00360
      COMMON /INPUT/ TEAR,PBAR,TEAR,STDEBA,SWT,VSWT,CEXSWT,AIDCST,ANOM, FPR00370
      2 DENSTY,COEFEX,ACCVAL,ARSTIN,ACKSTD,DESMAT,CBSEFV,VARBAL,ALCOM,T1PFPR00380
      3,T2F,P1P,P2P,H1F,H2P,CP1,CP2,CT1,CT2,CH1,CT2,OT1P,OT2P,CP1P,OP2P, FPR00390
      4 OH1P,CH2P,IOP,IEAL,NOES,NUNK,IRSTCU,IPRNT,ITPOS,ICKUSC,ICALDS, FPR00400
      5 LINVAR,N3,N4 FPR00410
      COMMON /UNITIC/ IF,Ih,IP,IPL,ITMP FPR00420
      COMMON /PCHOUT/ NTCP FPR00430
C*****FPR00440
C** TYPE STATEMENTS **FPR00450
C*****FPR00460
      DOUBLE PRECISION TRMASS,APPMAS,CORRB,TEMPAR FPR00470
      INTEGER TRMASX,BBBMAS,AAAMAS FPR00480
      IF (NTCP,NE,NOSER) GC TC 20 FPR00490
      IFLAG=999 FPR00500
      WRITE (ITMP) IFLAG,(IAP(I),I=2,5),TAP(1),IAP(6),TAP(2),IAP(7),IAP(FPR00510

```



```

28), (TAP(I), I=3, 5), IAP(9)                                FPR00
END FILE ITMP                                                FPR00
REWIND ITMP                                                  FPR00
10 READ (ITMP) (IAP(I), I=1, 5), TAP(1), IAP(6), TAP(2), IAP(7), IAP(8), (TAP(
2(I), I=3, 9), IAP(9)                                       FPR00
IF (IAP(1).EQ. IFLAG) GO TC 20                               FPR00
WRITE (IP, 200) (IAP(I), I=1, 5), TAP(1), IAP(6), TAP(2), IAP(7), IAP(8), (
2TAP(I), I=3, 9), IAP(9)                                     FPR00
GO TC 10                                                      FPR00
20 KKK=0                                                       FPR00
NOSER=200                                                      FPR00
IF (N3.EQ.0) GO TC 40                                         FPR00
C*****FPR00
C** MASS WAS GIVEN IN ENGLISH UNITS                          **FPR00
C** CCNVERT TO GRAMS                                         **FPR00
C*****FPR00
DO 30 I=1, NPRT                                              FPR00
TRMASS(I)=TRMASS(I)*(1.D0/453.59237D0)                      FPR00
APPMAS(I)=APPMAS(I)*(1.D0/453.59237D0)*1000.0D0            FPR00
CORRB(I)=CCRRB(I)*(1.D0/453.59237D0)*1000.0D0              FPR00
UNCERT(I)=UNCERT(I)*(1./453.59237)                          FPR00
VOLPRT(I)=VOLPRT(I)*.06102374                                FPR00
30 CCNTINUE                                                  FPR00
C*****FPR00
C** MASS WAS GIVEN IN METRIC UNITS                          **FPR00
C*****FPR00
40 IF (NPRT.EQ.1) GO TO 80                                    FPR00
DO 50 I=1, NPRT                                              FPR00
TEMPAR(I)=TRMASS(I)                                          FPR00
NNP(I)=I                                                      FPR00
50 CONTINUE                                                  FPR00
NNN=NPRT-1                                                    FPR00
DO 70 I=1, NNN                                              FPR00
IP1=I+1                                                       FPR00
DO 70 J=IP1, NPRT                                            FPR00
IF (TEMPAR(I)-TEMPAR(J)) 70, 70, 60                          FPR00
60 TEMP=TEMPAR(I)                                             FPR00
TEMPAR(I)=TEMPAR(J)                                          FPR00
TEMPAR(J)=TEMP                                                FPR00
TEMP=NNP(I)                                                   FPR00
NNP(I)=NNP(J)                                                 FPR00
NNP(J)=TEMP                                                   FPR00
70 CONTINUE                                                  FPR00
80 NN=1                                                       FPR00
IF (NPRT.NE.1) GO TO 90                                       FPR00
NNP(1)=1                                                       FPR00
90 IPAGE=IPAGE+1                                              FPR00
WRITE (IW, 260) (B1(K), K=1, 65), IPAGE                      FPR00
WRITE (IW, 270) (B2(K), K=1, 65), (IDATE(K), K=1, 3)        FPR00
WRITE (IW, 280) (B3(K), K=1, 65)                             FPR00
WRITE (IW, 290) (B7(K), K=1, 18)                             FPR00
IPGCT=8                                                       FPR00
IF (NN.EQ.2) GO TO 110                                         FPR00
IF (NN.EQ.3) GO TC 160                                         FPR00
JA=1                                                           FPR00
JB=12                                                          FPR00
IF (KKK.EQ.1) GO TC 100                                         FPR00
CALL TEXTS1

```

```

      KKK=1
100  CONTINUE
      IPAGE=IPAGE+1
      WRITE (IW,260) (B1(K),K=1,65),IPAGE
      WRITE (IW,270) (B2(K),K=1,65),(IDATE(K),K=1,3)
      WRITE (IW,280) (B3(K),K=1,65)
      WRITE (IW,290) (B7(K),K=1,18)
      CALL TEXTS2
      NN=2
      GO TO 90
110  CALL CHKLN (4)
      WRITE (IW,210)
      NN=3
      CALL CHKLN (4)
      IF (N3.EQ.0) GO TO 120
      WRITE (IW,300)
      GO TO 130
120  WRITE (IW,220)
130  DO 150 J=1,NPRT
      NNPF=NPRT+1-J
      I=NNPF(NNPP)
      CALL CHKLN (1)
      IF (N3.EQ.0) GO TO 140
C*****
C**  CONVERT DOUBLE PRECISION VALUE TO FLOATING POINT
C*****
      CALL DFFD (TRMASS(I),TRMASX,15,11)
      WRITE (IW,310) (AITEM(IU,I),IU=1,5),TRMASX,UNCERT(I),VOLPRT(I),COEFF
2PRT(I)
      GO TO 150
140  CALL DFFD (TRMASS(I),TRMASX,16,8)
      WRITE (IW,230) (AITEM(IU,I),IU=1,5),TRMASX,UNCERT(I),VOLPRT(I),COEFF
2PRT(I)
150  CONTINUE
      GO TO 90
160  CALL CHKLN (1)
      IF (N3.EQ.0) GO TO 170
      WRITE (IW,320)
      GO TO 180
170  WRITE (IW,240)
180  DO 190 J=1,NPRT
      NNPF=NPRT+1-J
      I=NNPF(NNPP)
      CALL DFFD (APPMAS(I),AAAMAS,13,5)
      CALL DFFD (CORRB(I),BBEMAS,13,5)
      CALL CHKLN (1)
      WRITE (IW,250) (AITEM(IU,I),IU=1,5),AAAMAS,BBEMAS
190  CONTINUE
      WRITE (IW,330)
      RETURN
C*****
C**  FORMAT STATEMENTS
C*****
200  FORMAT (3I2,I2,I3,F11.5,I3,F9.5,I2,I3,2F5.2,F6.2,F5.2,F7.4,F4.1,
2 F6.4,I2,1H$)
210  FORMAT (//20X,7HTABE 1//)
220  FORMAT (/24X,4HMASS,8X,11HUNCERTAINTY,2X,9HVCL AT 20,2X,
2 11HCOEF OF EXP/5X,4HITEM,16X,3H(G),13X,3H(G),6X,5H(CM3)/)

```

```

230  FORMAT (2X,5A3,1X,15A1,1X,F13.8,1X,F10.5,1X,F10.5)      FPR0
240  FORMAT (/2X,4HITEM,18X,22+CCR.A (MG)  CORR.C (MG)/)      FPR0
250  FORMAT (2X,5A3,1X,13A1,1X,13A1)      FPR0
260  FORMAT (1H1,65A1,4HPAGE,13)      FPR0
270  FORMAT (1X,65A1,12,1F/12,1H/12)      FPR0
280  FORMAT (1X,65A1)      FPR0
290  FORMAT (1X,11HTEST NUMBER,2X,18A1//)      FPR0
300  FORMAT (/24X,4HMASS,9X,11HUNCERTAINTY,2X,9HVOL AT 20,2X,
2  11HCOEF OF EXP/5X,4HITEM,15X,4H(LB),12X,4H(LB),6X,6H(IN3)/)  FPR0
310  FORMAT (2X,5A3,1X,15A1,1X,F13.11,1X,F10.6,1X,F8.6)      FPR0
320  FORMAT (12X,4HITEM,8X,5HCCR.A,7X,5HCCR.B/23X,10H(MICRO-LE),3X,
2  10H(MICRO-LE))      FPR0
330  FORMAT (1H1)      FPR0
    END      FPR0
--- TEXTS1 SUBPROGRAM ---
    SUBROUTINE TEXTS1      TS10
C*****TS10
C**  SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION **TS10
C**  PROGRAM VERSION OF SEPT.10,1971    WRITTEN BY R.C.RAYBOLD      **TS10
C**  AND MRS.R.N.VARNER      **TS10
C**  PRINT FIRST OF LAST TWO PAGES OF REPORT      **TS10
C**  MODIFIED BY R. N. VARNER SEPT 1979      **TS10
C*****TS10
    COMPCN/UNITIO/IR, IW, IP, IPL, ITMP      TS10
    WRITE (IW,10)      TS10
    WRITE (IW,20)      TS10
    WRITE (IW,30)      TS10
    WRITE (IW,40)      TS10
    WRITE (IW,50)      TS10
    WRITE (IW,60)      TS10
    WRITE (IW,70)      TS10
    WRITE (IW,80)      TS10
    WRITE (IW,90)      TS10
    WRITE (IW,100)      TS10
    RETURN      TS10
C*****TS10
C**  FORMAT STATEMENTS      **TS10
C*****TS10
10  FORMAT (1X,36H      SUMMARY      TS10
236H  THE ESTIMATED MASS VALUES/      TS10
31X,36H      TS10
436H  LISTED IN TABLE II ARE BASED ON AN/      TS10
51X,36H  FOR CONVENIENCE, THE RESULTS      TS10
636H  IMPLICIT TREATMENT OF DISPLACEMENTS/      TS10
71X,36HCF THIS WORK ARE SUMMARIZED IN      TS10
836H  VOLUMES, E.G., 'APPARENT MASS'/      TS10
91X,36HTABLES I AND II. THE VALUES      TS10
*36H  'APPARENT MASS VERSUS BRASS'.)      TS10
20  FORMAT (1X,36HASSIGNED ARE WITH REFERENCE TO THE      TS10
236H  'APPARENT MASS VERSUS DENSITY/      TS10
31X,36HSTANDARDS IDENTIFIED ON THE DATA      TS10
436H  8.0'. THE VALUES ARE LISTED AS/      TS10
51X,36HSHEETS. THE UNCERTAINTY FIGURE IS      TS10
636H  CORRECTIONS TO BE APPLIED TO THE/      TS10
71X,36HAN EXPRESSION OF THE OVERALL      TS10
836H  LISTED NOMINAL VALUE (A POSITIVE/      TS10
91X,36HUNCERTAINTY USING THREE STANDARD      TS10
*36H  CORRECTION INDICATES THAT THE MASS)

```

|    |                                                    |          |
|----|----------------------------------------------------|----------|
| 30 | FORMAT (1X,36HDEVIATIONS AS A LIMIT TO THE ,       | TS100440 |
|    | 236H IS LARGER THAN THE STATED NOMINAL/            | TS100450 |
|    | 31X,36HEFFECT OF RANDOM ERRORS OF THE ,            | TS100460 |
|    | 436H VALUE BY THE AMOUNT OF THE/                   | TS100470 |
|    | 51X,36HMEASUREMENT ASSOCIATED WITH THE ,           | TS100480 |
|    | 636H CORRECTION). THESE VALUES ARE/                | TS100490 |
|    | 71X,36HMEASUREMENT PROCESSES. THE MAGNI-           | TS100500 |
|    | 836H COMPUTED FROM THE VALUES BASED ON/            | TS100510 |
|    | 91X,36HTUDE OF SYSTEMATIC ERRORS FROM ,            | TS100520 |
|    | *36H AN EXPLICIT TREATMENT OF DISPLACE-)           | TS100530 |
| 40 | FORMAT (1X,36HSCURCES OTHER THAN THE USE OF ,      | TS100540 |
|    | 236H MENT VOLUMES USING THE FOLLOWING/             | TS100550 |
|    | 31X,36HACCEPTED VALUES FOR CERTAIN ,               | TS100560 |
|    | 436H DEFINING RELATIONS AND ARE/                   | TS100570 |
|    | 51X,36HSTARTING STANDARDS ARE CONSIDERED ,         | TS100580 |
|    | 636H UNCERTAIN BY THE AMOUNT SHOWN IN/             | TS100590 |
|    | 71X,36HNEGLECTIBLE. IT SHOULD BE NOTED ,           | TS100600 |
|    | 836H TABLE I. /                                    | TS100610 |
|    | 91X,36H THAT THE MAGNITUDE OF THE UNCER-           | TS100620 |
|    | *36H )                                             | TS100630 |
| 50 | FORMAT (1X,36HTAINTY REFLECTS THE PERFORMANCE OF , | TS100640 |
|    | 236H THE ADJUSTMENT OF WEIGHTS TO/                 | TS100650 |
|    | 31X,36H THE MEASUREMENT PROCESS USED TO ,          | TS100660 |
|    | 436H MINIMIZE THE DEVIATION FROM NCMI-/            | TS100670 |
|    | 51X,36H ESTABLISH THESE VALUES. THE MASS ,         | TS100680 |
|    | 636H NAL ON THE BASIS OF 'NORMAL BRASS'/           | TS100690 |
|    | 71X,36H UNIT, AS REALIZABLE IN ANOTHER ,           | TS100700 |
|    | 836H (IN ACCORDANCE WITH COR. A BELOW)/            | TS100710 |
|    | 91X,36H MEASUREMENT PROCESS, WILL BE ,             | TS100720 |
|    | *36H IS WIDESPREAD IN THIS COUNTRY AND)            | TS100730 |
| 60 | FORMAT (1X,36H UNCERTAIN BY AN AMOUNT WHICH IS A , | TS100740 |
|    | 236H IN MANY PARTS OF THE WORLD./                  | TS100750 |
|    | 31X,36H COMBINATION OF THE UNCERTAINTY OF ,        | TS100760 |
|    | 436H VALUES STATED ON EITHER BASIS ARE/            | TS100770 |
|    | 51X,36H THIS PROCESS AND THE PROCESS IN ,          | TS100780 |
|    | 636H INTERNALLY CONSISTENT AND/                    | TS100790 |
|    | 71X,36H WHICH THESE STANDARDS ARE USED. ,          | TS100800 |
|    | 836H DEFINITE. THERE IS, HOWEVER, A/               | TS100810 |
|    | 91X,36H ,                                          | TS100820 |
|    | *36H SYSTEMATIC DIFFERENCE BETWEEN THE)            | TS100830 |
|    | FORMAT (1X,36H THE ESTIMATED MASS VALUES ,         | TS100840 |
|    | 236H VALUES ASSIGNED ON EACH BASIS, THE/           | TS100850 |
|    | 31X,36H LISTED IN TABLE I ARE BASED ON AN ,        | TS100860 |
|    | 436H VALUE ON THE BASIS OF 'DENSITY/               | TS100870 |
|    | 51X,36H EXPLICIT TREATMENT OF DISPLACEMENT ,       | TS100880 |
|    | 636H 8.0' BEING 7 MICROGRAMS/GRAM LAR-/            | TS100890 |
|    | 71X,36H VOLUMES, E.G., 'TRUE MASS', 'MASS ,        | TS100900 |
|    | 836H GER THAN THE VALUE ON THE BASIS OF/           | TS100910 |
|    | 91X,36H IN VACUO', MASS IN THE NEWTONIAN ,         | TS100920 |
|    | *36H NORMAL BRASS. THIS SYSTEMATIC)                | TS100930 |
|    | FORMAT (1X,36H SENSE. THE DISPLACEMENT VOLUME ,    | TS100940 |
|    | 236H DIFFERENCE IS CLEARLY DETECTABLE/             | TS100950 |
|    | 31X,36H ASSOCIATED WITH EACH VALUE IS ,            | TS100960 |
|    | 436H ON MANY DIRECT READING BALANCES. /            | TS100970 |
|    | 51X,36H LISTED AS WELL AS THE VOLUMETRIC ,         | TS100980 |
|    | 636H /                                             | TS100990 |
|    | 71X,36H COEFFICIENT OF EXPANSION. THESE ,          | TS101000 |
|    | 836H CORRECTION A - 'APPARENT MASS/                | TS101010 |

```

          91X,36HVALUES SHOULD BE USED, TOGETHER ,
          *36H VERSUS BRASS' OR 'WEIGHT IN AIR)
90  FORMAT (1X,36HWITH APPROPRIATE CORRECTION FOR ,
          236H AGAINST BRASS' IS DETERMINED BY A/
          31X,36HTHE BUOYANT EFFECTS OF THE ,
          436H HYPOTHETICAL WEIGHING OF THE/
          51X,36HENvironment, TO ESTABLISH CONSIST- ,
          636H WEIGHT AT 20 CELSIUS IN AIR HAVING/
          71X,36HENT MASS VALUES FOR CEJECTS WHICH ,
          836H A DENSITY OF 1.2 MG/CM3, WITH A/
          91X,36HDIFFER SIGNIFICANTLY IN DENSITY ,
          *36H (NORMAL BRASS) STANDARD HAVING A)
100  FORMAT (1X,36HAND/CR FOR MEASUREMENTS WHICH MUST
          236H DENSITY OF 8.4 G/CM3 AT 0 CELSIUS/
          31X,36HBE MADE IN DIFFERING ENVIRONMENTS. ,
          436H WHOSE COEFFICIENT OF VOLUMETRIC/
          51X,36HTHE RELATION 1LB AVDP=.45359237KG ,
          636H EXPANSION IS 0.000054 PER DEGREE/
          71X,36HIS USED AS REQUIRED. ,
          836H CELSIUS, AND WHOSE VALUE IS BASED/
          91X,36H
          *36H
          END
--- TEXTS2 SUBPROGRAM ---
          SUBROUTINE TEXTS2
C*****
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION
C** PROGRAM VERSION OF SEPT.10,1971 WRITTEN BY R.C.RAYCLD
C** AND MRS.R.N.VARNER
C** PRINT LAST PAGE OF REPORT
C** MODIFIED BY R. N. VARNER SEPT 1979
C*****
COMMON/UNITID/IR,IL,IP,IPL,ITMP
WRITE (IW,10)
WRITE (IW,20)
RETURN
C*****
C** FORMAT STATEMENTS
C*****
10  FORMAT (
      236H
      31X,36HON ITS TRUE MASS OR WEIGHT IN ,
      436H WEIGHT, IN AIR HAVING A DENSITY OF /
      51X,36HVACUO. ,
      636H 1.2 MG/CM3, WITH A STANDARD HAVING /
      71X,36H
      836H A DENSITY OF 8.0 G/CM3 AT 20 )
20  FORMAT (1X,36H CORRECTION B - 'APPARENT MASS
      236H CELSIUS, AND WHOSE VALUE IS BASED /
      31X,36HVERSUS DENSITY 8.0' IS DETERMINED ,
      436H ON ITS TRUE MASS OR WEIGHT IN /
      51X,36HBY A HYPOTHETICAL WEIGHING OF THE ,
      636H VACUO.
      END
--- DPF SUBPROGRAM ---
          SUBROUTINE DPF (A,B,N,D)
C*****
C** SUBROUTINE OF THE NATIONAL BUREAU OF STANDARDS MASS CALIBRATION

```

```

C** PROGRAM VERSION CF SEPT.10.1971 WRITTEN BY R.C.RAYECLD **DPF00040
C** AND MRS.R.N.VARNER **CPF00050
C** MODIFIED BY R. N. VARNER SEPT 1979 **DPF00060
C***** **DPF00070
C** WRITTEN BY CLAYTON ALBRIGHT OF CSD **DPF00080
C** A ROUTINE TO CONVERT A DOUBLE PRECISION NUMBER TO A BLOCK OF **CPF00090
C** CHARACTERS WHICH WHEN OUTPUT WITH SUITABLE FORMAT (EG. 30A1) **DPF00100
C** WILL YIELD THE NUMBER IN F-TYPE FORMAT (NG EXPONENT). **DPF00110
C** THE OUTPUT NUMBER WILL BE RIGHT ADJUSTED IN THE FIELD **DPF00120
C** **DPF00130
C** A = THE DOUBLE PRECISION NUMBER TO BE CONVERTED **DPF00140
C** B = A BLOCK OF N WORDS TO CONTAIN THE RESULTS. THE CHARACTERS **CPF00150
C** ARE STORED ONE TO A WORD IN B. LEFT ADJUSTED IN THE WORD **DPF00160
C** N = FIELD WIDTH DESIRED **DPF00170
C** D = NUMBER OF PLACES DESIRED TO RIGHT OF DECIMAL POINT. **DPF00180
C** **DPF00190
C** IF NUMBER OVERFLOWS FIELD WIDTH, FIELD IS FILLED WITH '*'S **DPF00200
C** IF UNDERFLOW OCCURS RESULT WILL BE ZERO **DPF00210
C** **DPF00220
C** *RESTRICTION- AT LEAST ONE CHARACTER POSITION MUST BE ALLOWED **CPF00230
C** FOR SIGN REGARDLESS OF + OR -. IF USER OBJECTS **CPF00240
C** TO THIS RESTRICTION. HE NEED ONLY USE THE **DPF00250
C** RESULTING OUTPUT CHARACTERS BEGINNING AT E(2) **DPF00260
C** INSTEAD OF E(1). **DPF00270
C** **DPF00280
C***** **DPF00290
C** TYPE STATEMENTS **DPF00300
C***** **DPF00310
C** INTEGER B,D **DPF00320
C** DOUBLE PRECISION A,X **DPF00330
C***** **DPF00340
C** DIMENSION STATEMENT **DPF00350
C***** **DPF00360
C** DIMENSION E(1) **DPF00370
C** COMMON /DPFDVL/ KFD(18) **DPF00380
C** IF (D+1.GE.N) GO TO 60 **DPF00390
C***** **CPF00400
C** ROUND THE NUMBER AT DESIRED DECIMAL PLACE **DPF00410
C***** **DPF00420
C** X=DABS(A)+.5*10.**(-D) **CPF00430
C** MM=N-D-2 **DPF00440
C** X=X*10.**(-MM) **DPF00450
C** IF (X.GE.1.D0) GO TO 60 **CPF00460
C** IF (X.GE.0.1D0.AND.A.LT.0.D0) GO TO 60 **DPF00470
C** MM=MM+1 **DPF00480
C** DO 10 I=1,MM **DPF00490
C** B(I)=KFD(11) **DPF00500
C** K=IDINT(X*10.D0) **DPF00510
C** X=X*10.-FLOAT(K) **DPF00520
C** IF (K.NE.0) GO TO 20 **DPF00530
C** CONTINUE **DPF00540
C** I=MM **DPF00550
C***** **DPF00560
C** PREFIX MINUS SIGN IF A NEGATIVE **DPF00570
C***** **CPF00580
C** IF (A.LT.0.D0) B(I)=KFD(12) **DPF00590
C** IF (1.EQ.MM) GO TO 40 **DPF00600
C***** **DPF00610

```

```

C**      CONVERT INTEGER PART                                **DPF006
C*****DPF006
      I=I+1                                DPF006
      DO 30 J=I,MM                          DPF006
      B(J)=KFD(K+1)                        DPF006
      K=ICINT(X*10.D0)                    DPF006
30      X=X*10.-FLCAT(K)                    DPF006
C*****DPF006
C**      STORE DECIMAL FCINT                                **DPF00
C*****DPF00
40      MM=MM+2                                DPF00
      B(MM-1)=KFD(13)                    DPF00
C*****DPF00
C**      CONVERT FRACTIONAL PART                                **DPF00
C*****DPF00
      DO 50 I=1,C                                DPF00
      B(MM)=KFD(K+1)                        DPF00
      MM=MM+1                                DPF00
      K=ICINT(X*10.D0)                    DPF00
50      X=X*10.-FLCAT(K)                    DPF00
      RETURN                                DPF00
C*****DPF00
C**      STORE *'S IF CVERFLCW                                **DPF00
C*****DPF00
60      DO 70 I=1,N                                DPF00
70      B(I)=KFD(14)                        DPF00
      RETURN                                DPF00
      END                                    DPF00
--- CHKLN SUBPROGRAM ---
      SUBROUTINE CHKLN (N)                                CHK01
C*****CHK01
C**      SUBROUTINE TO CHECK BEGINNING CF A NEW PAGE                                **CHK01
C**      ADDED BY R. N. VARNER SEPT 1979                                **CHK01
C*****CHK01
      COMMON /PRT2/ IPAGE,NQSER,IPGCT                                CHK01
      COMMON /UNITIC/ IR,IR,IP,IFL,ITMP                                CHK01
      IF (IPGCT+N.GT.IPL) CALL FGCCNT                                CHK01
      IPGCT=IPGCT+N                                CHK01
      RETURN                                CHK01
      END                                    CHK01

```