

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Virginia

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2017.



2026

Scope

Mass Echelon II

20 kg to 1 mg

Volume Transfer, II

10 gal to 5 gal

Mass Echelon III

25 kg to 1 mg

Tuning Forks

2500 lb to 0.001 lb

10 kHz to 1 kHz

8 oz to 0.03125 oz

Weight Carts

6000 lb to 3000 lb

Katrice Lippa, Chief
NIST Office of Weights and Measures

Effective Dates: 2026-02-01 to 2027-01-31