

United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Indiana

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon III

1000 kg to 1 mg
2000 lb to 0.001 lb
2 oz to 0.03125 oz
50 oz t to 0.1 dwt

Weight Carts

6000 lb to 3000 lb

Volume Transfer, II


1000 gal to 5 gal
100 gal LPG

Time

Stopwatches
3 h

Length, Tape to Tape

100 ft to 1 ft



2016

CONDITIONAL

A handwritten signature in dark ink, reading "Carol T. Hockert".

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2016-01-01 to 2016-06-30

NOTES: Conditional Recognition will be issued only to meet weights and measures requirements and limitations will be stated in writing. Laboratory lacks adequate quality management system implementation.