

**United States Department of Commerce
National Institute of Standards and Technology**

**~ Certificate of Measurement Traceability ~
for**

Idaho

**a participant in the NIST Weights and Measures
Measurement Assurance Program for the Year(s)**

2009 - 2010

**The State standards are traceable to NIST and the laboratory has
demonstrated* measurement proficiency through training and
interlaboratory comparisons compliant to NIST HB 143:2007
(ISO/IEC 17025:2005).**



*Carol T. Hockert,
for the National Institute of Standards and Technology*

Chief, Weights and Measures Division
Title

January 1, 2010 to December 31, 2010
Effective Date Updated

***Demonstrated
Areas of Proficiency**

Mass

Echelon II
30 kg - 1 mg
50 lb - 1 μ lb
Echelon III
30 kg - 1 mg
2,500 lb - 0.001 lb

Volume

Gravimetric, I
5 gal
Volume transfer, II
750 gal - 5 gal
100 gal - 20 gal LPG

Length

Steel tapes, bench
100 ft - 0 ft
Rigid rules
72 in - 0 in

