

**United States Department of Commerce
National Institute of Standards and Technology**

**~ Certificate of Measurement Traceability ~
for**

Georgia

**a participant in the NIST Weights and Measures
Measurement Assurance Program for the Year(s)**

2011


**The State standards are traceable to NIST and the laboratory has
demonstrated* measurement proficiency through training and
interlaboratory comparisons compliant to NIST HB 143:2007
(ISO/IEC 17025:2005).**



*Carol T. Hockert
for the National Institute of Standards and Technology*

Chief, Weights and Measures Division
Title

January 1, 2011 to December 31, 2011
Effective Date



***Demonstrated
Areas of Proficiency**

Mass

Echelon II
30 kg - 1 mg
Echelon III
500 kg - 1 mg
1,000 lb - 0.001 lb
4 oz - 0.03125 oz
Weight Carts
5,000 lb - 3,000 lb

Volume

Volume transfer, II
500 gal - 1 gal
500 gal - 20 gal LPG

Length

Bench 100 ft - 0 ft
Rigid rules 36 in - 12 in

Grain Moisture

21 % - 12 %