United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Los Angeles County

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

2014

Scope

Mass Echelon III

10 kg to 100 g
5000 lb to 1 lb
8 oz to 4 oz

Volume Transfer, II

5 gal to 1 gal

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2014-01-01 to 2014-12-31

(Rev.: March 2013)