

United States Department of Commerce  
National Institute of Standards and Technology

~ Certificate of Measurement Traceability ~  
for

**Tennessee**

a participant in the NIST Weights and Measures  
Measurement Assurance Program for the Year(s)

**2010**

The State standards are traceable to NIST and the laboratory has  
*conditionally demonstrated\** measurement proficiency through  
training and interlaboratory comparisons.

**\*Demonstrated  
Areas of Proficiency**

**Mass**

Echelon III

100 g - 1 mg

1000 lb - 0.001 lb

8 oz - 0.03125 oz

**Volume**


Volume transfer, II

100 gal - 1 gal

*Conditional Status:*

*Facility has inadequate environ-  
mental controls (NIST HB 143  
5.3, 6.2). Conditional*

*Recognition is issued only so  
the laboratory can provide serv-  
ices to meet weights and meas-  
ures regulations.*



*Carol T. Hockert,  
for the National Institute of Standards and Technology*

Chief, Weights and Measures Division

*Title*

March 1, 2010 to December 31, 2010

*Effective Date*

