United States Department of Commerce

National Institute of Standards and Technology

Certificate of Metrological Traceability For:

West Virginia

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2005.

Scope

Mass Echelon III

25 kg to 1 mg
1000 lb to 0.001 lb
4 oz to 0.03125 oz

Weight Carts

5000 lb to 3000 lb

Volume Transfer, II

200 gal to 1 gal

100 gal to 25 gal LPG

2015

CONDITIONAL

Carol T. Hockert, Chief
NIST Office of Weights and Measures

Effective Dates: 2015-01-01 to 2015-12-31

NOTES: Conditional Recognition will be issued only to meet weights and measures requirements and limitation will be started in writing. Laboratory has unresolved corrective actions related to facility and environment.

(Rev.: March 2013)