


Erica Romsos & Peter Vallone

Applied Genetics Group, NIST

November 7, 2018



Steps in Forensic DNA Analysis




of Results



Sample Collection




DNA Extraction



DNA Quantitation



Multiplex PCR Amplification



Interpretation of Results

1-2 day process (a minimum of ~8 hours) with current laboratory procedures and technology



FORENSICS @ NIST

Commercial RDNA Instruments

- RapidHIT 200
 - PowerPlex 16HS
 - Globalfiler
- RapidHIT ID
 - Globalfiler



2 hour run time





1 Sample 90 minute run time

- ANDE/DNAScan
 - PowerPlex 16
- ANDE
 - FlexPlex (27 loci)



5 Samples 86 minute run time



Rapid DNA Instrument Testing

2012-2015: Ongoing communication with Developers

% S

2012

OVA


2016

DNA ACT Passs

2010: Apollo 200 Testing (**136 samples**)

09/2012: First prototypes delivered to NIST (DHS Instruments)

2015: NIST Participates in IXI PP16 Dev. Validation (**100** samples)







2016: NIST Participates in ANDE PP16 Dev. Validation (**150 samples**)

2016: NIST Participates in testing ParaDNA

2013: NIST Interlab Study (**350 samples**)

2017: NIST Participates in ANDE FlexPlex Dev Validation (**250 samples**)

Contents lists available at ScienceDirect

Forensic Science International: Genetics Supplement Series

journal homepage: www.elsevier.com/locate/FSIGSS

2014: NIST Maturity Assessment (**280 samples**)

Rapid DNA maturity assessment

Erica L. Romsos ^{a,*}, Sanae Lembirick ^b, Peter M. Vallone ^a

^a U.S. National Institute of Standards and Technology, NIST, 100 Bureau Drive, Gaithersburg, MD 20899-8314, USA



2018: NIST Rapid DNA Maturity Assessment (**240 samples**)

NIST: Rapid DNA Maturity Assessments

- Collection and distribution of samples to all participating laboratories
 - 2013: 3 labs, 150 samples
 - 2014: 7 labs, 280 samples
 - 2018: 9 participants, 240 samples
- Coordination of all testing sites to include return of all data to NIST for analysis and review
 - Includes IRB approval for swab collection and MTA for distribution
- Analysis and compilation of all data
- Summary of results presented across multiple meetings within the forensic and biometric communities



- Summer 2013 we felt confident that we could carry out an interlab assessment of the R-DNA <u>prototypes</u>
- Data was collected and analyzed in August 2013
 - All instruments had the same version of software and scripts for testing
- Three federal testing sites (7 individual instruments)
 - 50 single source reference buccal swabs were provided by NIST for each instrument tested
- 350 single source reference buccal swabs tested
 - Success defined as the automated calling of the 13 CODIS core STR loci
- Overall success = 88.3%



- Fall 2014 assessment of the current status of rapid DNA typing technology for the CODIS Core Loci
 - In support of lab and future booking station Rapid DNA implementation
- Many modifications to both hardware and software were made between 2013 and 2014
- 7 participating laboratories (11 independent instruments)
 - 20 single source reference buccal swabs tested
- 280 single source reference buccal swabs tested
 - Success defined as automated (lights out) calling of the 13 CODIS core loci
- Overall Success = 76.1%






 Goal: To measure the status of rapid DNA typing technology for the 20 CODIS core loci in support of booking station Rapid DNA implementation

 Rapid DNA instruments capable of genotyping the 20 CODIS core loci were eligible for participation

 20 single source reference buccal swabs were distributed to participating laboratories



2018 Rapid DNA Maturity Assessment Study Outline





FORENSICS @ NIST

Timeline of 2018 Maturity Assessment

February 2018-Present: Buccal samples collected at NIST and stored at RT (>900 swabs have been collected from >50 unique individuals)

May 2018: Official call for Participants in the 2018 Rapid DNA Maturity Assessment for the CODIS core 20 loci

July 2018: Samples shipped to participants

August 2018: Data returned to NIST for analysis



FORENSICS @ NIST

Participants	Instrument Platforms	Chemistry	Independent Instruments	Total Samples Tested	Analysis Method
Federal	ANDE 6C	FlexPlex	5	100	Rapid DNA Analysis
State	IntegenX RapidHIT 200				Modified
Police	Scriett 200	GlobalFiler Express	3	60	Rapid DNA Analysis
	IntegenX RapidHIT ID				
Vendor		GlobalFiler Express	4	80	Modified Rapid DNA Analysis
9 Participants	3 Platforms		12 Instruments	240 Sample:	S

Success Metrics

- Success was measured by complete and concordant genotypes produced for the 20 CODIS core loci
- Allele calls by the integrated rapid DNA devices were compared to lab generated profiles for concordance
 - Fusion 6C, PP21, GFE on a 3500xL
- Two interpretation parameters were implemented
 - Rapid DNA Analysis: Without human intervention
 - Modified Rapid DNA Analysis: Expert interpretation and analysis of electropherogram https://www.fbi.gov/file-repository/addendum-to-qas-for-rapid-dna.pdf/view



Rapid DNA Analysis-ANDE

Automated (lights-out) analysis without human intervention

Effective June 1, 2018, the following Rapid DNA system is approved for use at NDIS by an accredited forensic DNA laboratory:

Rapid DNA Analysis System for Accredited Laboratory Use

To date, ANDE 6C is the only rapid DNA system to be NDIS approved for automated rapid DNA analysis


Component	Name	Part/Version Number	
Rapid DNA Instrument	ANDE 6C Instrument	A0120001003	
Typing Kit	FlexPlex27	FlexPlex27	
Cartridge	ANDE A-Chip (FlexPlex)	A0210001057	
System Software	ANDE System Software	2.0.6	
Expert System Software	ANDE Expert System	2.0.5	

https://www.fbi.gov/services/laboratory/biometric-analysis/codis/rapid-dna



FORENSICS @ NIST

CODIS 20 Success: Automated Analysis



Average Success: 85%

Instruments




FORENSICS @ NIST

Modified Analysis Parameters

- Samples were manually reviewed (n=240)
 - ANDE profiles were analyzed in GeneMapper IDX v1.5
 - RapidHIT profiles were reviewed in GeneMarker HID v2.8.2
 - Heterozygote balance filter set at 0.25
- After manually interpreting a profile (PHR >0.25, low AT, recovery of "lost" data), concordance was checked against the laboratory generated reference profile
- Success was determined by complete and concordant profiles for the 20 CODIS core loci



CODIS 20 Success: Modified Analysis




Average Success: 90 %

Instruments



FORENSICS @ NIST

n=23





FORENSICS @ NIST

n=23

Instrument Related

Unknown

n=9

Data Transfer Failure

n=2

 Name
 Date modified
 Type

 □ DannoGUIState.xml
 8/21/2018 2:51 PM
 XML Document

 □ GelFill.HV_(14.23).png
 8/21/2018 2:51 PM
 PNG File

 □ StoryBoard_(14.23).txt
 8/21/2018 2:51 PM
 Text Document

 □ SyringePump_(14.23).csv
 8/21/2018 2:40 PM
 Microsoft Excel C...

No data was transferred from the instrument to analyze



n=23

Instrument Related


Partial Profile n=10

Unknown

n=9

Data Transfer Failure

n=2








n=23

Instrument Related Partial Profile n=10

Unknown n=9 Data
Transfer
Failure
n=2


Single Locus Dropout n=2





FORENSICS @ NIST

Summary of Success




Automated success rate of 85% (n=100)


- ANDE 6C is the only NDIS approved instrument for automated analysis
- xml files for the unsuccessful profiles were not generated by the instrument (15%)

Modified analysis success rate of 90%

- Manual interpretation of all samples (n=240 samples)
- Increase of success for manually reviewed profiles







Additional Metrics Analyzed

Base pair sizing precision


Heterozygote Balance




FORENSICS @ NIST

Precision-Base Pair Sizing

ANDE Size Pooled Standard Deviation




RapidHIT 200 Size Pooled Standard Deviation





ANDE




95% of the samples were greater than 52%



FORENSICS @ NIST

RapidHIT ID




95% of the samples were greater than 63%




FORENSICS @ NIST

RapidHIT 200



95% of the samples were greater than 52%





95% of the samples were greater than 59%

06 23 +++

FORENSICS @ NIST

Maturity Assessment Summary

- 12 instruments tested across 9 laboratories
- Total of 240 samples examined
 - 85% success rate for the CODIS 20 using Rapid DNA Analysis
 - 90% success rate for the CODIS 20 using Modified Rapid DNA Analysis
 - Success ranged from 60% to 100%
 - Precision was below 0.17 bp on for both ANDE 6C and RapidHIT 200
 - Combined heterozygote balance (all three instruments) was above 59%



Acknowledgements

Thank you to our participants

- ANDE
- Arizona Department of Public Safety
- Bensalem Police Department
- Federal Bureau of Investigation Laboratory
- Louisiana State Police Crime Laboratory
- Miami Beach Police Department
- Miami Dade Police Department
- NIST (DHS instruments, run at SNA Intl.)
- U.S. Army Criminal Investigation Laboratory

NIST – Applied Genetics Group Peter Vallone Steven Lund

Funding

FBI Biometrics Center of Excellence: Forensic DNA Typing as a Biometric tool.

Contact Information erica.romsos@nist.gov

