

United States Department of Commerce National Institute of Standards and Technology

Certificate of Metrological Traceability For:

Washington

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2017.



2020 to 2021

Scope

Mass Echelon I

20 kg to 1 mg
50 lb to 0.001 lb

Mass Echelon II

25 kg to 1 mg
1000 lb to 0.001 lb
4 oz to 0.03125 oz

Mass Echelon III

25 kg to 1 mg
4000 lb to 0.001 lb
4 oz to 0.03125 oz

Weight Carts

4000 lb to 3000 lb

Volume Gravimetric, I

5 gal

Volume Transfer, II

1000 gal to 5 gal

Volume Transfer, II LPG

200 gal to 19 gal

Time

Stopwatches, Timers
≤ 24 h

A handwritten signature in black ink, reading "Douglas A. Olson", is written over a horizontal line.

Douglas A. Olson, Chief
NIST Office of Weights and Measures

Effective Dates: 2020-02-01 to 2022-02-01