

# United States Department of Commerce

## National Institute of Standards and Technology

Certificate of Metrological Traceability For:

# Georgia

This laboratory has demonstrated evidence of an unbroken chain of metrological traceability of its standards to the international system of units (SI), documented measurement uncertainties, uses documented measurement procedures, successfully completed training and proficiency tests, documented calibration intervals, submitted a quality management system, and demonstrated suitable measurement assurance for the Scope listed on this certificate.

The Office of Weights and Measures Program assesses laboratories to NIST Handbook 143 - Program Handbook for State Weights and Measures Laboratories and ISO/IEC 17025:2017.



2022

### *Scope*

#### **Mass Echelon II**

20 kg to 1 mg

#### **Mass Echelon III**

500 kg to 1 mg

1000 lb to 0.001 lb

4 oz to 0.03125 oz

Weight Carts

5000 lb to 3000 lb

#### **Volume Transfer, II**

500 gal to 5 gal

#### **Volume Transfer, II LPG**

500 gal to 20 gal

A handwritten signature in black ink, appearing to read "K. Lippa", is written over a horizontal line.

*Katrice Lippa, Chief  
NIST Office of Weights and Measures*

*Effective Dates: 2022-02-01 to 2023-02-01*